



SLOVENSKI STANDARD
oSIST prEN ISO 3104:2022

01-december-2022

Naftni proizvodi - Prozorne in neprozorne tekočine - Določanje kinematične viskoznosti in izračun dinamične viskoznosti (ISO/DIS 3104:2022)

Petroleum products - Transparent and opaque liquids - Determination of kinematic viscosity and calculation of dynamic viscosity (ISO/DIS 3104:2022)

Mineralölerzeugnisse - Durchsichtige und undurchsichtige Flüssigkeiten - Bestimmung der kinematischen Viskosität und Berechnung der dynamischen Viskosität (ISO/DIS 3104:2022)

Produits pétroliers - Liquides opaques et transparents - Détermination de la viscosité cinématique et calcul de la viscosité dynamique (ISO/DIS 3104:2022)

Ta slovenski standard je istoveten z: prEN ISO 3104

ICS:

75.080 Naftni proizvodi na splošno Petroleum products in general

oSIST prEN ISO 3104:2022

en,fr,de

DRAFT INTERNATIONAL STANDARD

ISO/DIS 3104

ISO/TC 28

Secretariat: NEN

Voting begins on:
2022-10-21Voting terminates on:
2023-01-13

Petroleum products — Transparent and opaque liquids — Determination of kinematic viscosity and calculation of dynamic viscosity

Produits pétroliers — Liquides opaques et transparents — Détermination de la viscosité cinématique et calcul de la viscosité dynamique

ICS: 75.080

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN ISO 3104:2022](https://standards.iteh.ai/catalog/standards/sist/b69bd01a-9478-4fb2-8940-8c84598de438/osist-pren-iso-3104-2022)

<https://standards.iteh.ai/catalog/standards/sist/b69bd01a-9478-4fb2-8940-8c84598de438/osist-pren-iso-3104-2022>

This document is circulated as received from the committee secretariat.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO/DIS 3104:2022(E)

© ISO 2022

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 3104:2022](https://standards.iteh.ai/catalog/standards/sist/b69bd01a-9478-4fb2-8940-8c84598de438/osist-pren-iso-3104-2022)

<https://standards.iteh.ai/catalog/standards/sist/b69bd01a-9478-4fb2-8940-8c84598de438/osist-pren-iso-3104-2022>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Principle.....	2
5 Reagents and materials.....	2
6 Apparatus design and requirements.....	3
7 Verification.....	6
7.1 Viscometer.....	6
7.2 Liquid-in-glass thermometer.....	7
7.3 Digital contact thermometer.....	7
7.4 Timer.....	7
8 Re-calibration.....	7
9 Quality control.....	7
10 Sample preparation.....	8
10.1 Pre-analysis sample conditioning.....	8
10.2 Visual inspection and filtering.....	8
11 Procedure A — Manual equipment (Referee test method).....	8
12 Procedure B — Automated equipment.....	10
13 Cleaning of the viscometer tube.....	11
14 Calculation.....	12
14.1 Procedure A — Manual viscometers.....	12
14.2 Procedure B — Automated viscometers.....	13
15 Expression of results.....	13
16 Precision of procedure A.....	13
16.1 Determinability, <i>d</i>	13
16.2 Repeatability, <i>r</i>	14
16.3 Reproducibility, <i>R</i>	14
17 Precision of Procedure B.....	15
17.1 Determinability, <i>d</i>	15
17.2 Repeatability, <i>r</i>	15
17.3 Reproducibility, <i>R</i>	15
18 Test report.....	16
Annex A (normative) Viscometer types, calibration and verification.....	17
Annex B (normative) Thermometers for kinematic viscosity test.....	18
Annex C (normative) Conditioning of samples prior to manual or automated analysis.....	22
Annex D (normative) Calculation of acceptable tolerance zone (band) to determine conformance with a certified reference material (CRM).....	24
Bibliography.....	25

ISO/DIS 3104:2022(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 28, *Petroleum and related products, fuels and lubricants from natural or synthetic sources*.

This third edition cancels and replaces the second edition (ISO 3104:1994), which has been technically revised.

The main changes compared to the previous edition are as follows:

- precision data have been updated to all actual fuels on the market;
- biodiesel (FAME) blends and paraffinic diesel have been included in the scope;
- the procedure description and allowance of automated techniques have been included.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Many petroleum products, and some non-petroleum materials, are used as lubricants, and the correct operation of equipment depends upon the appropriate viscosity of the liquid being used. In addition, the viscosity of many petroleum fuels is important for the estimation of optimum storage, handling and operational conditions. Thus, the accurate measurement of viscosity is essential to many product specifications.

This document describes two test methods: Procedure A Manual and Procedure B Automated. Procedure A is the referee test method (or reference test method) to resolve doubts or dispute.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN ISO 3104:2022](https://standards.iteh.ai/catalog/standards/sist/b69bd01a-9478-4fb2-8940-8c84598de438/osist-pren-iso-3104-2022)

<https://standards.iteh.ai/catalog/standards/sist/b69bd01a-9478-4fb2-8940-8c84598de438/osist-pren-iso-3104-2022>

Petroleum products — Transparent and opaque liquids — Determination of kinematic viscosity and calculation of dynamic viscosity

WARNING — This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of users of this document to take appropriate measures to ensure the safety and health of personnel prior to the application of this document, and to determine the applicability of any other restrictions.

1 Scope

This document specifies Procedure A, using manual glass viscometers, and Procedure B, using glass capillary viscometers in an automated assembly, for the determination of the kinematic viscosity, ν , of liquid petroleum products, both transparent and opaque, by measuring the time for a volume of liquid to flow under gravity through a calibrated glass capillary viscometer. The dynamic viscosity, η , is obtained by multiplying the measured kinematic viscosity by the density, ρ , of the liquid. The range of kinematic viscosities covered in this test method is from 0,2 mm²/s to 300 000 mm²/s over the temperature range -20 °C to +150 °C.

NOTE The result obtained from this document is dependent upon the behaviour of the sample and is intended for application to liquids for which primarily the shear stress and shear rates are proportional (Newtonian flow behaviour). If, however, the viscosity varies significantly with the rate of shear, different results can be obtained from viscometers of different capillary diameters. The procedure and precision values for residual fuel oils, which under some conditions exhibit non-Newtonian behaviour, have been included.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3105, *Glass capillary kinematic viscometers — Specifications and operating instructions*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ASTM E1137, *Standard Specification for Industrial Platinum Resistance Thermometers*

ASTM E2877, *Standard Guide for Digital Contact Thermometers*

ISO 4259-4, *Petroleum and related products — Precision of measurement methods and results — Part 4: Use of statistical control charts to validate 'in-statistical-control' status for the execution of a standard test method in a single laboratory*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org>

ISO/DIS 3104:2022(E)

3.1 kinematic viscosity

 ν

resistance to flow of a fluid under gravity

Note 1 to entry: For gravity flow under a given hydrostatic head, the pressure head of a liquid is proportional to its density, ρ . For any particular viscometer, the time of flow of a fixed volume of fluid is directly proportional to its kinematic viscosity, ν :

$$\nu = \eta/\rho$$

where η is the dynamic viscosity coefficient.

3.2 dynamic viscosity

coefficient of dynamic viscosity
viscosity

 η

ratio between the applied shear stress and rate of shear of a liquid

Note 1 to entry: It is a measure of the resistance to flow or deformation of a liquid.

Note 2 to entry: The term dynamic viscosity is also used in a different context to denote a frequency-dependent quantity in which shear stress and shear rate have a sinusoidal time dependence.

3.3 density

 ρ

mass per unit volume of a substance at a given temperature

iTeh STANDARD PREVIEW
(standards.iteh.ai)

4 Principle

oSIST prEN ISO 3104:2022

<https://standards.iteh.ai/catalog/standards/sist/b69bd01a-9478-4fb2-8940->

The time is measured for a fixed volume of liquid to flow under gravity through the glass capillary of a calibrated viscometer under a reproducible driving head and at a known and closely controlled temperature. The kinematic viscosity is the product of the measured flow time and the calibration constant of the viscometer.

5 Reagents and materials

5.1 Cleaning solution, strongly-oxidizing cleaning solution or alkaline cleaning solutions can be used.

Alkaline cleaning solutions with a pH of greater than 10 are not recommended as they have been shown to change the viscometer calibration. If these are used, then the viscometer calibration should be verified to ensure there is no change.

5.2 Sample solvent, completely miscible a prewash with an aromatic solvent such as toluene or heptane might be necessary to remove asphaltenic material. When cleaning capillaries inside the bath, the boiling point of the cleaning solution shall be higher than the bath temperature.

5.3 Drying solvent, suitable and volatile at the used temperature. Filter before use. If moisture remains, use a drying solvent miscible with water (5.4).

NOTE When cleaning capillaries inside the bath and if the bath temperature is higher than 50 °C, acetone is not suitable.

5.4 Water, deionized or distilled, conforming to Grade 3 of ISO 3696:1987. Filter before use.

5.5 Certified viscosity reference standards (CRM), with data provided by an accredited calibration laboratory — traceable to the international agreed value of distilled water (1,003 4 mm²/s at 20 °C as specified in ISO/TR 3666) and calibrated in accordance with a standard practice for the basic calibration of master viscometers and viscosity oils, such as in ASTM D2162.^[8] ISO 17034 specifies the requirements for CRM producers.

6 Apparatus design and requirements

6.1 Drying tubes, consisting of a desiccant drying system, consisting of either externally mounted drying tubes or an integrated desiccant drying system designed to remove ambient moisture from the capillary tube. Ensure that they are packed loosely and that the desiccant is not saturated with water.

6.2 Sample filter, micron screen or fretted (sintered) glass filter, no more than 75 µm.

6.3 Reagent filter, micron screen or fretted (sintered) glass filter, no more than 11 µm.

6.4 Ultrasonic bath, unheated — with an operating frequency between 25 kHz to 60 kHz and a typical power output of ≤100 W, of suitable dimensions to hold container(s) placed inside of bath, for use in effectively dissipating and removing air or gas bubbles that can be entrained in viscous sample types prior to analysis. It is permissible to use ultra-sonic baths with operating frequencies and power outputs outside this range; however, it is the responsibility of the laboratory to conduct a data comparison study to confirm that results determined with and without the use of such ultrasonic baths does not materially impact results.

6.5 Manual apparatus

6.5.1 Glass capillary viscometer, calibrated in accordance with ISO 3105.

The viscometer shall have a certificate of calibration provided by an accredited laboratory that meets ISO/IEC 17025. The calibration constant should be checked before first use of the capillary and only changed if necessary.

The calibration constant, C , is dependent upon the gravitational acceleration at the place of calibration. The variation in the value of g across the earth's surface is about 0,5 % due to latitude plus approximately 0,003 % per 100 m altitude. Apply a gravity correction to the viscometer calibration constant as in [Formula \(1\)](#), if the acceleration of gravity of the testing laboratory differs by more than 0,1 % of the calibration laboratory.

$$C_2 = \left(\frac{g_2}{g_1} \right) C \quad (1)$$

where the subscripts 1 and 2 indicate, respectively, the calibration laboratory and the testing laboratory.

NOTE Calculation of acceleration of gravity values can be found at www.NPL.co.uk.

IMPORTANT — Viscometers used for silicone fluids, fluorocarbons and other liquids, which are difficult to remove using a cleaning agent, shall be reserved for the exclusive use of those fluids, except during their calibration. Subject such viscometers to calibration checks at frequent intervals. The solvent washings from these viscometers shall not be used for the cleaning of other viscometers. If the viscometer is cleaned using the material in [5.1](#) then the user shall verify the calibration before further use.

ISO/DIS 3104:2022(E)

6.5.2 Viscometer holder or mounting device within the temperature-controlled bath, enabling the glass viscometer to be suspended so that the upper meniscus is directly above the lower meniscus vertically within 1° in all directions.

Those viscometers whose upper meniscus is offset from directly above the lower meniscus shall be suspended vertically within 0,3° in all directions (see ISO 3105).

The proper alignment of vertical parts may be confirmed by using a plumb line, but for rectangular baths with opaque ends, this might not be possible.

6.5.3 Temperature-controlled bath, containing a transparent liquid of sufficient depth such that at no time during the measurement is any portion of the sample in the viscometer less than 20 mm below the surface of the bath liquid or less than 20 mm above the bottom of the bath.

Temperature control of the bath liquid shall be such that, for each series of flow-time measurements, within the range of 15 °C to 100 °C the temperature of the bath medium does not vary by more than $\pm 0,02$ °C from the selected temperature over the length of the viscometer, and/or between the position of each viscometer, and/or at the location of the temperature measuring device i.e. the temperature needs to be constant at the capillary and at the position of the temperature measuring device within a maximum difference of 0,04 °C. For temperatures outside this range, the deviation from the desired temperature shall not exceed $\pm 0,05$ °C.

Adjust and maintain the viscometer bath at the required test temperature within the limits given in 6.5.3, taking account of the conditions given in Annex B and of the corrections supplied on the certificates of calibration for the temperature measuring device. Maintain the bath temperature at the test temperature using the readings of the temperature measuring device with the corrections supplied by the certificate of calibration.

The temperature measuring device shall be held in an upright position under the same conditions of immersion as when calibrated.

6.5.4 Temperature-measuring device, for the range 0 °C to 100 °C, being either:

- a) a calibrated liquid-in-glass thermometer, as listed in Annex B with a calibration and measurement capability (CMC) of $\pm 0,04$ °C after correction or better, or
- b) a digital contact thermometer (DCT) as described in Table 1 for this temperature with equal or better CMC.

NOTE a DCT is preferred due to the lower uncertainty of measurement

6.5.4.1 The calibration data should be traceable to a calibration or metrology standards body and meet the uncertainty of measurement required. The calibration certificate shall include data covering the series of temperature test points which are appropriate for its intended use. When two temperature measuring devices are used in the same bath in this range, they shall agree within 0,04 °C.

If calibrated liquid-in-glass thermometers are used, the use of two thermometers is recommended.

Outside the range 0 °C to 100 °C, a calibrated liquid in-glass thermometer with a CMC of $\pm 0,1$ °C or better shall be used, and when two temperature measuring devices are used in the same bath they shall agree within $\pm 0,1$ °C.

When using liquid-in-glass thermometers, use a magnifying device to read the thermometer to the nearest 1/5 division (e.g. 0,01 °C or 0,02 °C) to ensure that the required test temperature and temperature control capabilities are met. It is recommended that thermometer readings (and any corrections supplied on the certificates of calibrations for the thermometers) be recorded on a periodic basis to demonstrate compliance with the test method requirements.

6.5.4.2 A DCT meeting the requirements in Table 1 is to be used and is dependent upon temperature range in use