
**Bunkering of marine fuel using the
Coriolis mass flow meter (MFM) system**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Full standard:
<https://standards.iteh.ai/catalog/standards/sist/290b44d9-0158-429f-815f-0e57cbad4eec/iso-prf-22192>

PROOF / ÉPREUVE



iTeh STANDARD PREVIEW
(standards.iteh.ai)
Full standard:
<https://standards.iteh.ai/catalog/standards/sist/290b44d9-0158-429f-815f-0e57cbad4eec/iso-prf-22192>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	v
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 General requirements (safety, health and the environment).....	7
5 Metrological requirements.....	7
5.1 General.....	7
5.2 Mass flow meter requirement.....	7
5.3 Mass flow meter system requirements.....	8
5.4 Post approval maintenance.....	9
5.4.1 Meter zero verification frequency.....	9
5.4.2 Zero verification procedure.....	9
5.4.3 Meter and ancillary devices verification and/or calibration frequency.....	9
5.4.4 Software upgrade/ software update.....	9
6 System integrity requirements.....	10
6.1 General.....	10
6.2 Metrological control.....	10
6.2.1 Documentation.....	10
6.2.2 Type approval and pattern evaluation.....	10
6.3 Security features.....	10
6.3.1 Equipment security.....	10
6.3.2 Software security.....	10
6.3.3 Data security.....	11
6.3.4 Critical alarm.....	11
6.4 Installation and commissioning.....	11
6.4.1 Pre-installation and MFM system sealing plan.....	11
6.4.2 Installation and re-installation.....	11
6.4.3 Commissioning.....	11
6.4.4 Re-commissioning.....	12
6.5 Operational security.....	12
6.6 Maintenance and control of MFM system.....	12
6.6.1 Inspection and verification.....	12
6.6.2 Breaking of seals and re-sealing of MFM system.....	12
7 Meter selection and installation requirements.....	12
7.1 General.....	12
7.2 Site survey onboard tankers.....	13
7.3 Meter selection.....	13
7.4 Meter installation.....	13
7.5 Meter commissioning.....	14
8 MFM system verification requirements.....	14
9 Metering procedures.....	14
9.1 General.....	14
9.2 Documentation.....	14
9.2.1 General.....	14
9.2.2 Pre-delivery documentation.....	15
9.2.3 Post-delivery documentation.....	15
9.3 Additional documentation for bunker surveyor.....	15
9.4 Additional documentation for bunker tanker.....	16
9.4.1 Meter totalizer log.....	16
9.4.2 Documents carried onboard the bunker tanker.....	16

9.5	Planning for bunkering operation.....	17
9.6	Pre-delivery procedures.....	17
9.6.1	Flow measurement conditions and checks on system integrity.....	17
9.6.2	Pre-delivery conference.....	17
9.6.3	Bunker requisition form (mass flow metering).....	17
9.6.4	Mass flow metering system seals checklist.....	18
9.6.5	Meter reading record form (delivery).....	18
9.7	Delivery procedures.....	19
9.7.1	General.....	19
9.7.2	Start of delivery.....	19
9.7.3	End of delivery.....	20
9.8	Post-delivery procedures and checks.....	21
9.8.1	Meter reading record form (delivery).....	21
9.8.2	Mass flow metering system seals checklist.....	21
9.8.3	Bunker metering ticket.....	21
9.8.4	Bunker delivery note.....	21
9.8.5	Custody transfer quantity.....	22
9.9	Others.....	23
9.9.1	MFM system failure.....	23
9.9.2	Quantity dispute.....	23
10	Sampling.....	23
	Annex A (informative) Safety, health and the environment.....	24
	Annex B (informative) Uncertainty budget table.....	28
	Annex C (informative) Metrological and system integrity requirements.....	29
	Annex D (informative) Procedures for zero verification.....	30
	Annex E (informative) Sealable bolts and nuts for blanks and ancillary device.....	31
	Annex F (informative) Request for information checklist.....	33
	Annex G (informative) Typical schematic diagram for MFM system (for delivery).....	35
	Annex H (normative) Example of bunker requisition form (mass flow metering).....	36
	Annex I (informative) Example of mass flow metering system seals checklist.....	37
	Annex J (informative) Example of meter reading record form (delivery).....	38
	Annex K (informative) Example of bunker metering ticket.....	39
	Annex L (informative) Bunkering pre-delivery safety checklist.....	41
	Annex M (informative) Example of a survey time log.....	44
	Annex N (informative) Example of statement of fact.....	46
	Annex O (informative) Example of meter totalizer log.....	47
	Annex P (informative) Examples of letter of protest.....	49
	Annex Q (informative) Quantity dispute procedures and documents.....	50
	Bibliography.....	51

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 28, *Petroleum and related products, fuels and lubricants from natural or synthetic sources*, Subcommittee SC 2, *Measurement of petroleum and related products*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document was developed for the benefit of the bunker industry comprising ship owners, operators, charterers, bunker suppliers, bunker craft operators and bunker surveyors and is intended to enhance the efficiency of bunkering operations and promote best practices in the measurement of bunker fuel delivered.

This document sets out the international best practices which documents principles, requirements and procedures in the application of mass flow metering to bunkering.

This document does not alter the contractual obligations of the parties involved in the bunker delivery.

[Figure 1](#) shows the application of MFM bunkering requirements for bunker custody transfer.




Figure 1 — Application of MFM bunkering requirements

Bunkering of marine fuel using the Coriolis mass flow meter (MFM) system

1 Scope

This document specifies procedures and requirements for the transfer of bunkers to vessels by bunker tankers using the Coriolis mass flow meter (MFM) system. It encompasses the process leading to the approval of the MFM system as installed on bunker tankers and post-approval bunkering operation. It covers terminology, specifications, requirements and procedures on metrology, system integrity, metering system selection and installation, acceptance test, bunker delivery and dispute handling.

NOTE Local and international regulations, such as the International Convention for the Prevention of Pollution from Ships (MARPOL) can apply.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM: 1995)*

ISO 13739, *Petroleum products — Procedures for the transfer of bunkers to vessels*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

International Recommendation OIML R117-1, *Dynamic measuring systems for liquids other than water*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <http://www.electropedia.org/>

3.1

accuracy of measurement

closeness of the agreement between the result of a measurement and the conventional, true value of the measurement

Note 1 to entry: Good accuracy implies small random and systematic errors.

Note 2 to entry: The quantitative expression of accuracy should be in terms of uncertainty of measurement.

3.2

adjustment

set of operations carried out on a measuring system to provide prescribed indications corresponding to given values of quantity to be measured

Note 1 to entry: Types of adjustment of a measuring system include zero adjustment of a measuring system, offset adjustment and span adjustment (sometimes called gain adjustment).

Note 2 to entry: Adjustment of a measuring system should not be confused with calibration, which is a prerequisite for adjustment.

Note 3 to entry: After an adjustment of a measuring system, the measuring system shall be recalibrated.

[SOURCE: JCGM 200]

3.3
air buoyancy correction

correction applied to obtain the conventional mass from true mass to take into account the reduction in true mass due to the buoyancy effect of air

3.4
ancillary device

device intended to perform a particular function, directly involved in elaborating, transmitting or displaying measurement results

EXAMPLE Zero adjustment device, repeating indicating device, printing device, memory device, totalising indicating device, correction device, conversion device, pre-setting device, self-service device.

3.5
bunker(s)

fuel supplied to a *vessel* (3.47) for its propulsion and/or operation

Note 1 to entry: Fuel with reference to Class F of ISO 8217.

3.6
bunker delivery note

BDN

proprietary document of the *bunker supplier* (3.9) providing details of the quality and quantity of the *bunker(s)* (3.5) delivered by the *bunker tanker* (3.11) to the *vessel* (3.47)

3.7
bunker metering ticket

BMT

ticket printed at the end of a *bunkering operation* (3.8)

3.8
bunkering operation

bunker delivery from *bunker tanker* (3.11) to *vessel* (3.47)

3.9
bunker supplier

company which contractually agrees with the buyer to deliver the product

3.10
bunker surveyor

person who inspects, measures, samples, investigates and reports as required on the bunkering operations

3.11
bunker tanker

bunker tanker supplying bunker(s) to the *vessel* (3.47)

3.12
bunker tanker operator

company which operates the *bunker tanker* (3.11)

3.13
bunker tanker representative

individual who represents the *bunker supplier* (3.9) and is responsible for *bunkering operations* (3.8) and documentations

3.14 calibration

operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and, in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

[SOURCE: JCGM 200]

3.15 calibration factor

numerical factor unique to each sensor derived during sensor *calibration* (3.14), which when programmed into the *transmitter* (3.44) ensures that the meter performs to its stated specification

[SOURCE: ISO 10790:2015,3.1.10, modified — the term has been changed from "calibrating factor" to "calibration factor" and Note 1 to entry has been merged in the definition.]

3.16 chief engineer

engineer of the *vessel* (3.47) who is responsible for receiving bunkers and documentation of the *bunkering operation* (3.8)

3.17 commissioning

process whereby the critical precision parameters impacting custody transfer are verified and checked

Note 1 to entry: Any setting changes during commissioning or re-commissioning is traceable to factory settings and justified *adjustments* (3.2) to meet the *measurement uncertainty* (3.30) or type classification.

3.18 conformity body

independent party or party accredited by national body that undertake conformity assessment activities such as verification, testing, inspection and certification

3.19 conventional mass

mass value of a body equal to the *mass* (3.24) of a standard that balances this body under conventionally chosen conditions

Note 1 to entry: The unit of a conventional mass is the kilogram. It is also known as mass in air.

[SOURCE: OIML D028]

3.20 custody transfer point

point at which, the *bunker* (3.5) is defined as being delivered or loaded

3.21 initial zero adjustment

setting of the indication of *mass flow rate* (3.27) to zero with the flowrate completely stopped and to fully filled flow meter according to approved procedure, before it is ready for custody transfer usage

3.22 linearity of MFM linearity of mass flow meter

consistency of change in the scaled output of a Coriolis flow meter, for a related, scaled change in the input of the flow meter

[SOURCE: ASME MFC-11]

3.23

low flow cut-off

transmitter (3.44) setting which sets the meter output(s) to zero flow if the flow rate falls below a preset value

Note 1 to entry: This setting inhibits the registration of flow when the flow meter is not properly filled with subject fluid that can lead to large measurement errors.

3.24

mass

physical quantity which can be ascribed to any material object and which gives a measure of its quantity of matter

Note 1 to entry: Also known as true mass.

[SOURCE: OIML D028, modified — Note 1 to entry has been added.]

3.25

mass flow meter

MFM

device consisting of a flow sensor (primary device) and a *transmitter* (3.44) (secondary device) which primarily measures the mass flow by means of the interaction between a flowing fluid and the oscillation of a tube or tubes; it may also provide measurements of the density and the process temperature of the fluid

3.26

mass flow meter system

MFM system

system that comprises the *mass flow meter* (3.25), its *ancillary devices* (3.4), pipelines and sealing points between the pump suction and the *custody transfer point* (3.20)

3.27

mass flow rate

flow rate at which the quantity of fluid which passes the *MFM* (3.25) is expressed as mass and denoted in MT/h

3.28

master

person in charge of the *bunker tanker* (3.11) or the vessel receiving bunker(s) as the case can be

3.29

maximum mass flow rate

Q_{\max}

maximum flow rate, up to which, the *MFM system* (3.26) has been qualified to operate in compliance with the required *accuracy* (3.1) Note 1 to entry: The maximum value is normally determined by the application

3.30

measurement uncertainty

non-negative parameter characterizing the dispersion of the quantity values being attributed to a measurand, based on the information used

[SOURCE: JCGM 200]

3.31

meter reading

value obtained from the *non-resettable totalizer(s)* (3.37)

3.32**meter stability**

property of a measuring instrument, whereby, its metrological properties remain constant over time

Note 1 to entry: Stability may be quantified in several ways:

- in terms of the duration of a time interval over which a metrological property changes by a stated amount;
- in terms of the change of a property over a stated time.

[SOURCE: JCGM 200]

3.33**metering**

measurement of quantity by the *MFM system* (3.26)

3.34**metering profile**

graphical overview of the process parameters recorded during a *bunkering operation* (3.8) and retained for purpose of providing transparent assessment

3.35**minimum mass flow rate**

Q_{\min}

lowest flow rate required to which the metering system has been qualified to operate, in compliance with the required *accuracy* (3.1)

Note 1 to entry: The minimum value is normally determined by the flow metering system.

3.36**minimum****MMQ**

smallest quantity of liquid for which the measurement is metrologically acceptable for that system or element

measured

quantity

3.37**non-resettable totalizer**

device that indicates the total cumulated flow quantity through the *MFM* (3.25) after it is secured for use in custody transfer such that its value is not resettable to zero or to other values

3.38**quantity delivered**

cumulative mass quantity measured between the start of delivery and end of delivery and transferred to the *vessel* (3.47)

3.39**repeatability**

proximity of a match among a series of results obtained with the same method on identical test material, under the same conditions (same operator, same apparatus, same laboratory and short intervals of time)

3.40**resettable totalizer**

device that indicates total flow quantity through the *MFM* (3.25) from the start to the end of each batch and its value can be reset to zero

3.41**sample**

bunker (3.5) specimen defined by time, location and method of sampling

3.42

stored zero value

value stored in the electronics after the zero-adjustment procedure

Note 1 to entry: Stored zero value is recorded during every zero-offset determination. Depending on manufacturer, the stored zero value can be in flow rate units or in time units or in % units.

3.43

traceability

metrological property of a measurement result, whereby the result can be related to a reference through a documented unbroken chain of calibrations verified by a national metrology institute, each contributing to the *measurement uncertainty* (3.30)

[SOURCE: JCGM 200]

3.44

transmitter

electronic control system that provides the drive and transforming the signals from the flow sensor, to give output(s) of measured and inferred parameters, and that also provides corrections derived from parameters such as temperature

3.45

update

installation of new system components, hardware or software, which have no significant effect on the metering result

Note 1 to entry: No testing is required after installation.

3.46

upgrade

installation of new system components, hardware or software, which can have a significant effect on the metering result

Note 1 to entry: New certification testing is required after installation.

3.47

vessel

ship that receives the *bunker(s)* (3.5)

3.48

zero offset

measurement output indicated under zero flow conditions

Note 1 to entry: A zero offset might be caused by stress being applied to the oscillating tubes by the surrounding pipework and by process conditions.

Note 2 to entry: A zero offset can be reduced by means of a zero-adjustment procedure.

3.49

zero offset limit

maximum allowable observed *zero offset* (3.48) in relation to the stored zero value, used to determine when to re-zero the flow meter, generally defined by the manufacturer

[SOURCE: API MPMS 5.6]

3.50**zero stability**

magnitude of the meter output deviation from the stored zero value at zero flow after the zero adjustment procedure has been completed, expressed by the manufacturer as an absolute value in mass per unit time

Note 1 to entry: The stated value for zero stability is valid for stable conditions where the fluid is free of bubbles and sediment.

3.51**zero verification procedure**

procedure to verify that the *zero offset* (3.48) does not exceed the *zero offset limit* (3.49)

4 General requirements (safety, health and the environment)

4.1 The requirements to be observed by all personnel for the safe transfer of bunker in port are set out in [Annex A](#). Internationally accepted safety standards, as appropriate, shall also be observed by the personnel of both the bunker tanker and the vessel and also the bunker surveyor (when engaged) for the safe transfer of bunkers in port.

NOTE Local requirements can also apply.

4.2 The respective masters of the bunker tanker and the vessel shall remain responsible for the safety of their vessel, crew, cargo and equipment at all times and should not permit safety to be prejudiced by the actions of others.

4.3 All parties involved in the bunkering processes shall equip themselves with the following minimum safety items:

- safety helmet;
- safety shoes;
- gloves;
- life jacket.

They shall wear personal protective equipment at all times while on board the vessel and the bunker tanker. They shall equip themselves with H₂S and O₂ monitors and use them throughout the operation.

4.4 All parties involved in the bunkering operation shall be free from the influence of any alcohol, drugs or other substances which impairs the safe and efficient execution of their work and personal health.

5 Metrological requirements**5.1 General**

[Clause 5](#) specifies the MFM's metrological traceability, calibration and re-calibration requirements for the approval and performance of the MFM system applicable to custody transfer bunkering. The MFM system shall be operated within rated conditions as set out in these requirements to meet the 0,5 % expanded measurement uncertainty.

5.2 Mass flow meter requirement

5.2.1 Every MFM shall be calibrated before custody transfer use for bunkering and shall include its adjustment device(s) and ancillary device(s).