
**Active implantable medical devices —
Four-pole connector system for
implantable cardiac rhythm
management devices — Dimensional
and test requirements**

*Dispositifs médicaux actifs implantables — Systèmes de branchement
à quatre pôles pour dispositifs implantables de gestion du rythme
cardiaque — Exigences de dimensions et d'essai*

(<https://standards.iteh.ai>)
Document Preview

ISO 27186:2020

<https://standards.iteh.ai/catalog/standards/iso/3ab0d701-601e-4a34-b2be-c379d01d1099/iso-27186-2020>



Reference number
ISO 27186:2020(E)

© ISO 2020

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 27186:2020

<https://standards.iteh.ai/catalog/standards/iso/3ab0d701-601e-4a34-b2be-c379d01d1099/iso-27186-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	5
4.1 General	5
4.2 Lead connector physical requirements	5
4.2.1 Dimensions	5
4.2.2 Materials	10
4.2.3 Lead connector electrical connections	10
4.2.4 Lead marking	10
4.2.5 Lead package labels and literature	12
4.3 Lead connector functional requirements	12
4.3.1 Functional dimensional check	12
4.3.2 Tensile loads	13
4.3.3 Deformation due to pin contact forces	13
4.3.4 Deformation due to ring contact forces	13
4.3.5 Seal zone requirement	14
4.3.6 Electrical isolation requirement	14
4.3.7 Dielectric strength requirement	15
4.3.8 Current-carrying requirement	15
4.3.9 Corrosion/environmental	15
4.4 Connector cavity physical requirements	15
4.4.1 Dimensions	15
4.4.2 Connector cavity electrical connections	18
4.4.3 Connector cavity/pulse generator marking	18
4.4.4 Pulse generator labels and literature	19
4.5 Connector cavity functional requirements	19
4.5.1 Insertion force	19
4.5.2 Retention force	20
4.5.3 Withdrawal force	21
4.5.4 Ring contact load	21
4.5.5 Seal zone load requirement	22
4.5.6 Electrical isolation requirement	22
4.5.7 Dielectric strength requirement	22
4.5.8 Current-carrying requirement (high-voltage connector cavity)	22
4.5.9 Contact resistance/stability	22
Annex A (normative) Electrical isolation test	23
Annex B (informative) Rational for Annex A	28
Annex C (normative) Dielectric strength test	30
Annex D (informative) Rational for Annex C	35
Annex E (normative) Current-carrying test high-voltage types	39
Annex F (informative) Rational for Annex E	44
Annex G (informative) Lead connector fatigue strength test	46
Annex H (informative) Lead connector seal zone materials	47
Annex I (informative) Seal zone creep	49
Annex J (informative) Contact resistance stability	54

Annex K (informative) Rational for Annex J	58
Annex L (informative) Selection of contact materials	60
Annex M (normative) Lead connector contact material requirements	62
Annex N (informative) Rational for Annex M	66
Annex O (informative) Rationale for requirements in this document	72
Annex P (informative) Connector products	79
Bibliography	81

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 27186:2020](#)

<https://standards.iteh.ai/catalog/standards/iso/3ab0d701-601e-4a34-b2be-c379d01d1099/iso-27186-2020>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared jointly by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 6, *Active implants*, and Technical Committee IEC/SC 62D, *Electromedical equipment*. The draft was circulated for voting to the national bodies of both ISO and IEC.

This second edition cancels and replaces the first edition (ISO 27186:2010), which has been technically revised.

The main changes compared to the previous edition are as follows:

- minor typographical errors have been corrected;
- the notch feature on lead connector pins has been made optional whereas previously it was required;
- the use of the notch feature for retention is no longer permitted;
- a clarification has been made to verify the functional sealing and functional contact zone requirements in [4.4.1.2](#) and [4.4.1.3](#).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The four-pole connector was created to reduce the number of individual lead connectors, to reduce pocket bulk associated with existing bifurcated or trifurcated leads, to reduce interaction of the lead bodies in the pocket and to reduce set screw connections.

The intent of this document is to define a four-pole connector assembly that provides interchangeability between implantable leads and pulse generators from different manufacturers.

This document establishes two types of connector assembly: a “high-voltage connector” and a “low-voltage only connector”, each of which has several configurations. The high-voltage connectors either have two low-voltage contacts combined with one or two high-voltage contacts, or they have only two high-voltage contacts. The low-voltage only connectors have either three or four low-voltage contacts.

The high-voltage and low-voltage only connectors and their voltage configurations are not intended to be interchangeable. This document specifies a dimensional lockout feature that prevents the low-voltage contacts of the lead connectors from contacting the high-voltage contacts of high-voltage connector cavities.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 27186:2020](https://standards.iteh.ai/catalog/standards/iso/3ab0d701-601e-4a34-b2be-c379d01d1099/iso-27186-2020)

<https://standards.iteh.ai/catalog/standards/iso/3ab0d701-601e-4a34-b2be-c379d01d1099/iso-27186-2020>

Active implantable medical devices — Four-pole connector system for implantable cardiac rhythm management devices — Dimensional and test requirements

WARNING — The low-voltage only connector cavity specified in this document is not to be used if the implantable pulse generator is capable of introducing dangerous non-pacing stimuli (e.g. defibrillation shocks) through the contacts of that connector cavity. Likewise, the high-voltage lead connector specified in this document is not to be used on leads intended for low-voltage only therapy.

1 Scope

This document specifies a four-pole connector system for implantable cardiac rhythm management (CRM) devices which have pacing, electrogram sensing and/or defibrillation functions. This document includes requirements for the connector portion of an implantable lead as well as for the mating connector cavity attached to an implantable pulse generator. Essential dimensions and performance requirements are specified together with appropriate test methods.

NOTE The safety, reliability, biocompatibility, biostability and function of any particular part are the responsibility of the manufacturer.

This document is not intended to replace or provide alternatives for unipolar or bipolar connector standards that currently exist (such as ISO 11318 and ISO 5841-3).

This document is not applicable to high-voltage systems with intended outputs greater than 1 000 V and/or 50 A. This document is not applicable to systems which include sensors or unique electrodes that are not capable of conventional pacing, electrogram sensing and/or defibrillation functions.

This document does not specify all connector features.

This document does not address all aspects of functional compatibility, safety or reliability of leads and pulse generators assembled into a system.

NOTE Lead and pulse generator connector systems not conforming to this document can be safe and reliable and can have clinical advantages.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7436, *Slotted set screws with cup point*

ASTM B348, *Standard Specification for Titanium and Titanium Alloy Bars and Billets*

ASTM F562, *Standard Specification for Wrought 35Cobalt-35Nickel-20Chromium-10Molybdenum Alloy for Surgical Implant Applications*

ASTM F746-04, *Standard Test Method for Pitting or Crevice Corrosion of Metallic Surgical Implant Materials*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1 axial pin movement

axial movement of a *lead connector pin* (3.18) with reference to the *lead connector* (3.16) body as present in some designs, particularly those with a rotating connector pin

3.2 bipolar

having two poles or electrodes

Note 1 to entry: See also *tripolar* (3.31), *integrated bipolar* (3.15), and *four-pole* (3.8).

3.3 connector system

assembly consisting of a *lead connector* (3.16) and a *connector cavity* (3.4) that are electrically and mechanically joined

3.4 connector cavity

cavity within the *pulse generator* (3.27) which is intended to receive a *lead connector* (3.16)

3.5 contact mechanism

conductive hardware within the *connector cavity* (3.4) provided for making electrical connection to corresponding contacts on a *lead connector* (3.16)

3.6 distal

farthest from a point of reference

Note 1 to entry: The point of reference for a lead is the *lead connector pin* (3.18). Therefore, the most distal electrode of a lead is the electrode that is farthest from the lead connector pin. See also *proximal* (3.26).

3.7 fixation zone

zone located on the *lead connector pin* (3.18) and within the *connector cavity* (3.4) where the *lead connector* (3.16) is mechanically secured within the connector cavity

3.8 four-pole

having four poles or electrodes

Note 1 to entry: Generally a four-pole ICD lead has two *low-voltage* (3.21) electrodes and two *high-voltage* (3.12) electrodes. A four-pole low-voltage only lead has four low-voltage electrodes. See also *bipolar* (3.2) and *tripolar* (3.31).

3.9 functional contact zone

zone in the *connector cavity* (3.4) where electrical contact with a *lead connector* (3.16) occurs

3.10 functional seal zone

zone within the *connector cavity* (3.4) where sealing contact with a *lead connector* (3.16) occurs

3.11**grip zone**

area of the *lead connector* (3.16) which is provided for grasping during insertion and withdrawal of the lead connector from the *connector cavity* (3.4)

3.12**high-voltage**

electrical potential greater than 20 V and that can go up to 1 000 V

Note 1 to entry: High-voltages are generally used for defibrillating the heart.

3.13**high-voltage connector**

lead connector (3.16) or *connector cavity* (3.4) that has *high-voltage* (3.12) contacts

Note 1 to entry: A high-voltage connector may also contain *low-voltage* (3.21) contacts. See also *low-voltage only connector* (3.22).

3.14**insertion indicator zone**

zone on the pin of the *lead connector* (3.16) allocated for manufacturers to provide a visual indicator for use in verifying full insertion of a lead connector into a *connector cavity* (3.4)

3.15**integrated bipolar**

two lead poles or *lead electrodes* (3.20) that are electrically common

Note 1 to entry: A typical integrated bipolar ICD lead has a *distal* (3.6) shock electrode that doubles as a *proximal* (3.26) pace/sense ring electrode and is electrically attached to two separate *lead connector contacts* (3.17).

3.16**lead connector**

part of a lead that is intended for insertion into the *connector cavity* (3.4) of a *pulse generator* (3.27)

3.17**lead connector contacts**

conductive elements on the *lead connector* (3.16) which include the *lead connector pin* (3.18) and *lead connector rings* (3.19)

3.18**lead connector pin**

most *proximal* (3.26) conductive element of a *lead connector* (3.16) provided for making electrical contact as well as for securing the lead connector within the *connector cavity* (3.4)

3.19**lead connector ring**

annular conductive element on the *lead connector* (3.16) intended for making electrical contact within the *connector cavity* (3.4)

Note 1 to entry: The *four-pole* (3.8) connector has three lead connector rings and a *lead connector pin* (3.18).

3.20**lead electrode**

distal (3.6) part of a lead through which electrical impulses are transmitted to or from cardiac tissue

Note 1 to entry: *High-voltage* (3.12) electrodes are capable of delivering high-voltage electrical impulses. *Low-voltage* (3.21) electrodes are used for transmitting and sensing low-voltage impulses and are generally not suitable for delivering high-voltage.

3.21

low-voltage

electrical potential less than or equal to 20 volts

Note 1 to entry: Low-voltage is generally used for pacing and sensing the heart. See also *high-voltage* (3.12).

3.22

low-voltage only connector

lead connector (3.16) or *connector cavity* (3.4) that has only *low-voltage* (3.21) contacts

Note 1 to entry: See also *high-voltage connector* (3.13).

3.23

pin visibility zone

zone within the *connector cavity* (3.4) which is allocated for visual verification that the *lead connector* (3.16) is fully inserted

Note 1 to entry: It corresponds to the *insertion indicator zone* (3.14) of the lead connector.

3.24

pristine contact zone

zone on the *lead connector* (3.16) which defines the minimum surface required for making electrical contact with the mating contact in the *connector cavity* (3.4)

Note 1 to entry: The pristine contact zones of the lead connector align with the *functional contact zones* (3.9) of the *connector cavity* (3.4) when the connectors are mated.

3.25

pristine seal zone

zone on the *lead connector* (3.16) which defines the minimum surface required for sealing with the mating seals in the *connector cavity* (3.4)

Note 1 to entry: The pristine seal zones of the lead connector align with the *functional seal zones* (3.10) of the *connector cavity* (3.4) when the connectors are mated.

3.26

proximal

nearest to a point of reference

Note 1 to entry: The point of reference for a lead is the *lead connector pin* (3.18). Therefore, the most proximal electrode of a lead is the electrode closest to the lead connector pin. See also *distal* (3.6).

3.27

pulse generator

device that delivers electrical energy to affect cardiac rhythms

3.28

sealing mechanism

circumferential barriers within the *connector cavity* (3.4) intended to maintain electrical isolation between electrically insulated parts of an assembled and implanted *connector system* (3.3)

3.29

securing mechanism

mechanism within the *connector cavity* (3.4) intended for mechanically securing the *lead connector* (3.16), typically a set screw

3.30**strain relief zone**

zone on the *lead connector* (3.16) provided for making a gradual transition from a more rigid section to a more flexible section

Note 1 to entry: The gradual transition results in an area over which strain is distributed so that concentrated mechanical forces do not occur when the lead is flexed.

3.31**tripolar**

three poles or electrodes

Note 1 to entry: See also *bipolar* (3.2) and *four-pole* (3.8).

4 Requirements**4.1 General**

Not all connector features or pulse generator features are specified nor do the requirements in this document address all aspects of functional compatibility, safety or reliability of leads and pulse generators assembled into a system. Each manufacturer is responsible for any requirements and tests necessary to address these as well as the biocompatibility and biostability of their material choices.

The test methods provided for the requirements are type (qualification) tests and are not intended to be used as routine production tests. Alternate test methods may be used, including those which result in equivalent or more stringent test conditions. However, in the event of dispute, the test methods described in this document shall be used.

The following tests should be conducted under ambient conditions unless otherwise specified. Each manufacturer is responsible for any preconditioning required to represent “as-shipped” configurations, as well as for selection of appropriate sample sizes.

Leads and pulse generators marked according to [Table 1](#) and [Table 2](#) shall comply with all requirements in this document.

4.2 Lead connector physical requirements**4.2.1 Dimensions****4.2.1.1 General**

Lead connectors shall have the dimensions specified in [Figure 1](#) and [Figure 2](#) and shall meet the requirements outlined in [4.2.1.2](#) to [4.2.1.11](#) according to each zone. See [Annex Q](#) for rationale.

4.2.1.2 Total axial pin movement, M

Total axial pin movement is the difference in lead connector pin length from when the connector pin is fully seated against datum A to when the connector pin is fully extended from datum A. Total axial pin movement shall not be greater than 0,25 mm.

4.2.1.3 Pristine contact zones

The minimum length of each of the pristine contact zones shall be 0,90 mm + M , where M is the total axial pin movement in millimetres.

Lead connectors shall have an electrically conductive contact surface over the entire length of each of the pristine contact zones. Contact surfaces may extend beyond the pristine contact zones.

The surface finish in these zones shall be Ra 0,8 μm maximum. The entire surface area shall be considered when measuring surface finish. No indentations, protrusions, gaps or steps exceeding surface finish allowance are allowed in these zones.

4.2.1.4 Pristine seal zones

The minimum length of each pristine seal zone shall be 1,81 mm + M , where M is the total axial pin movement in millimetres.

Lead connectors shall have a seal surface over the entire length of each of the pristine seal zones. Seal surfaces may extend beyond the pristine seal zones. No indentation, protrusions, gaps or steps exceeding surface finish allowance are allowed in these zones.

For surfaces of materials with hardness 75D or less, the surface finish in this zone shall be Ra 0,8 μm maximum. The entire surface shall be considered when measuring surface finish except that uniform linear protrusions, such as caused by mould parting lines, may be excluded from the measurement if they do not exceed 0,025 mm in height as measured radially or 0,12 mm in width.

For surfaces in this zone made from materials with hardness above 75D, the surface finish shall be Ra 0,4 μm maximum when the entire surface is considered, including any uniform linear protrusions.

4.2.1.5 Lead connector body

The diameter for all conductive components and surfaces within this zone shall be 3,2 mm \pm 0,03 mm.

The diameter for all non-conductive components and surfaces within this zone shall be 3,2 mm \pm 0,05 mm.

For all areas in this zone except pristine seal zones and pristine contact zones, the following requirements apply.

- a) Any radial steps or protrusions, such as can occur between two adjacent components or by welds, shall not exceed 0,05 mm (in height) and shall not cause the diameter to go outside the tolerance specified with the following exception. Uniform linear protrusions that do not exceed 0,025 mm in height as measured radially or 0,12 mm in width are allowed only for surfaces of materials that are at 75 Shore D or below.
- b) Any gap shall not exceed 0,1 mm in width when measured to include all edge breaks at the gap edge. There shall not be more than one gap between each pristine zone. For any gap that meets these requirements, the area within the gap need not meet the other requirements of this subclause, for example diameter and radial step requirements.
- c) Any indentations, such as holes or weld depressions, shall not exceed 0,5 mm in diameter.
- d) Surface finish for zone 3 of [Figure 1](#) shall be Ra 0,8 μm maximum from datum A to 16,04 mm and shall be Ra 1,6 μm maximum from 16,04 mm to the transition zone 7, excluding the chamfer area. Surface finish measurements need not include any surface features that meet the above requirements [4.2.1.5 a\) to c\)](#).

4.2.1.6 Strain relief zone

The diameter in this zone shall be 4,1 mm maximum and 3,8 mm minimum.

4.2.1.7 Grip zone

The diameter in this zone shall be 4,3 mm maximum.

4.2.1.8 Chamfer zone

The length of the chamfer in this zone shall be 0,35 mm minimally and 0,7 mm – M maximally, where M is the total axial pin movement in millimetres.

4.2.1.9 Transition zone

The transition between the Ø3,2 mm nominal and the Ø4,1 mm maximum shall occur within the theoretical envelope between datum B, which intersects the Ø4,1 mm at the 17,7 mm dimension, and Datum C.

The diameter in this zone shall not exceed 4,1 mm.

NOTE The 60° dimension defines datum B; transition geometry does not need to match this angle.

4.2.1.10 Insertion indicator zone

This zone is provided for an optional insertion indicator. If an insertion indicator is present it shall meet the following requirements.


- a) The indicator shall not extend beyond the zone.
- b) The proximal edge shall meet the 5,10 mm ± 0,10 mm dimension.
- c) The diameter shall fall within the nominal diameter specified and the tolerance of $\frac{+0,03}{-0,10}$ mm.
- d) Any gaps shall not exceed 0,10 mm in width. For any gap that meets this requirement, the area within the gap need not meet the other requirements of this section, for example diameter and radial step requirements.
- e) Any radial steps shall not exceed 0,05 mm.

4.2.1.11 Pin pristine contact zone

Lead connectors shall have an electrical contact surface over the entire length of this zone.

The surface finish in this zone shall be Ra 0,8 µm maximum.

Dimensions in millimetres



Key

- | | | | |
|---|------------------------|---|-------------------------------|
| 1 | pristine contact zones | 5 | grip zone |
| 2 | pristine seal zones | 6 | chamfer zone |
| 3 | lead connector body | 7 | transition zone |
| 4 | strain relief zone | 8 | total axial pin movement, M |

NOTE The diameter dimensions of the soft sections, in zone 4, zone 5 and zone 7, of the lead can be determined as the mean value of three measurements taken at locations oriented approximately 120° apart around the principal axis of the lead connector.

Figure 1 — Four-pole lead connector body