
**Non-destructive testing —
Characterization and verification of
ultrasonic test equipment —**

**Part 3:
Combined equipment**

*Essais non destructifs — Caractérisation et vérification de
l'appareillage de contrôle par ultrasons —*

Partie 3: Equipement complet

Document Preview

ISO 22232-3:2020

<https://standards.iteh.ai/catalog/standards/iso/7a7d09d4-851d-465e-956f-e6dc699442d0/iso-22232-3-2020>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 22232-3:2020

<https://standards.iteh.ai/catalog/standards/iso/7a7d09d4-851d-465e-956f-e6dc699442d0/iso-22232-3-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General requirements for conformity	2
5 Personnel qualification	2
6 Description of tests and reporting	2
6.1 Baseline measurements of characteristic values.....	2
6.2 Physical state and external aspects.....	3
6.2.1 Procedure.....	3
6.2.2 Acceptance criterion.....	3
6.2.3 Frequency of testing.....	3
6.3 Tests for angle-beam probes.....	3
6.3.1 General.....	3
6.3.2 Probe index point.....	3
6.3.3 Beam angle.....	4
6.3.4 Simultaneous determination of probe index point and beam angle.....	4
6.4 Vertical linearity.....	6
6.4.1 General.....	6
6.4.2 Procedure.....	6
6.4.3 Acceptance criteria.....	7
6.4.4 Frequency of testing.....	7
6.5 Sensitivity and signal-to-noise ratio.....	7
6.5.1 General.....	7
6.5.2 Procedure.....	8
6.5.3 Acceptance criterion.....	8
6.5.4 Frequency of testing.....	8
6.6 Pulse duration.....	8
6.6.1 General.....	8
6.6.2 Procedure.....	8
6.6.3 Acceptance criterion.....	9
6.6.4 Frequency of testing.....	9
Bibliography	10

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by ISO/TC 135, *Non-destructive testing*, Subcommittee SC 3, *Ultrasonic testing*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 138, *Non-destructive testing*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 22232 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Non-destructive testing — Characterization and verification of ultrasonic test equipment —

Part 3: Combined equipment

1 Scope

This document specifies methods, tolerances and acceptance criteria for verifying the performance of combined ultrasonic test equipment (i.e. instrument, probes and cables connected) by the use of appropriate standard calibration blocks.

These methods are specifically intended for manual test equipment, i.e. ultrasonic instruments according to ISO 22232-1, and for manual ultrasonic non-destructive testing with single- or dual-transducer probes according to ISO 22232-2. This document is also applicable for multi-channel instruments. For automated test equipment, different tests can be needed to ensure satisfactory performance.

The specified methods are intended for the use by operators working under site or shop floor conditions.

These methods are not intended to prove the suitability of the equipment for particular applications.

This document excludes ultrasonic instruments for continuous waves.

This document also excludes ultrasonic phased array systems, see e. g. ISO 18563-3. If a phased array instrument is used in combination with single- or dual-transducer probes, this document is applicable to this combination.

<https://standards.iteh.ai/catalog/standards/iso/7a7d09d4-851d-465e-956f-e6dc699442d0/iso-22232-3-2020>

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2400, *Non-destructive testing — Ultrasonic testing — Specification for calibration block No. 1*

ISO 5577, *Non-destructive testing — Ultrasonic testing — Vocabulary*

ISO 7963, *Non-destructive testing — Ultrasonic testing — Specification for calibration block No. 2*

ISO 9712, *Non-destructive testing — Qualification and certification of NDT personnel*

ISO 22232-1, *Non-destructive testing — Characterization and verification of ultrasonic test equipment — Part 1: Instruments*

ISO 22232-2, *Non-destructive testing — Characterization and verification of ultrasonic test equipment — Part 2: Probes*

3 Terms and definitions

For the purpose of this document, the terms and definitions given in ISO 5577 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

4 General requirements for conformity

The combined equipment complies with this document if it fulfils all of the following requirements:

- the ultrasonic instrument shall comply with ISO 22232-1;
- the probe shall comply with ISO 22232-2;
- the combined equipment shall comply with [Clause 6](#);
- the results of the baseline tests, [6.1](#), shall be reported including the instruments settings of the relevant parameters used for the baseline measurements.

The tests specified in this document, together with the frequency of testing, are summarized in [Table 1](#).

Table 1 — Tests to be performed

Title of test	Frequency of testing	Subclause
Physical state and external aspects	Daily	6.2
Probe index point	Daily	6.3.2/6.3.4
Beam angle	Daily	6.3.3/6.3.4
Vertical linearity	Weekly ^a	6.4
Sensitivity and signal-to-noise ratio	Weekly ^a	6.5
Pulse duration	Weekly ^a	6.6

^a It may be more convenient for the user to perform the weekly tests each time the equipment is used.

Non-compliance with the requirements in this document shall result in replacement, repair and/or verification according to ISO 22232-1 Group 2 tests or to ISO 22232-2 for the affected component.

5 Personnel qualification

Personnel performing the verifications in accordance with this document shall be qualified to an appropriate level in ultrasonic testing in accordance with ISO 9712 or equivalent.

6 Description of tests and reporting

6.1 Baseline measurements of characteristic values

For each set of combined equipment (instrument, cables and probe), base values shall be determined and reported. The later measured values are to be compared against the base values.

For angle-beam probes, initially the user shall establish base values for the probe index point ([6.3.2.2/6.3.4.2](#)) and the beam angle ([6.3.3.2/6.3.4.2](#)), unless these values are available for new probes.

For all systems, initially the user shall establish base values for the sensitivity, the signal-to-noise ratio and the pulse duration using the methods given in [6.5.2](#) and [6.6.2](#). These shall either be measured for the actual probe and instrument to be used for subsequent testing or for each combination of probe type and instrument type to be used.