
**Respiratory protective devices —
Human factors —**

Part 3:

**Physiological responses and
limitations of oxygen and limitations
of carbon dioxide in the breathing
environment**

iTech Standards
(<https://standards.iteh.ai>)

Appareils de protection respiratoire — Facteurs humains —

Document Precieux
*Partie 3: Réponses physiologiques et limitations en oxygène et en gaz
carbonique dans l'environnement respiratoire*

[ISO/TS 16976-3:2019](#)

<https://standards.iteh.ai/catalog/standards/iso/dc066132-2f2b-4408-a4e7-4169fe823518/iso-ts-16976-3-2019>



Reference number
ISO/TS 16976-3:2019(E)

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/TS 16976-3:2019](#)

<https://standards.iteh.ai/catalog/standards/iso/dc066132-2f2b-4408-a4e7-4169fe823518/iso-ts-16976-3-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols and abbreviated terms	5
5 Oxygen and carbon dioxide in the breathing environment: Physiological responses and limitations	5
5.1 General	5
5.2 Oxygen and carbon dioxide gas exchange in the human lung	6
5.3 Oxygen and carbon dioxide transport in the blood	6
5.4 Oxygen and carbon dioxide and the control of respiration	8
5.5 Hyperoxia: physiological effects	9
5.6 Hypoxia: physiological effects	10
5.7 Hypercarbia: Physiological effects	13
5.8 Relevance to the use of respiratory protective devices (RPD)	16
5.9 Interpretation of results	20
5.10 Significance of results	21
Bibliography	22

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO/TS 16976-3:2019](#)

<https://standards.iteh.ai/catalog/standards/iso/dc066132-2f2b-4408-a4e7-4169fe823518/iso-ts-16976-3-2019>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 94, *Personal safety — Personal protective equipment*, Subcommittee SC 15, *Respiratory protective devices*.

This second edition cancels and replaces the first edition (ISO/TS 16976-3:2011), which has been technically revised. The main changes compared to the previous edition are as follows:

- adding a new [Table 3](#) to give limits for CO₂ when breathing against imposed breathing resistance;
- explanation of the new limits for CO₂ derived from a study.

A list of all parts in the ISO/TS 16976 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Due to the nature of their occupations, millions of workers worldwide should wear respiratory protective devices (RPD). RPD vary considerably, from filtering devices, supplied breathable gas devices, and underwater breathing apparatus (UBA), to escape respirators used in emergency situations (self-contained self-rescuer or SCSR). Many of these devices protect against airborne contaminants without supplying air or other breathing gas mixtures to the user. Therefore, the user might be protected from particulates or other airborne toxins but still be exposed to an ambient gas mixture that differs significantly from that which is normally found at sea level. RPD that supply breathing air to the user, such as an SCBA or UBA, can malfunction or not adequately remove carbon dioxide from the breathing space, thus exposing the user to an altered breathing gas environment. In special cases, RPD intentionally expose the wearer to breathing gas mixtures that significantly differ from the normal atmospheric gas mixture of approximately 79 % nitrogen and 21 % oxygen with additional trace gases. These special circumstances occur in aviation, commercial and military diving, and in clinical settings.

Breathing gas mixtures that differ from normal atmospheric can have significant effects on most physiological systems. Many of the physiological responses to exposure to high or low levels of either oxygen or carbon dioxide can have a profound effect on the ability to work safely, to escape from a dangerous situation, and to make clear judgements about the environmental dangers. In addition, alteration of the breathing gas environment can, if severe enough, be dangerous or even fatal. Therefore, monitoring and controlling the breathing gas, and limiting user exposure to variations in the concentration or partial pressure of oxygen and carbon dioxide, is crucial to the safety and health of the worker.

This document discusses the gas composition of the Earth's atmosphere; the basic physiology of metabolism as the origin of carbon dioxide in the body, respiratory physiology and the transport of oxygen to the cells and tissues of the body; and the subsequent transport of carbon dioxide from the tissues to the lungs for removal from the body. Following the basic physiology of respiration, this document addresses the physiological responses to altered breathing environments (hyperoxia, hypoxia) and to the effects of excess carbon dioxide in the blood (hypercarbia). Examples are given from the relevant biomedical literature.

Finally, it deals with the impact of altered partial pressures/concentrations of oxygen and carbon dioxide on respirator use. The content of this Document is intended to serve as the basis for advancing research and development of RPD with the aim of minimizing the changes in the breathing environment, thus minimizing the physiological impact of RPD use on the wearer. If this can be accomplished, the health and safety of all workers recommended by their occupation to wear RPD will be enhanced.

Respiratory protective devices — Human factors —

Part 3:

Physiological responses and limitations of oxygen and limitations of carbon dioxide in the breathing environment

1 Scope

This document gives:

- a description of the composition of the Earth's atmosphere;
- a description of the physiology of human respiration;
- a survey of the current biomedical literature on the effects of carbon dioxide and oxygen on human physiology;
- examples of environmental circumstances where the partial pressure of oxygen or carbon dioxide can vary from that found at sea level.

This document identifies oxygen and carbon dioxide concentration limit values and the length of time within which they would not be expected to impose physiological distress. To adequately illustrate the effects on human physiology, this document addresses both high altitude exposures where low partial pressures are encountered and underwater diving, which involves conditions with high partial pressures. The use of respirators and various work rates during which RPD can be worn are also included.

[ISO/TS 16976-3:2019](https://www.iso.org/obp/obp.html#toc)

2 Normative references

<https://www.iso.org/obp/obp.html#toc>

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

alveoli

s. alveolus

terminal air sacs of the lungs in which respiratory gas exchange occurs between the alveolar air and the pulmonary capillary

Note 1 to entry: The alveoli are the anatomical and functional unit of the lungs.

3.2

ambient temperature pressure saturated

ATPS

standard condition for the expression of ventilation parameters related to expired air

Note 1 to entry: Actual ambient temperature and atmospheric pressure; saturated water pressure.

3.3

body temperature pressure saturated

BTPS

standard condition for the expression of ventilation parameters

Note 1 to entry: Body temperature (37 °C), atmospheric pressure 101,3 kPa (760 mmHg) and water vapour pressure (6,27 kPa) in saturated air.

3.4

carbaminohaemoglobin

HbCO₂

haemoglobin that has bound carbon dioxide at the tissue site for transport to the lungs

3.5

dead space

<anatomical> conducting regions of the pulmonary airways that do not contain alveoli and, therefore, where no gas exchange occurs

Note 1 to entry: These areas include the nose, mouth, trachea, large bronchia, and the lower branching airways. This volume is typically 150 ml in a male of average size.

3.6

(<https://standards.iteh.ai>)

dead space

<physiological> sum of all anatomical dead space as well as under-perfused (reduced blood flow) alveoli which are not participating in gas exchange

Note 1 to entry: The volume of the physiological dead space can vary with the degree of ventilation. Thus, the physiological dead space is the fraction of the tidal volume that does not participate in gas exchange in the lungs.

<https://standards.iteh.ai/catalog/standards/iso/dc066132-2f2b-4408-a4e7-4169fe823518/iso-ts-16976-3-2019>

3.7

dyspnoea

sense of air hunger, difficult or laboured breathing, or a sense of breathlessness

3.8

end-tidal carbon dioxide

ET CO₂

volume fraction of carbon dioxide in the breath at the mouth at the end of exhalation

Note 1 to entry: End-tidal carbon dioxide corresponds closely to alveolar carbon dioxide.

3.9

haemoglobin

Hb

specific molecules contained within all red blood cells that bind oxygen or carbon dioxide under normal physiological states and transport either oxygen or carbon dioxide to or from the tissues of the body

3.10

hypercarbia

hypercapnia

excess amount of carbon dioxide in the blood

3.11**hyperoxia**

volume fraction or partial pressure of oxygen in the breathing environment greater than that which is found in the Earth's atmosphere at sea level, which contributes to an excess of oxygen in the body

Note 1 to entry: This can occur when a person is under hyperbaric conditions (i.e. diving), subjected to breathing gas mixtures with an elevated oxygen fraction, or during certain medical procedures

3.12**hypoxia**

volume fraction or partial pressure of oxygen in the breathing environment below that which is found in the Earth's atmosphere at sea level

Note 1 to entry: Anaemic hypoxia is due to a reduction of the oxygen carrying capacity of the blood as a result of a decrease in the total haemoglobin or an alteration in the haemoglobin constituents.

3.13**hypocapnia**

volume fraction or partial pressure of carbon dioxide in the breathing environment or in the body that is lower than that which is found in the Earth's atmosphere at sea level

Note 1 to entry: This usually occurs under hyperventilation conditions (i.e. diving) or in medical settings that contribute to a reduction of carbon dioxide in the body.

3.14**inotropic**

affecting the force of muscle contraction

Note 1 to entry: A negative inotropic effect reduces and a positive inotropic effect increases the force of muscular contraction (e.g. both skeletal and heart muscle).

3.15**Document Preview****medulla oblongata, pons**

areas of the brain where the respiratory control centre is located

[ISO/TS 16976-3:2019](https://standards.iteh.ai/catalog/standards/iso/dc066132-2f2b-4408-a4e7-4169fe823518/iso-ts-16976-3-2019)

3.16 oxyhaemoglobin**HbO₂**

haemoglobin that has bound oxygen from the lungs for transport to the body tissues

3.17**partial pressure**

pressure exerted by each of the components of a gas mixture to form a total pressure

EXAMPLE Air is a mixture of oxygen, nitrogen, carbon dioxide, inert gases (argon, neon), and water vapour. The volume fraction of oxygen in air is about 20,9 %. At sea level, total atmospheric pressure is 101,3 kPa (760 mmHg). Water vapour pressure is 6,26 kPa (47 mmHg) (fully saturated in the lungs at a body temperature of approximately 37 °C). To find partial pressure of oxygen, subtract vapour pressure from total atmospheric pressure and then multiply the oxygen volume fraction by the dry atmospheric pressure. Thus, 101,3 – 6,3 = 95,1 kPa (760 mmHg – 47 mmHg = 713 mmHg); 0,21 × 95,1 kPa = 19,9 kPa (= 149 mmHg). If the ambient pressure increases (as in diving), the partial pressure of each component gas increases. Thus, at 2 atm absolute, the partial pressure of oxygen in dry gas is 101,3 × 2 = 202,6 kPa (760 mmHg × 2 = 1 520 mmHg); 0,21 × 202,6 = 42,6 kPa (0,21 × 1 520 mmHg = 319 mmHg) oxygen.

Note 1 to entry: Partial pressure is dependent on the volume fraction of the component gas.

Note 2 to entry: The partial pressure of a gas can increase or decrease while its relative volume fraction remains the same. Partial pressure drives the diffusion of gas across cell membranes and is, therefore, more important than relative volume fraction of the gas.

3.18

respiratory quotient

R_Q

ratio of volume of carbon dioxide exhaled to the volume of oxygen consumed

Note 1 to entry: R_Q will be calculated as follows:

$$R_Q = \frac{VCO_2}{VO_2}$$

where

VCO_2 is the volume of carbon dioxide exhaled;

VO_2 is the volume of oxygen consumed.

Note 2 to entry: R_Q gives an estimate of the content of substrate utilization during steady-state respiration and metabolism. At rest, $R_Q = 0,82$ reflecting a substrate utilization of a combination of carbohydrates and fats as the primary energy source.

3.19

respiratory system

tubular and cavernous organs (mouth, trachea, bronchi, lungs, alveoli, etc.) and structures which bring about pulmonary ventilation and gas exchange between ambient air and blood

3.20

standard temperature pressure dry

STPD

standard conditions for expression of oxygen consumption

Note 1 to entry: Standard temperature (0 °C) and pressure (101,3 kPa, 760 mmHg), dry air (0 % relative humidity).

3.21

ventilation (general)

process of exchange of air between the lungs and the ambient environment

[ISO/TS 16976-3:2019](https://standards.iteh.ai)

4 Symbols and abbreviated terms

APR	air purifying respirator
BSA	body surface area, expressed in m^2
PAPR	powered air purifying respirator
SAR	supplied air respirator
SCBA	self-contained breathing apparatus
UBA	underwater breathing apparatus
$p\text{CO}_2$	partial pressure of carbon dioxide
$p\text{ACO}_2$	alveolar partial pressure of carbon dioxide
$p_a\text{CO}_2$	arterial partial pressure of carbon dioxide
$p_v\text{CO}_2$	venous partial pressure of carbon dioxide
$p\text{O}_2$	partial pressure of oxygen
$p\text{AO}_2$	alveolar partial pressure of oxygen
$p_a\text{O}_2$	arterial partial pressure of oxygen
$p_i\text{O}_2$	partial pressure of inspired oxygen
$p_v\text{O}_2$	venous partial pressure of oxygen
V_E	minute ventilation (expired) total volume expired from the lungs in 1 min, in l/min (BTPS)
V_I	minute ventilation (inspired) total volume of air inspired into the lungs in 1 min, in l/min (BTPS)
VO_2	oxygen consumption volume of oxygen consumed by the human tissues, in l/min , derived from the difference between the minute volume of inhaled oxygen and the minute volume of exhaled oxygen.
$V\text{CO}_2$	carbon dioxide elimination rate volume of carbon dioxide produced per minute, derived from the product of minute ventilation and the difference between the fractional concentrations of exhaled and inhaled carbon dioxide

5 Oxygen and carbon dioxide in the breathing environment: Physiological responses and limitations

5.1 General

The Earth's atmosphere is composed primarily of nitrogen and oxygen along with some trace gases. Atmospheric carbon dioxide occurs in very low concentrations (approximately 0,03 %). Humans require oxygen as a primary element in the production of energy during aerobic cellular metabolism. Low atmospheric oxygen concentrations or partial pressures (such as occur at high altitude) can