



SLOVENSKI STANDARD SIST EN 50724:2024

01-februar-2024

Stacionarni ultrazvočni detektorji uhajanja plina (UGLD) - Splošne zahteve in preskusne metode

Fixed Ultrasonic Gas Leak Detectors (UGLD) - General requirements and test methods

Ortsfeste Ultraschall-Gasleckage-Detektoren (UGLD) - Allgemeine Anforderungen und Prüfverfahren

Détecteurs ultrasoniques fixes - Règles de performance et méthodes d'essai

Ta slovenski standard je istoveten z: **EN 50724:2023**

[SIST EN 50724:2024](https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-en-50724-2024)

<https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-en-50724-2024>

ICS:

13.320	Alarmni in opozorilni sistemi	Alarm and warning systems
29.260.20	Električni aparati za eksplozivna ozračja	Electrical apparatus for explosive atmospheres

SIST EN 50724:2024

en,fr,de

EUROPEAN STANDARD

EN 50724

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2023

ICS 13.320

English Version

Fixed Ultrasonic Gas Leak Detectors (UGLD) - General requirements and test methods

Détecteurs de fuites de gaz à ultrasons (DFGU) fixes -
Exigences générales et méthodes d'essai

Ortsfeste Ultraschall-Gasleckage-Detektoren (UGLD) -
Allgemeine Anforderungen und Prüfverfahren

This European Standard was approved by CENELEC on 2023-08-07. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Document Preview

[SIST EN 50724:2024](https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-en-50724-2024)

<https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-en-50724-2024>



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	4
Introduction.....	4
1 Scope	6
2 Normative references	6
3 Terms and definitions	7
3.1 Glossary.....	7
3.2 Types of equipment.....	7
3.3 Sensors.....	7
3.4 Signals and alarms	8
4 General requirements.....	9
4.1 Introduction.....	9
4.2 Design.....	10
4.2.1 General.....	10
4.2.2 Indication methods and devices	10
4.2.3 Alarm signal	11
4.2.4 Fault signals.....	11
4.2.5 Adjustments	11
4.2.6 Battery-powered equipment.....	12
4.2.7 Ultrasonic transmitter for use with separate control units	12
4.2.8 Equipment using software and/or digital technologies.....	12
4.3 Labelling and marking.....	12
4.4 Instruction manual.....	12
5 Test methods	14
5.1 Introduction.....	14
5.2 General requirements for tests	15
5.2.1 Samples and sequence of tests	15
5.2.2 Preparation of equipment before testing	15
5.3 Normal conditions for test	16
5.3.1 General.....	16
5.3.2 Test gas(es).....	16
5.3.3 Power supply	16
5.3.4 Temperature and humidity.....	16
5.3.5 Pressure.....	16
5.3.6 Rain	17
5.3.7 Ultrasonic background noise	17
5.3.8 Acclimation time	17
5.3.9 Orientation.....	17
5.3.10 Communications options.....	17
5.4 Test methods and performance requirements	17
5.4.1 Test conditions	17
5.4.2 Unpowered storage	19
5.4.3 Calibration, adjustment and repeatability	20
5.4.4 Stability	20
5.4.5 Alarm set point(s)	21
5.4.6 Temperature	21
5.4.7 Vibration	21
5.4.8 Response point.....	22
5.4.9 Time to alarm	22
5.4.10 Battery capacity.....	22
5.4.11 Power supply variations	23

5.4.12 Electromagnetic compatibility	23
5.4.13 Verification of software and digital components	23
5.4.14 False alarm sources	23
Annex A (informative) Mass flow rate	24
Annex B (informative) Test sheet	25
Annex C (informative) Temperature and humidity dB corrections	26
Annex D (normative) Hole typical drawings	27
Annex E (informative) Calibration	29
Annex F (informative) Background noise generation	31

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN 50724:2024](https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-en-50724-2024)

<https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-en-50724-2024>

EN 50724:2023 (E)**European foreword**

This document (EN 50724:2023) has been prepared by CLC/TC 31 “Electrical apparatus for potentially explosive atmospheres”.

The following dates are fixed:

- latest date by which this document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2024–08–07
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2026–08–07

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users’ national committee. A complete listing of these bodies can be found on the CENELEC website.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN 50724:2024](https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-cn-50724-2024)

<https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-cn-50724-2024>

Introduction

Ultrasonic Gas Leak Detectors (UGLDs) are being increasingly appointed to quickly detect gas leaks from pressurized systems to complement the use of point or line of sight detectors. The UGLD detects the acoustic emission generated by the leak, which propagates omni-directionally at the speed of sound. The leak can be detected at some distance without the gas itself needing to reach the position of the sensor. The main reason for implementing UGLDs is to improve the response time to detect dangerous gas leaks and to improve the gas leak detection coverage with the addition of a complementary detection method that is not affected by air flow.

In general terms, an UGLD functions by detecting the ultrasonic noise generated by gas escapes from a high-pressure area to a low-pressure area through a small aperture (leak). In practice the technique is of value for detecting leaks from tank/pipelines, etc running at pressures greater than 2 bar. The intensity of this airborne ultrasound generated by a gas leak is due to a number of factors including but not limited to gas type, gas pressure, leak size and gas temperature.

An UGLD does not detect specific gas types, measure percentage LFL or ppm concentration level, but instead responds to the specific ultrasonic sound generated by a pressurized gas leak. The reliable range coverage of an UGLD is mainly determined by the leak rate of the gas leak, the atmospheric transmission of the ultrasound, and the potential acoustic background noise that can interfere with the UGLD. The leak rate is mainly determined by gas pressure and leak size, but molecular weight and gas temperature also plays a role. The leak rate determines how fast a potentially dangerous gas cloud will be generated. In addition, physical obstructions between the location of the leak, and the UGLD will also have an influence on the detection range of the detector.

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[SIST EN 50724:2024](https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-en-50724-2024)

<https://standards.iteh.ai/catalog/standards/sist/1f1c2b4b-95b4-42f5-9d34-c3a7f4c98b82/sist-en-50724-2024>

EN 50724:2023 (E)**1 Scope**

This document refers to UGLDs (ultrasonic gas leak detectors) that work in a frequency beyond the audible range. This document is applicable to fixed ultrasonic gas leak detection equipment intended to provide an indication, alarm or other output function for the purpose of initiating automatic or manual protective action(s).

This document specifies general requirements for design, testing and performance, and describes test methods that apply to UGLD. The following items are considered in this document:

- Leak rates to be used to verify the detection range of UGLD,
- Test gas to be used (nitrogen or compressed air),
- Nozzle shape and size used at all tests leak rate tests,
- Gas pressure used at all leak rate tests,
- Time duration of each leak rate test,
- Test leak nozzle height from solid ground,
- Test leak nozzle angling relative to test UGLD,
- UGLD angle relative to the leak (field of coverage of the UGLD),
- Wind speed and direction, air temperature and humidity at day of test,
- Minimum distance to solid structures (walls, etc.) at test site,
- Installation height relative to the ground,
- Texture of solid ground between leak and UGLD,
- Background noise sources, known to interfere with UGLDs,
- Specification of detection radius in 3 dimensions,
- Operational requirements such as temperature, vibration, etc.

This document is also applicable when an equipment manufacturer makes any claims regarding any special features of construction or superior performance that exceed the minimum requirements of this document. This document prescribes that all such claims are verified, and that the test procedures are extended or supplemented, where necessary, to verify the claimed performance. The additional tests are agreed between the manufacturer and test laboratory and identified and described in the test report.

This document does not apply to portable gas detectors using ultrasonic measurements nor to gas detectors using non-ultrasonic measurements to detect a gas leak.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN IEC 61326-1, *Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements*

EN 50271:2018, *Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen - Requirements and tests for apparatus using software and/or digital technologies*

IEC 60068-2-6, *Environmental testing - Part 2-6: Tests - Test Fc: Vibration (sinusoidal)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1 Glossary

3.1.1

UGLD

Ultrasonic Gas Leak Detector

equipment that work normally in a frequency beyond the audible range (> 16 kHz)

3.1.2

leak rate

mass flow rate from the gas leak

3.2 Types of equipment

3.2.1

alarm-only equipment

equipment having an alarm but not having a proportional indication of sound pressure level

3.2.2

continuous duty equipment

equipment that is powered for long periods of time, but may have either continuous or intermittent sensing

3.2.3

fixed equipment

equipment that is intended to have all parts permanently installed

3.2.4

stand-alone leak detection equipment

fixed leak detection equipment that provides a conditioned electronic signal or output indication to a generally accepted industry standard (such as 4-20 mA), intended to be utilized with stand-alone control units, or signal processing data acquisition, central monitoring and similar systems which typically process information from various locations and sources including, but not limited to, leak detection equipment

3.2.5

stand-alone control unit

fixed leak detection control units intended to provide meter indication, alarm functions, output contacts and/or alarm signal outputs when utilized with stand-alone leak detection equipment

3.3 Sensors

3.3.1

integral sensor

sensor which is integral to the main body of the equipment

3.3.2

remote sensor

sensor which is not integral to the main body of the equipment

EN 50724:2023 (E)**3.3.3****sensor**

assembly in which the sensing element is housed and which may also contain associated circuit components

3.3.4**sensing element**

part of the sensor which is sensitive to the noise to be measured

3.4 Signals and alarms**3.4.1****sound pressure level****SPL**

level measured in dB defined as:

$$SPL = 10 \cdot \log_{10} \left[\frac{POW_{SIG}}{POW_{ref}} \right] = 20 \cdot \log_{10} \left[\frac{P_{SIG}}{P_{ref}} \right]$$

with the sound pressure P_{SIG} , the reference sound pressure $P_{ref} = 20\mu\text{Pa}$, the reference signal power POW_{ref} , and the signal power POW_{SIG} calculated from a time domain signal of N digits:

$$POW_{SIG} = \sum_{i=1}^{N/2} \left[\text{abs} \left(\text{FFT}_i(N) \right) \right]^2 \sim P_{SIG}^2$$

Note 1 to entry: FFT_i represents the signal within the i_{th} frequency bin generated by a Fast Fourier Transformation. No weighting filter is applied (like A, C or Z). The FFT window shall be chosen in a way that best represents the power of the ultrasonic signal.

Note 2 to entry: The reference signal power POW_{ref} , necessary to fixate the dB-scale to an absolute reference, is extremely difficult to measure, especially since the ultrasonic bandwidth ranges of UGLDs differ among each other. Therefore, the absolute calibration is traced back to the existing calibration in the audible range, the hearing threshold. A practical method and calibration set up is described in Annex E.

To generate a consistent dB-scale between the vendor devices, the calibration bandwidth is fixed from 16kHz to 64kHz, representing two full octave frequency bands.

3.4.2**reference noise**

small leak of air close to the UGLD – typically round nozzle, 300 μm , mass flow 0,06 g/s, 25 cm from the sensor, used for some specific small-scale tests

Note 1 to entry: The nozzle drawing is provided in Annex D.

3.4.3**reference leak**

25 g/s leak of test gas generated through a 2 mm nozzle using e.g. a mass flow controller

Note 1 to entry: An upstream pressure of 38-bar at 20°C may be necessary.

Note 2 to entry: The nozzle drawing is provided in Annex D.

3.4.4**low noise condition**

no external stimulus beyond 50 dB, measured with an external dedicated measurement equipment in the frequency range defined

3.4.5**measuring range**

measuring range that represents the interval between the lowest measured value (in dB) and the highest measured value

3.4.6**alarm set point**

fixed or adjustable setting of the equipment that is intended to pre-set the level of ultrasonic sound pressure level at which the equipment will automatically initiate an indication, alarm or other output function

3.4.7**fault signal**

audible, visible or other type of output, different from the alarm signal, permitting, directly or indirectly, a warning or indication that the equipment is not working faultless

3.4.8**latching alarm**

alarm that, once activated, requires deliberate action to be deactivated

3.4.9**repeatability**

level of agreement between the results of measurements of the same ultrasound noise level, carried out by the same method and conditions, on the same UGLD instrument type

3.4.10**drift**

variation in the equipment indication with time, at any fixed ultrasound noise level under constant ambient conditions

3.4.11**final indication**

indication given by the equipment after stabilisation

3.4.12**stabilisation**

state when three successive readings of an equipment, taken at 5 s intervals, indicate no changes greater than ± 1 dB

3.4.13**time to alarm**

time interval, with the equipment in a warmed-up condition, between the time when an instantaneous release is produced in the equipment surrounding and the time when the alarm is triggered

4 General requirements

4.1 Introduction

The equipment shall comply with the requirements of this document.

Electrical assemblies and components shall comply with the construction and test requirements of 4.2, where applicable. In addition, parts of the UGLD intended for use in potentially explosive atmospheres are expected to employ materials and comply with the construction and explosion protection as specified in the appropriate regulations for explosion protection.