
**Ugotavljanje zmogljivosti neprekinjeno delujočih zračnih nadzornikov - 2. del:
Zračni nadzorniki na podlagi tehnik vzorčenja pretoka zraka brez kopičenja
(ISO/TR 22930-2:2020)**

Evaluating the performance of continuous air monitors - Part 2: Air monitors based on flow-through sampling techniques without accumulation (ISO/TR 22930-2:2020)

Ermittlung der Leistungsfähigkeit kontinuierlicher Luftmonitore - Teil 2: Luftmonitore basierend auf Durchfluss-Sammeltechnik ohne Anreicherung (ISO/TR 22930-2:2020)

Évaluation des performances des dispositifs de surveillance de l'air en continu - Partie 2: Dispositifs de surveillance de l'air basés sur des techniques d'échantillonnage par circulation sans accumulation (ISO/TR 22930-2:2020)

Ta slovenski standard je istoveten z: CEN ISO/TR 22930-2:2021

ICS:

13.280 Varstvo pred sevanjem Radiation protection

SIST-TP CEN ISO/TR 22930-2:2021 en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST-TP CEN ISO/TR 22930-2:2021](https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021)

<https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>

TECHNICAL REPORT

CEN ISO/TR 22930-2

RAPPORT TECHNIQUE

TECHNISCHER BERICHT

August 2021

ICS 13.280

English Version

Evaluating the performance of continuous air monitors -
Part 2: Air monitors based on flow-through sampling
techniques without accumulation (ISO/TR 22930-2:2020)

Évaluation des performances des dispositifs de
surveillance de l'air en continu - Partie 2: Dispositifs de
surveillance de l'air basés sur des techniques
d'échantillonnage par circulation sans accumulation
(ISO/TR 22930-2:2020)

Ermittlung der Leistungsfähigkeit kontinuierlicher
Luftmonitore - Teil 2: Luftmonitore basierend auf
Durchfluss-Sammeltechnik ohne Anreicherung
(ISO/TR 22930-2:2020)

This Technical Report was approved by CEN on 16 August 2021. It has been drawn up by the Technical Committee CEN/TC 430.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

(standards.iteh.ai)

[SIST-TP CEN ISO/TR 22930-2:2021](https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021)

<https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST-TP CEN ISO/TR 22930-2:2021
<https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>

European foreword

The text of ISO/TR 22930-2:2020 has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" of the International Organization for Standardization (ISO) and has been taken over as CEN ISO/TR 22930-2:2021 by Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

Endorsement notice

The text of ISO/TR 22930-2:2020 has been approved by CEN as CEN ISO/TR 22930-2:2021 without any modification.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST-TP CEN ISO/TR 22930-2:2021](https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021)

<https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST-TP CEN ISO/TR 22930-2:2021](https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021)

<https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>

TECHNICAL REPORT

ISO/TR 22930-2

First edition
2020-05

Evaluating the performance of continuous air monitors —

Part 2:

Air monitors based on flow-through sampling techniques without accumulation

iTeh STANDARD PREVIEW

(standards.iteh.ai)
*Évaluation de la performance des dispositifs de surveillance de l'air
en continu —*

*Partie 2: Moniteurs d'air basés sur des techniques d'échantillonnage
par circulation sans accumulation*

<https://standards.iteh.ai/catalog/standards/sist/6095619-41c0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>



Reference number
ISO/TR 22930-2:2020(E)

© ISO 2020

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST-TP CEN ISO/TR 22930-2:2021

<https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 Symbols.....	4
5 Measuring principle.....	6
6 Study of dynamic behaviour.....	7
7 Evaluation of the characteristic limits.....	13
7.1 General.....	13
7.2 Single detector.....	13
7.2.1 General.....	13
7.2.2 Definition of the model.....	14
7.2.3 Standard uncertainty.....	14
7.2.4 Decision threshold.....	15
7.2.5 Detection limit.....	17
7.2.6 Limits of the coverage interval.....	17
7.3 Double detector.....	17
7.3.1 General.....	17
7.3.2 Definition of the model.....	18
7.3.3 Standard uncertainty.....	18
7.3.4 Decision threshold.....	19
7.3.5 Detection limit.....	20
7.3.6 Limits of the coverage interval.....	20
8 Alarms setup, minimum detectable concentration and potential missed exposure.....	20
Annex A (informative) Application example: Single detector with a proportional counter.....	23
Annex B (informative) Application example: Double detector in current mode.....	26
Bibliography.....	32

ISO/TR 22930-2:2020(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html. (standards.iteh.ai)

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

A list of all the parts in the ISO/TR 22930-series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Sampling and monitoring of airborne activity concentration in workplaces are critically important for maintaining worker safety at facilities where dispersible radioactive substances are used.

The first indication of a radioactive substance dispersion event comes, in general, from a continuous air monitor (CAM) and its associated alarm levels. In general, the response of a CAM is delayed in time compared to the actual situation of release.

The knowledge of a few factors is needed to interpret the response of a CAM and to select the appropriate CAM type and its operating parameters.

The role of the radiation protection officer is to select the appropriate CAM, to determine when effective release of radioactive substances occurs, to interpret measurement results and to take corrective action appropriate to the severity of the release.

The objective of ISO/TR 22930 series is to assist radiation protection officer in evaluating the performance of a CAM.

ISO/TR 22930 series describes the factors and operating parameters and how they influence the response of a CAM.

This document deals with monitoring systems based on flow-through sampling techniques without accumulation.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST-TP CEN ISO/TR 22930-2:2021](https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021)

<https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST-TP CEN ISO/TR 22930-2:2021

<https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-64a8ea7370c2/sist-tp-cen-iso-tr-22930-2-2021>

Evaluating the performance of continuous air monitors —

Part 2:

Air monitors based on flow-through sampling techniques without accumulation

1 Scope

The use of a continuous air monitor (CAM) is mainly motivated by the need to be alerted quickly and in the most accurate way possible with an acceptable false alarm rate when significant activity concentration value is exceeded, in order to take appropriate measures to reduce exposure of those involved.

The performance of this CAM does not only depend on the metrological aspect characterized by the decision threshold, the limit of detection and the measurement uncertainties but also on its dynamic capacity characterized by its response time as well as on the minimum detectable activity concentration corresponding to an acceptable false alarm rate.

The ideal performance is to have a minimum detectable activity concentration as low as possible associated with a very short response time, but unfortunately these two criteria are in opposition. It is therefore important that the CAM and the choice of the adjustment parameters and the alarm levels be in line with the radiation protection objectives.

This document describes [SIST-TP CEN ISO/TR 22930-2:2021](https://standards.iteh.ai/catalog/standards/sist/06953619-41e0-4bad-8d9d-04abca7570c2/sist-tp-cen-iso-tr-22930-2-2021)

- the dynamic behaviour and the determination of the response time,
- the determination of the characteristic limits (decision threshold, detection limit, limits of the coverage interval), and
- a possible way to determine the minimum detectable activity concentration and the alarms setup.

Finally the annexes of this document show actual examples of CAM data which illustrate how to quantify the CAM performance by determining the response time, the characteristics limits, the minimum detectable activity concentration and the alarms setup.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11929-1, *Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation — Fundamentals and application — Part 1: Elementary applications*

ISO 16639, *Surveillance of the activity concentrations of airborne radioactive substances in the workplace of nuclear facilities*

IEC 60761-1, *Equipment for continuous monitoring of radioactivity in gaseous effluents — Part 1: General requirements*