

#### SLOVENSKI STANDARD SIST EN 60601-1-1:2002

01-februar-2002

Nadomešča:

SIST EN 60601-1-1:1995

SIST EN 60601-1-1:1995/A1:1998

Medicinska električna oprema - 1-1. del: Splošne varnostne zahteve - Spremljevalni standard: Varnostne zahteve za medicinske električne sisteme (IEC 60601-1-1:2000)

Medical electrical equipment - Part 1-1: General requirements for safety - Collateral standard: Safety requirements for medical electrical systems (IEC 60601-1-1:2000)

Medizinische elektrische Geräte - Teil 1-1: Allgemeine Festlegungen für die Sicherheit - Ergänzungsnorm: Festlegungen für die Sicherheit von medizinischen elektrischen Systemen (IEC 60601-1-1:2000) eh ai/catalog/standards/sist/ba183c1f-d39c-4620-9480-e6266c879fd8/sist-en-60601-1-1-2002

Appareils électromédicaux - Partie 1-1: Règles générales de sécurité - Norme collatérale: Règles de sécurité pour systèmes électromédicaux (CEI 60601-1-1:2000)

Ta slovenski standard je istoveten z: EN 60601-1-1:2001

ICS:

11.040.01 Medicinska oprema na

Medical equipment in general

splošno

SIST EN 60601-1-1:2002 en

## iTeh STANDARD PREVIEW (standards.iteh.ai)

#### **EUROPEAN STANDARD**

#### EN 60601-1-1

#### NORME EUROPÉENNE

#### **EUROPÄISCHE NORM**

April 2001

ICS 11.040.01

Supersedes EN 60601-1-1:1993 + A1:1996

**English version** 

# Medical electrical equipment Part 1-1: General requirements for safety Collateral standard: Safety requirements for medical electrical systems (IEC 60601-1-1:2000)

Appareils électromédicaux Partie 1-1: Règles générales de sécurité -Norme collatérale: Règles de sécurité pour systèmes électromédicaux (CEI 60601-1-1:2000) Medizinische elektrische Geräte Teil 1-1: Allgemeine Festlegungen für die Sicherheit -Ergänzungsnorm: Festlegungen für die Sicherheit von medizinischen elektrischen Systemen (IEC 60601-1-1:2000)

This European Standard was approved by CENELEC on 2000-12-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

SIST EN 60601-1-1:2002

https://standards.iteh.ai/catalog/standards/sist/ba183c1f-d39c-4620-9480-

e6266c879fd CENEL-200C

European Committee for Electrotechnical Standardization Comité Européen de Normalisation Electrotechnique Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

#### **Foreword**

The text of document 62A/312/FDIS, future edition 2 of IEC 60601-1-1, prepared by SC 62A, Common aspects of electrical equipment used in medical practice, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-1-1 on 2000-12-01.

This European Standard supersedes EN 60601-1-1:1993 + A1:1996.

The following dates were fixed:

 latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement

(dop) 2001-10-01

 latest date by which the national standards conflicting with the EN have to be withdrawn

(dow) 2003-12-01

This European Standard is a Collateral Standard to EN 60601-1:1990, hereinafter referred to as the General Standard, and is the first of a series of Collateral Standards amplifying the General Standard.

In the 60601 series of publications, Collateral Standards specify general requirements for safety applicable to

- a group of MEDICAL ELECTRICAL EQUIPMENT (for example, radiological equipment);
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the General Standard (for example, electromagnetic compatibility).

The numbering of sections, clauses and subclauses of this Collateral Standard corresponds with that of the General Standard.

Subclauses and figures which are additional to those of the General Standard are numbered starting from 201; additional annexes are lettered AAA, BBB, etc.

In this Collateral Standard, the following print types are used:

- requirements, compliance with which can be tested and definitions: in roman type;
- explanations, advice, general statements, exceptions and references: in smaller roman type;
- test specifications: in italic type;
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD OR OF THIS COLLATERAL STANDARD: SMALL CAPITALS.

The requirements are followed by specifications for the relevant tests.

Some provisions or statements in the body of this Collateral Standard require additional information. Such information is presented in the informative annex AAA, General guidance and rationale. An asterisk (\*) at the left margin of a clause or subclause indicates the presence of additional information.

Annexes designated "normative" are part of the body of the standard.

Annexes designated "informative" are given for information only.

In this standard tannexes EEE and ZA gare normative land fannexes AAA,8 BBB, FFF and ZB are informative.

Annexes ZA and ZB replace annexes CCC and DDD of IEC 60601-1-1:2000.

#### **Endorsement notice**

The text of the International Standard IEC 60601-1-1:2000 was approved by CENELEC as a European Standard without any modification.

### Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

Publication	<u>Year</u>	<u>Title</u>	EN/HD	<u>Year</u>
IEC 60083	1997	Plugs and socket-outlets for domestic and similar general use standardized in member countries of IEC	-	-
IEC 60529	1989	Degrees of protection provided by enclosures (IP Code)	EN 60529	1991
IEC 60601-1	1988	Medical electrical equipment	EN 60601-1	1990
A1	1991	Part 1: General requirements for safety	+ corr. July A1	1994 1993
A2	1995		+ corr. July	1994
+ corr. June	1995		A2	1995
			A13	1996
IEC 60884-1	1994	Plugs and socket-outlets for household and similar purposes Part 1: General requirements	-	-
A1	1994	, and constant of the constant	-	-
A2	1995		-	-
IEC 60989	1991	Separating transformers, autotransformers, variable transformers	-	-
		and reactors	***	

iTeh STANDARD PREVIEW (standards.iteh.ai)

### Annex ZB (informative)

#### **Bibliography**

<u>Publication</u>	<u>Year</u>	<u>Title</u>	EN/HD	<u>Year</u>
IEC 60065 (mod)	1998	Audio, video and similar electronic apparatus - Safety requirements	EN 60065	1998
IEC 60335-1 (mod)	1991	Safety of household and similar electrical appliances	EN 60335-1 + corr. January	1994 1995
A1 (mod)	1994	Part 1: General requirements	A1	1996
IEC 60601-1-4	1996	Medical electrical equipment Part 1-4: General requirements for safety Collateral standard: Programmable electrical medical systems	EN 60601-1-4	1996
A1	1999	Systems	A1	1999
IEC 60825-1	1993	Safety of laser products Part 1: Equipment classification, requirements and user's guide	EN 60825-1 + corr. February	1994 1995
A1	1997	requirements and user s guide	-	-
IEC 60950 (mod) + corr. January	1999 2000	Safety of information technology equipment	EN 60950	2000
IEC 61010-1 (mod)	1990	Safety requirements for electrical equipment for measurement, control and laboratory use Part 1: General requirements		
+ A1 (mod) A2	1992 1995	Tart 1. Scholar requirements	EN 61010-11) A2 <sup>1)</sup>	1993 1995
ISO 7767	1997	Oxygen monitors for monitoring patient breathing mixtures - Safety requirements	-	-
ISO 8185	11997h	Humidifiers for medical use - General EV requirements for humidification systems Standards. Iteh. al	VEN ISO 8185	1997
ISO 8359	1996	Oxygen concentrators for medical use Safety requirements 1-1:2002	-	-
ISO 9918	ps://standar 1993	rds, iteh ai/catalog/standards/sist/ba183c1f-d39c-4620- Capnometers for use with humans - Requirements	9480-	-
ISO 10079-1	1991	Medical suction equipment Part 1: Electrically powered suction equipment - Safety requirements	EN ISO 10079-1	1996

<sup>1)</sup> EN 61010-1:1993 + A2:1995 are superseded by EN 61010-1:2001, which is based on IEC 61010-1:2001.

## **NORME** INTERNATIONALE INTERNATIONAL **STANDARD**

CEI **IEC** 60601-1-1

> Deuxième édition Second edition 2000-12

#### Appareils électromédicaux –

Partie 1-1:

The state of the state of

Règles générales de sécurité – Norme collatérale: Règles de sécurité

pour systèmes électromédicaux

Medical electrical equipment –

Part 1-1:

General requirements for safety – Collateral standard: Safety requirements for medical electrical systems

### iTeh STANDARD PREVIEW (standards.iteh.ai)

© IEC 2000 Droits de reproduction réservés — Copyright - all rights reserved

https://st.Aucune/partie-dej cette publication ne legul être reproduite ni o No part of this publication may be reproduced or utilized in utilisée sous quelque forme que ce soit et par aucun procédé, any form or by any means, electronic or mechanical, électronique ou mécanique y compris la photocopie et les microfilms, sans l'accord écrit de l'éditeur.

including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission Telefax: +41 22 919 0300 e-mail: inmail@iec.ch

3, rue de Varembé Geneva, Switzerland IEC web site http://www.iec.ch



Commission Electrotechnique Internationale International Electrotechnical Commission Международная Электротехническая Комиссия

CODE PRIX PRICE CODE



Pour prix, voir catalogue en vigueur For price, see current catalogue

#### CONTENTS

FOREW	Page /ORD
	SECTION ONE — GENERAL
Clause	Ocean and object
1	Scope and object
3	Terminology and definitions
6	Identification, marking and documents
	SECTION TWO — ENVIRONMENTAL CONDITIONS
10	Environmental conditions
	SECTION THREE — PROTECTION AGAINST ELECTRIC SHOCK HAZARDS
16	ENCLOSURES and PROTECTIVE COVERS
17	Separation
19	Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS
	SECTION FOUR — PROTECTION AGAINST MECHANICAL HAZARDS
22	Moving parts
	SECTION FIVE — PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION
	SECTION SIX — PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES
	SECTION SEVEN — PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS
44	Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization, disinfection and compatibility
49	Interruption of the power supply
	SECTION EIGHT — ACCURACY OF OPERATING DATA AND PROTECTION AGAINST-HAZARDOUS OUTPUT https://standards.iteh.ai/catalog/standards/sist/ba183c1f-d39c-4620-9480-
	e6266c879fd8/sist-en-60601-1-1-2002 SECTION NINE — ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS
52	Abnormal operation and fault conditions
	SECTION TEN — CONSTRUCTIONAL REQUIREMENTS
56	Components and general assembly
57	Mains parts, components and layout
58	Protective earthing - Terminals and connections
59	Construction and layout

		Pages
Figure 201	Example of PATIENT ENVIRONMENT	27
Annex AAA (ir	nformative) General guidance and rationale	29
Annex BBB (ir	nformative) Examples of combinations of MEDICAL ELECTRICAL EQU	JIPMENT and
	normative) Normative references	
Annex DDD (ir	nformative) Bibliography	49
Annex EEE (n	ormative) Requirements for MULTIPLE PORTABLE SOCKET-OUTLETS	51
Annex FFF (in	formative) Examples of application of MULTIPLE PORTABLE SOCKET	г-оитLETS 55

## iTeh STANDARD PREVIEW (standards.iteh.ai)

#### INTERNATIONAL ELECTROTECHNICAL COMMISSION

#### MEDICAL ELECTRICAL EQUIPMENT -

## Part 1-1: General requirements for safety – Collateral standard: Safety requirements for medical electrical systems

#### FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a world-wide organisation for standardisation comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardisation in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organisations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organisation for Standardisation (ISO) in accordance with conditions determined by agreement between the two organisations.
- 2) The formal decisions or agreements of the IEC on technical matters, prepared by technical committees on which all the National Committees having a special interest therein are represented, express, as nearly as possible, an international consensus of opinion on the subjects dealt with.
- 3) They have the form of recommendations for international use published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-1-1 has been prepared by subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition of 60601-1-1 cancels and replaces the first edition published in 1992 and its amendment 1(1995) and constitutes a technical revision.

This second edition is a Collateral Standard to 1EO 60601 1: Medical electrical equipment – Part 1: General requirements for safety, hereinafter referred to as the General Standard, and is the first of a series of Collateral Standards amplifying the General Standard.

https://standards.iteh.ai/catalog/standards/sist/ba183c1f-d39c-4620-9480-

The text of this Collateral Standard is based on the following documents:

FDIS	Report on voting	
62A/312/FDIS	62A/318/RVD	

Full information on the voting for the approval of this Collateral Standard can be found in the report on voting indicated in the above table.

In the 60601 series of publications, Collateral Standards specify general requirements for safety applicable to

- a group of MEDICAL ELECTRICAL EQUIPMENT (for example, radiological equipment);
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the General Standard (for example, electromagnetic compatibility).

The numbering of sections, clauses and subclauses of this Collateral Standard corresponds with that of the General Standard.

Subclauses and figures which are additional to those of the General Standard are numbered starting from 201; additional annexes are lettered AAA, BBB, etc.

In this Collateral Standard, the following print types are used:

- requirements, compliance with which can be tested and definitions: in roman type;
- explanations, advice, general statements, exceptions and references: in smaller roman type;
- test specifications: in italic type;
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD OR OF THIS COLLATERAL STANDARD: SMALL CAPITALS.

The requirements are followed by specifications for the relevant tests.

Some provisions or statements in the body of this Collateral Standard require additional information. Such information is presented in the informative annex AAA, General guidance and rationale. An asterisk (\*) at the left margin of a clause or subclause indicates the presence of additional information.

Annexes AAA, BBB, DDD and FFF are for information only.

Annexes CCC and EEE form an integral part of this Collateral Standard.

The committee has decided that the contents of this publication will remain unchanged until 2005. At this date, the publication will be

- reconfirmed;
- withdrawn; iTeh STANDARD PREVIEW
- replaced by a revised edition candards.iteh.ai)
- amended.

#### MEDICAL ELECTRICAL EQUIPMENT -

## Part 1-1: General requirements for safety – Collateral Standard: Safety requirements for medical electrical systems

#### SECTION ONE - GENERAL

#### 1 Scope and object

#### \*1.201 Scope

This standard applies to the safety of MEDICAL ELECTRICAL SYSTEMS, as defined in 2.201. It describes the safety requirements necessary to provide protection for the PATIENT, the OPERATOR and surroundings.

#### 2 Terminology and definitions

In this Collateral Standard, terms printed in small capitals are used in accordance with their definitions in IEC 60601-1.

Where the terms "voltage" and "current" are used, they mean the r.m.s. values of an alternating, direct or composite voltage or current.

For the purpose of this standard the following additional definitions apply:

#### 2 201

#### MEDICAL ELECTRICAL SYSTEM (hereinafter referred to as SYSTEM)

combination of items of equipment, at least one of which must be MEDICAL ELECTRICAL EQUIPMENT and inter-connected by FUNCTIONAL CONNECTION or use of a MULTIPLE PORTABLE SOCKET-OUTLET

NOTE Equipment, when mentioned in connection with a SYSTEM, should be taken to include EQUIPMENT. (See also examples given in annexes BBB and FFF.)

#### iTeh STANDARD PREVIEW

#### \*2.202

#### PATIENT ENVIRONMENT (standards.iteh.ai)

any volume in which intentional or unintentional contact can occur between PATIENT and parts of the SYSTEM or between PATIENT and other persons touching parts of the SYSTEM (see figure 201)

https://standards.iteh.ai/catalog/standards/sist/ba183c1f-d39c-4620-9480-e6266c879fd8/sist-en-60601-1-1-2002

#### \*2.203

SEPARATION DEVICE

a component or arrangement of components with input parts and output parts that, for safety reasons, prevent a transfer of unwanted voltage or current between parts of a SYSTEM

#### \*2.204

#### MULTIPLE PORTABLE SOCKET-OUTLET

a combination of two or more socket-outlets intended to be connected to, or integral with, flexible cables or cords, and which can easily be moved from one place to another while connected to the supply

NOTE A MULTIPLE PORTABLE SOCKET-OUTLET may be a separate item or an integral part of medical or non-medical equipment