

SLOVENSKI STANDARD
oSIST prEN ISO 20236:2021
01-september-2021

Kakovost vode - Določevanje celotnega organskega ogljika (TOC), raztopljenega organskega ogljika (DOC), celotnega vezanega dušika (TNb) in raztopljenega vezanega dušika (DNb) po katalitskem sežigu pri visoki temperaturi (ISO 20236:2018)

Water quality - Determination of total organic carbon (TOC), dissolved organic carbon (DOC), total bound nitrogen (TNb) and dissolved bound nitrogen (DNb) after high temperature catalytic oxidative combustion (ISO 20236:2018)

iTeh STANDARD PREVIEW

Wasserbeschaffenheit - Bestimmung des gesamten organischen Kohlenstoffs (TOC), des gelösten organischen Kohlenstoffs (DOC), des gebundenen Stickstoffs (TNb) und des gelösten gebundenen Stickstoffs (DNb) nach katalytischer oxidativer Hochtemperaturverbrennung (ISO 20236:2018)

Qualité de l'eau - Dosage du carbone organique total (COT), carbone organique (COD), azote lié total (TNb) et azote lié dissous (DNb) après combustion oxidatif catalytique à haute température (ISO 20236:2018)

Ta slovenski standard je istoveten z: prEN ISO 20236

ICS:

13.060.50	Preiskava vode na kemične snovi	Examination of water for chemical substances
-----------	---------------------------------	--

oSIST prEN ISO 20236:2021

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN ISO 20236:2021](https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-df465f7a2956/osist-pren-iso-20236-2021)
<https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-df465f7a2956/osist-pren-iso-20236-2021>

INTERNATIONAL
STANDARD

ISO
20236

First edition
2018-09

**Water quality — Determination of
total organic carbon (TOC), dissolved
organic carbon (DOC), total bound
nitrogen (TN_b) and dissolved bound
nitrogen (DN_b) after high temperature
catalytic oxidative combustion**

iTeh STANDARD PREVIEW

*Qualité de l'eau — Dosage du carbone organique total (COT),
carbone organique (COD), azote lié total (TN_b) et azote lié dissous
(DN_b) après combustion oxidatif catalytique à haute température*

[oSIST prEN ISO 20236:2021](https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-df465f7a2956/osist-pren-iso-20236-2021)

[https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-
df465f7a2956/osist-pren-iso-20236-2021](https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-df465f7a2956/osist-pren-iso-20236-2021)



Reference number
ISO 20236:2018(E)

© ISO 2018

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 20236:2021
https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-
df465f7a2956/osist-pren-iso-20236-2021](https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-df465f7a2956/osist-pren-iso-20236-2021)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	2
5 Interferences	3
5.1 General.....	3
5.2 TOC or DOC.....	3
5.3 TN _b or DN _b	4
6 Reagents	4
7 Apparatus	6
8 Quality requirements for the analytical system	8
8.1 System check.....	8
8.2 Recovery and variation of replicate determinations for particle processing control for TOC and TN _b	8
9 Sampling and sample preparation	8
10 Procedure	9
10.1 General.....	9
10.2 Calibration.....	9
10.3 Validity check of the calibration function.....	9
10.4 Measurement.....	9
10.4.1 General.....	9
10.4.2 Determination.....	10
11 Evaluation	11
12 Expression of results	11
13 Test report	11
Annex A (normative) Determination of TOC and TIC applying the difference method	12
Annex B (informative) Performance data for TOC or DOC and TN_b or DN_b	15
Annex C (informative) Alternative detection techniques for TN_b and DN_b	17
Bibliography	18

ISO 20236:2018(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Total organic carbon (TOC), dissolved organic carbon (DOC), total bound nitrogen (TN_b) and dissolved bound nitrogen (DN_b) are an analytical convention, the respective result of which is a parameter used for water quality control purposes. These parameters represent the sum of organically bound carbon as well as the sum of inorganic and organic nitrogen (but not nitrogen gas), which can be dissolved in water or bonded to dissolved or suspended matter under specified conditions and, if the sample is not filtered, includes that associated with suspended matter. It does not give information on the nature of the substances.

Details of an interlaboratory trial on the performance data for TOC or DOC and TN_b or DN_b are given in [Annex B](#).

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 20236:2021
https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-
df465f7a2956/osist-pren-iso-20236-2021](https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-df465f7a2956/osist-pren-iso-20236-2021)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN ISO 20236:2021](https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-df465f7a2956/osist-pren-iso-20236-2021)
<https://standards.iteh.ai/catalog/standards/sist/068f79df-03e9-4d09-9463-df465f7a2956/osist-pren-iso-20236-2021>

Water quality — Determination of total organic carbon (TOC), dissolved organic carbon (DOC), total bound nitrogen (TN_b) and dissolved bound nitrogen (DN_b) after high temperature catalytic oxidative combustion

WARNING — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices.

IMPORTANT — It is absolutely essential that tests conducted in accordance with this document be carried out by suitably qualified staff.

1 Scope

This document specifies a method for the determination of total organic carbon (TOC), dissolved organic carbon (DOC), total bound nitrogen (TN_b) and dissolved bound nitrogen (DN_b) in the form of free ammonia, ammonium, nitrite, nitrate and organic compounds capable of conversion to nitrogen oxides under the conditions described. The procedure is carried out instrumentally.

NOTE Generally the method can be applied for the determination of total carbon (TC) and total inorganic carbon (TIC), see [Annex A](#).

The method is applicable to water samples (e.g. drinking water, raw water, ground water, surface water, sea water, waste water, leachates).

The method allows a determination of TOC and DOC ≥ 1 mg/l and TN_b and DN_b ≥ 1 mg/l. The upper working range is restricted by instrument-dependent conditions (e.g. injection volume). Higher concentrations can be determined after appropriate dilution of the sample.

For samples containing volatile organic compounds (e.g. industrial waste water), the difference method is used, see [Annex A](#).

Cyanide, cyanate and particles of elemental carbon (soot), when present in the sample, can be determined together with the organic carbon.

The method is not appropriate for the determination of volatile, or purgeable, organic carbon under the conditions described by this method.

Dissolved nitrogen gas (N₂) is not determined.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO 20236:2018(E)

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1**total carbon****TC**

sum of organically and inorganically bound carbon present in water, including elemental carbon

3.2**total inorganic carbon****TIC**

sum of inorganic carbon present in water sample measured under the conditions of this method

Note 1 to entry: TIC is measured as CO₂ originating only from carbonates, hydrogen carbonates and dissolved carbon dioxide.

3.3**total organic carbon****TOC**

sum of organically bound carbon present in water, bonded to dissolved or suspended matter, including cyanate, thiocyanate and elemental carbon measured under the conditions of this method

Note 1 to entry: Volatile organic carbon cannot be guaranteed to be determined by the method.

Note 2 to entry: Generally, TOC includes organic compounds in water that cannot be purged under the conditions of this method, also known as non-purgeable organic carbon (NPOC).

3.4**dissolved organic carbon****DOC**

sum of organically bound carbon present in water originating from compounds passing through a membrane filter of 0,45 µm pore size, including cyanate and thiocyanate measured under the conditions of this method

3.5**total bound nitrogen****TN_b**

sum of organically bound and inorganically bound nitrogen present in water or suspended matter measured under the conditions of this method

3.6**dissolved bound nitrogen****DN_b**

sum of organically and inorganically bound nitrogen present in water originating from compounds passing through a 0,45 µm membrane filter measured under the conditions of this method

4 Principle

Thermal catalytic combustion of the sample containing organic carbon, and inorganic and organic nitrogen in an oxygen-containing atmosphere at ≥ 680 °C for TOC or DOC and ≥ 720 °C for TN_b or DN_b determinations.

The TOC or DOC determination is carried out in accordance with the direct measurement method.

Prior to combustion, remove inorganic carbon by acidification and purging with a carrier gas (6.7).

NOTE Platinum and cerium(IV), for example, can be used as catalyst material for combustion. The catalyst serves to accelerate the oxidation process of carbon containing water constituents in excess of oxygen to produce the required carbon dioxide gas for the detection process. Depending on combustion temperature and temperatures in the combustion zone, different catalysts can be used, e.g. metals or metal oxides for temperatures > 680 °C or sintered Alumina for temperatures around 1 200 °C, according to verifications of different suppliers.

Oxidation of organic carbon (TOC, DOC) with oxygen or synthetic air to carbon dioxide. Detection by means of infrared spectrometry (IR). Combustion of inorganic and organic nitrogen with oxygen or synthetic air and conversion to nitric oxide.

Reaction with ozone giving electronically excited nitrogen oxides. Detection by means of chemiluminescence (CLD) (see Annex C for alternative detection).

This document can be applied for the determination of TOC or DOC and TN_b or DN_b separately or for simultaneous TOC or DOC and TN_b or DN_b determinations, for example connecting the IR detector with a CLD in series.

Quality control is necessary to check the validity of the calibration function (see 10.3). Replicate determinations can be necessary. The method of standard addition can be required if matrix interferences are expected (see 5.3 and 10.4.2.1).

5 Interferences

iTeh STANDARD PREVIEW
(standards.iteh.ai)

5.1 General

Interferences with the determination of TOC or DOC and TN_b or DN_b can arise from memory effects. Replicate injections are necessary (see 10.4.1).

Detergents, oils and fats can influence the surface tension of the sample, causing erroneous data. A dilution of the sample can reduce such risk.

Samples with extreme pH values, highly buffered samples and samples with high salt contents can cause interference. Seek advice from the manufacturer to solve these interferences.

Suspended material can lead to a loss of quality of the analytical result. If a homogenized sample containing suspended material produces results obtained from replicate measurements that deviate by more than 10 %, an accurate TOC or TN_b result cannot be obtained on the sample.

5.2 TOC or DOC

Inorganic carbon (e.g. CO₂ or ions of carbonic acid) present in the sample interferes with the determination of TOC or DOC. Inorganic carbon is removed by acidification and purging with a gas that is free from CO₂ and organic compounds prior to the TOC or DOC determination (see 10.4.2.2 and 10.4.2.3).

NOTE 1 Alternatively, the differential method determining the TC and TIC separately can be applied (see Annex A). The TOC can be calculated by subtracting TIC from the TC. This calculation leads to correct results only as long as carbon monoxide, cyanide, cyanate and thiocyanate are present with negligible concentrations.

NOTE 2 Purgeable organic carbon substances, such as benzene, toluene, cyclohexane and chloroform, can partly escape upon stripping (see 10.4.2.2 and 10.4.2.3). In the presence of these substances, the TOC concentration can be determined separately, for example by applying the differential method (see Annex A).