
Varnostne zahteve za električno opremo za meritve, nadzor in laboratorijsko uporabo - 2-101. del: Posebne zahteve za diagnostično in vitro (IVD) medicinsko opremo

Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Safety requirements for in vitro diagnostic (IVD) medical equipment

Sicherheitsbestimmungen für elektrische Mess-, Steuer-, Regel- und Laborgeräte - Teil 2-101: Besondere Anforderungen an In-vitro-Diagnostik (IVD) Medizingeräte

(standards.iteh.ai)

Exigences de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire - Partie 2-101 : Exigences particulières pour le matériel médical de diagnostic in vitro (DIV)

Ta slovenski standard je istoveten z: prEN 61010-2-101:2017/prA11

ICS:

11.100.10	Diagnostični preskusni sistemi in vitro	In vitro diagnostic test systems
19.080	Električno in elektronsko preskušanje	Electrical and electronic testing
71.040.10	Kemijski laboratoriji. Laboratorijska oprema	Chemical laboratories. Laboratory equipment

oSIST prEN 61010-2-101:2018/oprA11:2021

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN 61010-2-101:2018/oprA11:2021](https://standards.iteh.ai/catalog/standards/sist/84262efb-52e2-4c89-a1d7-87bcdfe23c43/osist-pren-61010-2-101-2018-opra11-2021)
<https://standards.iteh.ai/catalog/standards/sist/84262efb-52e2-4c89-a1d7-87bcdfe23c43/osist-pren-61010-2-101-2018-opra11-2021>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

DRAFT
prEN IEC 61010-2-101:2017
prA11

October 2021

ICS 11.040.55; 19.080

English Version

Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Safety requirements for in vitro diagnostic (IVD) medical equipment

Exigences de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire - Partie 2-101 :
Exigences particulières pour le matériel médical de diagnostic in vitro (DIV)

Sicherheitsbestimmungen für elektrische Mess-, Steuer-, Regel- und Laborgeräte - Teil 2-101: Besondere Anforderungen an In-vitro-Diagnostik (IVD) Medizingeräte

This draft amendment prA11, if approved, will modify the European Standard prEN 61010-2-101:2017; it is submitted to CENELEC members for enquiry.

Deadline for CENELEC: 2022-01-21.

It has been drawn up by CLC/TC 66X.

If this draft becomes an amendment, CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this amendment the status of a national standard without any alteration.

This draft amendment was established by CENELEC in three official versions (English, French, German).

A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

prEN IEC 61010-2-101:2017/prA11 (E)

European foreword

This document (prEN IEC 61010-2-101:2017/prA11) has been prepared by CLC/TC 66X "Safety of measuring, control, and laboratory equipment".

This document is currently submitted to the Enquiry.

The following dates are proposed:

- latest date by which the existence of this document has to be announced at national level (doa) dor + 6 months
- latest date by which this document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) dor + 12 months
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) dor + 36 months (to be confirmed or modified when voting)

This document will amend prEN IEC 61010-2-101:2017.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN 61010-2-101:2018/oprA11:2021
https://standards.iteh.ai/catalog/standards/sist/84262efb-52e2-4c89-a1d7-87bcdfe23c43/osist-pren-61010-2-101-2018-opra11-2021](https://standards.iteh.ai/catalog/standards/sist/84262efb-52e2-4c89-a1d7-87bcdfe23c43/osist-pren-61010-2-101-2018-opra11-2021)

1 Modifications to 1.1.1, "Equipment included in scope"

Replace the title as follows:

"1.1.1 General"

Replace the second paragraph with the following:

"This part of IEC 61010 provides particular safety requirements to equipment intended for in vitro diagnostic (IVD) medical purposes, including self-test IVD medical purposes. It is intended to be used in conjunction of the manufacturer's risk management but not to replace it.

NOTE 1 A good design practice of an equipment starts from the beginning with a risk management process according to ISO 14971, which provides requirement and guidance for a comprehensive risk management process and identifies hazards and risks related with the equipment."

Replace the note with the following:

"

NOTE 2 A system, as specified by its manufacturer, is a combination of items of equipment, at least one of these is interconnected to another item. In the following text the term equipment is used for single equipment and systems.

It is possible that all or part of the equipment falls within the scope of one or more other Part 2 standards of IEC 61010 as well as within the scope of this document. In that case, the requirements of those other Part 2 standards will also apply."

2 Modifications to 1.1.2, "Equipment excluded from scope"

Replace the title as follows:

"1.1.2 Exclusions from the scope"

3 Modifications to 1.2.1, "Aspects included in scope"

Replace the first paragraph with the following:

The purpose of the requirements of this document is to ensure that HAZARDS to the OPERATOR, the SERVICE PERSONNEL and the surrounding area are reduced to a tolerable level."

Add the following item to the list:

"cc) any other energy sources (see Clause 101)"

4 Modifications to 1.2.2, "Aspects excluded from scope"

Delete items b) and c).

5 Modifications to Clause 2, "Normative references"

Add the following references:

"

EN IEC 61326-2-6:—¹, *Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment*

¹ To be published.

prEN IEC 61010-2-101:2017/prA11 (E)

EN 61326-3-1, *Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 3-1: Immunity requirements for safety-related systems and for equipment intended to perform safety-related functions (functional safety) - General industrial applications*

EN 62366-1, *Medical devices - Part 1: Application of usability engineering to medical devices*

ISO 13849-1, *Safety of machinery - Safety-related parts of control systems - Part 1: General principles for design*

"

6 Modifications to Clause 3.2, "Parts and accessories"

Add the following new term and definition.

"

3.2.201**CONTROL SYSTEM**

all the parts of the equipment forming a system to provide, for example, operational control, monitoring, interlocking, communications, protection or safety-related control functions

NOTE 1 to entry: These parts include electrical, electronic and programmable electronic parts and devices as well as the mechanical parts.

NOTE 2 to entry: Safety-related control functions can be performed by a control system that is either integral to or independent of those parts of a control system that performs non-safety-related functions.

"

(standards.iteh.ai)

7 Modifications to Clause 3.5, "Safety terms"

oSIST prEN 61010-2-101:2018/oprA11:2021

Add the following new term and definition.
<https://standards.iteh.ai/catalog/standards/sist/84262efb-52e2-4c89-a1d7-876cd1c23c43/osist-pren-61010-2-101-2018-opra11-2021>

"


3.5.201**SERVICE PERSONNEL**

person who is installing, changing or repairing the equipment, with the appropriate technical training, experience and awareness of HAZARDS and of measures to minimize danger to himself/herself, other persons or to the equipment.

"

8 Modifications to Table 1 in 5.1.3, "Mains supply"

Add the following item to Table 1:

201		ISO 7010 – W010 (modified)	Caution, risk of frostbite
-----	---	----------------------------	----------------------------

9 Modification to Clause 5.1, "Marking"

Add the following new subclause:

"5.1.201 Limited lifetime

If equipment has a limited lifetime it shall be marked on the equipment. The information can be given in time, date, or cycles whatever is appropriate for the equipment."

10 Modification to Clause 5.4,3 “Equipment installation”

Add the following item and note to the list:

"zz) Minimum hardware, software, and other related requirements to operate the equipment as intended and to connect it to a network to ensure safe operation.

NOTE Guidance can be found in IEC 80001-1."

11 Modifications to Clause 7, "Protection against mechanical HAZARDS"

In 7.1 “General”:

Add the following items to the list:

“

aaa) being trapped inside the equipment (see 7.201).

bbb) slipping, tripping or falling (see 7.202).

ccc) hand transmitted vibrations (see 7.203)

“

Replace the conformity statement with the following text:

“Conformity is checked as specified in 7.2 to 7.7 and 7.201 to 7.203.”

In 7.5.1 “General”:

Replace the first paragraph with the following text:

“Equipment or parts having a mass of 18 kg or more shall be provided with a means for lifting and carrying, or directions shall be given in the documentation to ensure safe transport, storage, installation, and disposal (also see 8.101).“

Add the following new subclauses

“

7.201 Prevention of being trapped in the equipment

If risk evaluation shows that the OPERATOR or professional SERVICE PERSONNEL can be exposed to a HAZARD of being trapped inside the equipment during normal use or during professional service, the design shall ensure that

- this situation is eliminated (inherent safety) or
- the trapped person is able to free herself or
- the trapped person can set an alarm to get help from another person.

In addition the instructions for use shall have a respective safety instruction.

Conformity is checked by inspection of the design and the instructions for use.

7.202 Prevention of slipping, tripping or falling

Surfaces belonging to the equipment on which OPERATOR or SERVICE PERSONNEL need to stand or step to carry out their duties shall be equipped with a non-slip surface compatible with normal work footwear. Where such surfaces are elevated by more than 0,3 m from the room floor or are narrower than 0,3 m (that is to say, they don't allow the whole foot to rest flat on the surface without twisting) handholds or handrails shall be provided. Any steps built into the equipment shall provide for adequate engagement of the foot on the step. A

“Cold touchable surfaces in normal condition shall not be below the values of Table 201.

Table 201 — Cold surface temperature limits in NORMAL CONDITION

Part	Limit °C
1 Touchable outer surface of enclosure (unintentional contact)	
a) metal, uncoated or anodized	-20
b) metal, coated (paint, non-metallic)	-30
c) plastics, plastic films, labels	-40
d) glass and ceramics	-30
2 Knobs (NORMAL USE contact for short periods up to 4s)	
a) metal	2
b) plastics	-40
c) glass and ceramics	-20
3 Handles (NORMAL USE contact for periods up to 120s)	5
NOTE ISO 13732-3 gives information about the effect of the duration of contact.	

(standards.iteh.ai)

Exception for temperatures lower than the values of Table 201 in NORMAL CONDITION and lower than -40°C in SINGLE FAULT CONDITION are permitted, provided that they are recognizable as such by appearance or function or are marked with symbol 201 (ISO 7010 – W010) of Table 1. Equipment cooled by its environment to temperature values lower than the values in Table 201 in NORMAL CONDITION and -40°C in SINGLE FAULT CONDITION need not to be marked with symbol 201 (ISO 7010 – 1818 W010).

Conformity is checked by measurements as specified in 10.4, and by inspection of barriers to check that they protect against accidentally touching surfaces that are at temperatures above the values of Table 19, or below the values of Table 201, and that they cannot be removed without a TOOL.

"

14 Modifications to Clause 12, "Protection against radiation, including laser sources, and against sonic and ultrasonic pressure"

In 12.1 "General":

Replace the first sentence with the following text:

“The equipment shall provide protection against the effects of internally generated ultraviolet, ionizing and microwave radiation, laser sources, sonic and ultrasonic pressure, and electromagnetic radiation.

NOTE Only safety aspects caused by the equipment are covered by this clause. General requirements for electromagnetic compatibility are excluded.

Conformity is checked as specified in 12.2 to 12.6 and 12.202.”

In 12.5.1 "Sound level":

Replace the first two paragraphs and note 1 and note 2 with the following text:

prEN IEC 61010-2-101:2017/prA11 (E)

"If the equipment can produce noise at a level which could cause harm to the OPERATOR, the manufacturer shall measure the maximum sound pressure level which the equipment can produce in NORMAL USE. The maximum sound pressure level does not apply to alarms and sound sources remote from the equipment.

If no level exceeds 70 dB(A) above a reference sound pressure of 20 μ Pa, this fact shall be stated in the documentation. For equipment which produces a maximum sound pressure level >70 dB(A) above a reference sound pressure of 20 μ Pa the following exposure threshold values limits shall not be exceeded.

$L_{EX,8h} = 80$ dB(A) and $L_{pC,peak} = 135$ dB(C) respectively

NOTE The daily noise exposure level ($L_{EX,8h}$) is the time-weighted average of the noise exposure levels on the basis of an eight-hour shift. It comprises all noise events occurring at the workplace. Attention is drawn to the RESPONSIBLE BODY that installation of equipment in a workplace with additional noise sources can create a HAZARDOUS SITUATION. The peak sound pressure level ($L_{pC,peak}$) is the maximum value of the instantaneous sound pressure level.

When applying the limit values, the attenuation provided by individual hearing protection worn by the operator shall not be considered.

Technical measures shall take precedence over organizational measures and shall have priority over the use of Personal Protection Equipment."

and replace the first sentence of the conformity statement with the following text:

"Conformity is checked by measuring the maximum A-weighted sound pressure level at the OPERATOR position and at bystander positions, but not further than 1 m from the equipment surfaces."

Add the following new subclauses (to Clause 12):

"

12.201 Safety aspect of electromagnetic disturbances

The protection for safety shall not be impaired by electromagnetic disturbances. The equipment shall comply with EN IEC 61326-2-6:—¹.

For equipment intended to be used in professional healthcare facility environment, the immunity requirements table 101 and 103 of EN IEC 61326-2-6:—¹ shall be applied.

For equipment intended to be used in home healthcare environment, the immunity requirements table 102, 103, 104 of EN IEC 61326-2-6:—¹ shall be applied.²

Conformity is checked by inspection of the EN IEC 61326-2-6:—¹ test documentation.

12.202 Safety aspect of electromagnetic emissions

EQUIPMENT that includes components which emits intentionally electromagnetic fields shall be safe for OPERATORS and SERVICE PERSONNEL. The emission requirements of EN IEC 61326-2-6 shall be applied.

NOTE 1 Such components could be wireless communication devices like cell phone, WLAN, RFID, Bluetooth and the like.

NOTE 2 Wireless devices fall under the RE directive 2014/53/EU and their harmonized standards.

Conformity is checked by inspection of the EN IEC 61326-2-6 test documentation.

"

15 Modifications to Clause 13, "Protection against liberated gases and substances, explosion and implosion"

Add the following new subclause:

² Editorial note regarding EN IEC 61326-2-6:—: The tables referenced in this paragraph refer to the next edition of EN IEC 61326-2-6 in preparation, expected to be published before the publication of this document.