



SLOVENSKI STANDARD
SIST-TS CEN/TS 17759:2023

01-februar-2023

Anorganska gnojila - Določanje pH-vrednosti raztopine gnojil iz amonijevega nitrata z veliko vsebnostjo dušika

Inorganic fertilizers - Determination of pH of a solution of ammonium nitrate fertilizers of high nitrogen content

Anorganische Düngemittel - Bestimmung des pH-Wertes in einer Lösung mit Ammoniumnitratdüngemitteln mit hohem Stickstoffgehalt

Engrais inorganiques - Détermination du pH d'une solution d'engrais à base de nitrate d'ammonium à forte teneur en azote

Ta slovenski standard je istoveten z: CEN/TS 17759:2022

ICS:

65.080 Gnojila Fertilizers

SIST-TS CEN/TS 17759:2023 en,fr,de

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

CEN/TS 17759

April 2022

ICS 65.080

English Version

Inorganic fertilizers - Determination of pH of a solution of ammonium nitrate fertilizers of high nitrogen content

Engrais inorganiques - Détermination du pH d'une solution d'engrais à base de nitrate d'ammonium à forte teneur en azote

Anorganische Düngemittel - Bestimmung des pH-Wertes in einer Lösung mit Ammoniumnitratdüngemitteln mit hohem Stickstoffgehalt

This Technical Specification (CEN/TS) was approved by CEN on 13 March 2022 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

[SIST-TS CEN/TS 17759:2023](https://standards.iteh.ai/catalog/standards/sist/3e431783-c0eb-4854-ab87-44e1ddf83d05/sist-ts-cen-ts-17759-2023)

<https://standards.iteh.ai/catalog/standards/sist/3e431783-c0eb-4854-ab87-44e1ddf83d05/sist-ts-cen-ts-17759-2023>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents		Page
European foreword		3
1	Scope	4
2	Normative references	4
3	Terms and definitions	4
4	Principle	4
5	Reagents	4
6	Apparatus and equipment	5
7	Sampling and sample preparation	5
8	Procedure	5
8.1	Calibration of the pH meter	5
8.2	Determination	5
9	Expression of the results	5
10	Test report	5
Bibliography		6

ITeH STANDARD PREVIEW
(standards.iteh.ai)

[SIST-TS CEN/TS 17759:2023](https://standards.iteh.ai/catalog/standards/sist/3e431783-c0eb-4854-ab87-44e1ddf83d05/sist-ts-cen-ts-17759-2023)

<https://standards.iteh.ai/catalog/standards/sist/3e431783-c0eb-4854-ab87-44e1ddf83d05/sist-ts-cen-ts-17759-2023>

European foreword

This document (CEN/TS 17759:2022) has been prepared by Technical Committee CEN/TC 260 “Fertilizers and liming materials” the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association.

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

(standards.iteh.ai)

[SIST-TS CEN/TS 17759:2023](https://standards.iteh.ai/catalog/standards/sist/3e431783-c0eb-4854-ab87-44e1ddf83d05/sist-ts-cen-ts-17759-2023)

<https://standards.iteh.ai/catalog/standards/sist/3e431783-c0eb-4854-ab87-44e1ddf83d05/sist-ts-cen-ts-17759-2023>

CEN/TS 17759:2022 (E)**1 Scope**

This document specifies a method for the determination of pH of a solution of ammonium nitrate fertilizer of high nitrogen content.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1482-2, *Fertilizers and liming materials — Sampling and sample preparation — Part 2: Sample preparation*

EN 12944-1, *Fertilizers and liming materials — Vocabulary — Part 1: General terms*

EN 12944-2, *Fertilizers and liming materials — Vocabulary — Part 2: Terms relating to fertilizers*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944-1 and EN 12944-2 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— IEC Electropedia: available at <https://www.electropedia.org/>

— ISO Online browsing platform: available at <https://www.iso.org/obp>

4 Principle

The measurement of the pH of an ammonium nitrate solution is carried out by means of a pH meter.

5 Reagents

Use only reagents of recognized analytical grade.

5.1 Distilled or demineralized water, free from carbon dioxide.

5.2 Buffer solution A, pH 6,88 at 20 °C or buffer solution B pH 4,00 at 20 °C.

For buffer solution A, dissolve 3,40 g ± 0,01 g of potassium dihydrogen phosphate (KH₂PO₄) in approximately 400 ml of water. Then dissolve 3,55 g ± 0,01 g of disodium hydrogen phosphate (Na₂HPO₄) in approximately 400 ml of water (5.1). Transfer the two solutions without loss into a 1 000-ml graduated flask (6.2), fill up to the mark and mix.

Keep this solution in an airtight vessel.

For buffer solution B, dissolve 10,21 g ± 0,01 g of potassium hydrogen phthalate (KHC₈O₄H₄) in water (5.1), transfer without loss into a 1 000-ml graduated flask (6.2), fill up to the mark and mix.

Keep this solution in an airtight vessel.

Alternatively, commercially available pH standard solutions may be used.

6 Apparatus and equipment

Usual laboratory glassware and equipment and, in particular, the following.

- 6.1 **pH meter**, equipped with glass and calomel electrodes or equivalent, sensitivity 0,05 pH unit.
- 6.2 **Graduated flask**, capacity 1 000 ml.
- 6.3 **Beaker**, capacity 250 ml.
- 6.4 **Balance**, capable of weighing to the nearest 0,01 g.
- 6.5 **Airtight vessel**.

7 Sampling and sample preparation

Sampling is not part of the method specified in this document. A recommended sampling method is given in EN 1482-1 [1].

Sample preparation shall be carried out in accordance with EN 1482-2.

8 Procedure

8.1 Calibration of the pH meter

Calibrate the pH meter (6.1) at a temperature of $20\text{ °C} \pm 1\text{ °C}$, using the buffer solutions (5.2). Pass a slow stream of nitrogen onto the surface of the solution and maintain this throughout the test.

8.2 Determination

Pour 100,0 ml of water onto $10\text{ g} \pm 0,01\text{ g}$ of the sample in a 250 ml beaker (6.3). Remove the insolubles by filtering, decanting or centrifuging the liquid. Measure the pH of the clear solution at a temperature of $20\text{ °C} \pm 1\text{ °C}$ according to the same procedure as for the calibration of the pH meter.

9 Expression of the results

Express the result in pH units, to the nearest 0,1 unit, and state the temperature used.

10 Test report

The test report shall contain at least the following information:

- a) all information necessary for the complete identification of the sample;
- b) test method used with reference to this document, CEN/TS 17759:2022;
- c) test result obtained expressed in pH units, to the nearest 0,1 unit, and state the temperature used;
- d) date of sampling and sampling procedure (if known);
- e) date when the analysis was finished;
- f) all operating details not specified in this document, or regarded as optional, together with details of any incidents that occurred when performing the method which might have influenced the test result(s).