



Designation: D3227 – 04a(Reapproved 2010)



Designation: 342/00

Standard Test Method for (Thiol Mercaptan) Sulfur in Gasoline, Kerosine, Aviation Turbine, and Distillate Fuels (Potentiometric Method)¹

This standard is issued under the fixed designation D3227; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope

1.1 This test method covers the determination of mercaptan sulfur in gasolines, kerosines, aviation turbine fuels, and distillate fuels containing from 0.0003 to 0.01 mass % of mercaptan sulfur. Organic sulfur compounds such as sulfides, disulfides, and thiophene, do not interfere. Elemental sulfur in amounts less than 0.0005 mass % does not interfere. Hydrogen sulfide will interfere if not removed, as described in 9.2.

1.2 The values in acceptable SI units are to be regarded as the standard. The values in parentheses are for information only.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. For specific warning statements, see Sections 6, 8, 9, and Appendix X1.*

2. Referenced Documents

2.1 *ASTM Standards:*²

D1193 Specification for Reagent Water

D1250 Guide for Use of the Petroleum Measurement Tables

D1298 Test Method for Density, Relative Density (Specific Gravity), or API Gravity of Crude Petroleum and Liquid Petroleum Products by Hydrometer Method

D4052 Test Method for Density, Relative Density, and API Gravity of Liquids by Digital Density Meter

D4057 Practice for Manual Sampling of Petroleum and Petroleum Products

D4177 Practice for Automatic Sampling of Petroleum and Petroleum Products

D6299 Practice for Applying Statistical Quality Assurance and Control Charting Techniques to Evaluate Analytical Measurement System Performance

3. Summary of Test Method

3.1 The hydrogen sulfide-free sample is dissolved in an alcoholic sodium acetate titration solvent and titrated potentiometrically with silver nitrate solution, using as an indicator the potential between a glass reference electrode and a silver/silver-sulfide indicating electrode. Under these conditions, the mercaptan sulfur is precipitated as silver mercaptide and the end point of the titration is shown by a large change in cell potential.

4. Significance and Use

4.1 Mercaptan sulfur has an objectionable odor, an adverse effect on fuel system elastomers, and is corrosive to fuel system components.

5. Apparatus

5.1 As described in 5.2-5.5; alternatively, any automatic titration system may be used that, using the same electrode pair described in 5.3, is capable of performing the titration as described in Section 9 and selecting the endpoint specified in 11.1 with a precision that meets or is better than that given in Section 14.

5.2 *Meter*—An electronic voltmeter, operating on an input of less than 9×10^{-12} A and having a sensitivity of ± 2 mV over a range of at least ± 1 V. The meter shall be electrostatically shielded, and the shield shall be connected to the ground.³

5.3 *Cell System*, consisting of a reference and indicating electrode. The reference electrode should be a sturdy, pencil-type glass electrode, having a shielded lead connected to ground. The indicating electrode shall be made from a silver

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.03 on Elemental Analysis.

Current edition approved May 1, 2010. Published May 2010. Originally approved in 1973. Last previous edition approved in 2004 as D3227-04A. DOI: 10.1520/D3227-04R10.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Any apparatus that will give equal or better precision is acceptable.

wire, 2 mm (0.08 in.) in diameter or larger, mounted in an insulated support. Silver billet electrodes can also be used.

5.4 *Buret*, 10-mL capacity, graduated in 0.05-mL intervals, with a tip that extends approximately 120 mm (5 in.) below the stopcock.

5.5 *Titration Stand*, preferably built as an integral part of the meter housing and provided with supports for the electrodes and electrical stirrer, all connected to ground. No permanent change in meter reading should be noticeable upon connecting or disconnecting the stirring motor.

6. Reagents and Materials

6.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁴ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

6.1.1 Commercially available solutions may be used in place of prepared laboratory solutions, when they are certified to meet the required concentrations.

6.1.2 Alternate volumes of solutions and solvents may be prepared, when an equivalent concentration is maintained.

6.2 *Water*—Reagent grade, Type I, Specification **D1193**.

6.3 *Cadmium Sulfate, Acid Solution* (150 g/L)—Dissolve 150 g of cadmium sulfate ($3\text{CdSO}_4 \cdot 8\text{H}_2\text{O}$) in water. (**Warning**—Poison. May be fatal if swallowed or inhaled. A known carcinogen (animal positive).) Add 10 mL of dilute H_2SO_4 (**Warning**—Poison. Causes severe burns. Harmful or fatal if swallowed or inhaled) and dilute to 1 L with water.

6.4 *Potassium Iodide, Standard Solution* (approximately 0.1 mol/L)—Dissolve 17 g of KI (weigh to 0.01 g) in 100 mL of water in a 1-L volumetric flask and dilute to 1 L. Calculate the exact molarity.

6.5 *Propan-2-ol*—(**Warning**—Flammable. **Warning**—Unless inhibited against it, peroxides can form in propan-2-ol when stored in the same container that is exposed to air. When this happens and the propan-2-ol evaporates to dryness, an explosion can occur. When peroxides are suspected, they may be removed by percolation through an activated alumina column.)

6.6 *Silver Nitrate, Standard Alcoholic Solution* (0.1 mol/L)—Dissolve 17 g of AgNO_3 in 100 mL of water in a 1-L volumetric flask and dilute to 1 L with propan-2-ol (99 %) (see **Note 1**). Store in a dark bottle and standardize at intervals frequent enough to detect a change of 0.0005 or greater in molarity.

NOTE 1—It is important to pass the propan-2-ol through a column of activated alumina to remove peroxides that may have formed on storage; failure to remove peroxides will lead to low results. It is not necessary to perform this step if the alcohol is tested and found free of peroxides.

6.6.1 *Standardization*—Add six drops of concentrated HNO_3 (relative density 1.42) (**Warning**—Poison. Causes severe burns. Harmful or fatal if swallowed or inhaled) to 100 mL of water in an appropriately sized beaker (for example, a 200-, 250-, or 300-mL size beaker is typically large enough). Remove oxides of nitrogen by boiling for 5 min. Cool to ambient temperature. Pipet 5 mL of 0.1 mol/L KI solution into the beaker and titrate with the AgNO_3 solution choosing the end point at the inflection of the titration curve.

6.7 *Silver Nitrate, Standard Alcoholic Solution* (0.010 mol/L)—Prepare daily when the test is being performed by dilution of the 0.1 N standard. Pipet 100 mL of the 0.1 mol/L standard into a 1-L volumetric flask and dilute to volume with propan-2-ol. Calculate the exact molarity.

6.8 *Sodium Sulfide Solution* (10 g/L)—Dissolve 10 g of Na_2S in water and dilute to 1 L with water. Prepare fresh as needed.

6.9 *Sulfuric Acid*, dilute. Cautiously dilute five volumes of water with one volume of sulfuric acid (relative density 1.84). (**Warning**—Adding the acid will generate heat: mix well. If water begins to boil, cool before adding more acid.) Note that only limited volumes are required because only 10 mL are needed for each litre of cadmium sulfate solution.

6.10 *Titration Solvent*—Low molecular weight mercaptans, as usually found in gasoline, are readily lost from the titration solution if an acidic titration solvent is used. For the determination of the higher molecular weight mercaptan as normally encountered in kerosines, aviation turbine fuels, and distillate fuels, the acidic titration solvent is used to achieve more rapid equilibrium between successive additions of the titrant.

6.10.1 *Alkaline Titration Solvent*—Dissolve 2.7 g of sodium acetate trihydrate ($\text{NaC}_2\text{H}_3\text{O}_2 \cdot 3\text{H}_2\text{O}$) or 1.6 g of anhydrous sodium acetate ($\text{NaC}_2\text{H}_3\text{O}_2$) in 25 mL of water free of dissolved oxygen and pour into 975 mL of propan-2-ol (99 %) (**Note 1**). When necessary, remove dissolved oxygen by purging the solution with a rapid stream of nitrogen for 10 min each day prior to use; keep protected from the atmosphere. To minimize oxygen from dissolving in the solution during storage, an option exists to nitrogen blanket the solution prior to sealing the solvent container.

6.10.2 *Acidic Titration Solvent*—Dissolve 2.7 g of $\text{NaC}_2\text{H}_3\text{O}_2 \cdot 3\text{H}_2\text{O}$ or 1.6 g of $\text{NaC}_2\text{H}_3\text{O}_2$ in 20 mL of water free of dissolved oxygen and pour into 975 mL of propan-2-ol (99 %) (**Note 1**) and add 4.6 mL of glacial acetic acid. When necessary, remove dissolved oxygen by purging the solution with a rapid stream of nitrogen for 10 min each day prior to use; keep protected from the atmosphere. To minimize oxygen from dissolving in the solution during storage, an option exists to nitrogen blanket the solution prior to sealing the solvent container.

6.11 *Polishing Paper or Cloth*, 6–20 μm average particle size abrasive.

⁴ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For Suggestions on the testing of reagents not listed by the American Chemical Society, see *Annual Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

7. Sampling

7.1 Take the sample in accordance with Practice **D4057** or Practice **D4177**.

7.2 Methanethiol (methyl mercaptan) has a boiling point of 6.2°C and may be expected to be present in light untreated gasolines. Therefore, when the presence of this low boiling thiol (mercaptan) is known or expected, specimen to be tested shall be kept below 4°C to prevent the loss of mercaptan from such samples.

8. Preparation of Apparatus

8.1 *Glass Electrode*—After each manual titration, or batch of titrations, in the case of automatic titration systems, wipe the electrode with a soft, clean tissue and rinse with water. Clean the electrode at frequent intervals (at least once a week) by stirring in cold chromic acid solution (**Warning**—Causes severe burns. A recognized carcinogen. Strong oxidizer—contact with other material may cause fire. Hygroscopic. An equivalent, chromium-free cleaning solution may be used) for a few seconds (10 s maximum). When not in use, keep lower half of the electrode immersed in water.

8.2 *Silver/Silver-Sulfide Electrode*—Each day prior to use, or as triggered by the analysis of a quality control (QC) sample (see Section 12), prepare a fresh silver sulfide coating on the electrode by the following method:

8.2.1 Burnish electrode with polishing paper or cloth until a clean, polished silver surface shows.

8.2.2 Place electrode in operating position and immerse it in 100 mL of titration solvent containing 8 mL of Na₂S solution.

8.2.3 Add slowly from a buret, with stirring, 10 mL of 0.1 mol/L AgNO₃ solution over a period from 10 to 15 min.

8.2.4 Remove electrode from solution, wash with water, and wipe with a soft, clean tissue.

8.2.5 Between manual titrations, or batches of titrations in the case of automatic titration systems, store the electrode a minimum of 5 min in 100 mL of titration solvent containing 0.5 mL of the 0.1 mol/L AgNO₃ solution.

9. Procedure

9.1 *Determination of Density*—If the sample is to be measured volumetrically, determine the density by Test Method **D1298** or Test Method **D4052** at the temperature at which the test portion will be taken, either directly or from the density determined at a reference temperature and converted to the transfer temperature by use of the Petroleum Measurement Tables (Guide **D1250**).

9.2 *Removal of Hydrogen Sulfide*—Test the sample qualitatively for hydrogen sulfide (H₂S) by mixing 5 mL of the sample with 5 mL of the acid CdSO₄ solution. If no precipitate appears, proceed with the analysis of the sample as described in 9.3. If a yellow precipitate appears, remove the H₂S in the following manner: Place a quantity of the sample, three to four times that required for the analysis, in a separatory funnel containing a volume of the acid CdSO₄ solution equal to one half that of the sample and shake vigorously. Draw off and discard the aqueous phase containing the yellow precipitate. Repeat the extraction with another portion of the CdSO₄

solution. Again draw off the aqueous phase, and wash the sample with three 25 to 30-mL portions of water, withdrawing the water after each washing. Filter the hydrocarbon through a rapid paper. Test a small portion of the washed sample, such as in a test tube or vial, with a few millilitres of the CdSO₄ solution as described previously. If no further precipitate is formed, proceed as directed in 9.3. If a precipitate appears, repeat the extraction with the CdSO₄ solution until all of the H₂S has been removed. (**Warning**—Certain straight run gasolines can contain significant amounts of both low molecular weight mercaptans and dissolved elemental sulfur, which, when heated under total reflux conditions, may produce H₂S by the inter-reaction of both. This phenomenon is particularly noticed in straight run gasolines produced from some natural gas condensates. Therefore, it is advisable that during the removal of H₂S (and after all H₂S has been extracted), that no heat should be applied to the sample.)

9.2.1 When the test results obtained are not for referee purposes and Quality Assurance/Quality Control (QA/QC) protocol permits, an alternative test for, and a procedure for the removal of, hydrogen sulfide can be used. This process uses lead acetate paper and sodium hydrogen carbonate and is described in **Appendix X1**.

9.2.2 There are available automated instruments that eliminate the necessity to remove H₂S when present, prior to performing the analysis, since the H₂S titration curve does not interfere with the titration curve of the thiol or mercaptan determination. Such equipment may be used when the test is not for referee purposes. The precision statement does not include data obtained from this alternate procedure.

9.2.3 For referee purposes, the protocol, as detailed in 9.2, shall be used. A possible non-referee situation may be the instance of routine refinery control.

9.3 Measure with a pipet (or other suitable transfer device) or weigh 20 to 50 mL of the original or treated sample into an appropriately sized beaker (for example, a 200, 250, or 300 mL size beaker is typically large enough) containing 100 mL of the appropriate titration solvent. Place the beaker on the titration stand or on the auto-sampler of an automatic titration system. If an automatic titration system is used, set up the system to reproduce the experimental conditions specified in 9.3.1-9.3.3. Adjust the position of the titration stand so that the electrodes are about half immersed. Fill the buret with 0.01 mol/L alcoholic AgNO₃ solution and position it in the titration assembly so that the tip extends approximately 25 mm (1 in.) below the surface of the liquid in the beaker. Adjust the speed of the stirrer to give vigorous stirring without spattering.

9.3.1 Record the initial buret and cell potential readings. The usual meter readings for mercaptan presence are in the -250 mV to -350 mV range. Add suitable small portions of 0.01 mol/L AgNO₃ solution and, after waiting until a constant potential has been established, record the buret and meter readings. Consider the potential constant if it changes less than 6 mV/min.

NOTE 2—If potential readings obtained with freshly prepared electrodes are erratic, it is possible that the electrodes are not properly conditioned. This difficulty usually disappears in succeeding titrations.