



# SLOVENSKI STANDARD SIST EN IEC 63305:2024

01-junij-2024

---

**Podvodna akustika - Kalibracija zvočnega vala vektorskih sprejemnikov v frekvenčnem območju od 5 Hz do 10 kHz (IEC 63305:2024)**

Underwater Acoustics - Calibration of acoustic wave vector receivers in the frequency range 5 Hz to 10 kHz (IEC 63305:2024)

Unterwasserakustik - Kalibrierung von Schallwellenvektorempfängern im Frequenzbereich 5 Hz bis 10 kHz (IEC 63305:2024)

Acoustique sous-marine - Étalonnage des récepteurs vectoriels d'ondes acoustiques dans la plage de fréquences de 5 Hz à 10 kHz (IEC 63305:2024)

**Ta slovenski standard je istoveten z: EN IEC 63305:2024**

[SIST EN IEC 63305:2024](https://standards.sist.si/catalog/standards/sist/63305-2024/iec-63305-2024)

<https://standards.sist.si/catalog/standards/sist/63305-2024/iec-63305-2024>

**ICS:**

17.140.50      Elektroakustika                      Electroacoustics

**SIST EN IEC 63305:2024**                      **en**



EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN IEC 63305**

April 2024

ICS 17.140.50

English Version

**Underwater acoustics - Calibration of acoustic wave vector  
receivers in the frequency range 5 Hz to 10 kHz  
(IEC 63305:2024)**

Acoustique sous-marine - Étalonnage des récepteurs  
vectoriels d'ondes acoustiques dans la plage de fréquences  
de 5 Hz à 10 kHz  
(IEC 63305:2024)

Unterwasserakustik - Kalibrierung von  
Schallwellenvektorempfängern im Frequenzbereich 5 Hz  
bis 10 kHz  
(IEC 63305:2024)

This European Standard was approved by CENELEC on 2024-04-02. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

[SIST EN IEC 63305:2024](https://standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024>



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## EN IEC 63305:2024 (E)

### European foreword

The text of document 87/839/FDIS, future edition 1 of IEC 63305, prepared by IEC/TC 87 "Ultrasonics" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 63305:2024.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2025-01-02
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2027-04-02

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

### Endorsement notice

The text of the International Standard IEC 63305:2024 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

IEC 60565-1:2020 NOTE Approved as EN IEC 60565-1:2020 (not modified)

IEC 60565-2:2019 NOTE Approved as EN IEC 60565-2:2019 (not modified)

<https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024>

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: [www.cencenelec.eu](http://www.cencenelec.eu).

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60500	2017	Underwater acoustics - Hydrophones - Properties of hydrophones in the frequency range 1 Hz to 500 kHz	EN 60500	2017
IEC 60565-1	2020	Underwater acoustics - Hydrophones - Calibration of hydrophones - Part 1: Procedures for free-field calibration of hydrophones	EN IEC 60565-1	2020
ISO 18405	2017	Underwater acoustics - Terminology	-	-
ISO 80000-8	2020	Quantities and units - Part 8: Acoustics	EN ISO 80000-8	2020
ISO/IEC Guide 98-3 -		Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)	-	-

[SIST EN IEC 63305:2024](https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024)

<https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024>





IEC 63305

Edition 1.0 2024-02

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



**Underwater acoustics – Calibration of acoustic wave vector receivers in the frequency range 5 Hz to 10 kHz**

**Acoustique sous-marine – Étalonnage des récepteurs vectoriels d'ondes acoustiques dans la plage de fréquences de 5 Hz à 10 kHz**

[SIST EN IEC 63305:2024](https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024)

<https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024>

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-8123-9

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope.....	8
2 Normative references .....	8
3 Terms and definitions .....	8
4 List of symbols .....	14
5 Relationship of vector quantities in sound field .....	16
6 General procedures for calibration.....	17
6.1 General calibration requirements .....	17
6.1.1 Types of calibration .....	17
6.1.2 Acoustic field requirements.....	17
6.2 Acoustic standing wave tube requirements.....	18
6.2.1 Requirements for standing wave tube [8] .....	18
6.2.2 Requirements for immersed depth of transducers .....	19
6.3 Acoustic travelling wave tube requirements.....	20
6.3.1 Requirements for driving signal .....	20
6.3.2 Requirements for the travelling wave tube .....	20
6.4 Equipment requirements .....	20
6.4.1 Calibration facility .....	20
6.4.2 Instrumentation.....	21
6.5 Positioning and alignment .....	23
6.5.1 Coordinate system.....	23
6.5.2 Reference direction .....	23
6.5.3 Transducer mounting and support.....	23
6.5.4 Alignment .....	24
6.6 Representation of the frequency response .....	25
6.7 Frequency limitations .....	25
6.7.1 High-frequency limit.....	25
6.7.2 Low frequency limit.....	25
6.8 Checks for acoustic interference .....	26
7 Electrical measurements.....	26
7.1 Signal type.....	26
7.2 Electrical earthing .....	26
7.3 Measurement of transducer output voltage.....	26
7.3.1 General .....	26
7.3.2 Signal analysis .....	27
7.3.3 Electrical loading by measuring instrument .....	27
7.3.4 Electrical loading by extension cables .....	27
7.3.5 Electrical noise .....	27
7.3.6 Cross-talk .....	28
7.3.7 Integral preamplifiers .....	28
7.4 Measurement of projector drive current.....	28
7.4.1 Instrumentation.....	28
7.4.2 Signal analysis .....	28
8 Preparation of measurement.....	28
8.1 Preparation of transducers.....	28



8.1.1	Soaking .....	28
8.1.2	Wetting .....	29
8.2	Environmental conditions (temperature and depth) .....	29
9	Free-field calibration .....	29
9.1	Free-field reciprocity calibration .....	29
9.1.1	General .....	29
9.1.2	Principle .....	30
9.1.3	Measurement .....	32
9.1.4	Uncertainty .....	32
9.2	Free-field calibration using optical interferometry .....	32
9.2.1	General .....	32
9.2.2	Principle .....	32
9.2.3	Measurement .....	33
9.2.4	Uncertainty .....	34
9.3	Free-field calibration using a reference hydrophone .....	34
9.3.1	General .....	34
9.3.2	Principle .....	34
9.3.3	Measurement .....	35
9.3.4	Uncertainty .....	35
10	Calibration in standing wave tube .....	35
10.1	Calibration using reference accelerometer .....	35
10.1.1	General .....	35
10.1.2	Principle .....	35
10.1.3	Measurement .....	37
10.1.4	Uncertainty .....	37
10.2	Comparison calibration using reference hydrophone in standing wave tube .....	37
10.2.1	General .....	37
10.2.2	Principle .....	37
10.2.3	Measurement .....	39
10.2.4	Uncertainty .....	39
10.3	Horizontal standing wave tube calibration .....	39
10.3.1	General .....	39
10.3.2	Principle .....	39
10.3.3	Measurement .....	41
10.3.4	Uncertainty .....	41
10.4	Calibration using optical interferometry in standing wave tube .....	41
10.4.1	General .....	41
10.4.2	Principle .....	41
10.4.3	Measurement .....	43
10.4.4	Uncertainty .....	43
11	Calibration in a travelling wave tube .....	43
11.1	General .....	43
11.2	Principle .....	44
11.2.1	General .....	44
11.2.2	Establishment of a unidirectional, plane progressive wave field .....	45
11.2.3	Sensitivity calculations .....	48
11.2.4	Uncertainty .....	48
12	Reporting of results .....	48

12.1	Sensitivity .....	48
12.2	Sensitivity level .....	49
12.3	Environmental considerations for calibration .....	49
12.4	Calibration uncertainties .....	49
12.5	Auxiliary metadata .....	49
13	Recalibration periods .....	50
Annex A (informative)	Directional response of a vector receiver .....	51
A.1	General principle .....	51
A.2	Types of measurement implementation .....	51
A.3	Coordinate system .....	51
A.4	Measurement of vector receiver directional response .....	51
A.5	Calculation of angular deviation loss .....	52
A.6	Uncertainty .....	52
Annex B (informative)	Inertial vector receiver calibration using optical interferometry in air .....	53
B.1	General .....	53
B.2	Principle .....	53
B.3	Procedure .....	53
B.4	Discussion .....	55
Annex C (informative)	Assessment of uncertainty of vector receiver calibration .....	56
C.1	General .....	56
C.2	Type A evaluation of uncertainty .....	56
C.3	Type B evaluation of uncertainty .....	56
C.4	Reported uncertainty .....	56
C.5	Common sources of uncertainty .....	57
Bibliography	.....	60
Figure 1	– The structure of the calibration chamber .....	19
Figure 2	– Co-vibrating vector receiver suspended on a mounting ring .....	24
Figure 3	– Measurement framework for free-field reciprocity calibration of the vector receiver .....	30
Figure 4	– Schematic diagram of free-field calibration for vector receiver using an optical interferometer .....	33
Figure 5	– Schematic diagram of free-field comparison calibration for vector receiver using reference hydrophone .....	34
Figure 6	– Schematic diagram of vertical standing wave tube calibration using reference accelerometer .....	36
Figure 7	– Schematic diagram of vertical standing wave tube calibration using reference hydrophone .....	38
Figure 8	– Schematic diagram of calibration principle and horizontal standing wave tube calibration .....	40
Figure 9	– Schematic diagram of calibration for vector receiver using optical interferometer in standing wave tube .....	42
Figure 10	– Schematic diagram of calibration for vector receiver in a travelling wave tube .....	44
Figure B.1	– Schematic diagram of calibration using optical interferometer in air for inertial vector receiver .....	54

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**UNDERWATER ACOUSTICS – CALIBRATION OF ACOUSTIC WAVE VECTOR RECEIVERS IN THE FREQUENCY RANGE 5 Hz TO 10 kHz**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 63305 has been prepared by IEC technical committee 87: Ultrasonics. It is an International Standard.

The text of this International Standard is based on the following documents:

Draft	Report on voting
87/839/FDIS	87/843/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

NOTE Words in **bold** in the text are defined in Clause 3.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

**IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## iTeh Standards (<https://standards.iteh.ai>) Document Preview

[SIST EN IEC 63305:2024](#)

<https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024>

## INTRODUCTION

Unlike traditional piezoelectric **hydrophones** which are sensitive to sound pressure, **vector receivers** measure **sound particle** motion (velocity, acceleration or displacement) or **sound pressure gradient**, and have strongly **directional response** in their working frequency range. The calibration of these **vector receivers** which measure **sound particle** motion or **sound pressure gradient** is considered in this document.

The output voltage of a **vector receiver** channel to be calibrated is proportional to the **sound particle** motion or **sound pressure gradient** at the reference centre of the receiver. The directivity of the **vector receiver** channel is independent of acoustical frequency, and the ratio of the output voltage of the receiver channel at angle  $\theta$  to the maximum output voltage on the axial direction is equal to  $|\cos\theta|$  [1]<sup>1</sup>.

Recent developments of **acoustic wave vector receivers** for ocean acoustics, such as those that measure **sound particle velocity**, have led to a number of commercial systems being made available on the market. In addition to providing sensors which possess some useful directivity for low-frequency applications, they are increasingly used for measurement of underwater noise exposure for marine fauna that are sensitive to sound particle motion rather than sound pressure (for example, fish and invertebrates). However, calibration of such sensors poses technical challenges, and is not covered by the existing international standards such as IEC 60565 [2], [3]. Building on work begun in China and Russia [4], where a successful bilateral comparison has recently been concluded, this work establishes an International Standard on calibration of **vector receivers** in the frequency range 5 Hz to 10 kHz.

# iTeh Standards (<https://standards.iteh.ai>) Document Preview

[SIST EN IEC 63305:2024](https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024)

<https://standards.iteh.ai/catalog/standards/sist/6ade758a-b626-442d-99b6-56d77d015f71/sist-en-iec-63305-2024>

---

<sup>1</sup> Numbers in square brackets refer to the Bibliography.

# UNDERWATER ACOUSTICS – CALIBRATION OF ACOUSTIC WAVE VECTOR RECEIVERS IN THE FREQUENCY RANGE 5 Hz TO 10 kHz

## 1 Scope

Usually, **acoustic wave vector receivers** are designed and constructed based on one of two principles. One is the sound pressure difference (gradient) principle. When measuring with this sensor, the **vector receiver** is rigidly fixed on a mount and supported in water. The other is the co-vibrating (**inertial**) principle. When measuring with this sensor, the **vector receiver** is suspended on a mount and supported in water in a non-rigid manner, which allows the **vector receiver** co-vibrate in the same direction as the **sound particle** in the sound wave field.

Many methods have been used to calibrate **vector receivers**, such as free-field calibration, calibration in standing wave tube and calibration in a travelling wave tube. This document specifies methods and procedures for calibration of **vector receivers** in the frequency range 5 Hz to 10 kHz, which are applicable to **vector receivers** based on the two different principles. In addition, it describes an absolute method of **inertial vector receiver** calibration in air using optical interferometry.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60500:2017, *Underwater acoustics – Hydrophones – Properties of hydrophones in the frequency range 1 Hz to 500 kHz*

IEC 60565-1:2020, *Underwater acoustics – Hydrophones – Calibration of hydrophones, Part 1: Procedures for free-field calibration of hydrophones*

ISO 80000-8:2020, *Quantities and units – Part 8: Acoustics*

ISO 18405:2017, *Underwater acoustics – Terminology*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60500:2017, IEC 60565-1:2020, ISO 80000-8:2020, ISO 18405:2017 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>