



Designation: ~~D2269-99 (Reapproved 2005)~~ Designation: D2269 – 10

Standard Test Method for Evaluation of White Mineral Oils by Ultraviolet Absorption¹

This standard is issued under the fixed designation D2269; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method describes a procedure for the examination and evaluation of NF and USP grade white mineral oils.

1.2 This test method is not applicable to oils containing additives soluble in dimethyl sulfoxide (DMSO) that exhibit fluorescence or fluorescence quenching properties.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific warning statements, see 7.1.1-7.1.3.

2. Referenced Documents

2.1 *ASTM Standards:*²

D1840 [Test Method for Naphthalene Hydrocarbons in Aviation Turbine Fuels by Ultraviolet Spectrophotometry](#)

E131 [Terminology Relating to Molecular Spectroscopy](#)

E275 [Practice for Describing and Measuring Performance of Ultraviolet and Visible Spectrophotometers](#)

2.2 *Other Standard:*

U.S. Pharmacopeia USP XIII/National Formulary (NF XVIII)³

3. Terminology

3.1 *Definitions:*

3.1.1 For definitions of terms and symbols relating to absorption spectroscopy see Terminology E131. Terms of particular significance are the following:

3.1.2 *radiant energy, n*—energy transmitted as electromagnetic waves.

3.1.3 *radiant power, P, n*—the rate at which energy is transported in a beam of radiant energy.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *absorbance, A, n*—the logarithm to the base 10 of the reciprocal of the transmittance, T . In symbols:

$$A = \log_{10} (1/T) = -\log_{10} T$$

where T is the transmittance as defined in 3.2.5.

3.2.2 *absorptivity, a, n*—the absorbance divided by the product of sample pathlength and concentration. In symbols:

$$a = A/bc$$

where A is the absorbance as defined in 3.2.1, b is the sample pathlength as defined in 3.2.4, and c is the concentration as defined in 3.2.3.

3.2.3 *concentration, c, n*—the quantity of sample expressed in grams per litre.

3.2.4 *sample pathlength, b, n*—the distance in centimetres, measured in the direction of propagation of the beam of radiant energy, between the surfaces of the specimen on which the radiant energy is incident and the surface of the specimen from which it is emergent.

3.2.5 *transmittance, T, n*—the ratio of the radiant power transmitted by the mineral oil extract in its cell to the radiant power transmitted by the solvent control in its cell. Expressed by:

$$T = P_e/P_c$$

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.04.0F on Hydrocarbon Analysis—Absorption Spectroscopic Methods.

Current edition approved June 1, 2005; 2010. Published August 2005; June 2010. Originally approved in 1964. Last previous edition approved in 1999; 2005 as ~~D2269-99~~^{ε1}. D2269-99(2005). DOI: 10.1520/D2269-99R05.10.1520/D2269-10.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from The United States Pharmacopeia (USP), 12601 Twinbrook Parkway, Rockville, MD 20852.

*A Summary of Changes section appears at the end of this standard.

where P_e is the radiant power transmitted by the mineral oil extract and P_c is the radiant power transmitted by the solvent control.

4. Summary of Test Method

4.1 A sample of oil is extracted with dimethyl sulfoxide and the ultraviolet absorbance of the extract is determined in the range from 260 to 350 nm. The absorbance is compared with that of a naphthalene standard.

5. Significance and Use

5.1 The ultraviolet absorption of white mineral oils is used to determine their suitability for use in food, drug, and cosmetic applications.

5.2 The U.S. Pharmacopeia and the National Formulary specifications for mineral oil require the measurement of ultraviolet absorption.

6. Apparatus

6.1 *Spectrophotometer*, equipped to handle liquid samples in 1-cm path length cells. The instrument shall be capable of measuring absorbance with a repeatability of $\pm 1.0\%$ or better from an average at the 0.4 absorbance level in the spectral region near 290 nm with a nominal band width of 1 nm or less.

NOTE 1—For recommended methods of testing spectrophotometers to be used in this test method refer to Practice E275.

6.2 *Fused Quartz Cells*, two, having path lengths of 1.00 ± 0.005 cm, or better. The distance in centimetres does not include the thickness of the cell in which the sample is contained.

6.3 *Separatory Funnels*, ~~glass-stoppered, 125-mL capacity, equipped with TFE-fluorocarbon stopcocks or other suitable stopcocks that will not contaminate the solvents used.~~, glass-stoppered, of sufficient capacity to perform the necessary extractions in the procedure, equipped with TFE-fluorocarbon stopcocks or other suitable stopcocks that will not contaminate the solvents used.

7. Reagents and Materials

7.1 Solvents:

7.1.1 *Normal Hexane*—Pure grade, having an ultraviolet light absorbance not exceeding 0.02 down to 260 nm when measured in a 1-cm cell (**Warning**—Normal hexane is extremely flammable, harmful if inhaled. May produce nerve cell damage.) The purity shall be such that the solvent control as defined in 8.3 shall have an absorbance curve compared to water showing no extraneous impurity peaks and no absorbance exceeding that of dimethyl sulfoxide compared to water at any wavelength in the range 260 to 350 nm, inclusive.

7.1.2 *Spectroscopic Grade 2,2,4-Trimethylpentane (Isooctane)*—(**Warning**—Isooctane is extremely flammable, harmful if inhaled.)

NOTE 2—For a suitable *isooctane* and a procedure for preparing spectroscopic solvents from commercially available stocks, see Test Method D1840.

7.1.3 *Dimethyl Sulfoxide*—(**Warning**—Dimethyl sulfoxide is combustible. Also it is rapidly absorbed through skin.) For use as spectroscopic solvent (see Note 3). Pure grade, clear, water white, 99.9 % dimethyl sulfoxide, m.p. 18.5°C, with an absorbance curve compared to water not exceeding 1.0 at 264 nm and showing no extraneous impurity peaks in the wavelength range up to 350 nm when measured through a path length of 1 cm.

NOTE 3—This solvent can be purified by percolation through a 1.2-m column of type CAL 1.68 by 0.420 mm (12 by 40 mesh) activated charcoal.⁴ The column is 25 mm in diameter, drawn to 6.4 mm in diameter at the bottom and has a reservoir at the top for containing the liquid. Glass wool is placed in the bottom of the column, and about 13 mm of 0.707 to 0.074 (25 to 200 mesh) or 0.0149 to 0.074 (100 to 200 mesh) silica gel is placed on top of it. The column is filled with activated charcoal and the spectroscopic solvent poured into the reservoir and allowed to percolate through the charcoal at atmospheric pressure. The purified solvent is collected at the bottom of the column and stored in glass-stoppered bottles as it is very hygroscopic and reacts with some metal containers in the presence of air.

7.2 *Naphthalene*—high-purity, 99+ %.

7.3 *Standard Reference Solution*—A solution containing 7.0 mg of naphthalene per litre of purified *isooctane*.

7.4 *Standard Reference Spectrum*—The absorbance curve obtained by scanning the standard reference solution in the range from 260 to 350 nm against *isooctane* of the same spectral purity as that used to prepare the standard (7.1.2).

7.5 *Standard Reference Absorbance*—The absorbance at 275 nm of the standard reference spectrum.

NOTE 4—This absorbance will be approximately 0.30.

8. Procedure

8.1 Transfer 25 mL of the mineral oil and 25 mL of hexane to a separatory funnel and mix. Add 5.0 mL of dimethyl sulfoxide

⁴ The sole source of supply of the activated charcoal known to the committee at this time is Calgon Carbon Dr., Robinson Township, PA 15205. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend.