

SLOVENSKI STANDARD SIST EN ISO 18563-3:2024

01-september-2024

Neporušitvene preiskave - Ugotavljanje značilnosti in preverjanje ultrazvočne opreme faznih sistemov - 3. del: Kombinirani sistemi (ISO 18563-3:2024)

Non-destructive testing - Characterization and verification of ultrasonic phased array equipment - Part 3: Complete systems (ISO 18563-3:2024)

Zerstörungsfreie Prüfung - Charakterisierung und Verifizierung der Ultraschall-Prüfausrüstung mit phasengesteuerten Arrays - Teil 3: Vollständige Prüfsysteme (ISO 18563-3:2024)

Essais non destructifs - Caractérisation et vérification de l'appareillage ultrasonore multiélément - Partie 3: Système complet (ISO 18563-3:2024)

Ta slovenski standard je istoveten z: EN ISO 18563-3:2024

ICS:

19.100 Neporušitveno preskušanje Non-destructive testing

SIST EN ISO 18563-3:2024 en,fr,de

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 18563-3:2024

https://standards.iteh.ai/catalog/standards/sist/3e27b539-2450-48ba-af9c-ced84559edce/sist-en-iso-18563-3-2024

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN ISO 18563-3

May 2024

ICS 19.100

Supersedes EN ISO 18563-3:2015

English Version

Non-destructive testing - Characterization and verification of ultrasonic phased array equipment - Part 3: Complete systems (ISO 18563-3:2024)

Essais non destructifs - Caractérisation et vérification de l'appareillage ultrasonore multiélément - Partie 3: Systèmes complets (ISO 18563-3:2024) Zerstörungsfreie Prüfung - Charakterisierung und Verifizierung der Ultraschall-Prüfausrüstung mit phasengesteuerten Arrays - Teil 3: Vollständige Prüfsysteme (ISO 18563-3:2024)

This European Standard was approved by CEN on 20 April 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN ISO 18563-3:2024 (E)

Contents	Page
European foreword	

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 18563-3:2024

https://standards.iteh.ai/catalog/standards/sist/3e27b539-2450-48ba-af9c-ced84559edce/sist-en-iso-18563-3-2024

European foreword

This document (EN ISO 18563-3:2024) has been prepared by Technical Committee ISO/TC 135 "Non-destructive testing" in collaboration with Technical Committee CEN/TC 138 "Non-destructive testing" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2024, and conflicting national standards shall be withdrawn at the latest by November 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 18563-3:2015.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 18563-3:2024 has been approved by CEN as EN ISO 18563-3:2024 without any modification.

https://standards.itell.ai/catalog/standards/sist/3e27b539-2450-48ba-af9c-ced84559edce/sist-en-iso-18563-3-2024

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 18563-3:2024

https://standards.iteh.ai/catalog/standards/sist/3e27b539-2450-48ba-af9c-ced84559edce/sist-en-iso-18563-3-2024



International Standard

ISO 18563-3

Non-destructive testing —
Characterization and verification
of ultrasonic phased array
equipment —

iTeh Standards

Part 3:

Complete systems

Essais non destructifs - Caractérisation et vérification de l'appareillage ultrasonore multiélément —

Partie 3: Systèmes complets

Second edition 2024-05

ttps://standards.<mark>i</mark>teh.ai)

talog/standards/sist/3e27b539-2450-48ba-at9c-ced84559edce/sist-en-iso-18563-3-2024

iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN ISO 18563-3:2024

https://standards.iteh.ai/catalog/standards/sist/3e27b539-2450-48ba-af9c-ced84559edce/sist-en-iso-18563-3-2024



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Contents	Pag	зe	
Foreword	i	iv	
1 Scope		1	
2 Normative references		1	
3 Terms and definitions		2	
Symbols			
•			
6 Qualification of test personnel		4	
7 Modes of operation for phased array techn	iques	4	
8 Equipment required for tests		6	
9 Tests to be performed		6	
9.3.1 General Section Section 1	dards itch ai)	.8	
9.3.3 Relative sensitivity of elements	reference amplitude and dead elements1	.0	
	1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1		
	18562.2:2004 1		
	n by using imaging		
9.4.4 Verification of correct operation	1 by using beams1	8	
	2		
	2		
	2		
Annex A (informative) Characterization of sound h			
	3		

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 135 *Non-destructive testing*, Subcommittee SC 3 *Ultrasonic testing*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 138, *Non-destructive testing*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 18563-3:2015), which has been technically revised.

The main changes are as follows:

- integration of matrix array probes;
- deletion of group 1 and 2 tests;
- addition of a clause on the use of imaging for complete system verification (9.4.3) as a simplification for a more functional standard (characterisation of beams moved to Annex A);
- addition of signal processing techniques using arrays (e.g. total focusing technique (TFM)) in the scope.

A list of all parts in the ISO 18563 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Non-destructive testing — Characterization and verification of ultrasonic phased array equipment —

Part 3:

Complete systems

1 Scope

This document addresses ultrasonic test systems implementing array probes, for contact technique (with or without wedge) or for immersion technique, with centre frequencies in the range of 0,5 MHz to 10 MHz.

This document provides methods and acceptance criteria for determining the compliance of the complete system (see 3.2). Its purpose is for the verification of the correct operation of the system prior to testing or verification of the absence of degradation of the system.

The methods are not intended to prove the suitability of the system for particular applications but are intended to prove the capability of the complete system (used for an application) to operate correctly according to the settings used. Tests can be performed on individual ultrasonic beams (for phased array technique, see 9.4.4) or on resulting images (for phased array technique and total focusing technique, see 9.4.3).

The tests can be limited to the functions that are intended to be used for a certain application.

This document does not cover the sensitivity setting of the system for a specific application. Nor does it apply to the characterization or verification of the mechanical scanning equipment. It is intended that these items will be covered by the test procedure.

This document does not address the phased array technique using tandem technique.

The characterization of beams, as recommended in case of dead elements or for more in-depth knowledge of 124 the beams, is presented in Annex A. It is not applicable for signal processing technology using arrays.

NOTE Unless stated otherwise, in this document 'TFM' and 'TFM technique' refer to the total focusing technique as defined in ISO 23243, and to related techniques, see for example ISO 23865 and ISO 23234.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5577, Non-destructive testing — Ultrasonic testing — Vocabulary

ISO 9712, Non-destructive testing — Qualification and certification of NDT personnel

ISO 18563-1, Non-destructive testing — Characterization and verification of ultrasonic phased array equipment — Part 1: Instruments

ISO 18563-2, Non-destructive testing — Characterization and verification of ultrasonic phased array equipment — Part 2: Probes

ISO 22232-2, Non-destructive testing — Characterization and verification of ultrasonic test equipment — Part 2: Probes

ISO 23243, Non-destructive testing — Ultrasonic testing with arrays — Vocabulary

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 5577, ISO 23243 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at https://www.electropedia.org/

3.1

combined equipment

connected set including the instrument, the array probe (with wedge if applicable) and connecting cables, including adapters

[SOURCE: ISO 23243:2020. Modified – array and wedge added]

3.2

complete system

combined equipment including the settings for a given mode of operation

Note 1 to entry: Settings are specific values or ranges of values, e.g. electronic scanning or steering range.

3.3

reference system

complete system, including an instrument according to ISO 18563-1 and an array probe initially according to ISO 18563-2, on which all of the applicable tests defined in <u>Clause 9</u> of this document have been performed successfully

3.4 identical system

complete system in which instrument, array probe, wedge, connecting cables and the settings for a given mode of operation are identical to those of the reference system

Note 1 to entry: Components are identical if from the same manufacturer and the same model.

http**3.5**standards.iteh.ai/catalog/standards/sist/3e27b539-2450-48ba-af9c-ced84559edce/sist-en-iso-18563-3-2024

system record sheet

document for reporting the results for a complete system which enables a comparison with the values obtained from the reference system

4 Symbols

For the purposes of this document, the symbols given in <a>Table 1 apply.

Table 1 — Symbols

Symbol	Unit	Definitions
$(X_C; Z_C)$	mm; mm	Coordinates of the position of the centre of the reference reflector
$(X_M; Z_M)$	mm; mm	Coordinates of the position of maximum amplitude of an indication
A_{el}	V or % FSH	Amplitude of one elementary signal
		Contact technique: reduced projected sound path length
$a_{\rm i}$	mm	<i>Immersion technique</i> : distance between the orthogonal projection of the axis of the sidedrilled hole on the test surface and the centre of the probe front surface
A_{\max}	V or % FSH	Maximum value of the amplitudes of all elementary signals
A_{\min}	V or % FSH	Minimum value of the amplitudes of all elementary signals, excluding the dead elements
$A_{\rm ref}$	V or % FSH	Median value of the amplitudes of all elementary signals