



Designation: F366 – 10

## Standard Specification for Fixation Pins and Wires<sup>1</sup>

This standard is issued under the fixed designation F366; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reappraisal.

### 1. Scope

1.1 This specification covers functional dimensions for fixation pins and wires.

1.2 In recognition of many broad and varied uses of such pins and wires, many options are included. A variety, but not necessarily all, of the options are illustrated in [Figs. 1-3](#).

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

### 2. Referenced Documents

2.1 *ASTM Standards*:<sup>2</sup>

[F67](#) Specification for Unalloyed Titanium, for Surgical Implant Applications (UNS R50250, UNS R50400, UNS R50550, UNS R50700)

[F75](#) Specification for Cobalt-28 Chromium-6 Molybdenum Alloy Castings and Casting Alloy for Surgical Implants (UNS R30075)

[F86](#) Practice for Surface Preparation and Marking of Metallic Surgical Implants

[F90](#) Specification for Wrought Cobalt-20Chromium-15Tungsten-10Nickel Alloy for Surgical Implant Applications (UNS R30605)

[F136](#) Specification for Wrought Titanium-6Aluminum-4Vanadium ELI (Extra Low Interstitial) Alloy for Surgical Implant Applications (UNS R56401)

[F138](#) Specification for Wrought 18Chromium-14Nickel-2.5Molybdenum Stainless Steel Bar and Wire for Surgical Implants (UNS S31673)

[F368](#) Specification for Fixation Pins-Knowles and Hagie Types (Withdrawn 1982)<sup>3</sup>

[F562](#) Specification for Wrought 35Cobalt-35Nickel-20Chromium-10Molybdenum Alloy for Surgical Implant Applications (UNS R30035)

[F563](#) Specification for Wrought Cobalt-20Nickel-20Chromium-3.5Molybdenum-3.5Tungsten-5Iron Alloy for Surgical Implant Applications (UNS R30563) (Withdrawn 2005)<sup>3</sup>

[F1314](#) Specification for Wrought Nitrogen Strengthened 22 Chromium-13 Nickel-5 Manganese-2.5 Molybdenum Stainless Steel Alloy Bar and Wire for Surgical Implants (UNS S20910)

[F1537](#) Specification for Wrought Cobalt-28Chromium-6Molybdenum Alloys for Surgical Implants (UNS R31537, UNS R31538, and UNS R31539)

[F1586](#) Specification for Wrought Nitrogen Strengthened 21Chromium-10Nickel-3Manganese-2.5Molybdenum Stainless Steel Alloy Bar for Surgical Implants (UNS S31675)

### 3. Materials

3.1 Fixation pins and wires shall be fabricated from material conforming to one of the following ASTM Specifications: [F67](#), [F75](#), [F90](#), [F136](#), [F138](#), [F562](#), [F563](#), [F1314](#), [F1537](#), and [F1586](#).

### 4. Performance Requirements

4.1 Factors considered to be important, but for which values and test methods have not been established, are bending strength, fatigue strength, breaking strength (Knowles Type only), torsion strength, and ductility.

### 5. Dimensions and Characteristics

5.1 Fixation pins and wires shall be fabricated in accordance with the dimensions illustrated in [Figs. 1-4](#).

5.2 Fixation pins and wires shall have surfaces prepared and marked in accordance with Practice [F86](#).

5.2.1 Optional marking on the fixation pins and wires shall identify the manufacturer or distributor.

### 6. Packaging and Labeling

6.1 Packaging shall be adequate to protect the fixation pins and wires during shipment.

6.2 Labeling for fixation pins and wires shall include:


6.2.1 Product name,

<sup>1</sup> This specification is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.21 on Osteosynthesis.

Current edition approved June 1, 2010. Published July 2010. Originally approved in 1973. Last previous edition approved in 2009 as F366 – 04 (2009). DOI: 10.1520/F0366-10.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>3</sup> The last approved version of this historical standard is referenced on [www.astm.org](http://www.astm.org).



ASTM F366-10

<https://standards.iteh.ai/catalog/standards/astm-f366-10> NOTE 1—Pins and wires may be smooth shank or threaded.

NOTE 2—Point angle and helix angle, where applicable, is as specified by manufacturer.

NOTE 3—On square or triangular shanks, flats are equal and corners are on the same circumference as the pin diameter. Shank diameters on pins larger than 3.2 mm may be reduced.

NOTE 4—Optional designs, both ends pointed or point with suture hole.

**FIG. 1 Fixation Pins and Wires**

6.2.2 Size, on the immediate container,

6.2.2.1 Length,

6.2.2.2 Diameter (if round) or cross-sectional size (if square of hexagonal), that is, 6.4 mm square, and

6.2.3 ASTM material specification Designation number.

## 7. Keywords

7.1 fixation materials; flexible surgical wire; orthopaedic medical devices; wire-surgical implants