



Designation: ~~C1287-03~~ Designation: **C1287 - 10**

# Standard Test Method for Determination of Impurities in Nuclear Grade Uranium Compounds by Inductively Coupled Plasma Mass Spectrometry<sup>1</sup>

This standard is issued under the fixed designation C1287; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method covers the determination of 67 elements in uranium dioxide samples and nuclear grade uranium compounds and solutions without matrix separation by inductively coupled plasma mass spectrometry (ICP-MS). The elements are listed in Table 1. These elements can also be determined in uranyl nitrate hexahydrate (UNH), uranium hexafluoride (UF<sub>6</sub>), triuranium octoxide (U<sub>3</sub>O<sub>8</sub>) and uranium trioxide (UO<sub>3</sub>) if these compounds are treated and converted to the same uranium concentration solution.

~~1.2 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. For a specific warning statement, see Note 1.~~

**NOTE 1—Warning:** The ICP-MS is a source of intense ultra-violet radiation from the radio frequency induced plasma. Protection from radio frequency radiation and UV radiation is provided by the instrument under normal operation.

~~1.3 The elements boron, sodium, silicon, phosphorus, potassium, calcium and iron can be determined using different techniques. The analyst's instrumentation will determine which procedure is chosen for the analysis.~~

~~1.4 The test method for technetium-99 is given in~~

1.2 The elements boron, sodium, silicon, phosphorus, potassium, calcium and iron can be determined using different techniques. The analyst's instrumentation will determine which procedure is chosen for the analysis.

1.3 The test method for technetium-99 is given in Annex A1.

1.4 The values stated in SI units are to be regarded as standard.

1.5 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. **Warning**—The ICP-MS is a source of intense ultra-violet radiation from the radio frequency induced plasma. Protection from radio frequency radiation and UV radiation is provided by the instrument under normal operation.

## 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

C753 [Specification for Nuclear-Grade, Sinterable Uranium Dioxide Powder](#)

C776 [Specification for Sintered Uranium Dioxide Pellets](#)

C787 [Specification for Uranium Hexafluoride for Enrichment](#)

C788 [Specification for Nuclear-Grade Uranyl Nitrate Solution or Crystals](#)

C967 [Specification for Uranium Ore Concentrate](#)

C996 [Specification for Uranium Hexafluoride Enriched to Less Than 5 % <sup>235</sup>U](#)

C1346 [Practice for Dissolution of UF<sub>6</sub> from P-10 Tubes](#) ~~from P-10 Tubes~~

C1347 [Practice for Preparation and Dissolution of Uranium Materials for Analysis](#)

D1193 [Specification for Reagent Water](#)

## 3. Summary of Test Method

3.1 The sample is dissolved in acid if it is not already a solution. A fixed quantity of internal standard is added to monitor and

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee C26 on Nuclear Fuel Cycle and is the direct responsibility of Subcommittee C26.05 on Methods of Test. Current edition approved July 10, 2003. Published August 2003. Originally approved in 1994. Last previous edition approved in 2001 as C1287-95(2001). DOI: 10.1520/C1287-03.

Current edition approved June 1, 2010. Published July 2010. Originally approved in 1994. Last previous edition approved in 2003 as C1287 - 03. DOI: 10.1520/C1287-10.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

correct for signal instability. The level of impurities in the solution is measured by ICP-MS. Customized software calculates the concentration of each element.

3.2 Uranium-concentration-matched standard solutions are used to calibrate the ICP-MS instrument. The calibration is linear up to at least 0.2 µg/ml (100 µg/g U) for each analyte.<sup>3,4</sup>

3.3 Microwave dissolution may be used as an alternate dissolution method.

#### 4. Significance and Use

4.1 This test method is capable of measuring the elements listed in Table 1, some of which are required by Specifications C753, C776, C787, C788, C967 and C996.

#### 5. Apparatus

~~5.1 ICP-MS, controlled by computer and fitted with the associated software and peripherals. May be fitted with cold plasma option.~~  
5.1 ICP-MS, controlled by computer and fitted with the associated software and peripherals. May be fitted with cold plasma option. Current instrumentation is available with dynamic reaction cell or collision cell options.

5.2 Autosampler, with tube racks and disposable plastic sample tubes compatible with 5.1 (optional).

5.3 Variable Micropipettes:

5.3.1 10 µL to 100 µL capacity.

5.3.2 100 µL to 1000 µL capacity.

5.3.3 1000 µL to 10.00 mL capacity.

5.4 Volumetric Flasks:

5.4.1 50 mL capacity—polypropylene.

5.4.2 100 mL capacity—polypropylene.

5.4.3 1 L capacity—glass.

5.5 Platinum Dish—100 mL capacity.

5.6 Silica Beaker—250 mL capacity.

5.7 Watch Glasses—75 mm diameter.

5.8 Polypropylene Tubes—50 mL, with graduation marks and with caps.

#### 6. Reagents

6.1 The sensitivity of the ICP-MS technique requires the use of ultra high purity reagents in order to be able to obtain the low levels of detection. All the reagents below are ultra high purity grade unless otherwise stated:

6.1.1 Element stock standards at 1000 µg/mL for all the elements in Table 1.

6.1.2 ~~Hydrofluoric acid (HF)~~Hydrofluoric Acid (HF), (40 g/100 g), 23 molar.

6.1.2.1 **Warning**—Hydrofluoric acid is highly corrosive acid that can severely burn skin, eyes, and mucous membranes. Hydrofluoric acid is similar to other acids in that the initial extent of a burn depends on the concentration, the temperature, and the duration of contact with the acid. Hydrofluoric acid differs from other acids because the fluoride ion readily penetrates the skin, causing destruction of deep tissue layers. Unlike other acids that are rapidly neutralized, hydrofluoric acid reactions with tissue may continue for days if left untreated. Due to the serious consequences of hydrofluoric acid burns, prevention of exposure or injury of personnel is the primary goal. Utilization of appropriate laboratory controls (hoods) and wearing adequate personnel protective equipment to protect from skin and eye contact is essential. Acute exposure to HF can cause painful and severe burns upon skin contact that require special medical attention. Chronic or prolonged exposure to low levels on the skin may cause fluorosis.

6.1.3 ~~Nitric acid~~Nitric Acid—Concentrated nitric acid (HNO<sub>3</sub>), 15 molar.

6.1.4 ~~Rhodium Stock Solution (1000 µg/mL Rh)~~—Commercially available solution (see ~~Note 2~~Note 1).

NOTE 21—Rhodium stock solution is commercially available supplied with a certificate of analysis for the element and a full range of trace impurities. The solutions are prepared by the manufacturer using a variety of media designed to keep each element in solution for a minimum of one year.

6.1.5 ~~Sulfuric acid~~Sulfuric Acid—Concentrated sulfuric acid (H<sub>2</sub>SO<sub>4</sub>), 18 molar.

6.1.6 Uranium Standard Base Solution—Uranyl nitrate solution to Specification C788, of known uranium (100 g/L) and aluminum content (≤ 2 µg/g U). The total metallic impurity (TMI) content must not exceed 50 µg/g U and no individual analyte must exceed 10 µg/g U.

6.1.7 Purity of Water—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification D1193, Type I.

6.1.8 Ammonia—Anhydrous, NH<sub>3</sub>, 99.9995 % minimum purity. Used with instruments fitted with dynamic reaction cell option.

#### 7. Standards

7.1 Four separate mixed standard solutions (A, B, C, and E) are prepared to prevent the precipitation of some elements (as

<sup>3</sup> “ICP-MS Versus Conventional Methods for the Analysis of Trace Impurities in Nuclear Fuel,” by Allenby, P., Clarkson, A. S., Makinson, P. R., presented at 2nd Surrey Conference on Plasma Source Mass Spectrometry, Guildford, UK, July 1987.

<sup>4</sup> “Trace Metals in NBL Uranium Standard CRM 124 Using ICP-MS,” by Aldridge, A. J., Clarkson, A. S., Makinson, P. R., Dawson, K. W., presented at 1st Durham International Conference on Plasma Source Mass Spectrometry, Durham, UK, September 1988.

**TABLE 1 Reporting Limits of Impurity Elements**

NOTE 1—The impurity elements were determined in 0.2 % uranium solutions, prepared following Section 8.

NOTE 2—Acquisition time = 10 s/isotope using peak jump mode.

NOTE 3—103 Rh was used as an internal standard. For the elements where the technique is identified as Perkin Elmer-El DRCII scan-5000A P-E-Edium was used as internal stan-5000Aard for boron, sodium and phosphorus. Rhodium was used as the internal standard for potassium, calcium and iron in Reaction Cell mode.

NOTE 4—The LRL is based on the within run standard deviation ( $S_b$ ) of 20 uranium-matched blank determinations for each analyte. This limit equals  $4 \times S_b$ , rounded up to a preferred value in the series 1, 1.5, 2, 3, 4, 6, multiplied or divided by the appropriate integer power of ten.

NOTE 5—The upper reporting limit can be increased by extending the calibration to 10  $\mu\text{g/mL}$  (5000  $\mu\text{g/g U}$ ) if the ICP-MS used has an extended dynamic range (EDR) accessory.

NOTE 6—For the elements where the technique is listed as P-E-Elan 5000A DRCII, the instrumentation may be specific to those elements. Alternatively cold plasma technique may be used and it is up to the analyst to perform testwork using spikes and reference materials and to determine the lower reporting levels. The impurity elements were determined in 0.16 % uranium solutions, prepared following Section 8. The dwell times are listed in 8.4.1.1.

NOTE 7—Some of the elements are not included in the material specifications and have been included only as a research record for the reader's interest.

Analyte	Mass Used	Analyte Group	Lower Reporting Limit (LRL), $\mu\text{g/g U}$	Upper Reporting Limit (URL), $\mu\text{g/g U}$	Technique
Lithium	7	A	0.01	100	normal plasma
Beryllium	9	A	0.04	100	normal plasma
Boron	11	E	0.3	100	P-E-Elan5000A
Boron	11	E	0.3	100	DRCII
Sodium	23	E	0.3	100	P-E-Elan5000A
Sodium	23	E	0.3	100	DRCII
Magnesium	24	A	4	100	normal plasma
Aluminum	27	D	2	1000	normal plasma
Silicon	28	E	1.5	100	P-E-Elan5000A
Phosphorus	31	E	1.5	100	P-E-Elan5000A
Phosphorus	31	E	1	100	DRCII
Potassium	39	E	2	100	P-E-Elan5000A
Potassium	39	E	2.0	100	DRCII
Calcium	40	E	3	100	P-E-Elan5000A
Calcium	40	E	3	100	DRCII
Scandium	45	A	4	100	normal plasma
Titanium	48	B	0.2	100	normal plasma
Vanadium	51	B	0.04	100	normal plasma
Chromium	52	B	0.1	100	normal plasma
Manganese	55	A	0.1	100	normal plasma
Iron	56	A	15	100	normal plasma
Iron	56	A	0.2	100	DRCII
Cobalt	59	A	0.02	100	normal plasma
Nickel	60	A	0.4	100	normal plasma
Copper	65	A	0.2	100	normal plasma
Zinc	66	A	0.3	100	normal plasma
Gallium	69	A	0.04	100	normal plasma
Germanium	74	A	0.2	100	normal plasma
Arsenic	75	A	0.2	100	normal plasma
Selenium	82	A	3	100	normal plasma
Rubidium	85	A	0.06	100	normal plasma
Strontium	88	A	0.06	100	normal plasma
Yttrium	89	A	0.04	100	normal plasma
Zirconium	90	B	0.02	100	normal plasma
Niobium	93	B	0.01	100	normal plasma
Molybdenum	95	B	0.04	100	normal plasma
Ruthenium	102	B	0.02	100	normal plasma
Palladium	106	B	0.2	100	normal plasma
Silver	107	A	0.1	100	normal plasma
Cadmium	111	A	0.03	100	normal plasma
Indium	115	A	0.04	100	normal plasma

**TABLE 1** *Continued*

Analyte	Mass Used	Analyte Group	Lower Reporting Limit (LRL), $\mu\text{g/g U}$	Upper Reporting Limit (URL), $\mu\text{g/g U}$	Technique
Tin	116	B	0.04	100	normal plasma
Antimony	121	B	0.02	100	normal plasma
Tellurium	130	B	0.4	100	normal plasma
Caesium	133	A	0.06	100	normal plasma
Barium	138	A	0.02	100	normal plasma
Lanthanum	139	C	0.1	100	normal plasma
Cerium	140	C	0.01	100	normal plasma
Praseodymium	141	C	0.01	100	normal plasma
Neodymium	146	C	0.01	100	normal plasma
Samarium	149	C	0.01	100	normal plasma
Europium	151	C	0.01	100	normal plasma
Gadolinium	158	C	0.01	100	normal plasma
Terbium	159	C	0.01	100	normal plasma
Dysprosium	163	C	0.01	100	normal plasma
Holmium	165	C	0.01	100	normal plasma
Erbium	166	C	0.01	100	normal plasma
Thulium	169	C	0.01	100	normal plasma
Ytterbium	174	C	0.01	100	normal plasma
Lutetium	175	C	0.01	100	normal plasma
Hafnium	178	B	0.01	100	normal plasma
Tantalum	181	B	0.01	100	normal plasma
Tungsten	184	B	0.01	100	normal plasma
Rhenium	187	A	0.02	100	normal plasma
Osmium	190	B	0.2	100	normal plasma
Iridium	193	B	0.2	100	normal plasma
Platinum	195	B	0.2	100	normal plasma
Gold	197	B	0.06	100	normal plasma
Mercury	202	A	0.4	100	normal plasma
Thallium	205	A	0.02	100	normal plasma
Lead	208	A	0.02	100	normal plasma
Bismuth	209	A	0.03	100	normal plasma
Thorium	232	B	0.01	100	normal plasma

insoluble chlorides, fluorides etc; see Table 1 for details of the analyte groups). Analyte group A contains element stock solutions prepared in  $\text{HNO}_3$  or  $\text{HNO}_3/\text{HF}$ , analyte group B contains element stock solutions prepared in  $\text{HCl}$  or  $\text{HCl}/\text{HF}$ , analyte group C contains the rare earth element stock solutions, and analyte group E contains boron sodium silicon, phosphorus, potassium and calcium. The mixed standard solutions should be prepared to contain only the analytes of interest. Other combinations of mixed standard solutions may be prepared to minimize the precipitation of the analytes.

7.1.1 Mixed standard solution A is prepared from stock solutions of each element from analyte group A. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL}$ ) of each element into a 50 mL polypropylene volumetric flask and add 500  $\mu\text{L}$  of concentrated nitric acid. Dilute to 50 mL with water and mix. This multi-element standard contains 20  $\mu\text{g}/\text{mL}$  of each analyte in 1 % nitric acid. This solution must be used on the day of preparation.

7.1.2 Mixed standard solution B is prepared from stock solutions of each element from analyte group B. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL}$ ) of each element into a 50 mL polypropylene volumetric flask and add 500  $\mu\text{L}$  of concentrated nitric acid. Dilute to 50 mL with water and mix. This multi-element standard contains 20  $\mu\text{g}/\text{mL}$  of each analyte in 1 % nitric acid. This solution must be used within one week of preparation.

7.1.3 Mixed standard solution C is prepared from stock solutions of each element from analyte group C. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL}$ ) of each element into a 50 mL polypropylene volumetric flask and add 500  $\mu\text{L}$  of concentrated nitric acid. Dilute to 50 mL with water and mix. This multi-element standard contains 20  $\mu\text{g}/\text{mL}$  of each analyte in 1 % nitric acid. This solution must be used within one week of preparation.

7.2 Standard solution D is prepared from the stock solution of aluminum from analyte group D. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL Al}$ ) into a 50 mL polypropylene volumetric flask and add 500  $\mu\text{L}$  of concentrated nitric acid. Dilute to 50  $\mu\text{L}$  with water and mix. This standard contains 20  $\mu\text{g}/\text{mL}$  of aluminum in 1 % nitric acid. This solution must be used within one week of preparation.

7.3 Mixed standard solution E is prepared from stock solutions of each element from analyte group E. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL}$ ) of each element into a 50 mL polypropylene volumetric flask and add 500  $\mu\text{L}$  of concentrated nitric acid. Dilute to 50 mL with water and mix. This multi-element standard contains 20  $\mu\text{g}/\text{mL}$  of each analyte in 1 % nitric acid. This solution must be used within one week of preparation.

7.4 Rhodium internal standard solution is prepared from the stock solution. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL Rh}$ ) into a 100 mL polypropylene volumetric flask and add 1000  $\mu\text{L}$  of concentrated nitric acid. Dilute to 100 mL with water and

mix. This internal standard solution contains 10 µg/mL Rh in a 1% nitric acid solution. Other internal standards such as scandium (used with B, Na, Si, P, K and Ca) may be used. With high mass elements the analyst may choose internal standards such as iridium or terbium. Other elements may be applicable as well but it is up to the analyst to conduct the appropriate testwork.

7.5 Diluent solution is prepared from rhodium stock standard solution. Transfer 1000 µL of the stock solution (1000 µg/mL Rh) into a 1 L volumetric flask and add 10.00 mL of concentrated nitric acid. Dilute to 1 L with water and mix. This diluent solution contains 1 µg/mL Rh in 1% nitric acid solution. Other internal standard diluent solutions may be used. Rhodium internal standard solution is prepared from the stock solution. Transfer 1000 µL of the stock solution (1000 µg/mL Rh) into a 100 mL polypropylene volumetric flask and add 1000 µL of concentrated nitric acid. Dilute to 100 mL with water and mix. This internal standard solution contains 10 µg/mL Rh in a 1 % nitric acid solution. Other internal standards such as scandium may be used. With high mass elements the analyst may choose internal standards such as iridium or terbium. Other elements may be applicable as well but it is up to the analyst to conduct the appropriate testwork.

NOTE 32—Throughout this standard, references to Rh internal standard solution will include all other internal standard elements that may be used.

## 8. Procedure

NOTE 4—A 3—A uranium-free reagent blank is used to eliminate bias due to the analyte concentrations in the uranium standard base solution. A uranium-matched reagent blank is necessary to provide a constant acid concentration in the nebulized solution.

### 8.1 ~~Sample Preparation for the Determination of All Elements Except Boron, Silicon, Potassium, and Calcium~~ Sample Preparation for the Determination of All Elements Except Boron:

8.1.1 Weigh a portion of uranium oxide containing between 2.45 and 2.55 g of uranium into a platinum dish. Record the weight to the nearest 0.001 g. For uranyl fluoride solutions prepared using Practice C1346 and uranyl nitrate solutions, aliquot between 2.45 and 2.55 g of uranium into a platinum dish. Use a variable volume plastic pipet for the transfer of uranyl fluoride solutions. Record the weight to the nearest 0.001 g.

8.1.2 Add 10 mL of water and 12.5 mL of concentrated nitric acid. Heat on a hotplate to assist dissolution.

8.1.3 Add 2.5 mL of hydrofluoric acid (40 g/100 g) and warm at about 80°C for 5 min.

8.1.4 Allow the solution to cool and transfer quantitatively to a 50 mL polypropylene volumetric flask. Dilute to 50 mL with water and mix. This solution contains 50 g of uranium per litre in 25 % nitric acid/5 % hydrofluoric acid.

8.1.5 Transfer 4.00 mL of the solution in 8.1.4 and 1.00 mL of the rhodium internal standard solution (see 7.4) into a 100 mL polypropylene volumetric flask. Dilute to 100 mL with water and mix. This solution contains 2 g of uranium per litre and 0.1 µg/mL Rh in 1 % nitric acid/0.2 % hydrofluoric acid.

8.1.6 A uranium-free reagent blank (see 8.3.1) and a control or recovery sample must be prepared with every run of samples.

8.1.7 Analyze these solutions as in 8.4 using the calibration solutions prepared in 8.3. The solutions must be analyzed within 8 h of preparation to minimize the effects of analyte precipitation.

### 8.2 ~~Sample Preparation for the Determination of Boron and Silicon Potassium and Calcium~~ Sample Preparation for the Determination of Boron:

8.2.1 Weigh a portion of uranium dioxide, uranium octoxide or uranium trioxide, containing between 0.095 and 0.105 g of uranium into a graduated 50 mL polypropylene tube (or alternative). The accuracy of the graduations on the tube must be verified. Record the weight to the nearest 0.001 g. For uranyl fluoride solutions prepared using Practice C1346 and uranyl nitrate solutions, aliquot between 0.095 and 0.105 g of uranium using variable volume plastic pipets. Record the weight to the nearest 0.001 g.

8.2.2 Add 1 mL of water and 1.25 mL of concentrated nitric acid. Cap. Heat in a hot water bath at about 80°C to assist dissolution. Heat until all the sample is dissolved.

8.2.3 Cool to room temperature. Add 0.1 mL of hydrofluoric acid (40 g/100 g) and cap. Heat in a hot water bath at about 80°C for 5 min.

8.2.4 Allow the solution to cool. Add 0.5 mL of scandium internal standard solution (see 7.4). Dilute to 50 mL with water and mix. This solution contains 2 g of uranium per litre and 0.1 µg/mL Sc in 2.5 % nitric acid/0.2 % hydrofluoric acid.

8.2.5 A uranium-free reagent blank and a control or recovery sample must be prepared with every run of samples.

8.2.6 Analyze these solutions as in 8.4 using the calibration solutions prepared in 8.3. The solutions must be analyzed within 8 h of preparation to minimize the effects of analyte precipitation.

### 8.3 Preparation of Blanks and Calibration Standard Solutions:

#### 8.3.1 ~~For the Determination of All Elements Except Boron Silicon Potassium and Calcium~~ For the Determination of All Elements Except Boron:

8.3.1.1 Uranium-free Reagent Blank—Transfer 12.5 mL of concentrated nitric acid and 2.5 mL of hydrofluoric acid (40 g/100 g) into a 50 mL polypropylene volumetric flask. Continue as instructed from 8.1.5 onwards.

8.3.1.2 Uranium-matched Calibration Blank—Transfer 2.00 mL of the uranium standard base solution (see 6.1.6; this is equivalent to 0.20 g of uranium) into a 100 mL polypropylene volumetric flask. Add 1000 µL of concentrated nitric acid, 200 µL of hydrofluoric acid (40 g/100 g) and 1000 µL of rhodium internal standard solution (see 7.4). Dilute to 100 mL with water and mix. This solution contains 2 g of uranium per litre and 0.1 µg/mL Rh in 1 % nitric acid/0.2 % hydrofluoric acid.

8.3.1.3 Uranium-matched Calibration Standard—Transfer 2.00 mL of the uranium standard base solution (see 6.1.6; this is

equivalent to 0.20 g of uranium) into a 100 mL polypropylene volumetric flask. Add 1000 µL of concentrated nitric acid, 200 µL of hydrofluoric acid (40 g/100 g), 1000 µL of each mixed standard solution (see 7.1.1, 7.1.2 and 7.1.3) and 1000 µL of rhodium internal standard solution (see 7.4). Dilute to 100 mL with water and mix. This solution contains 2 g of uranium per litre, 0.2 µg/mL of each analyte (equivalent to 100 µg/g U) and 0.1 µg/mL Rh in 1 % nitric acid/0.2 % hydrofluoric acid.

### 8.3.2 *For the Determination of Boron, Silicon, Potassium and Calcium (mass 44)* For the Determination of Boron:

8.3.2.1 *Uranium-matched Reagent/Calibration Blank*—Transfer 2.00 mL of the uranium standard base solution (see 6.1.6; this is equivalent to 0.20 g of uranium) into a 100 mL polypropylene volumetric flask. Add 2.5 mL of concentrated nitric acid, 200 µL of hydrofluoric acid (40 g/100 g), and 1000 µL of scandium internal standard solution (see 7.4). Dilute to 100 mL with water and mix. This solution contains 2 g of uranium per litre and 0.1 µg/mL Sc in ~~1%~~2.5 % nitric acid/0.2 % hydrofluoric acid.

8.3.2.2 *Uranium-matched Calibration Standard*—Transfer 2.00 mL of the uranium standard base solution (see 6.1.7; this is equivalent to 0.20 g of uranium) into a 100 mL polypropylene volumetric flask. Add 2.5 mL of concentrated nitric acid, 200 µL of hydrofluoric acid (40 g/100 g), 1000 µL of mixed standard solution (see 7.1 and 7.3), and 1000 µL of scandium internal standard solution (see 7.4). Dilute to 100 mL with water and mix. This solution contains 2 g of uranium per litre, 0.2 µg/mL of each analyte (equivalent to 100 µg/g U) and 0.1 µg/mL Sc in ~~1%~~2.5 % nitric acid/0.2 % hydrofluoric acid.

### 8.4 *Measurement of Elements by ICP-MS:*

8.4.1 To avoid contamination problems when nebulizing the samples, which contain hydrofluoric acid, the nebulizer system (that is, spray chamber and nebulizer) must be made from fluorinated plastic materials (for example, TFE-fluorocarbon or polychlorotrifluoroethylene). **Note 5**—~~For the analysis of boron and silicon, the alumina injector for the Perkin Elmer Elan 5000 was fitted with a TFE-fluorocarbon inner sleeve.~~

~~8.4.1.1 Set up the ICP-MS for the analysis using the parameters given in the manufacturer's operating manual. Nebulize the uranium-matched reagent/calibration blank solution to optimize conditions using the 103 Rh internal standard.~~

8.4.1.1 Set up the ICP-MS for the analysis using the parameters given in the manufacturer's operating manual. Nebulize the uranium-matched reagent/calibration blank solution to optimize conditions using the 103 Rh internal standard.

	Instrument Operating Conditions
	<b>Example Instrument Operating Conditions</b>
Solution Pumping Rate	Sample solution IN: 1.25 mL/min
ICP Incident Power	1400 watts
ICP Reflected Power	<10 watts
Plasma Argon Coolant	14 L/min at 70 psig
Plasma Argon Auxiliary	0.7 L/min at 70 psig
Plasma Argon Nebulizer	0.93 L/min at 40 psig
Integration Method	Valley Int.
Integration Area	0.80000 daltons (Atomic Mass Units—AMU)
Background Counts	10.0000 counts/s
Dead Time	100.000 µs
Safe Resting Mass	129.253 daltons (AMU) <small>Perkin-Elmer Elan-5000A Instrument Operating Conditions for Boron, Sodium, Silicon, Phosphorus, Potassium, Calcium-44</small>
Solution-pumping-rate	1.1 mL/min
ICP-incident-power	1000-W
Plasma-argon-coolant	15 L/min
Plasma-argon-auxiliary	0.87 L/min
Plasma-argon-nebulizer	1.1 L/min
Acquisition-method	Peak-jump
Number-of-points-across-mass-peak	1
Dead-time	35-nanoseconds
Dwell-time/mass	50-ms
Sweeps/reading	12
Readings/replicate	3
Sample-cone	platinum
Skimmer-cone	platinum

8.4.1.2 Acquire the data for all blank, calibration standard, control/recovery and sample solutions using the ICP-MS for the analytes required using the masses specified by the element menu and given in Table 1. The element menu must also contain the mass for the internal standard (normally 103 Rh). Uranium-matched calibration solutions are run at the start and end of each run. Recalibration during the run may be necessary.

## 9. Calculation

9.1 The use of a uranium-free reagent blank allows the analyte concentrations in the uranium-matched blank to be ignored. For ~~boron, silicon, potassium, calcium (mass-44), and iron~~ however, a uranium-matched reagent/calibration blank must be used. The concentration of ~~boron, silicon, potassium, calcium (mass-44)~~calcium, and iron in the uranium standard base solution (Z) must be added to the sample concentration to avoid reporting biased results (that is, if the ~~boron~~iron content of the uranium standard base solution is 2 µg/g U, the uranium-matched reagent/calibration blank contains 2 µg/g U and the uranium-matched calibration standard contains 102 µg/g U. The calculation software sets these values at zero and 100 µg/g U, respectively.)