



Designation: E2005-05^{e1} Designation: E2005 - 10

Standard Guide for Benchmark Testing of Reactor Dosimetry in Standard and Reference Neutron Fields¹

This standard is issued under the fixed designation E2005; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

^{e1}Note— Editorial changes were made in June 2005.

1. Scope

1.1 This guide covers facilities and procedures for benchmarking neutron measurements and calculations. Particular sections of the guide discuss: the use of well-characterized benchmark neutron fields to calibrate integral neutron sensors; the use of certified-neutron-fluence standards to calibrate radiometric counting equipment or to determine interlaboratory measurement consistency; development of special benchmark fields to test neutron transport calculations; use of well-known fission spectra to benchmark spectrum-averaged cross sections; and the use of benchmarked data and calculations to determine the uncertainties in derived neutron dosimetry results.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

2. Referenced Documents

2.1 ASTM Standards:²

- E170 Terminology Relating to Radiation Measurements and Dosimetry
- E261 Practice for Determining Neutron Fluence, Fluence Rate, and Spectra by Radioactivation Techniques
- E263 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Iron
- E264 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Nickel
- E265 Test Method for Measuring Reaction Rates and Fast-Neutron Fluences by Radioactivation of Sulfur-32
- E266 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Aluminum
- E343 Test Method for Measuring Reaction Rates by Analysis of Molybdenum-99 Radioactivity From Fission Dosimeters
- E393 Test Method for Measuring Reaction Rates by Analysis of Barium-140 From Fission Dosimeters
- E482 Guide for Application of Neutron Transport Methods for Reactor Vessel Surveillance, E706 (IID)
- E523 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Copper
- E526 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Titanium
- E704 Test Method for Measuring Reaction Rates by Radioactivation of Uranium-238
- E705 Test Method for Measuring Reaction Rates by Radioactivation of Neptunium-237
- E854 Test Method for Application and Analysis of Solid State Track Recorder (SSTR) Monitors for Reactor Surveillance, E706(IIIB)
- E910 Test Method for Application and Analysis of Helium Accumulation Fluence Monitors for Reactor Vessel Surveillance, E706 (IIC)
- E1297 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Niobium
- E2006 Guide for Benchmark Testing of Light Water Reactor Calculations

3. Significance and Use

3.1 This guide describes approaches for using neutron fields with well known characteristics to perform calibrations of neutron sensors, to intercompare different methods of dosimetry, and to corroborate procedures used to derive neutron field information from measurements of neutron sensor response.

¹ This guide is under the jurisdiction of ASTM Committee E10 on Nuclear Technology and Applications and is the direct responsibility of Subcommittee E10.05 on Nuclear Radiation Metrology.

Current edition approved Jan. 1, 2005. Published February 2005. Originally approved in 1999. Last previous edition approved in 1999 as E2005-99. DOI: 10.1520/E2005-05E01 on Nuclear Radiation Metrology.

Current edition approved June 1, 2010. Published August 2010. Originally approved in 1999. Last previous edition approved in 2005 as E2005 - 05 ^{e1}. DOI: 10.1520/E2005-10.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.

3.2 This guide discusses only selected standard and reference neutron fields which are appropriate for benchmark testing of light-water reactor dosimetry. The Standard Fields considered are neutron source environments that closely approximate the unscattered neutron spectra from ^{252}Cf spontaneous fission and ^{235}U thermal neutron induced fission. These standard fields were chosen for their spectral similarity to the high energy region ($E > 2 \text{ MeV}$) of reactor spectra. The various categories of benchmark fields are defined in Terminology E170.

3.3 There are other well known neutron fields that have been designed to mockup special environments, such as pressure vessel mockups in which it is possible to make dosimetry measurements inside of the steel volume of the “vessel.” When such mockups are suitably characterized they are also referred to as benchmark fields. A variety of these engineering benchmark fields have been developed, or pressed into service, to improve the accuracy of neutron dosimetry measurement techniques. These special benchmark experiments are discussed in Guide E2006, and in Refs (1)³ and (2).

4. Neutron Field Benchmarking

4.1 To accomplish neutron field “benchmarking,” one must perform irradiations in a well-characterized neutron environment, with the required level of accuracy established by a sufficient quantity and quality of results supported by a rigorous uncertainty analysis. What constitutes sufficient results and their required accuracy level frequently depends upon the situation. For example:

4.1.1 Benchmarking to test the capabilities of a new dosimeter;

4.1.2 Benchmarking to ensure long-term stability, or continuity, of procedures that are influenced by changes of personnel and equipment;

4.1.3 Benchmarking measurements that will serve as the basis of intercomparison of results from different laboratories;

4.1.4 Benchmarking to determine the accuracy of newly established benchmark fields; and

4.1.5 Benchmarking to validate certain ASTM standard methods or practices which derive exposure parameters (for example, fluence $> 1 \text{ MeV}$ or dpa) from dosimetry measurements and calculations.

5. Description of Standard and Reference Fields

5.1 There are a few facilities which can provide certified “free field” fluence irradiations. The following provides a list of such facilities. The emphasis is on facilities that have a long-lived commitment to development, maintenance, research, and international interlaboratory comparison calibrations. As such, discussion is limited to recently existing facilities.

5.2 ^{252}Cf Fission Spectrum—Standard Neutron Field:

5.2.1 The standard fission-spectrum fluence from a suitably encapsulated ^{252}Cf source is characterized by its source strength, the distance from the source, and the irradiation time. In the U.S., neutron source emission rate calibrations are all referenced to source calibrations at the National Institute of Standards and Technology (NIST) accomplished by the MnSO_4 technique (3). Corrections for neutron absorption, scattering, and other than point-geometry conditions may, by careful experimental design, be held to less than 3 %. Associated uncertainties for the NIST ^{252}Cf irradiation facility are discussed in Ref (4). The principal uncertainties, which only total about 2.5 %, come from the source strength determination, scattering corrections, and distance measurements. Extensive details of standard field characteristics and values of measured and calculated spectrum-averaged cross sections are all given in a compendium, see Ref (5).

5.2.2 The NIST ^{252}Cf sources have a very nearly unperturbed spontaneous fission spectrum, because of the light-weight encapsulations, fabricated at the Oak Ridge National Laboratory (ORNL), see Ref (6).

5.2.3 For a comprehensive view of the calibration and use of a special (32 mg) ^{252}Cf source employed to measure the spectrum-averaged cross section of the $^{93}\text{Nb}(n,n')$ reaction, see Ref (7).

5.3 ^{235}U Fission Spectrum—Standard Neutron Field:

5.3.1 Because ^{235}U fission is the principal source of neutrons in present nuclear reactors, the ^{235}U fission spectrum is a fundamental neutron field for benchmark referencing or dosimetry accomplished in reactor environments. This remains true even for low-enrichment cores which have up to 30 % burnup.

5.3.2 There are currently two ^{235}U standard fission spectrum facilities available, facilities, one in the thermal column of the NIST Research Reactor (8) and one at CEN/SCK, Mol, Belgium (9).

5.3.3 A standard ^{235}U neutron field is obtained by driving (fissioning) ^{235}U in a field of thermal neutrons. Therefore, the fluence rate depends upon the power level of the driving reactor, which is frequently not well known or particularly stable. Time dependent fluence rate, or total fluence, monitoring is necessary in the ^{235}U field. Certified fluence irradiations are monitored with the $^{58}\text{Ni}(n,p)^{58}\text{Co}$ activation reaction. The fluence-monitor calibration must be benchmarked.

5.3.4 For ^{235}U , as for ^{252}Cf irradiations, small (nominally $< 3 \%$) scattering and absorption corrections are necessary. In addition, for ^{235}U , gradient corrections of the measured fluence which do not simply depend upon distance are necessary. The scattering and gradient corrections are determined by Monte Carlo calculations. Field characteristics of the NIST ^{235}U Fission Spectrum Facility and associated measured and calculated cross sections are given in Ref (5).

5.4 *Materials Dosimetry Reference Facility (MDRF)—Reference Neutron Field:*

5.4.1 A high-intensity reference-neutron field for reactor dosimetry has been placed into operation, by NIST, in the pool adjacent to the Ford Nuclear Reactor (FNR) at the University of Michigan. This facility has recently been decommissioned but there is an

³ The boldface numbers given in parentheses refer to a list of references at the end of the text.

ongoing discussion on the need to resurrect the field at another research reactor. Neutron spectrum characterizations have been accomplished by measurement of spectral indexes (see Section 7) and DORT neutron transport calculations. The fast neutron fluence rate has been determined by means of the fluence transfer procedure (see 6.2) from a NIST ²⁵²Cf standard neutron field and by other means (Ref

5.4 There are several additional facilities that can provide free field fluence irradiations that qualify as reference fields. The following is a list of some of the facilities that have characterized reference fields:

5.4.1 Annular Core Research Reactor (ACRR) Central Cavity – Reference Neutron Field (10,11)—MDRF-certified neutron fluences were monitored with the ⁵⁸Ni(n,p)⁵⁸Co activation reaction. The fast neutron spectrum was similar to that at the inside surface of the pressure vessel in a PWR reactor with a thermal shield. Spectral monitoring measurements during changes in nearest neighbors in the pool and fuel element changes at the nearest edge of the core indicated fluence rate level variations but very small (<5%) changes in the spectrum (for example, no significant changes in various reaction rate ratios).

5.4.2 The MDRF was employed for calibration and validation experiments in support of materials neutron dosimetry for the nuclear reactor industry and for the metallurgical community engaged in estimating radiation damage in steel. Field characterization and user operation of the facility was a joint effort by the NIST and the Phoenix Memorial Laboratory of the University of Michigan. This reference field was a natural extension of a long-term NIST program to develop standard and reference neutron fields for measurement assurance applications and for testing new detectors and techniques.

5.4.3 The MDRF complemented the Cavity Fission Source at the NIST reactor by providing a tenfold increase in the fast neutron fluence rate, a much larger irradiation volume with modest gradients, and a neutron spectrum similar to that at the inner surface of a PWR pressure vessel. An option to alter the intermediate-energy neutron rate was afforded by installation of a ¹⁰B filter liner. The fluence rate (E > 1 MeV) was 2.7×10^{11} neutrons cm⁻²s⁻¹ and the neutron-to-gamma fluence ratio is 0.35 (dimensionless).

- 2
- 5.4.2 ACRR Lead-Boron Cavity Insert – Reference Neutron Field (11),
 - 5.4.3 YAYOI fast neutron field – Reference Neutron Field (12,13),
 - 5.4.4 SIGMA-SIGMA neutron field – Reference Neutron Field (12,13).

6. Applications of Benchmark Fields

6.1 *Notation—Reaction Rate, Fluence Rate, and Fluence*—The notation employed in this section will follow that in E261 (Standard Practice for Determining Neutron Fluence Rate, and Spectra by Radioactivation Techniques) except as noted. The reaction rate, *R*, for some neutron-nuclear reaction {reactions/[(dosimeter target nucleus)(second)]} is given by:

$$R = \int_0^{\infty} \sigma(E) \phi(E) dE \quad (1)$$

or:

$$R = \bar{\sigma} \phi \quad (2)$$

where:

- $\sigma(E)$ = the dosimeter reaction cross section at energy *E* (typically of the order of 10⁻²⁴ cm²),
- $\phi(E)$ = the differential neutron fluence rate, that is the fluence per unit time and unit energy for neutrons with energies between *E* and *E* + *dE* (neutrons cm⁻²s⁻¹ MeV⁻¹),
- ϕ = the total fluence rate (neutrons cm⁻²s⁻¹), the integral of $\phi(E)$ over all *E*, and
- $\bar{\sigma}$ = the spectral-averaged value of $\sigma(E)$, R/ϕ .

NOTE 1—Neutron fluence and fluence rate are defined formally in Terminology E170 under the listing “particle fluence.” Fluence is just the time integral of the fluence rate over the time interval of interest. The fluence rate is also called the flux or flux density in many papers and books on neutron transport theory.

6.1.1 The reaction rate is found experimentally using an active instrument such as a fission chamber (see Ref (12)(14)) or a passive dosimeter such as a solid state track recorder (see Test Method E854), a helium accumulation fluence monitor (see Test Method E910), or a radioactivation dosimeter (see Practice E261). For the radioactivation method, there are also separate standards for many particularly important dosimetry nuclides, for example, see Test Methods E263, E264, E265, E266, E343, E393, E523, E526, E704, E705, and E1297.

6.2 *Fluence Rate Transfer*: Note that if one determines $\phi = R/\bar{\sigma}$ from Eq 2, then the uncertainty in ϕ will be a propagation of the uncertainties in both *R* and $\bar{\sigma}$. The uncertainty in $\bar{\sigma}$ is frequently large, leading to a less accurate determination of ϕ than desired. However, if one can make an additional irradiation of the same type of dosimeter in a standard neutron field with known fluence rate, then one may apply Eq 2 to both irradiations and write

$$\phi_A = \phi_B (R_A/R_B) (\bar{\sigma}_B/\bar{\sigma}_A) \quad (3)$$

where “A” denotes the field of interest and “B” denotes the standard neutron field benchmark. In Eq 3 the ratios of spectral average cross section, will have a small uncertainty if the spectral shapes $\phi_A(E)$ and $\phi_B(E)$ are fairly similar. There may also be important cancellation of poorly known factors in the ratio R_A/R_B , which will contribute to the better accuracy of Eq 3. Whether ϕ is better determined by Eq 3 or Eq 2 must be evaluated on a case by case basis. Often the fluence rate from Eq 3 is substantially