



SLOVENSKI STANDARD SIST EN ISO 80601-2-80:2024

01-november-2024

Nadomešča:

SIST EN ISO 80601-2-80:2019

Medicinska električna oprema - 2-80. del: Posebne zahteve za osnovno varnost in bistvene lastnosti pomožne ventilacijske opreme pri nezadostnem prezračevanju (ISO 80601-2-80:2024)

Medical electrical equipment - Part 2-80: Particular requirements for basic safety and essential performance of ventilatory support equipment for ventilatory insufficiency (ISO 80601-2-80:2024)

Medizinische elektrische Geräte - Teil 2-80: Besondere Festlegungen für die grundlegende Sicherheit und die wesentlichen Leistungsmerkmale von Heimbeatmungsgeräten zur Atemunterstützung von Patienten mit Atmungsinsuffizienz (ISO 80601-2-80:2024)

Appareils électromédicaux - Partie 2-80: Exigences particulières pour la sécurité de base et les performances essentielles des équipements d'assistance ventilatoire en cas d'insuffisance ventilatoire (ISO 80601-2-80:2024)

Ta slovenski standard je istoveten z: EN ISO 80601-2-80:2024

ICS:

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
-----------	--	--

SIST EN ISO 80601-2-80:2024

en,fr,de

EUROPEAN STANDARD

EN ISO 80601-2-80

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2024

ICS 11.040.10

Supersedes EN ISO 80601-2-80:2019

English Version

Medical electrical equipment - Part 2-80: Particular requirements for basic safety and essential performance of ventilatory support equipment for ventilatory insufficiency (ISO 80601-2-80:2024)

Appareils électromédicaux - Partie 2-80: Exigences particulières pour la sécurité de base et les performances essentielles des équipements d'assistance ventilatoire en cas d'insuffisance ventilatoire (ISO 80601-2-80:2024)

Medizinische elektrische Geräte - Teil 2-80: Besondere Festlegungen für die grundlegende Sicherheit und die wesentlichen Leistungsmerkmale von Heimbeatmungsgeräten zur Atemunterstützung von Patienten mit Atmungsinsuffizienz (ISO 80601-2-80:2024)

This European Standard was approved by CEN on 5 August 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

[SIST EN ISO 80601-2-80:2024](https://standards.itih.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024)

<https://standards.itih.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024>

European foreword

This document (EN ISO 80601-2-80:2024) has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" in collaboration with Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2025, and conflicting national standards shall be withdrawn at the latest by March 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 80601-2-80:2019.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

(<https://standards.iteh.ai>)

Endorsement notice

The text of ISO 80601-2-80:2024 has been approved by CEN as EN ISO 80601-2-80:2024 without any modification.

<https://standards.iteh.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024>



**International
Standard**

ISO 80601-2-80

**Medical electrical equipment —
Part 2-80:
Particular requirements for basic
safety and essential performance of
ventilatory support equipment for
ventilatory insufficiency**

Appareils électromédicaux —

*Partie 2-80: Exigences particulières pour la sécurité de base
et les performances essentielles des équipements d'assistance
ventilatoire en cas d'insuffisance ventilatoire*

**Second edition
2024-08**

ISO 80601-2-80:2024(en)

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN ISO 80601-2-80:2024](https://standards.iteh.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024)

<https://standards.iteh.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

ISO 80601-2-80:2024(en)

Contents

Foreword.....	v
Introduction.....	vii
201.1 Scope, object and related standards.....	1
201.2 Normative references	4
201.3 Terms and definitions.....	5
201.4 General requirements.....	24
201.5 General requirements for testing of <i>ME equipment</i>	28
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>	29
201.7 <i>ME equipment</i> identification, <i>marking</i> and documents	29
201.8 Protection against electrical <i>hazards</i> from <i>ME equipment</i>	37
201.9 Protection against <i>mechanical hazards</i> of <i>ME equipment</i> and <i>ME systems</i>	37
201.10 Protection against unwanted and excessive radiation <i>hazards</i>	39
201.11 Protection against excessive temperatures and other <i>hazards</i>	39
201.12 Accuracy of controls and instruments and protection against hazardous outputs	43
201.13 <i>Hazardous situations</i> and fault conditions for <i>ME equipment</i>	55
201.14 <i>Programmable electrical medical systems (PEMS)</i>	56
201.15 Construction of <i>ME equipment</i>	57
201.16 <i>ME systems</i>	58
201.17 Electromagnetic compatibility of <i>ME equipment</i> and <i>ME systems</i>	58
201.101 Gas connections.....	59
201.102 Requirements for the <i>VBS</i> and <i>accessories</i>	62
201.103 Spontaneous breathing during loss of power supply	63
201.104 Indication of duration of operation	64
201.105 <i>Functional connection</i>	64
201.106 Display loops	65
201.107 <i>Ventilatory support equipment</i> security	65
202 Electromagnetic disturbances — Requirements and tests	66
206 Usability.....	67
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	69
211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment.....	71
Annex C (informative) Guide to <i>marking</i> and labelling requirements for <i>ME equipment</i> and <i>ME systems</i>	72

ISO 80601-2-80:2024(en)

Annex D (informative) <i>Symbols on marking</i>.....	79
Annex AA (informative) Particular guidance and rationale.....	81
Annex BB (informative) Data interface requirements	102
Annex CC (informative) Reference to the IMDRF <i>essential principles</i> and labelling guidances..	109
Bibliography.....	113
Terminology — Alphabetized index of defined terms.....	116

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN ISO 80601-2-80:2024](https://standards.iteh.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024)

<https://standards.iteh.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024>

ISO 80601-2-80:2024(en)

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives or www.iec.ch/members_experts/refdocs).

ISO and IEC draw attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO and IEC take no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO and IEC had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents and <https://patents.iec.ch>. ISO and IEC shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html. In the IEC, see www.iec.ch/understanding-standards.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Medical equipment, software, and systems*, Subcommittee SC D, *Particular medical equipment, software, and systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215 *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 80601-2-80:2018), which has been technically revised.

The main changes are as follows:

- alignment with IEC 60601-1:2005+AMD1:2012+AMD2:2020, IEC 60601-1-2:2014+AMD1:2020 IEC 60601-1-6:2010+AMD1:2013+AMD2:2020, IEC 60601-1-8:2006+AMD1:2012+AMD2:2020 and IEC 60601-1-11:2015+AMD1:2020;
- clarified *maximum limited pressure* requirements;
- clarified high *airway pressure alarm condition* requirements;
- added requirements for *ventilatory support equipment system recovery*; and
- harmonization with ISO 20417, where appropriate.

ISO 80601-2-80:2024(en)

A list of all parts in the ISO 80601 series and the IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html and www.iec.ch/national-committees.

iTeh Standards (<https://standards.itih.ai>) Document Preview

[SIST EN ISO 80601-2-80:2024](https://standards.itih.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024)

<https://standards.itih.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024>

ISO 80601-2-80:2024(en)

Introduction

This document specifies requirements for *ventilatory support equipment* that is intended for use in the *home healthcare environment* for *patients* who are not dependent for *ventilation* for their life support. *Ventilatory support equipment* is frequently used in locations where *supply mains* is not reliable. *Ventilatory support equipment* is often supervised by non-healthcare personnel (*lay operators*) with varying levels of training. *Ventilatory support equipment* conforming with this document can be used elsewhere (i.e. in healthcare facilities).

Varying levels of ventilatory support is often used for *patients* who have stable ventilatory needs and in some cases, changing needs as their disease worsens. This document addresses *patients* who typically have severe enough respiratory function to prohibit certain activities that the *patient* might normally pursue, and to interfere with daily living, occurring in association with measurements of respiratory mechanics or gas exchange that are markedly abnormal. This is best characterised by *lung* functions worse than^[36]

- $FEV_1/FVC^1 < 70 \%$, or
- $FEV_1 < 50 \%$ predicted

where

FEV_1 is the forced expiratory volume in 1 s, and

FVC is the forced vital capacity.

Examples of diseases that require ventilatory support are:

- moderate to severe Chronic Obstructive Pulmonary Disease (COPD);
- moderate Amyotrophic Lateral Sclerosis (ALS)^[44];
- severe bronchopulmonary dysplasia; and
- muscular dystrophy.

Ventilatory support equipment intended for this group of *patients* typically can require *technical alarm conditions* in the event that *essential performance* is absent. The most fragile of these *patients* would likely experience injury, but not serious injury or death, with the loss of this *artificial ventilation*. For these *patients*, it is likely that ventilatory support is needed during waking hours while *patients* are moving inside or outside the home in order to facilitate mobility and functional independence in the activities of daily living.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *terms defined in Clause 3 of the general standard², in this document or as noted: italic type*; and
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

In referring to the structure of this document, the term:

¹ This is also known as the Tiffeneau-Pinelli index.

² The general standard is IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*.

ISO 80601-2-80:2024(en)

- “clause” means one of the five numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7, 201.8, etc.); and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or capability; and
- “must” is used to express an external constraint.

Annex C contains a guide to the *marking* and labelling requirements in this document.

Annex D contains a summary of the *symbols* referenced in this document.

Requirements in this document have been decomposed so that each requirement is uniquely delineated. This is done to support automated requirements tracking.

[SIST EN ISO 80601-2-80:2024](https://standards.iteh.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024)

<https://standards.iteh.ai/catalog/standards/sist/c5baa9c9-5748-4b38-a44c-ea714e0dd7a5/sist-en-iso-80601-2-80-2024>