



SLOVENSKI STANDARD
oSIST prEN IEC 60601-2-40:2023
01-maj-2023

Medicinska električna oprema - 2-40. del: Posebne zahteve za osnovno varnost in bistvene lastnosti za elektromiografe in opremo za izzvane odzive

Medical electrical equipment - Part 2-40: Particular requirements for the basic safety and essential performance of electromyographs and evoked response equipment

Medizinische elektrische Geräte - Teil 2-40: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Elektromyographen und Geräten für evozierte Potentiale

Appareils électromédicaux - Partie 2-40: Exigences particulières pour la sécurité de base et les performances essentielles des électromyographes et des appareils à potentiel évoqué

Ta slovenski standard je istoveten z: prEN IEC 60601-2-40:2023

ICS:

11.040.50 Radiografska oprema Radiographic equipment

oSIST prEN IEC 60601-2-40:2023 en



62D/2019/CDV

COMMITTEE DRAFT FOR VOTE (CDV)

PROJECT NUMBER: IEC 60601-2-40 ED3	
DATE OF CIRCULATION: 2023-03-17	CLOSING DATE FOR VOTING: 2023-06-09
SUPERSEDES DOCUMENTS: 62D/1901/CD, 62D/1928A/CC	

IEC SC 62D : PARTICULAR MEDICAL EQUIPMENT, SOFTWARE, AND SYSTEMS	
SECRETARIAT: United States of America	SECRETARY: Ms Ladan Bulookbashi
OF INTEREST TO THE FOLLOWING COMMITTEES:	PROPOSED HORIZONTAL STANDARD: <input type="checkbox"/> Other TC/SCs are requested to indicate their interest, if any, in this CDV to the secretary.
FUNCTIONS CONCERNED: <input type="checkbox"/> EMC <input type="checkbox"/> ENVIRONMENT <input type="checkbox"/> QUALITY ASSURANCE <input checked="" type="checkbox"/> SAFETY	
<input checked="" type="checkbox"/> SUBMITTED FOR CENELEC PARALLEL VOTING Attention IEC-CENELEC parallel voting The attention of IEC National Committees, members of CENELEC, is drawn to the fact that this Committee Draft for Vote (CDV) is submitted for parallel voting. The CENELEC members are invited to vote through the CENELEC online voting system.	<input type="checkbox"/> NOT SUBMITTED FOR CENELEC PARALLEL VOTING

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of

- any relevant patent rights of which they are aware and to provide supporting documentation,
- any relevant "in some countries" clauses to be included should this proposal proceed. Recipients are reminded that the enquiry stage is the final stage for submitting "in some countries" clauses. See AC/22/2007.

TITLE:

Medical electrical equipment - Part 2-40: Particular requirements for the basic safety and essential performance of electromyographs and evoked response equipment

PROPOSED STABILITY DATE: 2029

NOTE FROM TC/SC OFFICERS:

IEC 60601-2-40 is revised to align to the Amendment projects of the IEC 60601-1 series and to incorporate several significant technical changes. Please refer to 62D/1792/DC and 62D/1808/INF for maintenance decision. National Committees are invited to vote on the Committee Draft for Vote at this stage.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN IEC 60601-2-40:2023](https://standards.iteh.ai/catalog/standards/sist/788e88a0-6b7b-429f-93e9-635fbf25752b/osist-pren-iec-60601-2-40-2023)

<https://standards.iteh.ai/catalog/standards/sist/788e88a0-6b7b-429f-93e9-635fbf25752b/osist-pren-iec-60601-2-40-2023>

CONTENTS

1		
2	FOREWORD.....	4
3	INTRODUCTION.....	7
4	201.1 Scope, object and related standards	8
5	201.2 Normative references.....	10
6	201.3 Terms and definitions.....	10
7	201.4 General requirements	12
8	201.5 General requirements for testing of ME EQUIPMENT	12
9	201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	12
10	201.7 ME EQUIPMENT identification, marking and documents	12
11	201.8 Protection against electrical HAZARDS from ME EQUIPMENT	14
12	201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	15
13	201.10 Protection against unwanted and excessive radiation HAZARDS	15
14	201.11 Protection against excessive temperatures and other HAZARDS	15
15	201.12 Accuracy of controls and instruments and protection against hazardous	
16	outputs	16
17	201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	18
18	201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	19
19	201.15 Construction of ME EQUIPMENT.....	19
20	201.16 ME SYSTEMS	19
21	201.17 ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS	19
22	202 * ELECTROMAGNETIC DISTURBANCES – Requirements and tests	19
23	Annexes https://standards.iteh.ai/catalog/standards/sist/788e88a0-6b7b-429f-93e9-333b257571e1/prn-iec-60601-2-40-2023	24
24	Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT	
25	and ME SYSTEMS.....	24
26	Annex AA (informative) Particular guidance and rationale	25
27	Bibliography.....	31
28	Index of defined terms used in this particular standard.....	32
29		
30	Figure AA.1 – Suggested test layout for EMISSION and IMMUNITY testing.....	28
31	Figure AA.2 – Example of test setup for protection against the effects of HF SURGICAL	
32	ME EQUIPMENT	29
33	Figure AA.3 – Example of test setup for protection against the effects of HF SURGICAL	
34	ME EQUIPMENT	30
35		
36	Table 202.101 – Pass/fail criteria for Table 4 of IEC 60601-1-2:2014/AMD1:2020.....	21
37	Table 202.102 – Pass/fail criteria for Table 7 of IEC 60601-1-2:2014/AMD1:2020.....	22
38	Table 202.103 – Pass/fail criteria for Table 8 of IEC 60601-1-2:2014/AMD1:2020.....	23
39	Table 202.104 – Pass/fail criteria for Table 5, 6, 9 of IEC 60601-1-2:2014/AMD1:2020.....	23
40		
41	Table 201.C.101 – Marking on the outside of ELECTROMYOGRAPHES and EVOKED	
42	RESPONSE EQUIPMENT or its parts.....	24
43		
44		

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-40: Particular requirements for the basic safety and essential performance of electromyographs and evoked response equipment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-40 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition of IEC 60601-2-40 published in 2016. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) added requirements for constant voltage stimulators ;
- b) clarified requirements for VISUAL STIMULATORS.

The text of this standard is based on the following documents:

FDIS	Report on voting
62D/XXXX/FDIS	62D/XXXX/RVD

98
99 Full information on the voting for the approval of this standard can be found in the report on
100 voting indicated in the above table.

101 This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

102 In this standard, the following print types are used:

- 103 – Requirements and definitions: roman type.
104 – *Test specifications: italic type.*
105 – Informative material appearing outside of tables, such as notes, examples and references: in smaller type.
106 Normative text of tables is also in a smaller type.
107 – TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS
108 NOTED: SMALL CAPITALS.

109 In referring to the structure of this document, the term

- 110 – “clause” means one of the seventeen numbered divisions within the table of contents,
111 inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
112 – “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all
113 subclauses of Clause 7).

114 References to clauses within this document are preceded by the term “Clause” followed by
115 the clause number. References to subclauses within this particular standard are by number
116 only.

117 In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any
118 combination of the conditions is true.

119 The verbal forms used in this document conform to usage described in Clause 7 of the
120 ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- 121 – “shall” means that compliance with a requirement or a test is mandatory for compliance
122 with this document;
123 – “should” means that compliance with a requirement or a test is recommended but is not
124 mandatory for compliance with this document;
125 – “may” is used to describe a permissible way to achieve compliance with a requirement or
126 test.

127 An asterisk (*) as the first character of a title, or at the beginning of a paragraph or table title,
128 indicates that there is guidance or rationale related to that item in Annex AA.

129 A list of all parts of the IEC 60601 series, published under the general title *Medical electrical*
130 *equipment*, can be found on the IEC website.

131

132 The committee has decided that the contents of this publication will remain unchanged until
133 the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data
134 related to the specific publication. At this date, the publication will be

- 135 • reconfirmed,
- 136 • withdrawn,
- 137 • replaced by a revised edition, or
- 138 • amended.

139

140

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN IEC 60601-2-40:2023](https://standards.iteh.ai/catalog/standards/sist/788e88a0-6b7b-429f-93e9-635fbf25752b/osist-pren-iec-60601-2-40-2023)

<https://standards.iteh.ai/catalog/standards/sist/788e88a0-6b7b-429f-93e9-635fbf25752b/osist-pren-iec-60601-2-40-2023>

141

INTRODUCTION

142 This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of
143 ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT. It amends and supplements
144 IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for basic safety
145 and essential performance* (IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, and IEC
146 60601-1:2005/AMD2:2020), hereinafter referred to as the general standard.

147 The aim of this revision is to bring this particular standard up to date with reference to edition
148 3.2 of the general standard.

149 The requirements of this particular standard take priority over those of the general standard.

150 A “General guidance and rationale” for the more important requirements of this
151 particular standard is included in Annex AA. It is considered that knowledge of the
152 reasons for these requirements will not only facilitate the proper application of the
153 document but will, in due course, expedite any revision necessitated by changes in
154 clinical practice or as a result of developments in technology. However, Annex AA
155 does not form part of the requirements of this document.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN IEC 60601-2-40:2023](https://standards.iteh.ai/catalog/standards/sist/788e88a0-6b7b-429f-93e9-635fbf25752b/osist-pren-iec-60601-2-40-2023)

<https://standards.iteh.ai/catalog/standards/sist/788e88a0-6b7b-429f-93e9-635fbf25752b/osist-pren-iec-60601-2-40-2023>

156 **MEDICAL ELECTRICAL EQUIPMENT –**
157 **Part 2-40: Particular requirements for the basic safety and essential**
158 **performance of electromyographs and evoked response equipment**
159
160

161 **201.1 Scope, object and related standards**

162 Clause 1 of the general standard¹ applies, except as follows:

163 **201.1.1 Scope**

164 *Replacement:*

165 This particular standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of
166 ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT, hereafter referred to as ME EQUIPMENT.

167 NOTE 1 Myofeedback equipment, where the capturing of muscle contraction is based on electromyography, is
168 within the scope of this particular standard.

169 NOTE 2 EMG/EP equipment is intended for diagnostic and monitoring applications.

170 NOTE 3 If the ME EQUIPMENT supports both ELECTROMYOGRAPHY and EVOKED RESPONSE STIMULATION,
171 clauses for electrical, auditory, and visual stimulators are applicable. In case the equipment supports
172 ELECTROMYOGRAPHY, but not EVOKED RESPONSE STIMULATION, clauses concerning solely requirements for
173 stimulators are NOT within scope.

174 If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to
175 ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the
176 case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

177 The following ME EQUIPMENT are excluded: <https://standards.iteh.ai/catalog/standards/sist/788e88a0-6b7b-429f-93e9-429f-93e9-pren-iec-60601-2-40-2023>

178 ME EQUIPMENT intended for therapeutic application

179 ME EQUIPMENT intended for transcutaneous electrical nerve stimulators and electrical muscle
180 stimulators (ME EQUIPMENT covered by IEC 60601-2-10.)

181 **201.1.2 Object**

182 *Replacement:*

183 The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL
184 PERFORMANCE requirements for ELECTROMYOGRAPHS and EVOKED RESPONSE EQUIPMENT [as
185 defined in 201.3.201 and 201.3.202.]

186 **201.1.3 Collateral standards**

187 *Addition:*

188 This particular standard refers to those applicable collateral standards that are listed in
189 Clause 2 of the general standard and Clause 201.2 of this particular standard.

¹ The general standard is IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

190 IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020 apply as modified in Clause 202.
191 IEC 60601-1-3, IEC 60601-1-8 and IEC 60601-1-10 do not apply. All other published collateral
192 standards in the IEC 60601-1 series apply as published.

193 **201.1.4 Particular standards**

194 *Replacement:*

195 In the IEC 60601 series, particular standards may modify, replace or delete requirements
196 contained in the general standard and collateral standards as appropriate for the particular
197 ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL
198 PERFORMANCE requirements.

199 A requirement of a particular standard takes priority over the general standard.

200 For brevity, IEC 60601-1 is referred to in this particular standard as the general standard.
201 Collateral standards are referred to by their document number.

202 The numbering of clauses and subclauses of this particular standard corresponds to that of
203 the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content
204 of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x",
205 where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this
206 particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral
207 standard, 203.4 in this particular standard addresses the content of Clause 4 of the
208 IEC 60601-1-3 collateral standard, etc.) The changes to the text of the general standard are
209 specified by the use of the following words:

210 "Replacement" means that the clause or subclause of the general standard or applicable
211 collateral standard is replaced completely by the text of this particular standard.

212 "Addition" means that the text of this particular standard is additional to the requirements of
213 the general standard or applicable collateral standard.

214 "Amendment" means that the clause or subclause of the general standard or applicable
215 collateral standard is amended as indicated by the text of this particular standard.

216 Subclauses, figures or tables which are additional to those of the general standard are
217 numbered starting from 201.101. However, due to the fact that definitions in the general
218 standard are numbered 3.1 through 3.154, additional definitions in this standard are
219 numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and
220 additional items aa), bb), etc.

221 Subclauses, figures or tables which are additional to those of a collateral standard are
222 numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for
223 IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

224 The term "this standard" is used to make reference to the general standard, any applicable
225 collateral standards and this particular standard taken together.

226 Where there is no corresponding clause or subclause in this particular standard, the clause or
227 subclause of the general standard or applicable collateral standard, although possibly not
228 relevant, applies without modification; where it is intended that any part of the general
229 standard or applicable collateral standard, although possibly relevant, is not to be applied, a
230 statement to that effect is given in this particular standard.

231 **201.2 Normative references**

232 The following documents, in whole or in part, are normatively referenced in this document and
 233 are indispensable for its application. For dated references, only the edition cited applies. For
 234 undated references, the latest edition of the referenced document (including any
 235 amendments) applies. NOTE Informative references are listed in the bibliography beginning on page 31.

236 *Addition:*

237 IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic*
 238 *safety and essential performance*
 239 Amendment 1:2012
 240 Amendment 2:2020

241
 242 IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic*
 243 *safety and essential performance – Collateral standard: Electromagnetic disturbances –*
 244 *Requirements and tests*
 245 Amendment 1:2020

246

247 IEC 60318 (all parts), *Electroacoustics – Simulators of human head and ear*

248 ISO 15004-2, *Ophthalmic instruments – Fundamental requirements and test methods – Part 2:*
 249 *Light hazard protection*

250 **201.3 Terms and definitions**

251 For the purposes of this document, the terms and definitions given in IEC 60601-1:2005,
 252 IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and the following apply.

253 ISO and IEC maintain terminological databases for use in standardization at the following
 254 addresses:

- 255 • IEC Electropedia: available at <http://www.electropedia.org/>
- 256 • ISO Online browsing platform: available at <http://www.iso.org/obp>

257 NOTE An index of defined terms is found beginning on page 32.

258 *Addition:*

259 **201.3.201**

260 **ELECTROMYOGRAPH**

261 ME EQUIPMENT for the detection or recording of biopotentials accompanying nerve and muscle
 262 action, either spontaneously, intentionally or evoked by electrical or other stimulation

263 **201.3.202**

264 **EVOKED RESPONSE EQUIPMENT**

265 ME EQUIPMENT for the detection or recording of biopotentials resulting from an evoking
 266 stimulus

267 Note 1 to entry: The stimulus may be electrical, tactile, auditory, visual, olfactory, etc.

268 **201.3.203**

269 **ELECTRICAL STIMULATOR**

270 part of ME EQUIPMENT for the application of electric currents via ELECTRODES in direct contact
 271 with the PATIENT, for the evoking of biopotentials