

---

**Medicinska električna oprema - Karakteristike digitalnih naprav za rentgensko slikanje - 2-1. del: Ugotavljanje učinkovitosti dvoenergijskega odštevanja - Detektorji, ki se uporabljajo pri radiografskem slikanju z dvojno energijo (IEC 62220-2-1:2023)**

Medical electrical equipment - Characteristics of digital X-ray imaging devices - Part 2-1: Determination of dual-energy subtraction efficiency - Detectors used for dual-energy radiographic imaging (IEC 62220-2-1:2023)

Medizinische elektrische Geräte - Merkmale digitaler Röntgenbildgeräte - Teil 2-1: Bestimmung des Wirkungsgrades der Zwei-Energie-Subtraktion - Detektoren für die Zwei-Energie-Röntgenbildgebung (IEC 62220-2-1:2023)

Appareils électromédicaux - Caractéristiques des dispositifs d'imagerie à rayonnement X - Partie 2-1: Détermination de l'efficacité de soustraction à double énergie - Détecteurs utilisés en imagerie radiographique à double énergie (IEC 62220-2-1:2023)

**Ta slovenski standard je istoveten z: EN IEC 62220-2-1:2023**

---

**ICS:**

11.040.50      Radiografska oprema      Radiographic equipment

**SIST EN IEC 62220-2-1:2024**      en



EUROPEAN STANDARD

EN IEC 62220-2-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2023

ICS 11.040.50

English Version

Medical electrical equipment - Characteristics of digital X-ray  
imaging devices - Part 2-1: Determination of dual-energy  
subtraction efficiency - Detectors used for dual-energy  
radiographic imaging  
(IEC 62220-2-1:2023)

Appareils électromédicaux - Caractéristiques des dispositifs  
d'imagerie à rayonnement X - Partie 2-1: Détermination de  
l'efficacité de soustraction à double énergie - Détecteurs  
utilisés en imagerie radiographique à double énergie  
(IEC 62220-2-1:2023)

Medizinische elektrische Geräte - Merkmale digitaler  
Röntgenbildgeräte - Teil 2-1: Bestimmung des  
Wirkungsgrades der Zwei-Energie-Subtraktion - Detektoren  
für die Zwei-Energie-Röntgenbildgebung  
(IEC 62220-2-1:2023)

This European Standard was approved by CENELEC on 2023-09-13. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

**EN IEC 62220-2-1:2023 (E)****European foreword**

The text of document 62B/1288/CDV, future edition 1 of IEC 62220-2-1, prepared by SC 62B "Diagnostic imaging equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 62220-2-1:2023.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2024-06-13
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2026-09-13

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

**Endorsement notice**

The text of the International Standard IEC 62220-2-1:2023 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

IEC 60601-2-54	NOTE	Approved as EN 60601-2-54
IEC 60601-1-3:2008	NOTE	Approved as EN 60601-1-3:2008 (not modified) + A11:2016
IEC 61674:2012	NOTE	Approved as EN 61674:2013 (not modified)
IEC 62220-1-1:2015	NOTE	Approved as EN 62220-1-1:2015 (not modified)
IEC 60601-2-68:2014	NOTE	Approved as EN 60601-2-68:2015 (not modified)

## Annex ZA (normative)

### Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: [www.cencenelec.eu](http://www.cencenelec.eu).

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60336	-	Medical electrical equipment - X-ray tube assemblies for medical diagnosis - Focal spot dimensions and related characteristics	EN IEC 60336	-
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[SIST EN IEC 62220-2-1:2024](https://standards.iteh.ai/catalog/standards/sist/e62d7725-c4cd-47a7-8d08-978b6d0dd693/sist-en-iec-62220-2-1-2024)

<https://standards.iteh.ai/catalog/standards/sist/e62d7725-c4cd-47a7-8d08-978b6d0dd693/sist-en-iec-62220-2-1-2024>





IEC 62220-2-1

Edition 1.0 2023-08

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Medical electrical equipment – Characteristics of digital X-ray imaging devices –  
Part 2-1: Determination of dual-energy subtraction efficiency – Detectors used  
for dual-energy radiographic imaging**

**Appareils électromédicaux – Caractéristiques des dispositifs d'imagerie à  
rayonnement X –  
Partie 2-1: Détermination de l'efficacité de soustraction à double énergie –  
DéTECTEURS utilisés en imagerie radiographique à double énergie**

[SIST EN IEC 62220-2-1:2024](https://standards.iteh.ai/SIST/EN/IEC/62220-2-1:2024)

<https://standards.iteh.ai/catalog/standards/sist/e62d7725-c4cd-47a7-8d08-978b6d0dd693/sist-en-iec-62220-2-1-2024>

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.50

ISBN 978-2-8322-7305-0

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references .....	6
3 Terms and definitions .....	6
4 Requirements .....	7
4.1 Operating conditions .....	7
4.2 X-RAY EQUIPMENT.....	7
4.3 RADIATION QUALITY.....	8
4.4 TEST DEVICE .....	8
4.5 Geometry.....	9
4.6 IRRADIATION conditions .....	9
4.6.1 General conditions.....	9
4.6.2 IRRADIATION for the determination of the dual-energy contrast.....	10
4.6.3 AIR KERMA measurement .....	10
5 Corrections of RAW DATA .....	10
6 Definitions of the REGIONS OF INTEREST .....	11
7 Calculation of derived images.....	11
7.1 Calculation of high-energy and low-energy images .....	11
7.2 Calculation of TISSUE-SUBTRACTED IMAGES .....	12
8 Determination of the dual-energy subtraction efficiency.....	13
8.1 Definition and formula of the <b>DSE</b> .....	13
8.2 Determination of the dual-energy contrast.....	13
9 Format of conformance statement .....	14
Annex A (informative) Dual-energy subtraction efficiency interpretation .....	15
A.1 Metric description .....	15
A.2 Sample device comparisons.....	16
Bibliography.....	19
Index of defined terms used in this particular standard.....	20
Figure 1 – TEST DEVICE for the determination of the dual-energy contrast.....	8
Figure 2 – Geometry for all IRRADIATIONS of the TEST DEVICE .....	9
Figure 3 – Feature and background REGIONS OF INTEREST defined about circular image of features .....	11
Figure A.1 – Sample <b>DSE<sub>s</sub></b> results for a simulated MULTI-EXPOSURE DEVICE.....	15
Figure A.2 – Sample <b>DSE<sub>s</sub></b> results for a simulated MULTI-EXPOSURE DEVICE at increasing SCINTILLATOR thickness .....	17
Figure A.3 – Sample <b>DSE<sub>s</sub></b> results for a simulated MULTI-EXPOSURE DEVICE at differing X-RAY TUBE VOLTAGE used for their high-energy IRRADIATIONS .....	17
Figure A.4 – Sample <b>DSE<sub>s</sub></b> results for a simulated MULTI-EXPOSURE DEVICES at increasing PIXEL pitch.....	18



## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –  
CHARACTERISTICS OF DIGITAL X-RAY IMAGING DEVICES –**

**Part 2-1: Determination of dual-energy subtraction efficiency –  
Detectors used for dual-energy radiographic imaging**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62220-2-1 has been prepared by subcommittee 62B: Medical imaging equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

The text of this document is based on the following documents:

Draft	Report on voting
62B/1288/CDV	62B/1316/RVC

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

A list of all parts in the IEC 62220 series, published under the general title *Medical electrical equipment – Characteristics of digital X-ray imaging devices*, can be found on the IEC website.

In this document, terms printed in SMALL CAPITALS are used as defined in IEC 60788, in Clause 3 of this document or in other IEC publications referenced in the Index of defined terms. Where a defined term is used as a qualifier in another defined or undefined term, it is not printed in SMALL CAPITALS, unless the concept thus qualified is defined or recognized as a “derived term without definition”.

NOTE Attention is drawn to the fact that, in cases where the concept addressed is not strongly confined to the definition given in one of the publications listed above, a corresponding term is printed in lower-case letters.

In this document, certain terms that are not printed in SMALL CAPITALS have particular meanings, as follows:

- "shall" indicates a requirement that is mandatory for compliance;
- "should" indicates a strong recommendation that is not mandatory for compliance;
- "may" indicates a permitted manner of complying with a requirement or of avoiding the need to comply;
- "specific" is used to indicate definitive information stated in this document or referenced in other standards, usually concerning particular operating conditions, test arrangements or values connected with compliance;
- "specified" is used to indicate definitive information stated by the manufacturer in accompanying documents or in other documentation relating to the equipment under consideration, usually concerning its intended purposes, or the parameters or conditions associated with its use or with testing to determine compliance.

<https://standards.iteh.ai> The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.