
**Nickel alloys — Determination
of tantalum — Inductively
coupled plasma optical emission
spectrometric method**

*Alliages de nickel — Détermination du tantal — Méthode par
spectrométrie d'émission optique avec source à plasma induit par
haut fréquence*

ITEH Standards

(<https://standards.iteh.ai>)
Document Preview

[ISO 23166:2018](#)

<https://standards.iteh.ai/catalog/standards/iso/757a7d88-cb58-4a0f-83e2-8faaded58d15/iso-23166-2018>



Reference number
ISO 23166:2018(E)

© ISO 2018

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 23166:2018](#)

<https://standards.iteh.ai/catalog/standards/iso/757a7d88-cb58-4a0f-83e2-8faaded58d15/iso-23166-2018>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	2
6 Apparatus	3
7 Sampling and sample preparation	3
8 Procedure	4
8.1 Test portion	4
8.2 Preparation of test solution, T_{TA}	4
8.3 Optimization of the spectrometer	4
8.4 Pre-determination of the test solution	5
8.5 Preparation of calibration solutions for bracketing, $K_{I,Ta}$ and $K_{h,Ta}$	5
8.6 Measurement of test solutions	6
9 Expression of results	6
9.1 Method of calculation	6
9.2 Precision	6
9.2.1 Laboratory tests	6
9.2.2 Wavelength for measurement	6
9.2.3 Statistical evaluation	6
9.3 Trueness	7
10 Test report	7
Annex A (normative) Checking the performance of an inductively coupled plasma optical emission spectrometer	8
Annex B (informative) Notes on the test programme	10
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 155, *Nickel and nickel alloys*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

[ISO 23166:2018](http://www.iso.org/iso/23166:2018)

<https://standards.iteh.ai/catalog/standards/iso/757a7d88-cb58-4a0f-83e2-8faaded58d15/iso-23166-2018>

Nickel alloys — Determination of tantalum — Inductively coupled plasma optical emission spectrometric method

1 Scope

This document specifies an inductively coupled plasma optical emission spectrometric method for the determination of tantalum contents between 0,1 % and 5 % in nickel alloys.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 648, *Laboratory glassware — Single-volume pipettes*

ISO 1042, *Laboratory glassware — One-mark volumetric flasks*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <http://www.electropedia.org/> ISO 23166:2018

4 Principle

Dissolution of a test portion in a mixture of hydrofluoric, hydrochloric, nitric and phosphoric acid and fuming after addition of perchloric acid. Addition of hydrofluoric acid and, if desired, of an internal reference element and dilution of the solution to known volume. Nebulization of the solution into an inductively coupled plasma optical emission spectrometer and measurement of the intensity of the emitted light from tantalum, and, where appropriate, from the internal reference element, simultaneously.

An example of the wavelength for tantalum is given in [Table 1](#).

The method uses a calibration based on a very close matrix-matching of the calibration solutions to the sample and bracketing between 0,75 and 1,25 of the approximate content of tantalum in the sample to be analysed. The content of all elements in the sample has, therefore, to be approximately known. If the contents are not known, the sample has to be analysed by some semi-quantitative method. The advantage of this procedure is that all possible interferences from the matrix will be compensated, which will result in high accuracy. This is most important for spectral interferences, which can be severe in very highly alloyed matrixes. All possible interferences shall be kept at a minimum level. Therefore, it is essential that the spectrometer used meets the performance criteria specified in the method for the selected wavelengths.

The line corresponding to 240,06 nm has been investigated. If other lines are used, they shall be carefully checked. The wavelength for the internal reference element should be selected carefully. The

use of scandium at 363,07 nm is recommended. This wavelength is interference-free for the elements and contents generally found in nickel alloys.

Table 1 — Example of wavelength for tantalum

Element	Wavelength	Interferences
Tantalum	240,06	Co - Fe - Hf

NOTE The use of an internal reference element is not essential since no relevant differences between laboratories operating with or without internal reference elements were found when the precision test was carried out (see [9.2.2](#)).

5 Reagents

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade and only grade 2 water as specified in ISO 3696.

5.1 Hydrofluoric acid, HF, $\rho = 1,14$ g/ml, or $\rho = 1,17$ g/ml.

WARNING — Hydrofluoric acid is extremely irritating and corrosive to skin and mucous membranes producing severe skin burns which are slow to heal. In the case of contact with skin, wash well with water, apply a topical gel containing 2,5 % (mass fraction) calcium gluconate, and seek immediate medical treatment.

5.2 Hydrochloric acid, HCl, $\rho = 1,19$ g/ml.

5.3 Nitric acid, HNO₃, $\rho = 1,40$ g/ml.

5.4 Phosphoric acid, H₃PO₄, $\rho = 1,70$ g/ml.

5.5 Perchloric acid, HClO₄, $\rho = 1,54$ g/ml, or $\rho = 1,67$ g/ml.

<https://standards.iteh.ai/catalog/standards/iso/757a7d88-cb58-4a0f-83e2-8faaded58d15/iso-23166-2018>

WARNING — Perchloric acid vapour may cause explosions in the presence of ammonia, nitrous fumes or organic material in general. The use of fume hoods (water scrubbed) when using perchloric acid is highly recommended.

5.6 Internal reference element solution, 100 mg/l.

Choose a suitable element to be added as internal reference and prepare a 100 mg/l solution.

5.7 Tantalum standard solution, 10 g/l.

Weigh, to the nearest 0,001 g, 1 g of tantalum [minimum 99,9 % (mass fraction)], place it in a beaker ([6.1](#)), cover with an appropriate lid and dissolve it in a mixture of 10 ml of hydrofluoric acid ([5.1](#)) and 10 ml of nitric acid ([5.3](#)).

Cool and transfer quantitatively into a 100 ml one-mark volumetric flask. Dilute to the mark with water and mix.

1 ml of this solution contains 10 mg of tantalum.

5.8 Tantalum standard solution, 1 g/l.

Weigh, to the nearest 0,000 5 g, 0,1 g of tantalum [minimum 99,9 % (mass fraction)], place it in a beaker ([6.1](#)), cover with an appropriate lid and dissolve it in a mixture of 10 ml of hydrofluoric acid ([5.1](#)) and 10 ml of nitric acid ([5.3](#)).