



SLOVENSKI STANDARD
SIST EN IEC 60601-2-39:2025

01-marec-2025

Medicinska električna oprema - 2-39. del: Posebne zahteve za osnovno varnost in bistvene lastnosti opreme za trebušno dializo

Medical electrical equipment - Part 2-39: Particular requirements for basic safety and essential performance of peritoneal dialysis equipment

Medizinische elektrische Geräte - Teil 2-39: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von Peritoneal-Dialyse-Geräten

Appareils électromédicaux - Partie 2-39: Exigences particulières pour la sécurité de base et les performances essentielles des appareils de dialyse péritonéale

Ta slovenski standard je istoveten z: EN IEC 60601-2-39:2025

[SIST EN IEC 60601-2-39:2025](http://standards.sist.si/catalog/standards/sist/60601-2-39:2025/iec/60601-2-39:2025)

ICS:

11.040.99 Druga medicinska oprema Other medical equipment

SIST EN IEC 60601-2-39:2025

en

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN IEC 60601-2-39

January 2025

ICS 11.040.99

Supersedes EN IEC 60601-2-39:2019

English Version

**Medical electrical equipment - Part 2-39: Particular requirements
for the basic safety and essential performance of peritoneal
dialysis equipment
(IEC 60601-2-39:2024)**

Appareils électromédicaux - Partie 2-39: Exigences
particulières pour la sécurité de base et les performances
essentielles des appareils de dialyse péritonéale
(IEC 60601-2-39:2024)

Medizinische elektrische Geräte - Teil 2-39: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Peritoneal-Dialyse-
Geräten
(IEC 60601-2-39:2024)

This European Standard was approved by CENELEC on 2025-01-15. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

EN IEC 60601-2-39:2025 (E)**European foreword**

The text of document 62D/2162/FDIS, future edition 4 of IEC 60601-2-39, prepared by SC 62D "Particular medical equipment, software, and systems" of IEC/TC 62 "Medical equipment, software, and systems" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-39:2025.

The following dates are fixed:

- latest date by which the document has to be implemented at national (dop) 2026-01-31 level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with the (dow) 2028-01-31 document have to be withdrawn

This document supersedes EN IEC 60601-2-39:2019 and all of its amendments and corrigenda (if any).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

iTeh Standards

(<https://standards.itih.ai>)

Endorsement notice

The text of the International Standard IEC 60601-2-39:2024 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

IEC 60601-2-16	NOTE	Approved as EN IEC 60601-2-16
ISO 23500-3:2024	NOTE	Approved as EN ISO 23500-3:2024 (not modified)
ISO 23500-1:2024	NOTE	Approved as EN ISO 23500-1:2024 (not modified)
ISO 23500-5:2024	NOTE	Approved as EN ISO 23500-5:2024 (not modified)
ISO 80369-1:2018	NOTE	Approved as EN ISO 80369-1:2018 (not modified)
ISO 23500-4:2024	NOTE	Approved as EN ISO 23500-4:2024 (not modified)
ISO 11197:2019	NOTE	Approved as EN ISO 11197:2019 (not modified)
ISO 17664-1:2021	NOTE	Approved as EN ISO 17664-1:2021 (not modified)
ISO 17664-2:2021	NOTE	Approved as EN ISO 17664-2:2023 (not modified)
ISO 15883 series	NOTE	Approved as EN ISO 15883 series
IEC 80001-1:2021	NOTE	Approved as EN IEC 80001-1:2021 (not modified)
IEC 60601-1-9:2007	NOTE	Approved as EN 60601-1-9:2008 (not modified)
IEC 60601-1-9:2007/A1:2013	NOTE	Approved as EN 60601-1-9:2008/A1:2013 (not modified)
IEC 60601-1-9:2007/A2:2020	NOTE	Approved as EN 60601-1-9:2008/A2:2020 (not modified)

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Annex ZA of EN 60601-1:2006¹, applies, except as follows:

Add:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1	2006
-	-	(https://standards.iteh.ai)	+ AC	2010
+ A1	2012	Document Preview	+ A1	2013
-	-		+ AC	2014
-	-		+ A12	2014
+ A2	2020	SIST EN IEC 60601-2-39:2025	+ A2	2021
-	-		+ AC	2022
-	-		+ A13	2024
IEC 60601-1-10	2007	Medical electrical equipment - Part 1-10: General requirements for basic safety and essential performance - Collateral Standard: Requirements for the development of physiologic closed-loop controllers	EN 60601-1-10	2008
+ A1	2013		+ A1	2015
+ A2	2020		+ A2	2021

¹ As impacted by EN 60601-1:2006/AC:2010, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A1:2013/AC:2014, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021, EN 60601-1:2006/AC:2022-12 and EN 60601-1:2006/A13:2024.

EN IEC 60601-2-39:2025 (E)

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60601-1-11	2015	Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	EN 60601-1-11	2015
+ A1	2020		+ A1	2021
IEC 61672-1	2013	Electroacoustics - Sound level meters - Part 1: Specifications	EN 61672-1	2013
ISO 3744	2010	Acoustics - Determination of sound power levels and sound energy levels of noise sources using sound pressure - Engineering methods for an essentially free field over a reflecting plane	EN ISO 3744	2010

iTeh Standards
 (<https://standards.iteh.ai>)
 Document Preview

[SIST EN IEC 60601-2-39:2025](https://standards.iteh.ai/catalog/standards/sist/684e1137-3e30-40be-876f-7e3edc8da8d9/sist-en-iec-60601-2-39-2025)

<https://standards.iteh.ai/catalog/standards/sist/684e1137-3e30-40be-876f-7e3edc8da8d9/sist-en-iec-60601-2-39-2025>



IEC 60601-2-39

Edition 4.0 2024-12

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-39: Particular requirements for the basic safety and essential
performance of peritoneal dialysis equipment**

**Appareils électromédicaux –
Partie 2-39: Exigences particulières pour la sécurité de base et les
performances essentielles des appareils de dialyse péritonéale**

<https://standards.iteh.ai/catalog/standards/sist/684e1137-3e30-40be-876f-7e3edc8da8d9/sist-en-iec-60601-2-39-2025>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.99

ISBN 978-2-8327-0010-5

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	11
201.4 General requirements	13
201.5 General requirements for testing ME EQUIPMENT	17
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	17
201.7 ME EQUIPMENT identification, marking and documents	17
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	22
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	22
201.10 Protection against unwanted and excessive radiation HAZARDS	22
201.11 Protection against excessive temperatures and other HAZARDS	22
201.12 Accuracy of controls and instruments and protection against hazardous outputs	25
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	28
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	28
201.15 Construction of ME EQUIPMENT.....	30
201.16 * ME SYSTEMS.....	31
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	32
202 Electromagnetic disturbances – Requirements and tests.....	32
208 * General requirements, tests and guidance for ALARM SYSTEMS in MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS.....	33
209 Requirements for environmentally conscious design	35
210 Requirements for the development of PHYSIOLOGIC CLOSED-LOOP CONTROLLERS	35
211 * Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the HOME HEALTHCARE ENVIRONMENT	35
Annexes	36
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures.....	37
Annex AA (informative) Particular guidance and rationale.....	38
Bibliography.....	52
Index of defined terms used in this document	54
Figure AA.1 – Powered by SUPPLY MAINS only	44
Figure AA.2 – Alarm at depletion of battery for limited functionality.....	45
Figure AA.3 – Alarm before battery for limited functionality gets depleted (10 min minimum).....	45
Figure AA.4 – Alarm before battery for limited functionality gets depleted (battery lasting for equal or less than 10 min)	46

Figure AA.5 – Alarm at battery depletion.....	46
Figure AA.6 – Alarm before battery gets depleted (10 min minimum)	47
Figure AA.7 – Example of a PD ME SYSTEM	49
Table 201.101 – ESSENTIAL PERFORMANCE requirements	13

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN IEC 60601-2-39:2025](https://standards.iteh.ai/catalog/standards/sist/684e1137-3e30-40be-876f-7e3edc8da8d9/sist-en-iec-60601-2-39-2025)

<https://standards.iteh.ai/catalog/standards/sist/684e1137-3e30-40be-876f-7e3edc8da8d9/sist-en-iec-60601-2-39-2025>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-39: Particular requirements for the basic safety and essential performance of peritoneal dialysis equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60601-2-39 has been prepared by subcommittee 62D: Particular medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This fourth edition cancels and replaces the third edition published in 2018. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) update of references to IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, of references to IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020, of references to IEC 60601-1-8:2006, IEC 60601-1-8:2006/AMD1:2012 and IEC 60601-1-8:2006/AMD2:2020, of references to IEC 60601-1-9:2007, IEC 60601-1-9:2007/AMD1:2013 and IEC 60601-1-9:2007/AMD2:2020, of references to IEC 60601-1-10:2007, IEC 60601-1-10:2007/AMD1:2013 and IEC 60601-1-10:2007/AMD2:2020 and of references to IEC 60601-1-11:2015 and IEC 60601-1-11:2015/AMD1:2020;
- b) consideration of ESSENTIAL PERFORMANCE in SINGLE FAULT CONDITION regarding IEC 60601-1:2005/AMD1:2012/ISH1:2021;
- c) including the information given in the document 62D/1771A/INF regarding 201.11.8;
- d) including the information given in the document 62D/1734/INF regarding technical issues of the previous edition;
- e) including SECURITY (CYBERSECURITY) requirements;
- f) additions related to online PD SOLUTION generation (ONLINE PD);
- g) improvements regarding the definition of the APPLIED PART;
- h) improvement of the essential performance requirements clause/subclauses;
- i) improvements for labelling;
- j) other minor technical improvements;
- k) editorial improvements.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62D/2162/FDIS	62D/2182/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- terms defined in Clause 3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, in this document or as noted: SMALL CAPITALS.

In referring to the structure of this document, the term:

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

NOTE The attention of the users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

The minimum safety requirements specified in this document are considered to provide for a practical degree of safety in the operation of PERITONEAL DIALYSIS ME EQUIPMENT.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST EN IEC 60601-2-39:2025](https://standards.iteh.ai/catalog/standards/sist/684e1137-3e30-40be-876f-7e3edc8da8d9/sist-en-iec-60601-2-39-2025)

<https://standards.iteh.ai/catalog/standards/sist/684e1137-3e30-40be-876f-7e3edc8da8d9/sist-en-iec-60601-2-39-2025>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-39: Particular requirements for the basic safety and essential performance of peritoneal dialysis equipment

201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of PERITONEAL DIALYSIS ME EQUIPMENT, hereafter referred to as PD EQUIPMENT. It applies to PD EQUIPMENT intended for use either by medical staff or under the supervision of medical experts, including PD EQUIPMENT operated by the PATIENT, regardless of whether the PD EQUIPMENT is used in a hospital or domestic environment.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

This document does not take into consideration specific safety details of the DIALYSING SOLUTION control system of PD EQUIPMENT using regeneration of DIALYSING SOLUTION or CENTRAL DELIVERY SYSTEMS for DIALYSING SOLUTION. It does, however, take into consideration the specific safety requirements of such PD EQUIPMENT concerning electrical safety and PATIENT safety.

This document specifies the minimum safety requirements for PD EQUIPMENT. These PD EQUIPMENT are intended for use either by medical staff or for use by the PATIENT or other trained personnel under medical supervision.

This document includes all ME EQUIPMENT that is intended to deliver a PERITONEAL DIALYSIS treatment to a PATIENT, independent of the treatment duration and location.

These particular requirements do not apply to:

- PRE-MANUFACTURED DIALYSING SOLUTION bags,
- DIALYSING SOLUTION CIRCUITS,
- DIALYSING SOLUTION CONCENTRATE,
- DIALYSIS WATER supply systems (see ISO 23500-2 [1])¹,
- CENTRAL DELIVERY SYSTEMS for DIALYSING SOLUTION CONCENTRATES, described as systems for bulk mixing concentrate at a dialysis facility,
- equipment used to perform HAEMODIALYSIS (see IEC 60601-2-16 [2]).

¹ Numbers in square brackets refer to the Bibliography.