
Optični senzorji - 1-2. del: Merjenje deformacij - Porazdeljeno zaznavanje na podlagi Brillouinovega sipanja

Fibre Optic Sensors - Part 1-2: Strain measurement - Distributed sensing based on Brillouin scattering

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Capteurs fibroniques - Partie 1-2: Mesure de déformation - Détection répartie basée sur la diffusion de Brillouin

[oSIST prEN IEC 61757-1-2:2022](https://standards.iteh.ai/catalog/standards/sist/0a75a4b5-f319-4247-b41d-444444444444/sist-pr-en-iec-61757-1-2-2022)

<https://standards.iteh.ai/catalog/standards/sist/0a75a4b5-f319-4247-b41d-444444444444/sist-pr-en-iec-61757-1-2-2022>

Ta slovenski standard je istoveten z: prEN IEC 61757-1-2:2022

ICS:

33.180.99

Druga oprema za optična
vlakna

Other fibre optic equipment

oSIST prEN IEC 61757-1-2:2022

en



86C/1798/CDV

COMMITTEE DRAFT FOR VOTE (CDV)

PROJECT NUMBER:

IEC 61757-1-2 ED1

DATE OF CIRCULATION:

2022-06-24

CLOSING DATE FOR VOTING:

2022-09-16

SUPERSEDES DOCUMENTS:

86C/1768/CD, 86C/1796/CC

IEC SC 86C : FIBRE OPTIC SYSTEMS AND ACTIVE DEVICES

SECRETARIAT:

United States of America

SECRETARY:

Mr Fred Heismann

OF INTEREST TO THE FOLLOWING COMMITTEES:

TC 17,TC 18,TC 20,TC 38,TC 45,TC 65,TC 85

PROPOSED HORIZONTAL STANDARD:

☐

Other TC/SCs are requested to indicate their interest, if any, in this CDV to the secretary.

FUNCTIONS CONCERNED:

☐ EMC☐ ENVIRONMENT☒ QUALITY ASSURANCE☐ SAFETY☒ SUBMITTED FOR CENELEC PARALLEL VOTING☐ NOT SUBMITTED FOR CENELEC PARALLEL VOTING**Attention IEC-CENELEC parallel voting**

The attention of IEC National Committees, members of CENELEC, is drawn to the fact that this Committee Draft for Vote (CDV) is submitted for parallel voting.

The CENELEC members are invited to vote through the CENELEC online voting system.

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

TITLE:

Fibre Optic Sensors - Part 1-2: Strain measurement - Distributed sensing based on Brillouin scattering

PROPOSED STABILITY DATE: 2026

NOTE FROM TC/SC OFFICERS:

CONTENTS

1	FOREWORD	3
2	INTRODUCTION	5
3	1 Scope	6
4	2 Normative references	6
5	3 Terms, definitions, abbreviated terms and quantity symbols	6
6	3.1 Terms and definitions	6
7	3.2 Abbreviated terms	10
8	3.3 Quantity symbols	10
9	4 General test setups for measurement of performance parameters	10
10	4.1 General and test setup requirements	10
11	4.2 General documentation requirements	15
12	5 Measurement procedures for performance parameters	15
13	5.1 Strain measurement error	15
14	5.1.1 Test procedure and conditions	15
15	5.1.2 Parameter calculation and reporting	16
16	5.2 Spatial resolution	16
17	5.2.1 Test procedure and conditions	16
18	5.2.2 Parameter calculation and reporting	16
19	5.3 Strain repeatability	17
20	5.3.1 Test procedure and conditions	17
21	5.3.2 Parameter calculation and reporting	17
22	5.4 Spatial strain uncertainty	18
23	5.4.1 Test procedure and conditions	18
24	5.4.2 Parameter calculation and reporting	18
25	5.5 Warm-up time	18
26	5.5.1 Test procedure and conditions	18
27	5.5.2 Parameter calculation and reporting	19
28	5.6 Attenuation range	19
29	5.6.1 General	19
30	5.6.2 Method A based on strain repeatability	19
31	5.6.3 Method B based on spatial strain uncertainty	20
32	Bibliography	22
33	Figure 1 – Optical fibre strain profile and related strain sample points	8
34	Figure 2 – General test setup for single-ended configuration	11
35	Figure 3 – General test setup for loop configuration	12
36	Figure 4 – Measured vs applied strain (typical curve)	14
37	Figure 5 – Brillouin frequency shift as a function of elongation of a standard telecommunication fibre	14

INTERNATIONAL ELECTROTECHNICAL COMMISSION

FIBRE OPTIC SENSORS –

Part 1-2: Strain measurement – Distributed sensing based on Brillouin scattering

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61757-1-2 has been prepared by subcommittee SC 86C: Fibre optic systems and active devices, of IEC technical committee TC 86: Fibre optics. It is an International Standard.

The text of this International Standard is based on the following documents:

Draft	Report on voting
XX/XX/FDIS	XX/XX/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement,

- 96 available at www.iec.ch/members_experts/refdocs. The main document types developed by
97 IEC are described in greater detail at www.iec.ch/standardsdev/publications.
- 98 A list of all parts in the IEC 61757 series, published under the general title *Fibre optic*
99 *sensors*, can be found on the IEC website.
- 100 The committee has decided that the contents of this document will remain unchanged until the
101 stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to
102 the specific document. At this date, the document will be
- 103 • reconfirmed,
 - 104 • withdrawn,
 - 105 • replaced by a revised edition, or
 - 106 • amended.
- 107

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN IEC 61757-1-2:2022](https://standards.iteh.ai/catalog/standards/sist/0a75a4b5-f319-4247-b41d-504848a98546/osist-pren-iec-61757-1-2-2022)

<https://standards.iteh.ai/catalog/standards/sist/0a75a4b5-f319-4247-b41d-504848a98546/osist-pren-iec-61757-1-2-2022>

108

INTRODUCTION

109 This International Standard is part of the IEC 61757 series, which is dedicated to fibre optic
110 sensors. Generic specifications for fibre optic sensors are defined in IEC 61757.

111 The individual parts of the IEC 61757 series are numbered as IEC 61757-M-T, where M
112 denotes the measure and T the technology of the fibre optic sensor. The IEC 61757-1-T
113 series is concerned with strain measurements.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN IEC 61757-1-2:2022](https://standards.iteh.ai/catalog/standards/sist/0a75a4b5-f319-4247-b41d-504848a98546/osist-pren-iec-61757-1-2-2022)

<https://standards.iteh.ai/catalog/standards/sist/0a75a4b5-f319-4247-b41d-504848a98546/osist-pren-iec-61757-1-2-2022>

FIBRE OPTIC SENSORS –

Part 1-2: Strain measurement – Distributed sensing based on Brillouin scattering

1 Scope

This part of IEC 61757 defines detailed specifications for distributed strain measurements with a fibre optic sensor, also known as fibre optic distributed strain sensing. It is applicable to distributed strain sensing systems (DSS) based on spontaneous or stimulated Brillouin scattering in the optical fibre sensor (strain sensitive element), that is, to sensors capable of measuring absolute strain. This International Standard specifies the most important DSS performance parameters and defines the procedures for their determination.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 50244:2016, *Electrical apparatus for the detection of combustible gases in domestic premises. Guide on the selection, installation, use and maintenance*

IEC 60050 (all parts), *International Electrotechnical Vocabulary*

IEC 61757:2018, *Fibre optic sensors – Generic specification*

IEC 61757-2-2:2016, *Fibre optic sensors – Part 2-2: Temperature measurement – Distributed sensing*

IEC 61757-3-2:2022, *Fibre optic sensors – Part 3-2: Acoustic sensing and vibration measurement – Distributed sensing*

ISO/IEC GUIDE 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 99, *International vocabulary of metrology – Basic and general concepts and associated terms (VIM)*

3 Terms, definitions, abbreviated terms and quantity symbols

3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 61757, IEC 61757-2-2, IEC 61757-3-2, IEC 60050 (all parts), and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

Note: For the following definitions, the relevant test procedures and parameters are defined in Clause 4.

3.1.1

attenuation range

total accumulated optical loss (one way) tolerated by the DSS without affecting the specified measurement performance more than a given factor for a given location, spatial resolution, and measurement time

Note 1 to entry: The total accumulated loss can include fibre attenuation as well as point defect losses introduced by components such as connectors, splices, fibre kinks, and attenuators.

Note 2 to entry: The attenuation range is usually expressed in dB.

Note 3 to entry: For fibre loop configurations, the attenuation range is specified by the loss between the output and input connector of the interrogation unit.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.2

distributed fibre optic strain sensing system

DSS

measurement set-up consisting of a distributed fibre optic sensor connected to an interrogation unit, including processor, data archive, and user interface, which provides a spatially resolved strain measurement

[SOURCE: IEC 61757-3-2:2022, modified – adapted to distributed strain measurement]

3.1.3

distance measurement range

maximum distance from the DSS interrogation unit output connector along the fibre optic sensor within which the DSS measures strain with specified measurement performance under defined conditions

Note 1 to entry: Defined conditions are spatial resolution (3.1.9), spatial strain uncertainty (3.1.10) and measurement time (3.1.6).

Note 2 to entry: This supporting parameter is closely related to the attenuation range of the interrogation unit. In test cases used to prove or verify the reported specifications, the total fibre length shall be equal to or greater than the specified distance measurement range, for the specified attenuation range.

Note 3 to entry: The distance measurement range is usually expressed in km.

Note 4 to entry: For fibre loop configurations, the distance measurement range is given by half the fibre length between the output and input connector of the interrogation unit.

[SOURCE: IEC 61757-2-2:2016 and ISO/IEC Guide 99, 4.7, modified – adapted to distributed strain measurement]

3.1.4

strained spot

ΔL

length of fibre optic sensor that experiences a small elongation (δL), which causes strain that is significantly bigger than the strain repeatability of the interrogation unit and which is confirmed by a reference strain measurement

Note 1 to entry: The applied strain ε is equal to $(\delta L / \Delta L)$.

Note 2 to entry: It is useful to define strain in $\mu\varepsilon$, where 1 $\mu\varepsilon$ corresponds to a δL of 1 μm over a ΔL of 1 m.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.5

location

L

optical distance from the DSS interrogation unit output connector to a desired strain sample point along the fibre optic sensor

Note 1 to entry: The farthest location from the DSS interrogation unit output connector for the particular test is quantified as $L_{F, \text{long}}$ km and is often chosen to be the same as the distance measurement range for purposes of comparing the measurement results with quoted specifications.

Note 2 to entry: The location is usually expressed in km.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.6

measurement time

time between independent strain measurements when making successive measurements on a single fibre optic sensor

Note 1 to entry: Equivalently, it is the time interval between successive strain trace timestamps under these conditions.

Note 2 to entry: This parameter includes acquisition time and processing time for the measured data. This parameter is typically selectable by the user in some limited fashion. Multiple independent strain measurements may be averaged together to provide an overall measurement time.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.7

point defect

local deviation of a fibre optic sensor from its nominal optical and mechanical properties occurring at a single location, or over a length substantially less than the DSS spatial resolution

Note 1 to entry: The definition of a point defect encompasses a wide range of situations, which can produce similar effects on the strain trace. Examples include:

- a point loss, like a bad fibre splice,
- a back reflection (or return loss), as can be introduced by a fibre connector,
- a localized region of high loss, such as a bend or kink in the fibre,
- a physical discontinuity in the fibre, like a splice between two fibres of different core diameters.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.8

sample spacing

distance between two consecutive strain sample points in a single strain trace

Note 1 to entry: Sample spacing can be a user-selectable parameter in the interrogation unit.

Note 2 to entry: The sample spacing is usually expressed in m.

Note 3 to entry: See Figure 1.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]




Figure 1 – Optical fibre strain profile and related strain sample points

3.1.9

spatial resolution

smallest length of strain-affected fibre optic sensor for which a DSS can measure and confirm the reference strain of a defined strained spot within the specified strain measurement error of the DSS

Note 1 to entry: The spatial resolution is usually expressed in m.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.10

spatial strain uncertainty

uncertainty of the location of strain data in a single strain trace, expressed by twice the standard deviation of a specified number of adjacent strain sample points, with the fibre optic sensor held at constant strain and temperature

Note 1 to entry: Due to a potential cross-sensitivity of DSS to temperature, it can be necessary to stabilize the temperature of the fibre optic sensor.

Note 2 to entry: The spatial strain uncertainty is usually expressed in units of $\mu\epsilon$ and noted as a tolerance (e.g., $\pm xx \mu\epsilon$), where $1 \mu\epsilon$ corresponds to a δL of $1 \mu\text{m}$ over a ΔL of 1m .

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.11

strain dead zone

limited zone of a strain trace, where the strain sample points deviate from the undisturbed parts of the trace by a specified limit due to a point defect

Note 1 to entry: The strain dead zone is usually expressed in m.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.12

strain measurement error

maximum difference between a centred and uniformly weighted moving average of the measured strain and a reference strain for all data points of the fibre optic sensor over the full operating temperature range and all acquisition times

Note 1 to entry: Single value (worst case) is expressed like a tolerance in units of $\mu\epsilon$ (e.g., $\pm xx \mu\epsilon$).

Note 2 to entry: The number of elements used for the moving average is defined later in the document. In practical applications other methods of smoothing might be applicable.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.13

strain repeatability

precision of strain data based on repeated strain traces at a given location expressed by twice the standard deviation of corresponding strain sample points in each strain trace, with the fibre optic sensor held at constant strain and temperature

Note 1 to entry: The strain repeatability is expressed like a tolerance in units of $\mu\epsilon$ (e.g., $\pm xx \mu\epsilon$).

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.14

strain sample point

measured strain value associated with a single point at a known location along a fibre optic sensor

Note 1 to entry: Due to signal averaging effects, the measured value represents the strain along a very small section of the fibre optic sensor that includes the strain sample point.

Note 2 to entry: See Figure 1.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.15

strain trace

set of strain sample points distributed along a fibre optic sensor and spaced by the sample spacing

Note 1 to entry: All sample points are associated with a common time of measurement, often called the trace timestamp. The measured values represent the strain during a time period that includes the timestamp.

Note 2 to entry: All sample points in a strain trace are measured values produced by the DSS, and not interpolated or smoothed values produced by subsequent processing outside the interrogation unit.

[SOURCE: IEC 61757-2-2:2016, modified – adapted to distributed strain measurement]

3.1.16

total fibre length

$L_{F,tot}$

distance from the DSS interrogation unit output connector to the final end of the fibre optic sensor