

SLOVENSKI STANDARD SIST-TP CEN/TR 17911:2023

01-april-2023

Emisije nepremičnih virov - Smernica za pripravo standardiziranih merilnih metod - Priporočila za strukturo in vsebino

Stationary source emissions - Guideline for the elaboration of standardized measurement methods - Recommendations for the structure and content

Emissionen aus stationären Quellen - Leitfaden zur Erarbeitung von standardisierten Messverfahren - Empfehlungen für die Struktur und den Inhalt

Émissions de sources fixes - Cadre directeur pour l'élaboration de méthodes de mesurage normalisées - Recommandations pour la structure et le contenu

CEN/TR 17911:2023 Ta slovenski standard je istoveten z:

ICS:

13.040.40 Emisije nepremičnih virov Stationary source emissions

SIST-TP CEN/TR 17911:2023 en SIST-TP CEN/TR 17911:2023

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST-TP CEN/TR 17911:2023

https://standards.iteh.ai/catalog/standards/sist/322d92a8-3fa4-4d72-b5bf-20c6c0426837/sist-tp-cen-tr-17911-2023

TECHNICAL REPORT RAPPORT TECHNIQUE **CEN/TR 17911**

TECHNISCHER REPORT

February 2023

ICS 13.040.40

English Version

Stationary source emissions - Guideline for the elaboration of standardized measurement methods - Recommendations for the structure and content

Émissions de sources fixes - Cadre directeur pour l'élaboration de méthodes de mesurage normalisées - Recommandations pour la structure et le contenu

Emissionen aus stationären Quellen - Leitfaden zur Erarbeitung von standardisierten Messverfahren -Empfehlungen für die Struktur und den Inhalt

This Technical Report was approved by CEN on 6 February 2023. It has been drawn up by the Technical Committee CEN/TC 264.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

standards.iteh.ai

https://standards.iteh.ai/catalog/standards/sist/322d92a8-3fa4-4d72-b5bf 20c6c0426837/sist-tp-cen-tr-17911-2023



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Cont	ents	Page
Europ	ean foreword	4
Introd	uction	5
1	Scope	<i>6</i>
2	Normative references	<i>6</i>
3	Terms and definitions	<i>6</i>
4	Symbols and abbreviations	7
4.1	Symbols	
4.2	Abbreviations	7
5	Overview	7
6	Guidance on the elaboration of documents	11
6.1	Type of document	11
6.2	Title	12
6.3	Foreword	12
6.4	Introduction	12
6.5	Scope	12
6.6	Normative references	
6.7	Terms and definitionsSymbols and abbreviations	13
6.8	Symbols and abbreviations	13
6.9	Principle	13
6.9.1	General SIST-TP CFN/TR 17911:2023	13
6.9.2	Measuring principle and itah ai/antalog/standards/sist/20040008-2564-4470-h5hf	13
6.10	Description of the measuring system	14
6.11	Performance characteristics	
	Performance characteristics and performance criteria	
	Calculation of measurement uncertainty	15
	Planning	
	Measurement planning	
	Sampling strategy	
	Sampling and analysis	
	Choice of the measuring system	
	Field operation	
	Preparations	
	Sampling	
	Data recording	
	QA/QC procedures	
	Handling, storage and transportation of samples for the analytical determination	
6.14	Analytical procedure	
6.15	Ongoing quality control	
	General	
	QA/QC procedures and frequency of checks	
6.16	Calculations and expression of results	
	Calculation of the result of measurement	
	Conversion to oxygen reference condition	
6.16.3	Assessment and expression of results of measurement	19

6.17	Equivalence of an alternative method	19
6.18	Measurement report	19
	Annex "Validation of the method"	
6.20	Annex "Measurement uncertainty"	19
6.21	Annex "Calculation of the uncertainty associated with a concentration expressed	on
	dry gas and at an oxygen reference concentration"	19
Annex	A (informative) Electronic template	20
Biblio	graphy	21

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST-TP CEN/TR 17911:2023</u> https://standards.iteh.ai/catalog/standards/sist/322d92a8-3fa4-4d72-b5bf 20c6c0426837/sist-tp-cen-tr-17911-2023

European foreword

This document (CEN/TR 17911:2023) has been prepared by Technical Committee CEN/TC 264 "Air Quality", the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST-TP CEN/TR 17911:2023</u> https://standards.iteh.ai/catalog/standards/sist/322d92a8-3fa4-4d72-b5bf 20c6c0426837/sist-tp-cen-tr-17911-2023

Introduction

This document has been prepared by WG 46 "Task Force Emissions" of the Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This document is intended to assist working groups of CEN/TC 264 when drafting new or revising existing documents on measurement methods for the determination of stationary source emissions. A document describing a standardized measurement method can be a

- European standard (EN);
- European Technical Specification (CEN/TS);
- European Technical Report (CEN/TR).

This document is designed to ensure that these documents are

- consistent regarding structure, definitions and general aspects;
- in accordance and properly linked with other relevant documents such as EN 15259;
- appropriate technical references giving no rise to significant misunderstandings and/or significant differences of interpretation by testing laboratories and/or technical auditors.

It is considered that such significant differences, misunderstandings and/or inconsistencies would, on the one hand, impair the quality and comparability of data produced according to European emission measurement methods and, on the other hand, result in unfair competition among European laboratories in the field of emission measurements.

This document is applicable to manual and automated measurement methods.

This document is supplemented by an electronic template providing a uniform structure and common elements and texts. The uniform structure is based on the requirements and recommendations specified in the CEN-CENELEC Internal Regulations Part 3.

This document can be applicable to other air quality fields.

1 Scope

This document supports the elaboration of standardized measurement methods for the determination of stationary source emissions by manual or automated measurement methods.

This document describes the basic elements of standardized measurement methods for the determination of stationary source emissions.

This document is supplemented by an electronic template providing a uniform structure and common elements and texts.

NOTE Detailed information on the electronic template is given in Annex A.

This document is addressed to working groups of CEN/TC 264 dealing with stationary source emissions. It aims at facilitating in the working groups the elaboration and the harmonization of documents produced by CEN/TC 264. Such documents can be European standards (EN), European Technical Specifications (CEN/TS) or European Technical Reports (CEN/TR).

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at https://www.electropedia.org/

3.1 standard reference method

standard reference method SRM

reference method prescribed by European or national legislation

[SOURCE: EN 15259:2007, 3.9, modified – Note 1 to entry removed]

3.2

reference method

RM

measurement method taken as a reference by convention, which gives the accepted reference value of the measurand

Note 1 to entry: A reference method is fully described.

Note 2 to entry: A reference method can be a manual or an automated method.

Note 3 to entry: Alternative methods can be used if equivalence to the reference method has been demonstrated.

[SOURCE: EN 15259:2007, 3.8]

3.3

measurement method

method described in a written procedure containing all the means and procedures required to sample and analyse, namely field of application, principle and/or reactions, definitions, equipment, procedures, presentation of results, other requirements and measurement report

[SOURCE: EN 14793:2017, 3.4]

3.4

alternative method

AM

measurement method which complies with the criteria given by EN 14793:2017 with respect to the reference method

Note 1 to entry: An alternative method can consist of a simplification of the reference method.

[SOURCE: EN 14793:2017, 3.3]

3.5

demonstration of equivalence

act of subjecting a measurement method to a study, which is based on a standardised and/or recognised protocol and which provides proof that, for its field of application, the measurement method satisfies preestablished performance criteria

Note 1 to entry: In the framework of EN 14793:2017, the demonstration of equivalence of a measurement method is mainly based on an "in field" study that includes comparison to a reference method.

[SOURCE: EN 14793:2017, 3.7]

4 Symbols and abbreviations log/standards/sist/322d92a8-3fa4-4d72-b5bf-

4.1 Symbols

This document does not include symbols.

4.2 Abbreviations

AM alternative method

AMS automated measuring system

P-AMS portable automated measuring system QA/QC quality assurance and quality control

RM reference method

SRM standard reference method

5 Overview

This document applies to the elaboration of new or the revision of existing documents on emission measurement methods. It provides a uniform structure and common elements for such documents. The uniform structure is based on the requirements and recommendations specified in the CEN-CENELEC Internal Regulations Part 3. The use of a uniform structure and common elements is intended to ensure uniform application of emission measurement methods. Elements can be added, modified or removed as needed for the specific emission measurement method.

Table 1 shows the sequence of elements of standardized measurement methods for the determination of stationary source emissions and their respective content. It indicates the elements applicable to manual or automated measurement methods as well as the reference to the description in this document and the associated clause number in the electronic template.

Table 1 — Elements of standardized measurement methods for the determination of stationary source emissions and their respective content, relevance for manual (M) or automated (A) measurement methods, reference to the description in this document and associated clause number in the electronic template

Ref.	Clause number	Element	Content	M	A
6.2		Title	clear and concise description of the subject matter covered by the document	×	×
6.3		Foreword	revision statement including any amendments and technical corrigenda and a list of changes with respect to previous edition, in case of a revision of the document	×	×
6.4		Introduction iTeh STANDA	background information and general information on the type and content of the document	×	×
6.5	1	Scope (standa)	measured component	×	×
			measurand (measured quantity)	×	×
		SIST-TP CE	principle of method	×	×
		https://standards.iteh.ai/catalog/sta 20c6c0426837/sis	type of method (manual or automated)	×	×
			field of application	×	×
			measurement range	×	×
			validation range	×	×
			exclusions	×	×
6.6	2	Normative references	list of all normatively referenced documents	×	×
6.7	3	Terms and definitions	list of all relevant terms used in the document	×	×
6.8	4	Symbols and abbreviations	list of all symbols and abbreviations used in the document	×	×
6.9	5	Principle	general information on the measurement method and the content of the document	×	×
			description of the measuring principle of the measurement method	×	×

Ref.	Clause number	Element	Content	M	A
6.10	6	Description of the measuring system	equipment	×	×
			materials	×	×
			reagents	×	×
6.11	7	Performance characteristics	performance characteristics and performance criteria	×	×
			performance test of the P-AMS for automated measurement methods according to EN 15267-4		×
			calculation of measurement uncertainty	×	×
6.12	8	Planning	measurement planning	×	×
			sampling strategy	×	×
			sampling and analysis	×	×
			choice of the measuring system	×	×
6.13	9 1	Field operation	preparations / R	×	×
		(standards	sampling	×	×
		(standards.	continuous data recording		×
		SIST-TP CEN/TR 1	QA/QC procedures	×	×
	https://	standards.iteh.ai/catalog/standards 20c6c0426837/sist-tp-cer	handling, storage and transportation of samples for the analytical determination	×	
6.14	10	Analytical procedure	preparations	×	
			analysis	×	
			QA/QC procedures	×	
			presentation of the analytical result including uncertainty and level of confidence	×	
6.15	11	Ongoing quality control	QA/QC procedures	×	×
			frequency of checks	×	×
6.16	12	Calculations and expression of results	calculations and expression of the results of measurement including measurement uncertainty and level of confidence	×	×
			conversion to oxygen reference conditions	×	×
			assessment and reporting of results of measurement	×	×