

FINAL
DRAFT

INTERNATIONAL
STANDARD

ISO/FDIS
23256

ISO/TC 147/SC 2

Secretariat: DIN

Voting begins on:
2023-01-19

Voting terminates on:
2023-03-16

Water quality — Detection of selected congeners of polychlorinated dibenzo-p-dioxins and polychlorinated biphenyls — Method using a flow immunosensor technique

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 23256

<https://standards.iteh.ai/catalog/standards/sist/359d07c6-36b3-4711-9d42-97f59de8a397/iso-fdis-23256>

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 23256:2023(E)

© ISO 2023

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 23256

<https://standards.iteh.ai/catalog/standards/sist/359d07c6-36b3-4711-9d42-97f59de8a397/iso-fdis-23256>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	v
Introduction.....	vi
1 Scope.....	1
2 Normative references.....	1
3 Terms, definitions, and abbreviated terms.....	1
3.1 Terms and definitions.....	1
3.2 Abbreviated terms.....	3
4 Principle.....	4
4.1 Flow immunosensor.....	4
4.2 Specific antibody.....	4
5 Interferences.....	4
6 Reagents.....	5
7 Apparatus and materials.....	7
8 Sampling.....	7
8.1 General.....	7
8.1.1 General.....	7
8.1.2 Bottles and material for sampling.....	7
8.1.3 Bottles and material pre-cleaning.....	7
8.1.4 Sampling.....	8
8.2 Preparation of sample.....	8
8.2.1 Extraction.....	8
8.2.2 Clean-up.....	8
8.2.3 Sample solutions.....	8
9 Procedure.....	9
9.1 General.....	9
9.2 Sample set up.....	9
9.3 Measuring procedure.....	9
10 Data processing.....	10
10.1 Measurement data.....	10
10.2 Calibration curve.....	11
10.3 Calculation of measured concentration.....	12
11 Validation.....	12
11.1 General.....	12
11.2 Accuracy profile.....	12
11.3 Range of quantitation.....	13
11.4 Limit of detection.....	14
11.5 Evaluation of data.....	14
11.6 Quality control.....	14
12 Test report.....	15
Annex A (informative) Specificity of mouse monoclonal antibodies used for the flow immunosensor.....	16
Annex B (informative) Example of the automate sample clean-up and concentration device.....	17
Annex C (informative) Example of the flow immunosensor.....	19
Annex D (informative) Recoveries, calibration curves, ranges of quantitation of the standard analytes 2,3,7,8-TCDD and 3,3',4,4',5-PeCB in drinking water.....	25
Annex E (informative) Measurement of 2,3,7,8-TCDD and 3,3',4,4',5-PeCB in river waters.....	29

Annex F (informative) Interlaboratory trial of the measurement of 2,3,7,8-TCDD and 3,3',4,4',5-PeCB in water samples by the use of flow immunosensor method	32
Bibliography	35

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 23256

<https://standards.iteh.ai/catalog/standards/sist/359d07c6-36b3-4711-9d42-97f59de8a397/iso-fdis-23256>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Persistent organic pollutants (POPs) including dibenzo-*p*-dioxins (PCDDs), dibenzofurans (PCDFs) and dioxin-like polychlorinated biphenyls (DL-PCBs) in water and wastewater are analysed by instrumental methods such as GC-MS and HRGC/HRMS. These methods are accurate and precise, but labour-intensive, time-consuming, and costly. Alternatively, immunoassays are used for monitoring of POPs. These methods have similar sensitivity and selectivity, but are more cost-effective and able to manage large loads of sample well. The use of immunoassays is suitable for timely detection of selected congeners of PCDDs and PCBs in water and wastewater in advance of subsequent confirmatory methods.

Recently, automated immunoassay methods including a flow immunosensor have been developed for detection of POPs in environmental samples. These methods reduced manual operations such as pipetting, time-consumption and coefficient of variation. Therefore, practical use of these methods can play an important role in timely, continuous cost-saving monitoring of water and wastewater in both developed and developing countries. This method can be applicable to timely continuous monitoring of selected congeners of PCDDs and PCBs in water and wastewater to prioritize those for subsequent confirmatory determination.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 23256

<https://standards.iteh.ai/catalog/standards/sist/359d07c6-36b3-4711-9d42-97f59de8a397/iso-fdis-23256>

Water quality — Detection of selected congeners of polychlorinated dibenzo-*p*-dioxins and polychlorinated biphenyls — Method using a flow immunosensor technique

WARNING — The PCDDs, PCDFs and PCBs are among the most hazardous chemicals. Therefore, all work with PCDDs, PCDFs and PCBs require the utmost care; the national safety measures which correspond to those for hazardous substances shall be strictly adhered to. The responsibility of the user is to establish appropriate safety and health practices.

IMPORTANT — It is absolutely essential that tests conducted in accordance with this document be carried out by suitably qualified staff.

1 Scope

This document specifies methods and principles for detection of selected congeners of polychlorinated dibenzo-*p*-dioxins (PCDDs) and polychlorinated biphenyls (PCBs) in water and wastewater using a flow immunosensor. The flow immunosensor utilizes antibodies specific to 2,3,7,8-TCDD and 3,3',4,4',5-PeCB, which have the highest toxic equivalent factor (TEF) value among the congeners of each of PCDDs and PCBs. The method is applicable to timely monitoring of selected congeners of 4,3,7,8-TCDD and 3,3',4,4',5-PeCB in water and wastewater to prioritize those for subsequent confirmatory determination.

This document specifies practical methods and procedures for sampling, extraction, clean-up, measurement in a flow immunosensor, data processing and validation of measurement results. The combined use of automated instruments for extraction, clean-up, and flow immunosensing can reduce time-consumption and labour-intensity, while providing reproducible precise data. This method can provide the lower LOQ for 2, 3, 7, 8-TCDD and 3,3',4,4',5-PeCB of 28 pg/l and 152 pg/l, respectively at 20 % or less of coefficient variation (CV) depending on sampling, extraction, clean-up and measurement conditions.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO 5667-6, *Water quality — Sampling — Part 6: Guidance on sampling of rivers and streams*

ISO 5667-10, *Water quality — Sampling — Part 10: Guidance on sampling of waste water*

3 Terms, definitions, and abbreviated terms

3.1 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1.1

analyte

selected congeners of polychlorinated dibenzo-*p*-dioxins and polychlorinated biphenyls which bind to specific monoclonal *antibodies* (3.1.2)

Note 1 to entry: [Annex A](#) gives information of the specific of mouse monoclonal antibodies used for the flow immunosensor.

3.1.2

antibody

class of serum proteins that are induced by exposing to an immunogen and will bind specifically to *antigen/analyte* (3.1.3)/(3.1.1) forming an antibody-antigen complex

Note 1 to entry: An antibody with a fluorescent label is used as a secondary antibody for binding to an antibody specific to antigen/analyte.

3.1.3

antigen

molecule which selectively binds to an *antibody* (3.1.2)

Note 1 to entry: Antigens are analytes in the case of selected congeners of polychlorinated dibenzo-*p*-dioxins and polychlorinated biphenyls.

3.1.4

conjugate

large molecule covalently coupled with a small molecule

Note 1 to entry: An antigen-mimic BSA conjugate is used for coating polymer beads.

3.1.5

cross-reactivity

ability of an *antibody* (3.1.2) to bind to certain molecules which structurally related to an *antigen* (3.1.3)

3.1.6

flocculation

physical process of contact and adhesions wherein the aggregates form larger-size clusters called flocs being excluded from suspension for water treatment

3.1.7

flow immunosensor

immunosensor based on a *kinetic exclusion assay* (3.1.11) using a monoclonal *antibody* (3.1.2) in a heterogeneous assay system for rapid detection of an *analyte* (3.1.1) in a number of samples

Note 1 to entry: The system of a flow immunosensor is operated automatically according to a program.

3.1.8

IgG

immunoglobulin-G

antibody molecule, consisting of two identical heavy(H)-chains and two identical light(L)-chains held together with disulfide bonds

Note 1 to entry: Both H- and L-chains consist of the variable and constant regions. An antigen binding site is presented in the variable region.

3.1.9**immunoassay**

immunochemical detection procedure based on specific *antibody-antigen* (3.1.2)-(3.1.3) binding theory often using a tracer for the detection of a free or bound antibody

Note 1 to entry: In a heterogeneous immunoassay, one of immunoreagents is immobilized on a solid support and requires a washing step to separate the bound and free immunoreagents.

3.1.10**inhibition concentration**

analyte concentration which reduces the measuring signal of the *zero standard* (3.1.14), in the case of 50 % inhibition of the zero standard that expresses 50 values

3.1.11**kinetic exclusion assay**

binding between *antibody* (3.1.2) and *antigen* (3.1.3) reaches equilibrium in solution, and then three species such as the specific antibody, the antigen and the antibody-antigen complex are present

Note 1 to entry: A kinetic exclusion assay measures the concentration of the free antibody without perturbing the equilibrium. The format of the assay is to measure a free specific antibody with a label.

3.1.12**monoclonal antibody**

antibody (3.1.2) population, possessing identical selectivity and affinity produced by a single antibody-producing cell line

Note 1 to entry: Monoclonal antibodies specific to selected congeners of polychlorinated dibenzo-*p*-dioxins and polychlorinated biphenyls are prepared for measurement of these analytes.

3.1.13**procedure blank**

solution prepared in the laboratory, using reagent water or other blank matrix

3.1.14**zero standard**

analyte-free standard (blank) which is used for calibration

3.2 Abbreviated terms

BSA	bovine serum albumin
CV	coefficient of variation
DMSO	dimethyl sulfoxide
GC-MS	gas chromatography/mass spectrometry
HRGC	high-resolution gas chromatography
HRMS	high-resolution mass spectrometry
IgG	immunoglobulin G
LED	light emitting diode
LLE	liquid-liquid extraction
LOD	limit of detection
LOQ	limit of quantification

PBS	phosphate-buffered saline
PCDD	polychlorinated dibenzo- <i>p</i> -dioxin
PCDF	polychlorinated dibenzofuran
PCB	polychlorinated biphenyl
POP	persistent organic pollutant
PTFE	polytetrafluoroethylene
SD	standard deviation
SPE	solid phase extraction
TEF	toxic equivalent factor

4 Principle

4.1 Flow immunosensor

A flow immunosensor based on kinetic exclusion assay principles consists of a flow cell containing a capture reagent immobilized on a solid phase matrix such as azlactone, sepharose, polymethyl methacrylate, polystyrene, or the like, along with an antibody specific for the analyte and a fluorescent label used for detection. In many cases, the fluorescent label is attached to a second-anti-species antibody that recognizes and binds to the analyte specific antibody. In practice, the capture reagent used on the solid phase is frequently either the analyte itself, an analogue of the analyte or a protein conjugate of either the analyte or an analogue. To perform the assay a suitable sample (which can be extracted and/or concentrated from a large environmental sample) is mixed with the specific antibody and the secondary antibody (C.2.6) with a fluorescent label, and then incubated for a period of time to allow binding to occur. After incubation, the antibody mixture is flowed over the solid phase matrix and free (not bound to analyte) antibody bound to the secondary antibody with a fluorescent label from the solution binds to the solid phase capture reagent. A fluorescent signal is measured from the captured antibody. The largest fluorescent signal occur when there is zero analyte present and the presence of analyte results in reduced signal levels^{[1][3]}.

4.2 Specific antibody

Specific antibodies (IgG) are selected for measurement of analytes in a flow immunosensor. Monoclonal anti-2,3,7,8-TCDD antibody and anti-3,3',4,4',5-PeCB antibody show the highest affinity to 2,3,7,8-TCDD and 3,3',4,4',5-PeCB, respectively as shown in Annex A. Each of these congeners is known to be present in the environment and has the highest TEF value among the congeners of each of PCDDs and PCBs. Thus, both antibodies are used as the representatives for detection of selected congeners of PCDDs and PCBs in water and wastewater using a flow immunosensor. When antibodies specific for other congeners or other related compounds (e.g. PCDFs) are available, the application of the immunosensor described in 4.1 are readily expanded for detection of those related compounds or congeners.

5 Interferences

PCDDs, PCBs and other dioxin-like compounds are hydrophobic and largely present in water and wastewater as adsorbed on solid particles. When these compounds are spiked into water samples in containers, these compounds are quickly adsorbed on the surface of containers. The use of inappropriate sampling devices and/or sampling flask can influence the test result because of the possible adsorption of these compounds leading to false-negative results. On the other hand, these compounds can be released into the sample from sampling flasks, especially when wares used are contaminated with these compounds, and false-positive results can be generated. If filtered samples are tested in order

to remove the sample solid particles, the dioxins and dioxin-like compounds which are adsorbed on particles cannot be detected.

Measurement conditions, for instance pH or sample components such as humic acids, salinity and solvents influencing the detection which so-called matrix effects can interfere with antibody binding. The interference of matrix effects shall be assessed by spiking samples or reference samples with known amounts of the analyte. In any case, extraction and clean-up are necessary to avoid interference. In addition, air bubbles in measuring solutions and in a flow cell interfere not only antibody binding but also fluorescence detection. Therefore, it shall be taken care to prepare measuring solutions particularly by stirring gently at a room temperature.

Monoclonal antibodies specific to antigen/analyte are used for a flow immunosensor as shown in [Annex A](#). These antibodies cross-react against structurally related compounds and also bind to non-specific compounds to show matrix effects. The extracted samples are each submitted to clean-up process in an automate sample clean-up and concentration device as shown in [Annex B](#), in which a multiphase silica gel column works to eliminate polynuclear aromatic hydrocarbons and sulfur compounds. In addition, the clean-up samples are also submitted to HRGC/HRMS analysis to confirm removal of matrices

6 Reagents

Use reagents with a high purity for instance the grade “for residue analysis”. Information on type and origin of antibodies as well as the cross-reactivity are stated in [Annex A](#). Information on storage and stability of the used reagents shall be requested according to the information of the supplier. If sensitivity in a flow immunosensor is lower than expected, analyte-interfering substances such as matrices and air bubbles shall be removed from samples and measuring solutions by suitable procedures.

6.1 Water, complying with grade 3 as defined in ISO 3696.

For instance, a Milli-Q water is recommended for preparation of buffer and regeneration solution.

6.2 Sodium chloride, NaCl, > 995 g/kg purity.

6.3 Potassium chloride, KCl, > 990 g/kg purity.

6.4 Sodium hydroxide, NaOH, > 970 g/kg purity.

6.5 Monosodium phosphate, NaH₂PO₄, > 980 g/kg purity.

6.6 Sodium dihydrogen phosphate dodecahydrate, Na₂HPO₄·12H₂O, >980 g/kg purity.

6.7 Organic solvents

Solvents used for spiking, extraction, clean-up, and measurement in the following:

6.7.1 Propanone (acetone), >995 ml/l purity.

6.7.2 Toluene, >990 ml/l purity.

6.7.3 Hexane, >960 ml/l purity.

6.7.4 Dimethyl sulfoxide, DMSO, >995 ml/l purity.

6.7.5 Ethanol, >995 ml/l purity.

6.8 Standard solution.

The standard acetone solutions of 2,3,7,8-TCDD and 3,3',4,4',5-PeCB are each diluted with acetone for spiking into a 3 l or more water sample and for confirmation of spiking concentration by HRGC/HRMS analysis. For preparation of sample solutions, a part of each acetone solution of both standard chemicals is diluted with DMSO.

6.9 Flocculant.

A flocculant consisting of activated charcoal powder, poly aluminium chloride, silica gel and sodium carbonate is added into a water sample adjusted to pH 7 to pH 8 at a rate of 0,1 g/l. Then, flocs formed are collected by filtration with a 0,45 µm of pore size of a glass fibre filter and then air-dried overnight for extraction^[4].

6.10 Phosphate-buffered saline.

Prepare 50 mmol/l of PBS at pH 7,4 by dissolving 2,9 g of Na₂HPO₄·12H₂O, 0,2 g of KH₂PO₄, 8,0 g of NaCl and 0,2 g of KCl in 1 000 ml of water (6.1).

6.11 Sample buffer.

Used for sample preparation, with PBS (6.15) containing 1 g/l of BSA and 0,2 g/l of NaN₃, passed through a 0,45 µm of pore size of a polyvinylidene difluoride filter. The buffer prepared in a glass bottle is kept in a refrigerator.

6.12 Measuring buffer.

PBS containing 1 g/l of BSA, 0,2 g/l of NaN₃ and 55 g/l of DMSO, passed through a 0,45 µm polyvinylidene difluoride filter. The buffer in a glass bottle is kept in a refrigerator.

6.13 Correcting solution.

Prepare 0,6 ng/ml of 2,3,7,8-TCDD in DMSO and 1,6 ng/ml of 3,3',4,4',5-PeCB in DMSO by the dilution of each of the standard chemicals in DMSO. The solution prepared in a glass bottle is kept in a refrigerator.

An antigen-mimic, 3-[6-(2,4,5-trichlorophenoxy) hexanoylamino] propionic acid, is also used as a reference at a concentration of 9,5 ng/l in DMSO.

6.14 Regeneration solution.

Prepare 1 g/l of NaOH (6.4) and 55 g/l of DMSO in water passed through a 0,45 µm polyvinylidene difluoride filter. The solution prepared in a glass bottle is kept in refrigerator.

6.15 Cleaning solution, Prepare 20 ml/l ethanol (volume fraction) in a glass bottle.

6.16 Antibody solution.

Antibody solution is prepared to mix with each antibody which is bind specifically to 2,3,7,8-TCDDs or 3,3',4,4',5-PeCB and the second with a fluorescent label as a secondary antibody (C.2.6) in sample buffer (6.16). A fluorescent label should have a high intensity of emission and absorption at certain wavelength which can be detected by a photodetector of an immunosensor.

Anti-2,3,7,8-TCDD antibody solution (D48 kit¹⁾) contains 0,12 µg/ml IgG of anti-2,3,7,8-TCDD antibody and 0,06 µg/ml of secondary antibody in sample buffer (6.16). Anti-3,3',4,4',5-PeCB antibody solution

1) D48 kit is the trade name a product supplied by Seeds Tec Co., Ltd. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used if they can be shown to lead to the same results.

[P126 kit²⁾] contains 0,10 µg IgG/ml of anti-3,3',4,4',5-PeCB antibody and 0,05 µg/ml of secondary antibody (C.2.6) in sample buffer. D48 kit and P126 kit are each used for detection of 2,3,7,8-TCDD and 3,3',4,4',5-PeCB, respectively. Each of both solutions in a glass bottle is covered with an aluminium foil for protection against light and kept in a refrigerator.

7 Apparatus and materials

All water sample containers are rinsed with acetone, toluene and hexane in turn, and dried prior to use. Apparatus used for extraction, clean up and concentration are referred to in ISO 18073^[5] and ISO 17858^[6].

7.1 Flow immunosensor (4.1).

7.2 **Vial**, add a containing mixed antibody solution (6.16) and either samples or correcting solution (6.13) and then mix gently.

7.3 Flow cell.

Stuffing with polymer beads coated with the antigen-mimic BSA conjugate (C.2.7) in a column. There are two kinds of flow cell, a disassembly type and assembly types. A disassembly type flow cell is supplied pre-filled with polymer beads coated with the antigen-mimic BSA conjugate (C.2.7) and the entire flow cell is removed and replaced after several measurements. For an assembly types, the flow cell remains in service much longer and the beads are automatically filled for each measurement.

8 Sampling

8.1 General

8.1.1 General

This document describes specific requirements for the sampling and samples with respect to the determination of dioxins or dioxin-like compounds in water samples. These compounds are hydrophobic and are present in water and wastewater as adsorbed on solid particles as well as the surface of wares used. Thus, collection, extraction and clean-up procedures suitable for these compounds are requested. For general information about sampling and samples, ISO 5667-1, ISO 5667-3, ISO 5667-6 and ISO 5667-10 shall be considered.

8.1.2 Bottles and material for sampling

Use clean glass bottles preferably with PTFE-lined caps. To avoid photodegradation of compounds of interest, preferably use amber glass bottles. If transparent glass bottles are used, wrap the bottles in aluminium foil or store them in a dark container.

8.1.3 Bottles and material pre-cleaning

After the routine cleaning procedure, additionally clean bottles and bottles and caps as follows: rinse the clean bottles and the caps three times with a minimum amount of acetone. Let the residual acetone evaporate (e.g. drying oven). Close the bottles immediately after drying. Rinse all glassware, spatulas etc. getting in contact with the sample three times with a minimum amount of acetone. Let the residual acetone evaporate.

2) P126 kit is the trade name of a product supplied by Seeds Tec Co., Ltd. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used if they can be shown to lead to the same results.