

SLOVENSKI STANDARD oSIST prEN ISO 2613-2:2023

01-april-2023

Analiza zemeljskega plina - Biometan - 2. del: Določevanje siloksana s plinsko kromatografijo s spektrometrijo na osnovi mobilnosti ionov (GC-IMS) (ISO/DIS 2613-2:2023)

Analysis of natural gas - Biomethane - Part 2: Determination of siloxane content by gas chromatography ion mobility spectrometry (ISO/DIS 2613-2:2023)

Analyse von Erdgas - Siliziumanteil von Biomethan - Teil 2: Bestimmung des Siloxananteils durch gaschromatographische Ionenmobilitätsspektroskopie (ISO/DIS 2613-2:2023)

Analyse du gaz naturel - Biométhane - Partie 2: Détermination de la teneur en siloxanes par chromatographie en phase gazeuse spectrométrie de mobilité ionique (ISO/DIS 2613-2:2023)

Ta slovenski standard je istoveten z: prEN ISO 2613-2

ICS:

75.060 Zemeljski plin Natural gas

oSIST prEN ISO 2613-2:2023 en,fr,de

oSIST prEN ISO 2613-2:2023

iTeh STANDARD PREVIEW (standards.iteh.ai)

oSIST prEN ISO 2613-2:2023
https://standards.iteh.ai/catalog/standards/sist/71284514-d2e2-400a-bef8-fee5628c1193/osist-pren-iso-2613-2-2023

DRAFT INTERNATIONAL STANDARD ISO/DIS 2613-2

ISO/TC **193**/SC **1**

Secretariat: NEN

Voting begins on: **2023-02-13**

Voting terminates on:

2023-05-08

Analysis of natural gas — Biomethane —

Part 2:

Determination of siloxane content by gas chromatography ion mobility spectrometry

ICS: 75.060

iTeh STANDARD PREVIEW (standards.iteh.ai)

oSIST_prEN ISO 2613-2:2023 https://standards.iteh.ai/catalog/standards/sist/71284514-d2e2-400a-bef8fee5628c1193/osist-pren-iso-2613-2-2023

This document is circulated as received from the committee secretariat.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

ISO/CEN PARALLEL PROCESSING



Reference number ISO/DIS 2613-2:2023(E)

iTeh STANDARD PREVIEW (standards.iteh.ai)

oSIST prEN ISO 2613-2:2023 https://standards.iteh.ai/catalog/standards/sist/71284514-d2e2-400a-bef8fee5628c1193/osist-pren-iso-2613-2-2023



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Coı	ntents	Page
Fore	eword	iv
Intr	oduction	v
1	Scope	1
2	Normative references	1
3	Terms and definitions	
4	Principle	
5	Materials 5.1 Calibration gases 5.2 Diluent gas 5.3 Carrier & drift gas	3 3
6	Apparatus 6.1 GC-IMS analyser 6.2 Pressure regulator 6.3 Dilution device 6.4 Auxiliary valves, tubing and accessories	3 3 3 3
7	Sampling 7.1 General 7.2 Safety precautions 7.3 Temperature control 7.4 Construction materials 7.5 Cleanness 7.6 Sampling of biomethane into vessels 7.7 Installation of the calibration gas cylinder 7.8 Pressure control 7.9 Purging of wetted flow path 7.10 Flow control 7.11 Diffusion control	4 4 4 4 5 5 5 5 5
8	Calibration 8.1 Calibration procedure 8.2 Frequency of calibration	6
9	Measurement procedure 9.1 Safety precautions 9.2 Analysis	6
10 Ann	Expression of results 10.1 Quantities and units 10.2 Uncertainty 10.3 Test report	7 7
	nex A (informative) Example of application	δ

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 193, Natural gas, Subcommittee SC 1, Analysis of natural gas.

A list of all parts in the ISO 2613 series can be found on the ISO website. 14-d2e2-400a-bef8-

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document describes a method for the determination of the siloxane content of biomethane that is applicable to both laboratory and online analysis.

Siloxanes are a common contaminant in biogas, caused by their extensive use in common household products such as cosmetics and detergents. These products eventually become entrained in waste streams that are used to generate biogas. Additionally, siloxanes are sometimes added to digesters as anti-foaming agents to improve generation efficiency.

If not effectively removed, siloxanes can cause damage to gas processing and vehicle injection equipment through formation of silica deposits at high temperatures. Many gas distributors require measurement reports to demonstrate that the resultant biomethane conforms with their requirements, which typically includes siloxane content due to its potential to cause damage. Additionally, producers may wish to monitor siloxane content for process optimisation purposes.

The technique described in this standard is commercially available and tailored specifically to measuring siloxanes within biomethane at low concentrations (for instance, as specified in EN 16723-1^[1] and EN 16723-2^[2]). It is suitable for use both in the laboratory and field.

iTeh STANDARD PREVIEW (standards.iteh.ai)

oSIST prEN ISO 2613-2:2023 https://standards.iteh.ai/catalog/standards/sist/71284514-d2e2-400a-bef8fee5628c1193/osist-pren-iso-2613-2-2023 **oSIST** prEN ISO 2613-2:2023

iTeh STANDARD PREVIEW (standards.iteh.ai)

https://standards.iteh.ai/catalog/standards/sist/71284514-d2e2-400a-bef8-fee5628c1193/osist-pren-iso-2613-2-2023

Analysis of natural gas — Biomethane —

Part 2:

Determination of siloxane content by gas chromatography ion mobility spectrometry

1 Scope

This document describes a Gas Chromatography – Ion Mobility Spectroscopy (GC-IMS) method for the determination of the concentration of siloxanes in biomethane. The method is applicable to the following siloxanes:

- hexamethyldisiloxane (L2)
- octamethyltrisiloxane (L3)
- decamethyltetrasiloxane (L4)
- dodecamethylpentasiloxane (L5)
- hexamethylcyclotrisiloxane (D3)
- octamethylcyclotetrasiloxane (D4)
- octamethylcyclotetrasiloxane (D5)
- dodecamethylcyclohexasiloxane (D6)

This document describes suitable calibration and measurement strategies to quantify siloxanes in (bio) methane around and above the 0,3 mg m⁻³ (14 μ mol mol⁻¹) level and applies to analysis within absolute pressure ranges of 1 bar – 2 bar, temperatures of 0 °C – 40 °C and relative humidity < 90 %.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6142-1, Gas analysis — Preparation of calibration gas mixtures — Part 1: Gravimetric method for Class I mixtures

ISO 6143, Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures

ISO 6144, Gas analysis — Preparation of calibration gas mixtures — Static volumetric method

ISO 6145-1, Gas analysis — Preparation of calibration gas mixtures using dynamic methods — Part 1: General aspects

ISO 6974-1, Natural gas — Determination of composition and associated uncertainty by gas chromatography — Part 1: General guidelines and calculation of composition

ISO 6974-2, Natural gas — Determination of composition and associated uncertainty by gas chromatography — Part 2: Uncertainty calculations

ISO 7504, Gas analysis — Vocabulary

ISO 10715, Natural gas — Gas sampling

ISO 10723, Natural gas — Performance evaluation for analytical systems

ISO 14532, Natural gas — Vocabulary

ISO 14912, Gas analysis — Conversion of gas mixture composition data

ISO/IEC guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 7504, ISO/IEC guide 98-3:2008, ISO 14532 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at https://www.electropedia.org/

3.1 ion mobility spectrometry IMS

analytical method used to separate and identify ionised molecules in the gas phase based on their mobility in a carrier gas under an electric field

4 Principle

oSIST prEN ISO 2613-2:2023

A gaseous biomethane sample is drawn into a sample loop via a vacuum pump. The sample is injected onto a GC column, which separates the individual components in the gas mixture. As they exit the column, the components are ionised and passed into a drift tube. The drift tube applies an electroc field and a flow of inert nitrogen drift gas that causes the ions to drift at different rates based on their mass and geometric structure onto a detector.

The instrument is calibrated by sampling an appropriate range of siloxane concentrations and using the detector output to generate a calibration function. The function can then be applied to unknown samples to calculate the concentration of siloxanes. This can then be summed to calculate total siloxanes and total silicon.

The method is applicable to siloxanes within the concentration ranges provided in Table 1.

Component	Abbreviation	Formula	Lower limit (mg m ⁻³)	Upper limit (mg m ⁻³)
hexamethyldisiloxane	L2	C ₆ H ₁₈ Si ₂ O	0,1	10,0
octamethyltrisiloxane	L3	$C_8H_{24}Si_3O_2$	0,1	10,0
decamethyltetrasiloxane	L4	$C_{10}H_{30}Si_4O_3$	0,1	10,0
dodecamethylpentasiloxane	L5	$C_{12}H_{36}Si_5O_4$	0,1	10,0
hexamethylcyclotrisiloxane	D3	$C_{12}H_{18}O_3Si_3$	0,1	10,0
octamethylcyclotetrasiloxane	D4	C ₈ H ₂₄ O ₄ Si ₄	0,1	10,0
octamethylcyclotetrasiloxane	D5	$C_{10}H_{30}O_{5}Si_{5}$	0,1	10,0
dodecamethylcyclohexasiloxane	D6	$C_{12}H_{36}O_{6}Si_{6}$	0,3	10,0

Table 1 — Application ranges