
**Tehnični pripomočki za celovitost tkiv v ležečem položaju - 1. del: Splošne zahteve
(ISO/FDIS 20342-1:2022)**

Assistive products for tissue integrity when lying down - Part 1: General requirements
(ISO/FDIS 20342-1:2022)

Unterstützende Produkte zur Gewebeintegrität im Liegen - Teil 1: Allgemeine
Festlegungen (ISO/FDIS 20342-1:2022)

Produits d'assistance pour l'intégrité des tissus en position allongée - Partie 1: Exigences
générales (ISO/FDIS 20342-1:2022)

Ta slovenski standard je istoveten z: prEN ISO 20342-1

oSIST prEN ISO 20342-1:2022
<https://standards.itec.ai/catalog/standards/sist/057c1724-0214-4ee3-8397-407be15dd1fe/osist-pren-iso-20342-1-2022>

ICS:

11.180.01	Pripomočki za onesposobljene in hendikepirane osebe na splošno	Aids for disabled and handicapped persons in general
-----------	---	--

oSIST prEN ISO 20342-1:2022**en,fr,de**

**iTeh STANDARD
PREVIEW
(standards.iteh.ai)**

oSIST prEN ISO 20342-1:2022

<https://standards.iteh.ai/catalog/standards/sist/057c1724-0214-4ee3-8397-407be15dd1fe/osist-pren-iso-20342-1-2022>

FINAL
DRAFT

INTERNATIONAL
STANDARD

ISO/FDIS
20342-1

ISO/TC 173

Secretariat: SIS

Voting begins on:
2022-03-02

Voting terminates on:
2022-05-25

Assistive products for tissue integrity when lying down —

Part 1: General requirements

*Produits d'assistance pour l'intégrité des tissus en position allongée —
Partie 1: Exigences générales*

(standards.iteh.ai)

oSIST prEN ISO 20342-1:2022

<https://standards.iteh.ai/catalog/standards/sist/057c1724-0214-4ee3-8397-407be15dd1fe/osist-pren-iso-20342-1-2022>

IMPORTANT — Please use this updated version dated 2022-02-17, and discard any previous version of this FDIS.

ISO/CEN PARALLEL PROCESSING

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.



Reference number
ISO/FDIS 20342-1:2022(E)

© ISO 2022

iTeh STANDARD PREVIEW (standards.iteh.ai)

oSIST prEN ISO 20342-1:2022

<https://standards.iteh.ai/catalog/standards/sist/057c1724-0214-4ee3-8397-407be15dd1fe/osist-pren-iso-20342-1-2022>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	2
3 Terms and definitions	3
4 General requirements and safety	7
4.1 General requirements	7
4.2 Intended use	8
4.2.1 General requirements	8
4.2.2 Consideration regarding intended use	8
4.2.3 Intended use statement	8
4.3 APTI risk management	9
4.4 APTI usability	9
4.4.1 General	9
4.4.2 Design requirements in relation to persons with cognitive impairment	9
4.5 Design controls	9
4.6 Clinical evaluation	9
4.7 Foreseeable misuse	9
4.8 Test conditions	9
4.9 Lifting and carrying means	10
5 Safety requirements	10
5.1 Requirements for information supplied by the manufacturer	10
5.1.1 General	10
5.1.2 APTI traceability	11
5.1.3 Education and training	11
5.1.4 Pre-sale information	11
5.1.5 User information	12
5.1.6 Service information and inspection	13
5.1.7 Labelling	13
5.1.8 Marking of user weight and maximum load	13
5.1.9 Packaging	13
5.2 APTI that can be dismantled	14
5.2.1 General requirements	14
5.2.2 Small parts	14
5.2.3 Fasteners and connections	14
5.3 Resistance to corrosion	14
5.4 Noise and vibration	14
5.5 Sound audible acoustic energy	14
5.6 Default indicators	15
5.7 Feedback	16
6 Flammability	16
6.1 General	16
6.2 Flammability	16
6.3 Moulded parts used as enclosures for electrical equipment	17
7 Mechanical safety	17
7.1 Prevention of traps for the human body	17
7.2 Safety of moving and folding parts	17
7.3 V-shaped openings	19
7.4 Surfaces, corners, edges and protruding parts	19
7.5 Folding and adjusting mechanisms	19
7.6 Instability hazard	19
7.7 Temperature of parts that come into contact with human skin	20

ISO/FDIS 20342-1:2022(E)

7.8	Ergonomic principles.....	21
7.9	Additional consideration.....	21
8	Safety of electrical equipment.....	21
8.1	General electrical requirements.....	21
8.2	Electromagnetic compatibility.....	21
8.2.1	General.....	21
8.2.2	Emissions.....	21
8.2.3	Immunity.....	21
8.2.4	Power frequency magnetic field immunity.....	22
8.3	Liquid ingress.....	22
8.4	Interruption of power supply/supply mains to an APTI.....	22
8.5	Hold to run activation.....	22
8.6	Emergency stop functions.....	23
9	Biocompatibility.....	23
9.1	Biocompatibility and toxicity.....	23
9.2	Animal tissue.....	23
10	Contamination.....	23
10.1	Liquid ingress.....	23
10.2	Cleaning and disinfection.....	24
10.3	Cross infection and microbial contamination.....	24
Annex A (informative)	General information.....	26
Annex B (informative)	Environmental and consumer related guidance.....	30
Annex C (informative)	Periodic inspection.....	34
Bibliography		35

ITih STANDARD
PREVIEW
(standards.iteh.ai)

oSIST prEN ISO 20342-1:2022
<https://standards.iteh.ai/catalog/standards/sist/057c1724-0214-4ee3-8397-407be15dd1fe/osist-pren-iso-20342-1-2022>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 173, *Assistive products*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 293, *Assistive products and accessibility*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 20342-1:2019), which has been technically revised.

The main changes are as follows:

- the Scope was clarified;
- [Clause 2](#) was updated;
- [Clause 3](#) was updated;
- [subclause 7.3](#) about V-shaped openings was amended;
- [subclause 7.7](#) and Table 4 were amended (regarding surface temperature);
- the bibliography was updated.

A list of all parts in the ISO 20342 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document addresses Assistive Products for Tissue Integrity (APTI). As some devices can be used/reused in more than one application environment, different requirements and test methods can apply to the same APTI, depending on the application environment.

APTI play a very important role in the prevention and treatment of pressure injuries. Another important role in the prevention and treatment of pressure injury is the clinical practice and the clinical evaluation. Guidance can be found in the NPUAP/EPUAP/PPPIA Guidelines^[25].

Surfaces applied on operating theatre tables can also impact in the process of patient management and might need to be taken into consideration. It should be recognized however, patient stability and specialist equipment used during an operation often create conflicting priorities to those of an APTI.

Using this document, clinicians and manufacturers should consider the impact of other items (including additional APTI) used in conjunction with an APTI on tissue integrity and safety.

This document only covers general requirements to ensure safety of users.

iTeh STANDARD PREVIEW (standards.iteh.ai)

oSIST prEN ISO 20342-1:2022

<https://standards.iteh.ai/catalog/standards/sist/057c1724-0214-4ee3-8397-407be15dd1fe/osist-pren-iso-20342-1-2022>

Assistive products for tissue integrity when lying down —

Part 1: General requirements

1 Scope

This document specifies general requirements and related test methods that are relevant to assistive products for tissue integrity (APTI) in the lying position in different application environments such as hospitals, home care and institutions. This document applies to the safety of APTI that are intended to remain in situ during periods of lying, and to prevent and/or treat pressure injuries.

This document covers a range of different lying support surfaces intended to be used in combination with the appropriate support platform (adjustable included) or as a whole integrated system.

This document does not apply to medical beds.

This document also covers assistive products primarily intended for tissue integrity for changing a lying position and assistive products for maintaining a lying position.

This document does not apply to lying support surfaces used in combination with incubators or operating/surgical tables.

It also covers safety and performance test methods to ensure protection against injuries to the user.

This document addresses the combination of a full body support surface and an adjustable mattress support platform. It also covers safety and performance test methods to ensure protection against injuries to the user. <https://standards.iteh.ai/catalog/standards/sist/057c1724-0214-4ee3-8397-407be15dd1fe/osist-pren-iso-20342-1-2022>

This document specifies requirements and test methods for APTI within the following classifications of ISO 9999:—:

04 33 06 Assistive products for tissue integrity when lying down such as but not limited to

- mattresses and mattress overlays for pressure injury prevention, and
- mattress coverings for pressure injury prevention mattresses.

12 31 03 Assistive products for sliding and turning such as but not limited to the following:

Devices for changing position or direction of a person using sliding or turning techniques. The only products included are those intended to be used in a lying position and remain in situ as part of the lying support surface. They are the following:

- sliding products that glide one way and lock the other way;
- sheets and underlays in flexible materials with low friction;
- fabric sold by the metre, cut as required for repositioning use;
- powered turning product;

This excludes sliding boards unless the product is intended to be left in situ.

09 07 06 Positioning pillows, positioning cushions and positioning systems such as but not limited to

- leg positioners,

ISO/FDIS 20342-1:2022(E)

- arm positioners, and
- multipurpose body positioners.

18 12 15 Bedding such as but not limited to

- draw sheets.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 554, *Standard atmospheres for conditioning and/or testing — Specifications*

ISO 3746, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Survey method using an enveloping measurement surface over a reflecting plane*

ISO 9614-1, *Acoustics — Determination of sound power levels of noise sources using sound intensity — Part 1: Measurement at discrete points*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 13732-1, *Ergonomics of the thermal environment — Methods for the assessment of human responses to contact with surfaces — Part 1: Hot surfaces*

ISO 14155, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 22442-1, *Medical devices utilizing animal tissues and their derivatives — Part 1: Application of risk management*

IEC 60601-1:2006, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-2, *Medical electrical equipment — Part 1-2: General requirements for basic safety and essential performance — Collateral standard: Electromagnetic compatibility — Requirements and tests*

IEC 60601-1-6, *Medical electrical equipment — Part 1-6: General requirements for basic safety and essential performance — Collateral standard: Usability*

IEC 60601-1-8, *Medical electrical equipment — Part 1-8: General requirements for basic safety and essential performance — Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*

IEC 60601-1-11, *Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance — Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 60529, *Degrees of protection provided by enclosures (IP Code)*

IEC 60695-11-10, *Fire hazard testing — Part 11-10: Test flames — 50 W horizontal and vertical flame test methods*

IEC 61000-3-2, *Electromagnetic compatibility (EMC) — Part 3-2: Limits — Limits for harmonic current emissions (equipment input current ≤16 A per phase)*

IEC 61000-3-3, *Electromagnetic compatibility (EMC) — Part 3-3: Limits — Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤16 A per phase and not subject to conditional connection*

IEC 61000-4-3, *Electromagnetic compatibility (EMC) — Part 4-3: Testing and measurement techniques — Radiated, radio-frequency, electromagnetic field immunity test*

IEC 61000-4-8, *Electromagnetic compatibility (EMC) — Part 4-8: Testing and measurement techniques — Power frequency magnetic field immunity test*

IEC 61672-1, *Electroacoustics — Sound level meters — Part 1: Specifications*

IEC 61672-2, *Electroacoustics — Sound level meters — Part 2: Pattern evaluation tests*

IEC 62366, *Medical devices — Part 1: Application of usability engineering to medical devices*

IEC 80601-2-35, *Medical electrical equipment — Part 2-35: Particular requirements for the basic safety and essential performance of heating devices using blankets, pads or mattresses and intended for heating in medical use*

EN 716-2:2017, *Furniture — Children's cots and folding cots for domestic use — Part 2: Test methods*

EN 1041, *Information supplied by the manufacturer of medical devices*

CISPR 11, *Industrial, scientific and medical (ISM) radio-frequency equipment — Electromagnetic disturbance characteristics — Limits and methods of measurement*

European Commission, MEDDEV 2.7/1 CLINICAL EVALUATION: A GUIDE FOR MANUFACTURERS AND NOTIFIED BODIES UNDER DIRECTIVES 93/42/EEC and 90/385/EEC

(standards.iteh.ai)

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

application environment 4

care provided in a domestic area where the *assistive product for tissue integrity* (3.5) is used to alleviate or compensate for an injury, disability or disease

Note 1 to entry: This excludes use in all other application environments (e.g. nursing homes, rehabilitation and geriatric facilities) when an *assistive product for tissue integrity* (3.5) is purely designed for application environment 4.

[SOURCE: IEC 60601-2-52:2009+AMD1:2015, 201.3.204, modified — "assistive product for tissue integrity" replaced "ME equipment".]

3.2

applied part

part of the *assistive product for tissue integrity* (3.5) that, in normal use, comes into physical contact with the user of the assistive product for tissue integrity or a medical system to perform its function

[SOURCE: IEC 60601-1:2006, 3.8, modified — "assistive product for tissue integrity" replaced "ME Equipment", "of the" replaced "for", "user" replaced "patient", and "necessary" and notes not included.]

ISO/FDIS 20342-1:2022(E)

3.3

assistant

person who is helping a *user* (3.30) of the *assistive product for tissue integrity* (3.5)

EXAMPLE The ways assistants help persons with a *disability* (3.11) can be reposition in bed, bed ingress and egress, operating hoists and assisting with transferring in/out of seats.

Note 1 to entry: An assistant can be a health care professional or a non-professional, e.g. a relative.

3.4

assistive product

instrument, equipment or technical system intended by the manufacturer used for the prevention, treatment or alleviation of or compensation for *impairment* (3.13)

3.5

assistive product for tissue integrity**APTI**

surface intended to protect body tissue, designed to interface with the body when *lying down* (3.15) or in adjusted position

3.6

bedding

items normally placed on a *mattress* (3.16)

EXAMPLE Mattress covers, underlays, sheets, blankets, quilts (duvets) and their covers, cushions, pillows, bolsters and pillow cases.

3.7

body mass index**BMI**

value derived from the mass (weight in kilograms) and height (in metres) of an individual, defined as the body mass divided by the square of the body length, expressed in units of kg/m², calculated by the following:

$$BMI = m/l^2$$

<https://standards.iteh.ai/catalog/standards/sist/057c1724-0214-4ee3-8397-407be15dd1fe/osist-pren-iso-20342-1-2022>

where

m is the mass in kg;

l is the length in m.

3.8

clinical evaluation

assessment and analysis of clinical data pertaining to a *medical device* (3.21) to verify the clinical safety and performance of the device when used as intended by the manufacturer

Note 1 to entry: Can include a compilation of clinical data, any scientific literature and the results of any *clinical investigations* (3.9), taking into account any relevant harmonized standards.

Note 2 to entry: Guidance for clinical data evaluation is given in MEDDEV 2.7/1.

[SOURCE: ISO 13485:2016, 3.3, modified — Notes to entry added.]

3.9

clinical investigation

systematic investigation in one or more human subjects, undertaken to assess the clinical performance, effectiveness, or safety of a medical device

[SOURCE: ISO 14155:2020, 3.8, modified — Note 1 to entry was removed.]

3.10**detachable part**

part designed to be unfastened or disconnected without damage to the part or the whole

3.11**disability**

impairments (3.13), activity limitations, and participation restrictions denoting the negative aspects of the interaction between an individual (with a health condition) and that individual's contextual factors (environmental and personal factors)

[SOURCE: ICF 2001, WHO]

3.12**expected service life**

time specified by the manufacturer during which the *assistive product for tissue integrity* (3.5) is expected to remain safe for use (i.e. maintain basic safety and claimed performance)

Note 1 to entry: Maintenance can be necessary during expected service life.

[SOURCE: IEC 60601-1:2006, 3.28, modified — "assistive product for tissue integrity" replaced "ME equipment or ME system", "may" replaced "can" and "claimed" replaced "essential".]

3.13**impairment**

problem in body function or structure, such as a significant deviation or loss

[SOURCE: ICF 2001, WHO]

3.14**intended use**

use of a product, process or service in accordance with the specifications, instructions and information provided by the manufacturer

Note 1 to entry: This information includes pre-sale information.
[SOURCE: ISO 14971: 2019, 3.6, modified — "of" replaced "for which", note to entry modified.]

3.15**lying down**

position when the *user* (3.30) is in prone, supine, semi-recumbent or lateral on a full body support surface

3.16**mattress**

full body support surface designed to be placed directly on the existing bed frame

[SOURCE: RESNA SS-1: 2014, Section 1]

3.17**mattress overlay**

additional support surface designed to be placed directly on top of an existing support surface

[SOURCE: RESNA SS-1:2014, modified — "mattress" changed to "support surface".]

3.18**maximum load**

safe working load

SWL

greatest permissible load specified by the manufacturer

Note 1 to entry: This load is related to safety of the product; e.g., strength and durability, and covers the mass of the *user* (3.30), accessories, and other loads placed on the *assistive product for tissue integrity* (3.5).