



Designation: E1511 – 93 (Reapproved 2010)

Standard Practice for Testing Conductivity Detectors Used in Liquid and Ion Chromatography¹

This standard is issued under the fixed designation E1511; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers the testing of the performance of conductivity detectors used as the detection component of a liquid or ion chromatography system.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

2. Referenced Documents

2.1 *ASTM Standards*:²

E1151 Practice for Ion Chromatography Terms and Relationships

3. Terminology

3.1 See Practice **E1151**.

3.2 *Definitions*:

3.2.1 *cell constant*—the cell constant (K) of a conductivity cell is equal to $1/A$, so $\kappa = G K$.

3.2.1.1 *Discussion*—If the cell constant of the flow-through cell used is equal to one, then the conductivity equals the conductance. Although the cell constant is often specified for conductivity detectors, there is little practical value in knowing the constant as long as the detector is properly calibrated for conductivity.

3.2.2 *conductance*—the conductance (G) of a solution is the inverse of the resistance measured between two electrodes in a cell, expressed in units of siemens (S), equal to inverse ohms.

3.2.2.1 *Discussion*—The term resistance refers specifically to the dc resistance to ionic current, independent of the capacitive reactance at the interfaces between the electrodes and the solution.

¹ This practice is under the jurisdiction of ASTM Committee **E13** on Molecular Spectroscopy and Separation Science and is the direct responsibility of Subcommittee **E13.19** on Separation Science.

Current edition approved Nov. 1, 2010. Published November 2010. Originally approved in 1993. Last previous edition approved in 2005 as E1511 - 93(2005). DOI: 10.1520/E1511-93R10.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

3.2.3 *conductivity*—since the conductance is dependent on both the conductive properties of the solution and on the dimensions of the electrodes and the cell, the conductivity (κ) of the solution is defined to be independent of electrode and cell dimensions. Specifically,

$$\kappa = G \frac{1}{A} \quad (1)$$

where:

l = the distance between two planer disk electrodes, and

A = the electrode's surface area.

3.2.3.1 *Discussion*—In liquid and ion chromatography, cell dimensions are commonly measured in centimetres, so the units of κ are S/cm. (Alternatively, the SI units of S/m may be used. S/m = 100 S/cm.)

3.2.4 *drift*—the average slope of the noise envelope expressed in nano siemens per centimetre per hour as measured over a period of 1 h.

3.2.5 *equivalent conductivity*—of an ionic solute, the contribution of the solute to the total conductivity of the solution, measured in microsiemens per centimetre, divided by its concentration in milliequivalents/litre.

3.2.6 *flow dependence rate*—the change in measured conductivity as a function of flow rate.

3.2.7 *limiting equivalent conductivity*—of an ionic solute, its equivalent conductivity extrapolated to infinite dilution.

3.2.8 *linear range*—of a conductivity detector for a given solute in a specific solvent, the concentration range of solute for which the detector response factor is within 5 % of the response factor in the middle of the range as determined from the linearity plot specified in Section 11.

3.2.8.1 *Discussion*—The lower limit may be limited by noise, and the upper limit by deviation from linearity. (The upper limit may instead be limited by the maximum full-scale deflection on the detector's least sensitive output range.)

3.2.9 *long-term noise*—the maximum amplitude in nano siemens per centimetre for all random variations of the detector output of frequencies between 2 and 60 cycles per hour.

3.2.9.1 *Discussion*—Long-term noise represents noise that can be mistaken for eluting peaks.

3.2.10 *minimum detectability*—of a conductivity detector, that concentration of solute in a specific solvent that corresponds to twice the short-term noise.

3.2.10.1 *Discussion*—Because of the difficulty of pumping solvents through the chromatographic system without any contamination of the solvents from the system, this quantity can only be measured with solutes retained by a column. Since minimum detectability is dependent on the chromatographic system used, it is not measured in this practice. However, if the minimum detectability of a solute is measured on one system with one detector, the minimum detectability can be predicted when other detectors are tested on the same system by comparing the measured values of short-term noise.

3.2.11 *response factor*—of a conductivity detector, the measured conductivity response of a solute divided by the solute concentration.

3.2.12 *response time of the detector*—the time required for the output of the detector to change from 10 to 90 % of the new equilibrium value when the composition of the eluent is changed in a stepwise manner, within the linear range of the detector.

3.2.12.1 *Discussion*—A slow response time has the effect of limiting resolution for efficient peaks such as early eluting peaks and those from highly efficient columns or microbore columns. Response time is generally dependent on three factors: (a) cell volume, (b) volume of heat transfer tubing leading to the cell, and (c) electronic filtering of the output.

3.2.13 *sensitivity*—the detector response divided by concentration, which is also the response factor (11.1.1).

3.2.13.1 *Discussion*—Sensitivity is therefore by definition the same for all properly calibrated conductivity detectors. (Sensitivity is often confused with minimum detectability, which is dependent on both sensitivity and noise.) Therefore, the calibration of the detector should be measured, and if necessary, adjusted. Follow the manufacturer's procedure for calibrating the detector. The procedure in Section 9 is used by many manufacturers and is useful for the tests in this practice.

3.2.14 *short-term noise*—the maximum amplitude in nano siemens per centimetre for all random variations of the detector output of a frequency greater than one cycle per minute.

3.2.14.1 *Discussion*—Short-term noise determines the smallest signal detectable by a conductivity detector, limits the precision available for the determination of trace samples, and may set the lower limit of linearity.

4. Summary of Practice

4.1 Four different tests are performed to characterize a detector.

4.1.1 *Noise and drift* are measured while a solution is flowing through the detector cell. The test is performed using two different solutions: deionized water (DI) and 1 mM potassium chloride (KCl).

4.1.2 *Linear range* is determined by preparing a plot of response factor versus the log of solute concentration using standard solutions of KCl and hydrochloric acid (HCl) as solutes.

4.1.3 *Dependence of response on flow rate* is measured by pumping 1 mM KCl through the conductivity cell at several flow rates and measuring the detector output.

4.1.4 *Response time* is measured by measuring the time required for the detector output to change from that measured with DI water to that measured with 1 mM KCl.

5. Significance and Use

5.1 This practice is intended to describe the performance of a conductivity detector independent of the chromatographic system in terms that the analyst can use to predict overall system performance when the detector is coupled to the column and other chromatography system components.

5.2 Although it is possible to observe each of the several characteristics of a detector under different and unique conditions, it is the intent of this practice that a complete set of detector specifications should be obtained at the same operating conditions, including the setup used for testing, flow rates, and temperatures. It should be noted that to specify a detector's capability completely, its performance should be measured at several sets of conditions within the useful range of the detector. The terms and tests described in this practice are sufficiently general so that they may be used at whatever conditions may be chosen for other reasons.

6. Reagents

6.1 Reagent chemicals are reagent grade or better.

6.1.1 *Deionized Water*, (DI water), 18 M-ohm.

6.1.2 *Potassium Chloride*, (KCl) dry powder.

6.1.3 *Hydrochloric Acid*, (HCl) standard 0.1000 N solution.

7. Preparation of Standards

7.1 *Potassium Chloride Standards*:

7.1.1 Prepare a 10-mM KCl standard stock solution. Weigh out 0.7455 g KCl (desiccated) and dissolve it in 18 M-ohm DI water in a 1-L plastic volumetric flask. Fill the flask to 1 L with DI water.

7.1.2 Prepare KCl standards from the 10-mM KCl standard stock solution. Using accurate Class A pipettes, pipette the volumes of the 10-mM standard stock solution listed below into 100-mL plastic volumetric flasks. For the 1-mM KCl standard, fill a 100-mL plastic volumetric flask with the 10-mM KCl solution and transfer to a 1-L plastic volumetric flask. Fill to the line with DI water.

KCl Concentration, mM	Volume in 100 mL DI Water, mL
0.05	0.5
0.1	1
0.2	2
0.5	5
1	10 mL in 1 L
2	20
5	50
10 mM	No dilution

7.2 *Hydrochloric Acid Standards*:

7.2.1 Prepare a 2.00-mM HCl standard stock solution by diluting 20.0 mL of standard 0.1000 N HCl into a 1-L plastic volumetric flask and filling to the line with DI water. If standard 0.1000 N HCl is not available, a 0.10-mM HCl

solution can be prepared by diluting 8.3 mL of 12 *N* (37 %) concentration HCl into 1 L of DI water. (The concentration of this solution will be less accurate than that prepared from 0.1000 *N* HCl standard.)

7.2.2 Prepare the following HCl calibration standards from the 2.00-mM HCl standard stock solution. Use accurate Class A pipettes and 100-mL plastic volumetric flasks.

HCl Concentration, mM	Volume in 100 mL DI Water, mL
0.02	1
0.04	2
0.1	5
0.2	10
0.4	20
1	50
2	No dilution

8. Instrumentation Set-Up

8.1 Set up the chromatographic system according to the manufacturer's recommendation. Also, passivate the conductivity cell using the manufacturer's recommended procedure. Set the flow rate on the pump to 1.0 mL/min or to the flow rate normally used in your application. Fill the eluent bottle with 18 M-ohm DI water. Connect the outlet of the pump to the injection valve, and the outlet of the injection valve directly to the conductivity cell using as short a length of tubing as is practical. (Standard 0.25-mm (0.01-in.) internal diameter HPLC tubing may be used.) Do not install any columns or suppressors. To ensure smooth operation of the pump, it is necessary to supply more pressure than that normally provided by the detector cell and the standard tubing alone. This is accomplished by installing a 1-m coil of 0.25-mm internal diameter narrow bore tubing between the pump and the injection valve. Increase the length or decrease the diameter of the tubing if the pressure is not high enough to produce smooth pump operation. Generally, 500 to 1 000 psi will be sufficient. The waste line connected to the cell outlet should be of sufficient length to provide enough backpressure on the cell to prevent the formation of bubbles inside the cell. Inserting 20 cm of 0.25-mm internal diameter tubing between the cell and waste line should provide sufficient backpressure.

8.2 Install a sample injection loop of approximately 200 μ L on the injection valve. This can be constructed from 1 m of 0.5-mm (0.02-in.) internal diameter tubing. During the tests described in Sections 6 and 7, observe the recorder trace and verify that a plateau is reached after injection of the standard solutions. If no plateau is reached, then a larger sample injection loop is needed.

8.3 If the conductivity detector has a setting for temperature compensation, set it to 2.0. If not, the DI water eluent and all of the test solutions should be thermostated as close as possible to 25°C. Or, the detector cell may be thermostated at a higher temperature but be calibrated as if the cell were at 25°C. If the cell is thermostated, ensure that the cell temperature has stabilized. Refer to the manufacturer's procedure for cell temperature stabilization. Turn off any output filtering on the detector. The output from the detector should be monitored on a strip-chart recorder, integrator, or computer. The calibration and linearity tests can be performed with a voltmeter monitor-

ing the detector output or, on some detectors, the output is monitored on the front panel readout.

9. Calibration

9.1 *Method*—The detector is calibrated by adjusting the detector output to 147.0 μ S/cm for a 1-mM solution of potassium chloride flowing through the conductivity cell at 1 mL/min.

9.1.1 Set up the chromatographic system according to the instructions in Section 8. Turn on the pump and ensure that the pump is pumping smoothly.

9.1.2 Monitor the detector output. The conductivity should be below 1 μ S/cm. If it is higher, continue flushing out the system to remove leftover salts until the conductivity stabilizes below 1 μ S/cm. (A higher reading is an indication of either an incompletely cleaned flow system or of poor deionized water quality and may compromise noise and linearity tests.) Fill the sample injection loop with DI water and inject. Note the minimum conductivity reported during the elution of the DI water through the detector cell. Either calibrate the detector to zero using the injected DI water or subtract the conductivity of injected DI water from all subsequent measurements.

9.1.3 Fill the sample injection loop with the 1-mM KCl standard and inject the standard. Note the maximum conductivity reported during the elution through the detector cell of the 1-mM KCl standard. It should be 147.0 μ S/cm. If it is not, follow the manufacturer's procedure for calibrating the conductivity detector so that the reading will be 147.0 μ S/cm.

9.1.4 Some conductivity detectors do not report conductivity directly in siemens, but instead provide a voltage output proportional to conductivity. Instead of adjusting the detector output, calibrate these detectors by recording the known conductivity of the calibration solution (147.0 μ S/cm for 1 mM KCl), the detector sensitivity range, and the measured voltage output. (Be sure to subtract the voltage output for a blank of DI water.) Divide the known conductivity by the net voltage output reading and multiply all subsequent voltage output readings by this value.

10. Noise and Drift

10.1 *Method of Measurement*—Noise and drift are measured under two conditions. Pure DI water is pumped through the conductivity cell at 1 mL/min and the noise and drift measured. The procedure is then repeated using 1 mM KCl. The detector output may be sensitive to temperature changes. It is worthwhile to perform this test twice: once with the temperature of the eluents and of the laboratory held as constant as possible, and once with controlled changes in laboratory temperature of approximately 5°C. This may be accomplished by cycling on and off a room heater or air conditioner.

10.1.1 Set up the chromatographic system as described in Section 8. Calibrate the detector according to the procedure in Section 9. Pump DI water through the system for at least 1 h or until the detector output has stabilized below 1 μ S/cm. Set the detector to its maximum sensitivity setting and adjust the detector output to approximately midscale on the recorder. Adjust the response time or time constant of the detector output filter to 1 s or to the setting closest to 1 s. Record the baseline for at least 1 h. If the detector output goes off scale during the