



# SLOVENSKI STANDARD

## SIST EN 17864:2024

01-junij-2024

Nadomešča:  
SIST EN 15705:2010

---

### Anorganska gnojila - Določanje dušika v IBDU (izobutilidendiurea) in CDU (krotonilidendiurea)

Inorganic fertilizers - Determination of nitrogen content in IBDU (isobutylidenediurea) and CDU (crotonylidenediurea)

Anorganische Düngemittel - Bestimmung des Stickstoffgehalts in IBDU (Isobutylidendiharnstoff) und CDU (Crotonylidendiharnstoff)

Engrais inorganiques - Détermination de la teneur en azote dans l'IBDU (isobutylidène diurée) et la CDU (crotonylidène diurée)

Ta slovenski standard je istoveten z: **EN 17864:2023**

<https://standards.itech.ai/catalog/standards/sist/971146f6-ddb8-48cb-a065-a6a53738b438/sist-en-17864-2024>

---

#### **ICS:**

65.080                      Gnojila    Fertilizers

**SIST EN 17864:2024**    **en,fr,de**



EUROPEAN STANDARD

EN 17864

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2023

ICS 65.080

Supersedes EN 15705:2010

English Version

## Inorganic fertilizers - Determination of nitrogen content in IBDU (isobutylidenediurea) and CDU (crotonylidenediurea)

Engrais inorganiques - Détermination de la teneur en  
azote dans l'IBDU (isobutylidène diurée) et la CDU  
(crotonylidène diurée)

Anorganische Düngemittel - Bestimmung des  
Stickstoffgehalts in IBDU (Isobutylidendiharnstoff) und  
CDU (Crotonylidendiharnstoff)

This European Standard was approved by CEN on 27 November 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

[SIST EN 17864:2024](https://standards.iteh.ai/catalog/standards/sist/971146f6-ddb8-48cb-a065-a6a53738b438/sist-en-17864-2024)

<https://standards.iteh.ai/catalog/standards/sist/971146f6-ddb8-48cb-a065-a6a53738b438/sist-en-17864-2024>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

<b>Contents</b>	<b>Page</b>
European foreword .....	3
<b>1 Scope</b> .....	<b>4</b>
<b>2 Normative references</b> .....	<b>4</b>
<b>3 Terms and definitions</b> .....	<b>4</b>
<b>4 Principle</b> .....	<b>4</b>
<b>5 Reagents</b> .....	<b>4</b>
<b>6 Apparatus and equipment</b> .....	<b>4</b>
<b>7 Sampling and sample preparation</b> .....	<b>5</b>
<b>8 Procedure</b> .....	<b>5</b>
<b>8.1 General</b> .....	<b>5</b>
<b>8.2 Calibration</b> .....	<b>5</b>
<b>8.2.1 Stock solution IBDU, mass concentration <math>\rho(\text{IBDU}) = 100 \text{ mg/l}</math></b> .....	<b>5</b>
<b>8.2.2 Stock solution CDU, <math>\rho(\text{CDU}) = 100 \text{ mg/l}</math></b> .....	<b>5</b>
<b>8.2.3 Calibration solutions</b> .....	<b>5</b>
<b>8.3 Preparation of the test solution</b> .....	<b>6</b>
<b>8.4 Measurement</b> .....	<b>6</b>
<b>9 Calculation and expression of the result</b> .....	<b>6</b>
<b>10 Precision</b> .....	<b>7</b>
<b>10.1 Inter-laboratory tests</b> .....	<b>7</b>
<b>10.2 Repeatability</b> .....	<b>7</b>
<b>10.3 Reproducibility</b> .....	<b>7</b>
<b>11 Test report</b> .....	<b>7</b>
<b>Annex A (informative) Results of the inter-laboratory tests</b> .....	<b>8</b>
<b>Annex B (informative) Example chromatogram and calibration curves</b> .....	<b>9</b>
<b>B.1 Example chromatogram</b> .....	<b>9</b>
<b>B.2 Example calibration curves</b> .....	<b>9</b>
<b>Annex C (informative) Example conditions for a suitable HPLC system</b> .....	<b>11</b>
<b>Bibliography</b> .....	<b>12</b>

## European foreword

This document (EN 17864:2023) has been prepared by Technical Committee CEN/TC 260 “Fertilizers and liming materials”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2024, and conflicting national standards shall be withdrawn at the latest by June 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document, together with EN 15705:2023, supersedes EN 15705:2010.

In comparison with EN 15705:2010, the following technical modifications have been made:

- EN 15705:2010 is split into two documents:
  - Method A of EN 15705:2010 is given in this document;
  - Method B of EN 15705:2010 is transferred to EN 15705:2023.

This document has been prepared under a standardization request addressed to CEN by the European Commission. The Standing Committee of the EFTA States subsequently approves these requests for its Member States.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## EN 17864:2023 (E)

### 1 Scope

This document specifies a method for the determination of nitrogen content in IBDU (isobutylidenediurea) and CDU (crotonylidenediurea) using high-performance liquid chromatography (HPLC).

The method is applicable to all fertilizers which do not contain interfering organic compounds.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1482-2, *Fertilizers and liming materials — Sampling and sample preparation — Part 2: Sample preparation*

EN 12944-1, *Fertilizers and liming materials — Vocabulary — Part 1: General terms*

EN 12944-2, *Fertilizers and liming materials — Vocabulary — Part 2: Terms relating to fertilizers*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944-1 and EN 12944-2 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

### 4 Principle

The sample is extracted with water and, after appropriate dilution, analyzed using a suitable HPLC system.

### 5 Reagents

Use only reagents of recognized analytical grade and distilled or demineralized water, free from carbon dioxide and all nitrogenous compounds and having an electric conductivity < 0,5 mS/m.

**5.1 Acetonitrile**, HPLC-grade.

**5.2 Isobutylidenediurea (IBDU)**, in pure form, as standard.

**5.3 Crotonylidenediurea (CDU)**, in pure form, as standard.

**5.4 Eluent**, acetonitrile + water (10 + 90,  $V + V$ ).

Mix 100,0 ml of acetonitrile (5.1) with 900,0 ml of water and homogenize well.

### 6 Apparatus and equipment

Usual laboratory glassware and equipment and, in particular, the following.