



SLOVENSKI STANDARD
oSIST ISO/DIS 13794:2018
01-september-2018

Zunanji zrak - Določevanje azbestnih vlaken - Metoda transmisijske elektronske mikroskopije s posrednim prenosom

Ambient air - Determination of asbestos fibres - Indirect-transfer transmission electron microscopy method

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Air ambient - Dosage des fibres d'amiante - Méthode par microscopie électronique à transmission par transfert indirect

<https://standards.iteh.ai/catalog/standards/sist/735984c4-204f-4c6c-b33b-ad73280404ad/sist-iso-13794-2018>

Ta slovenski standard je istoveten z: ISO/DIS 13794

ICS:

13.040.20 Kakovost okoljskega zraka Ambient atmospheres

oSIST ISO/DIS 13794:2018

en

DRAFT INTERNATIONAL STANDARD

ISO/DIS 13794

ISO/TC 146/SC 3

Secretariat: ANSI

Voting begins on:
2018-05-23Voting terminates on:
2018-08-15

Ambient air — Determination of asbestos fibres — Indirect-transfer transmission electron microscopy method

Air ambient — Dosage des fibres d'amiante — Méthode par microscopie électronique à transmission par transfert indirect

ICS: 13.040.20

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST ISO 13794:2019](https://standards.iteh.ai/catalog/standards/sist/735984c4-204f-4c6c-b33b-ad73280404ad/sist-iso-13794-2019)

<https://standards.iteh.ai/catalog/standards/sist/735984c4-204f-4c6c-b33b-ad73280404ad/sist-iso-13794-2019>

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

This document is circulated as received from the committee secretariat.



Reference number
ISO/DIS 13794:2018(E)

© ISO 2018

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 13794:2019

<https://standards.iteh.ai/catalog/standards/sist/735984c4-204f-4c6c-b33b-ad73280404ad/sist-iso-13794-2019>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents	Page
Foreword.....	vii
Introduction.....	viii
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
4 Symbols and abbreviated terms.....	6
5 Type of sample.....	7
6 Range.....	7
7 Limit of detection.....	7
8 Principle.....	8
9 Reagents.....	9
9.1 Water, fibre-free	9
9.2 Chloroform.....	9
9.3 1-Methyl-2-pyrrolidone	9
9.4 1,2-Diaminoethane (Ethylene diamine)	9
9.5 Dimethylformamide	9
9.6 Glacial acetic acid.....	9
9.7 Acetone.....	9
10 Apparatus.....	9
10.1 Air sampling	9
10.1.1 Filter cassette.....	9
10.1.2 Sampling pump.....	10
10.1.3 Stand.....	10
10.1.4 Personal Sampling.....	10
10.1.5 Flowmeter	10
10.2 Specimen preparation laboratory	10
10.3 Equipment for analysis.....	11
10.3.1 Transmission electron microscope.....	11
10.3.2 Energy dispersive X-ray analyzer	12
10.3.3 Plasma asher	13
10.3.4 Vacuum coating unit.....	13
10.3.5 Sputter coater	13
10.3.6 Beakers.....	13
10.3.7 Vacuum source	13
10.3.8 Glass filtration apparatus	13
10.3.9 Solvent washer (Jaffe washer)	14
10.3.10	Condensation
washer	15
10.3.11	Slide warmer or
oven	16
10.3.12	Ultrasonic
bath.....	16
10.3.13	Carbon grating
replica.....	16

ISO/DIS 13794:2018(E)

10.3.14.....	Calibration specimen grids for EDXA	16
10.3.15.....	Carbon rod sharpener.....	16
10.3.16.....	Disposable tip micropipettes	16
10.3.17.....	Thermometer	16
10.3.18.....	Stopwatch	16
10.4	Consumable supplies	16
10.4.1	Copper or nickel electron microscope grids	16
10.4.2	Gold or nickel electron microscope grids	17
10.4.3	Aluminium foil	17
10.4.4	Carbon rod electrodes	17
10.4.5	Routine electron microscopy tools and supplies.....	17
10.4.6	Reference asbestos samples.....	17
10.4.7	Reference samples of mineral fibres other than asbestos	18
11	Air sample collection	18
11.1	Calculation of analytical sensitivity.....	18
11.2	Sample collection procedure	18
12	Procedure for analysis	19
12.1	General.....	19
12.2	Cleaning of sample cassettes.....	20
12.3	Preparation of analytical filters.....	20
12.3.1	Selection of filter area for ashing	20
12.3.2	Ashing of sample collection filters.....	20
12.3.3	Aqueous dispersal of residual ash from sample collection filters	21
12.3.4	Assembly of system for filtration of aqueous dispersions	21
12.3.5	Filtration of aqueous dispersions	21
12.4	Preparation of TEM specimens from PC analytical filters.....	22
12.4.1	Selection of filter area for carbon-coating	22
12.4.2	Carbon-coating of filter portions.....	22
12.4.3	Preparation of the Jaffe washer	22
12.4.4	Placing of specimens into the Jaffe washer.....	22
12.5	Preparation of TEM specimens from cellulose ester analytical filters.....	23
12.5.1	Selection of area of filter for preparation	23
12.5.2	Preparation of solution for collapsing cellulose ester filters.....	23
12.5.3	Filter-collapsing procedure.....	23
12.5.4	Plasma etching of the filter surface	23
12.5.5	Carbon-coating.....	23
12.5.6	Preparation of the Jaffe washer	24
12.5.7	Placing of specimens in the Jaffe washer.....	24
12.6	Criteria for acceptable TEM specimen grids	24
12.7	Procedure for structure counting by TEM	25
12.7.1	General.....	25
12.7.2	Measurement of mean grid opening area	25
12.7.3	TEM alignment and calibration procedures.....	25
12.7.4	Determination of stopping point.....	26
12.7.5	General procedure for structure counting and size analysis	26
12.7.6	Estimation of mass concentration of asbestos fibres and bundles	27
12.7.7	Magnification requirements	27

12.8	Blank and quality control determinations	29
12.9	Calculation of results.....	30
13	Performance characteristics	30
13.1	General	30
13.2	Interferences and limitations of fibre identification.....	30
13.3	Precision and accuracy.....	31
13.3.1	Precision	31
13.3.2	Accuracy.....	31
13.3.3	Inter- and intra-laboratory analyses.....	31
13.4	Limit of detection.....	32
14	Test report	32
Annex A (normative)	Determination of operating conditions for plasma asher.....	36
A.1	General	36
A.2	Procedure	36
Annex B (normative)	Determination and standardization of operating conditions for ultrasonic bath.....	37
B.1	General	37
B.2	Procedure.....	37
Annex C (normative)	Calibration procedures	39
C.1	Calibration of the TEM	39
C.1.1	Calibration of TEM screen magnification.....	39
C.1.2	Calibration of ED camera constant.....	39
C.2	Calibration of the EDXA system.....	39
Annex D (normative)	Structure counting criteria.....	42
D.1	General	42
D.2	Structure definitions and treatment.....	42
D.2.1	Fibre	42
D.2.2	Bundle.....	42
D.2.3	Cluster.....	43
D.2.4	Matrix.....	46
D.2.5	Asbestos structure larger than 5 μm	46
D.2.6	Asbestos fibre or bundle longer than 5 μm	46
D.2.7	PCM equivalent structure	46
D.2.8	PCM equivalent fibre	46
D.3	Other structure counting criteria	46
D.3.1	Structures which intersect grid bars.....	46
D.3.2	Fibres which extend outside the field of view.....	47
D.4	Procedure for data recording.....	50
D.4.1	General	50
D.4.2	Fibres	50
D.4.3	Bundles.....	50
D.4.4	Disperse clusters (type D)	50
D.4.5	Compact clusters (type C)	50
D.4.6	Disperse matrices (type D).....	50
D.4.7	Compact matrices (type C)	51
D.4.8	Procedure for recording of partially obscured fibres and bundles	51
D.5	Special considerations for counting of PCM equivalent fibres.....	51
D.6	Special considerations for determination of an estimate of the mass concentration of asbestos fibres and bundles.....	52

ISO/DIS 13794:2018(E)

Annex E (normative) Fibre identification procedure.....	53
E.1 General.....	53
E.2 ED and EDXA techniques	53
E.2.1 General.....	53
E.2.2 ED techniques.....	53
E.2.3 EDXA measurements.....	55
E.3 Interpretation of fibre analysis data.....	56
E.3.1 Chrysotile.....	56
E.3.2 Amphiboles.....	56
E.4 Fibre classification categories.....	57
E.4.1 Procedure for classification of fibres with tubular morphology suspected to be chrysotile.....	58
E.4.2 Procedure for classification of fibres without tubular morphology, suspected to be amphibole.....	60
E.5 Routine fibre identification for investigation of known sources of the regulated asbestos varieties.....	63
Annex F (normative) Determination of the concentration of asbestos fibres and bundles longer than 5 µm, and PCM equivalent asbestos fibres.....	70
Annex G (normative) Calculation of results.....	71
G.1 General.....	71
G.2 Test for uniformity of distribution of fibrous structures on TEM grids.....	71
G.3 Calculation of the analytical sensitivity	71
G.4 Calculation of the mean and confidence interval of the structure concentration.....	72
G.4.1 Calculation of the mean structure concentration.....	72
G.4.2 Calculation of confidence intervals	72
G.4.3 Example of calculation of Poissonian 95 % confidence intervals	72
G.4.4 Example of calculation of Gaussian 95 % confidence intervals	73
G.4.5 Summary of procedure for calculation of results.....	73
G.5 Calculation of structure length, width, and aspect ratio distributions.....	74
G.5.1 General.....	74
G.5.2 Calculation of structure length cumulative number distribution.....	74
G.5.3 Calculation of structure width cumulative number distribution	75
G.5.4 Calculation of structure aspect ratio cumulative number distribution.....	75
G.6 Calculation of an estimate of mass concentration of asbestos fibres and bundles	75
Annex H (normative) Test procedure to determine suitability of cellulose ester sample collection filters.....	77
H.1 General.....	77
H.2 Test procedure	77
Annex I (informative) Strategies for collection of air samples.....	78
I.1 General.....	78
I.2 Air sample collection in the outdoors environment	78
I.3 Air sample collection inside buildings	78
Bibliography.....	79

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/XXX

This second/third/... edition cancels and replaces the first/second/... edition (), [clause(s) / subclause(s) / table(s) / figure(s) / annex(es)] of which [has / have] been technically revised.

ISO XXXX consists of the following parts. [Add information as necessary.]

ISO/DIS 13794:2018(E)

Introduction

This International Standard is applicable to the measurement of airborne asbestos in a wide range of ambient air situations, including the interior atmospheres of buildings, and for detailed evaluation of any atmosphere in which asbestos fibres are likely to be present. Because the best available medical evidence indicates that the numerical fibre concentration and the fibre size and type are the relevant parameters for evaluation of the inhalation hazards, a fibre counting and measuring technique is the only logical approach. Most fibres in ambient atmospheres are not asbestos, and therefore there is a requirement for fibres to be identified. Many airborne asbestos fibres in ambient atmospheres have diameters below the resolution limit of the optical microscope. This International Standard is based on transmission electron microscopy, which has adequate resolution to allow detection of small fibres and is currently the only technique capable of unequivocal identification of the majority of individual fibres of asbestos. The fibres found suspended in an ambient atmosphere can often be identified unequivocally, if sufficient measurement effort is expended. However, if each fibre were to be identified in this way, the analysis becomes prohibitively expensive. Because of instrumental deficiencies or because of the nature of the particulate, some fibres cannot be positively identified as asbestos, even though the measurements all indicate that they could be asbestos. Subjective and instrumental factors therefore contribute to this measurement, and consequently a very precise definition of the procedure for identification and enumeration of asbestos fibres is required.

In addition to single fibres and bundles, asbestos is often found in air samples as very complex, aggregated structures which may or may not be also aggregated with other particles. The number of asbestos fibres and bundles incorporated in these complex structures often exceeds the number of isolated fibres and bundles observed, and many of them may be completely obscured in direct-transfer TEM preparations. The method defined in this International Standard incorporates specimen preparation procedures that result in selective concentration of asbestos fibres, and removal of organic, water-soluble and acid-soluble materials. These procedures have the effect of dispersing the majority of the complex clusters and aggregates of fibres into their component fibres and bundles so that the asbestos in the air sample can be more accurately quantified. All of the feasible specimen preparation techniques result in some modification of the airborne particulate. Even the collection of particles from a three-dimensional airborne dispersion on to a two-dimensional filter surface can be considered a modification of the particulate, and some of the particles in most samples are modified by the specimen preparation procedures. Although this method results in dispersal of complex clusters and aggregates, it minimizes other effects on the size distribution of fibres and fibre bundles.

This International Standard requires that a very detailed and logical procedure be used to reduce the subjective aspects of the measurement. The method of data recording specified in the standard is designed to allow re-evaluation of the fibre counting data as new medical evidence becomes available.

This International Standard describes the method of analysis for a single air filter. However, one of the largest potential errors in characterizing asbestos in ambient atmospheres is associated with the variability between filter samples. For this reason, it is necessary to design a replicate sampling scheme in order to determine the standard's accuracy and precision.

Comparison of results obtained using this indirect-transfer procedure with those from the direct-transfer procedure may not be done a priori. A site-specific inter-comparison study must be done which takes into account the fibre size and type of asbestos, and also the nature of the source of the airborne asbestos.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 13794:2019

<https://standards.iteh.ai/catalog/standards/sist/735984c4-204f-4c6c-b33b-ad73280404ad/sist-iso-13794-2019>

Ambient air — Determination of asbestos fibres — Indirect-transfer transmission electron microscopy method

1 Scope

This International Standard specifies a reference method using transmission electron microscopy for the determination of airborne asbestos fibres and structures in a wide range of ambient air situations, including the interior atmospheres of buildings, and for a detailed evaluation for asbestos structures in any atmosphere. The specimen preparation procedure incorporates ashing and dispersion of the collected particulate, so that all asbestos is measured, including the asbestos originally incorporated in particle aggregates or particles of composite materials. The lengths, widths and aspect ratios of the asbestos fibres and bundles are measured, and these, together with the density of the type of asbestos, also allow the total mass concentration of airborne asbestos to be calculated. The method allows determination of the type(s) of asbestos fibres present. The method cannot discriminate between individual fibres of the asbestos and non-asbestos analogues of the same amphibole mineral.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4225:1994, *Air Quality — General aspects — Vocabulary*

ISO 4226:2007, *Air quality — General aspects — Units of measurement*

ISO 10312:1995, *Ambient air — Determination of asbestos fibres — Direct-transfer transmission electron microscopy method*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply (see also ISO 4225).

3.1


acicular

shape shown by an extremely slender crystal with cross-sectional dimensions which are small relative to its length, i.e. needle-like. [SOURCE: ISO 10312:1995, definition 3.1]

3.2

amphibole

group of rock-forming ferromagnesium silicate minerals, closely related in crystal form and composition, and having the nominal formula:



where:

A = K, Na;

ISO/DIS 13794:2018(E)

B = Fe²⁺, Mn, Mg, Ca, Na;

C = Al, Cr, Ti, Fe³⁺, Mg, Fe²⁺;

T = Si, Al, Cr, Fe³⁺, Ti.

[SOURCE: ISO 10312:1995, definition 3.2]

NOTE In some varieties of amphibole, these elements can be partially substituted by Li, Pb, or Zn. Amphibole is characterized by a cross-linked double chain of Si-O tetrahedra with a silicon:oxygen ratio of 4:11, by columnar or fibrous prismatic crystals and by good prismatic cleavage in two directions parallel to the crystal faces and intersecting at angles of about 56° and 124°.

3.3**amphibole asbestos**

amphibole in an asbestiform habit. [SOURCE: ISO 10312:1995, definition 3.3]

3.4**analytical filter**

filter through which an aqueous dispersion of ash from the sample collection filter is passed, and from which TEM specimen grids are prepared. [SOURCE: ISO 13794:1999, definition 2.4]

3.5**analytical sensitivity**

calculated airborne asbestos structure concentration in structures/litre, equivalent to counting of one asbestos structure in the analysis. [SOURCE: ISO 10312:1995, definition 3.4]

NOTE 1 It is expressed in structures/litre.

NOTE 2 This method does not specify a unique analytical sensitivity. The analytical sensitivity is determined by the needs of the measurement and the conditions found on the prepared sample.

3.6**asbestiform**

specific type of mineral fibrosity in which the fibres and fibrils possess high tensile strength and flexibility. [SOURCE: ISO 10312:1995, definition 3.5]

3.7**asbestos**

term applied to a group of silicate minerals belonging to the serpentine and amphibole groups which have crystallized in the asbestiform habit, causing them to be easily separated into long, thin, flexible, strong fibres when crushed or processed. [SOURCE: ISO 10312:1995, definition 3.6]

NOTE The Chemical Abstracts Service Registry Numbers of the **most common** asbestos varieties are: chrysotile (12001-29-5), crocidolite (12001-28-4), grunerite asbestos (Amosite) (12172-73-5), anthophyllite asbestos (77536-67-5), tremolite asbestos (77536-68-6) and actinolite asbestos (77536-66-4). Other varieties of asbestiform amphibole, such as richterite asbestos and winchite asbestos [11], are also found in some products such as vermiculite and talc.

3.8**asbestos structure**

term applied to an individual fibre, or any connected or overlapping grouping of asbestos fibres or bundles, with or without other particles. [SOURCE: ISO 10312:1995, definition 3.7]

3.9**ashed filter blank**

fibre count made on TEM specimens prepared by the indirect procedure from a blank membrane filter of the type used for collection of air samples. [SOURCE: ISO 13794:1999, definition 2.9]

3.10**aspect ratio**

ratio of length to width of a particle. [SOURCE: ISO 10312:1995, definition 3.8]

3.11**blank**

structure count made on TEM specimens prepared from an unused filter, to determine the background measurement. [SOURCE: ISO 10312:1995, definition 3.9]

3.12**camera length**

equivalent projection length between the specimen and its electron diffraction pattern, in the absence of lens action. [SOURCE: ISO 10312:1995, definition 3.10]

3.13**chrysotile**

fibrous mineral of the serpentine group which has the nominal composition: $Mg_3Si_2O_5(OH)_4$ [SOURCE: ISO 10312:1995, definition 3.11]

NOTE Most natural chrysotile deviates little from this nominal composition. In some varieties of chrysotile, minor substitution of silicon by Al^{3+} may occur. Minor substitution of magnesium by Al^{3+} , Fe^{2+} , Fe^{3+} , Ni^{2+} , Mn^{2+} and Co^{2+} may also be present. Chrysotile is the most prevalent type of asbestos.

3.14**cleavage**

breaking of a mineral along one of its crystallographic directions. [SOURCE: ISO 10312:1995, definition 3.12]

3.15**cleavage fragment**

fragment of a crystal that is bounded by cleavage faces. [SOURCE: ISO 10312:1995, definition 3.13]

NOTE Crushing of non-asbestiform amphibole generally yields elongated fragments that conform to the definition of a fibre, but rarely have aspect ratios exceeding 30:1.

3.16**cluster**

structure in which two or more fibres, or fibre bundles, are randomly oriented in a connected grouping. [SOURCE: ISO 10312:1995, definition 3.14]

3.17**direct-transfer blank**

a structure count made on TEM specimens prepared by the direct-transfer procedure from a blank filter of the type used for filtration of aqueous dispersions of ash. [SOURCE: ISO 13794:1999, definition 2.17]

3.18**d-spacing**

The distance between identical adjacent and parallel planes of atoms in a crystal. [SOURCE: