

## SLOVENSKI STANDARD **SIST EN 455-2:2024**

01-julij-2024

Medicinske rokavice za enkratno uporabo - 2. del: Zahteve in preskusi za ugotavljanje fizikalnih lastnosti

Medical gloves for single use - Part 2: Requirements and testing for physical properties

Medizinische Handschuhe zum einmaligen Gebrauch - Teil 2: Anforderungen und Prüfung der physikalischen Eigenschaften

Gants médicaux non réutilisables - Partie 2 : Exigences et essais pour propriétés physiques

Ta slovenski standard je istoveten z:

ICS:

11.140 Oprema bolnišnic Hospital equipment

SIST EN 455-2:2024 en,fr,de

# iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN 455-2:2024

https://standards.iteh.ai/catalog/standards/sist/7e3dbe54-11d8-4276-93ad-e0d691cd0f35/sist-en-455-2-2024

EUROPEAN STANDARD NORME EUROPÉENNE

**EUROPÄISCHE NORM** 

EN 455-2

May 2024

ICS 11.140

Supersedes EN 455-2:2015

### **English Version**

# Medical gloves for single use - Part 2: Requirements and testing for physical properties

Gants médicaux non réutilisables - Partie 2 : Exigences et essais pour propriétés physiques

Medizinische Handschuhe zum einmaligen Gebrauch -Teil 2: Anforderungen und Prüfung der physikalischen Eigenschaften

This European Standard was approved by CEN on 15 April 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

#### SIST FN 455\_2-2024

https://standards.iteh.ai/catalog/standards/sist/7e3dbe54-11d8-4276-93ad-e0d691cd0f35/sist-en-455-2-202-



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

## EN 455-2:2024 (E)

Cont	Pag	зe
European foreword		
1	Scope	. 5
2	Normative references	. 5
3	Terms and definitions	. 5
4 4.1	DimensionsGeneral	
4.2 4.3	LengthWidth	. 7
5 5.1 5.2 5.3	Strength  General  Force at break  Force at break after challenge testing	. 8 . 8
6	Test report	11
7	Labelling	11
	ZA (informative) Relationship between this European Standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered	12 14
_10110	D 4 D	

#### SIST EN 455-2:2024

https://standards.iteh.ai/catalog/standards/sist/7e3dbe54-11d8-4276-93ad-e0d691cd0f35/sist-en-455-2-2024

## **European foreword**

This document (EN 455-2:2024) has been prepared by Technical Committee CEN/TC 205 "Non-active medical devices", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2024, and conflicting national standards shall be withdrawn at the latest by November 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 455-2:2015.

Compared to the previous edition EN 455-2:2015, the following main changes have been introduced:

- a) normative references have been revised;
- b) subclause 4.2 has been updated with regard to recording the measured length ("median" has been removed);
- c) Clause 5 has been updated;
- d) Clause 6 has been updated;
- e) Annex ZA has been updated for harmonization under Medical Device Regulation (EU) 2017/745 (MDR).

This document has been prepared under a standardization request addressed to CEN by the European Commission. The Standing Committee of the EFTA States subsequently approves these requests for its Member States.

For the relationship with EU Legislation, see informative Annex ZA, which is an integral part of this document.

The EN 455 series consists of the following parts under the general title "Medical gloves for single use":

- Part 1: Requirements and testing for freedom from holes;
- Part 2: Requirements and testing for physical properties;
- Part 3: Requirements and testing for biological evaluation;
- Part 4: Requirements and testing for shelf life determination.

The following part is under development:

Part 5: Extractable chemical residues.

A list of all parts in a series can be found on the CEN website.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

#### EN 455-2:2024 (E)

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

# iTeh Standards (https://standards.iteh.ai) Document Preview

SIST EN 455-2:2024

https://standards.iteh.ai/catalog/standards/sist/7e3dbe54-11d8-4276-93ad-e0d691cd0f35/sist-en-455-2-2024