

FINAL
DRAFT

INTERNATIONAL
STANDARD

ISO/FDIS
19980

ISO/TC 172/SC 7

Secretariat: DIN

Voting begins on:
2021-03-10

Voting terminates on:
2021-05-05

Ophthalmic instruments — Corneal topographers

Instruments ophtalmiques — Topographes de la cornée

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/FDIS 19980](https://standards.iteh.ai/catalog/standards/sist/6d508bdb-229d-47fd-bbc3-df0c6fc8b3d4/iso-fdis-19980)

<https://standards.iteh.ai/catalog/standards/sist/6d508bdb-229d-47fd-bbc3-df0c6fc8b3d4/iso-fdis-19980>

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO/FDIS 19980:2021(E)

© ISO 2021

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/FDIS 19980

<https://standards.iteh.ai/catalog/standards/sist/6d508bdb-229d-47fd-bbc3-df0c6fc8b3d4/iso-fdis-19980>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	8
4.1 Area measured.....	8
4.2 Measurement sample density.....	9
4.3 Measurement and report of performance.....	9
4.4 Colour presentation of results.....	9
5 Test methods and test devices	9
5.1 Tests.....	9
5.1.1 Accuracy test.....	9
5.1.2 Repeatability test.....	9
5.2 Test surfaces.....	9
5.2.1 Reflection-based and Placido ring corneal topographers.....	9
5.2.2 Luminous surface corneal topographer.....	9
5.2.3 Optical-sectioning corneal topographer.....	9
5.2.4 Specification of test surfaces.....	10
5.2.5 Verification of test surfaces.....	10
5.2.6 Type testing of surfaces.....	10
5.3 Data collection — Test surfaces.....	11
5.4 Analysis of the data.....	11
5.4.1 General.....	11
5.4.2 Structure of the accuracy data set.....	12
5.4.3 Analysis of the paired data sets.....	12
5.4.4 Report of accuracy performance.....	13
6 Accompanying documents	13
7 Marking	13
Annex A (informative) Test surfaces for corneal topographers (CTs)	14
Annex B (informative) Standardized displays for corneal topographers (CTs)	16
Annex C (normative) Calculation of area-weighting values	19
Annex D (normative) Test methods for measuring human corneas	21
Bibliography	22

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 170, *Ophthalmic optics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 19980:2012), which has been technically revised. The main changes compared to the previous edition are as follows:

- a) normative references were updated;
- b) [5.2.6](#) regarding requirements for test surfaces and requirement for testing of accuracy was changed;
- c) in [5.4.3](#), formulae for data analysis have been updated;
- d) Table 4 was deleted;
- e) document editorially revised.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Ophthalmic instruments — Corneal topographers

1 Scope

This document specifies minimum requirements for instruments and systems that fall into the class of corneal topographers (CTs). It also specifies tests and procedures to verify that a system or instrument complies with this document and thus qualifies as a CT according to this document. It also specifies tests and procedures that allow the verification of capabilities of systems that are beyond the minimum requirements for CTs.

This document defines terms that are specific to the characterization of the corneal shape so that they may be standardized throughout the field of vision care.

This document is applicable to instruments, systems and methods that are intended to measure the surface shape of the cornea of the human eye.

NOTE The measurements can be of the curvature of the surface in local areas, three-dimensional topographical measurements of the surface or other more global parameters used to characterize the surface.

This document is not applicable to ophthalmic instruments classified as ophthalmometers.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005 + A1:2012, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

corneal apex

location on the corneal surface where the mean of the local principal curvature is greatest

Note 1 to entry: See Figure 1.

3.2

corneal eccentricity

e_c

eccentricity, e , of the conic section that best fits the *corneal meridian* (3.3) of interest

Note 1 to entry: If the meridian is not specified, the corneal eccentricity is that of the flattest corneal meridian (see [Table 1](#) and [Annex A](#)).

3.3
corneal meridian
 θ

curve created by the intersection of the corneal surface and a plane that contains the corneal topographer axis

Note 1 to entry: A meridian is identified by the angle θ , that the plane creating it makes to the horizontal (see ISO 8429).

Note 2 to entry: The value of θ , for a full meridian, ranges from 0° to 180°.

3.4
corneal shape factor
 E

value that specifies the type of conic section that best fits a *corneal meridian* (3.3), given by Formula (1):

$$E = 1 - p \tag{1}$$

where

p is the value that specifies a conic section such as a circle, ellipse, hyperbola, or parabola

value p is given by Formula (2):

$$p = \pm \frac{a^2}{b^2} \tag{2}$$

iTeh STANDARD PREVIEW
(standards.iteh.ai)

where

a and b are the semi-diameters of the axes of the conic section;
+ indicates a circle or ellipse;
- indicates a hyperbola

a conic section is specified by Formula (3):

$$\frac{z^2}{b^2} \pm \frac{x^2}{a^2} = 1 \tag{3}$$

value E also is the square of the *eccentricity* (3.9) of the conic section, given by Formula (4):

$$E = e^2 \tag{4}$$

Note 1 to entry: Unless otherwise specified, E refers to the meridian with least curvature (flattest meridian). See Table 1 and Annex A.

Note 2 to entry: Although the magnitude of E is equal to the square of the eccentricity and so is always positive, the sign of E is a convention to signify whether an ellipse takes a prolate or oblate orientation.

Note 3 to entry: The negative value of E is defined by ISO 10110-12 as the conic constant designated by the symbol K . The negative value of E has also been called asphericity and given the symbol Q .

Table 1 — Conic section descriptors

Conic section	Value of p	Value of E	Value of e
Hyperbola	$p < 0$	$E > 1$	$e > 1$
^a Eccentricity, e , does not distinguish between prolate and oblate orientations of an ellipse (see 3.9 and Annex A).			

Table 1 (continued)

Conic section	Value of p	Value of E	Value of e
Parabola	0,0	1,0	1,0
Prolate ellipse	$0 < p < 1$	$1 > E > 0$	$1 > e > 0^a$
Circle	1,0	0,0	0,0
Oblate ellipse	$p > 1$	$E < 0$	$0 < e < 1^b$

^a Eccentricity, e , does not distinguish between prolate and oblate orientations of an ellipse (see 3.9 and Annex A).

3.5 corneal topographer

CT

instrument or system that measures the shape of corneal surface in a non-contact manner

Note 1 to entry: A corneal topographer that uses a video camera system and video image processing to measure the corneal surface by analysing the reflected image created by the corneal surface of a luminous target is also referred to as a videokeratograph.

3.5.1 optical-sectioning corneal topographer

corneal topographer (3.5) that measures the corneal surface by analysing multiple optical sections of that surface

3.5.2 Placido ring corneal topographer

corneal topographer (3.5) that measures the corneal surface by analysing the reflected image of a Placido ring target created by the corneal surface

3.5.3 reflection-based corneal topographer

corneal topographer (3.5) that measures the corneal surface using light reflected from the air/pre-corneal tear film interface

3.5.4 luminous surface corneal topographer

corneal topographer (3.5) that measures the corneal surface using light back-scattered from a target projected onto the pre-corneal tear film or the corneal anterior tissue surface

Note 1 to entry: Back-scattering is usually introduced in these optically clear substances by the addition of a fluorescent material into the pre-corneal tear film. A target may include a slit or scanning slit of light or another projecting pattern of light. Other methods are possible.

3.6 corneal topographer axis


CT axis

line parallel to the optical axis of the instrument and often coincident with it, that serves as one of the coordinate axes used to describe and define the corneal shape

3.7 corneal vertex

point of tangency of a plane perpendicular to the *corneal topographer* (3.5) axis with the corneal surface

Note 1 to entry: See Figure 1.



Key

- 1 corneal vertex
- 2 apex
- 3 radius of curvature at the apex
- 4 centre of meridional curvature point
- 5 cross-section of the corneal surface
- 6 plane perpendicular to the CT axis
- 7 CT axis

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 19980
<https://standards.iteh.ai/catalog/standards/sist/6d508bdb-229d-47fd-bbc3-df0c6fc8b3d4/iso-fdis-19980>

Figure 1 — Illustration of the corneal vertex and the apex

3.8 Curvature

3.8.1 Axial curvature

Note 1 to entry Axial curvature is expressed in reciprocal millimetres.

3.8.1.1 axial curvature sagittal curvature

K_a
<calculated using the axial radius of curvature> reciprocal of the distance from a point on a surface to the *corneal topographer* (3.5) axis along the *corneal meridian* (3.3) normal at the point and given by [Formula \(5\)](#):

$$K_a = \frac{1}{r_a} \tag{5}$$

where r_a is the axial radius of curvature

Note 1 to entry: See [Figure 2](#).

3.8.1.2 axial curvature

K_a
<calculated using the meridional curvature> average of the value of the tangential curvature from the corneal vertex to the meridional point and given by [Formula \(6\)](#):


$$K_a = \frac{\int_0^{x_p} K_m(x) dx}{x_p} \quad (6)$$

where

x is the radial position variable on the meridian;

x_p is the radial position at which K_a is evaluated;

K_m is the meridional curvature.



Key

- 1 normal to meridian at point P
- 2 P, a point on the meridian where curvature is to be found
- 3 centre of meridional curvature point
- 4 intersection normal — CT axis
- 5 meridian (a cross-section of the corneal surface)
- 6 CT axis

Figure 2 — Illustration of axial curvature, K_a , axial radius of curvature, r_a , meridional curvature, K_m , and meridional radius of curvature, r_m

3.8.2 Gaussian curvature

product of the two principal normal curvature values at a surface location

Note 1 to entry: Gaussian curvature is expressed in reciprocal square millimetres.

3.8.3 meridional curvature tangential curvature

K_m
local surface curvature measured in the meridional plane and defined by [Formula \(7\)](#):

$$K_m = \frac{\partial^2 M(x) / \partial x^2}{\{1 + [\partial M(x) / \partial x]^2\}^{3/2}} \quad (7)$$

where $M(x)$ is a function giving the elevation of the meridian at any perpendicular distance, x , from the *corneal topographer* (3.5) axis

Note 1 to entry: The meridional plane includes the surface point and the chosen axis. The meridional normal is a line passing through the surface point perpendicular to the tangent to the meridional curve at that point and lying in the meridional plane.

Note 2 to entry: Meridional curvature is in general not a normal curvature. It is the curvature of the corneal meridian at a point on a surface.

Note 3 to entry: See [Figure 2](#).

3.8.4 normal curvature

curvature at a point on the surface of the curve created by the intersection of the surface with any plane containing the normal to the surface at that point

3.8.4.1 mean curvature

arithmetic average of the principal curvatures at a point on the surface

3.8.4.2 principal curvature

maximum or minimum curvature at a point on the surface

ITEH STANDARD PREVIEW
(standards.iteh.ai)

ISO/FDIS 19980
<https://standards.iteh.ai/catalog/standards/sist/6d508bdb-229d-47fd-bbc3-df0c6fe8b3d4/iso-fdis-19980>

3.9 eccentricity

e
value descriptive of a conic section and the rate of curvature change away from the apex of the curve, i.e. how quickly the curvature flattens or steepens away from the apex of the surface

Note 1 to entry: Eccentricity ranges from zero to positive infinity for the group of conic sections. In order to signify use of an oblate ellipse, e is sometimes given a negative sign that is not used in computations. Otherwise, use of the prolate ellipse is assumed. See [Table 1](#).

3.10 elevation

distance between a corneal surface and a defined reference surface, measured in a defined direction from a specified position

3.10.1 axial elevation

elevation as measured from a selected point on the corneal surface in a direction parallel to the *corneal topographer* (3.5) axis

3.10.2 normal elevation

elevation as measured from a selected point on the corneal surface in a direction along the normal to the corneal surface at that point

3.10.3**reference normal elevation**

elevation as measured from a selected point on the corneal surface in a direction along the normal to the reference surface

3.11**corneal power**

refracting power of the cornea based on the radius of curvature of the front surface
Note 1 to entry: Corneal power is expressed in *keratometric dioptres* (3.12).

3.12**keratometric dioptres**

unit of corneal power based on the radius of curvature of the front surface and the keratometric constant, 337,5, using [Formula \(8\)](#):

$$\text{keratometric dioptres} = 337,5/\text{radius of curvature} \quad (8)$$

Note 1 to entry: The radius of curvature is expressed in millimetres.

3.13**surface normal**

line passing through a surface point of the surface perpendicular to the plane tangent to the surface at that point

3.14**ellipsoid of revolution**

surface of revolution that results when the generating arc is non-circular

3.15**radius of curvature**

reciprocal of the curvature

[ISO/FDIS 19980](https://standards.iteh.ai/catalog/standards/sist/6d508bdb-229d-47fd-bbc3-df0c6fe8b3d4/iso-fdis-19980)

<https://standards.iteh.ai/catalog/standards/sist/6d508bdb-229d-47fd-bbc3-df0c6fe8b3d4/iso-fdis-19980>

Note 1 to entry: The radius of curvature is expressed in millimetres.

3.15.1**axial radius of curvature****sagittal radius of curvature**

r_a

distance from a surface point, P, to the axis along the normal to corneal meridian at that point, and defined by [Formula \(9\)](#):

$$r_a = \frac{x}{\sin \varphi(x)} \quad (9)$$

where

x is the perpendicular distance from the axis to the meridian point, in millimetres;

$\varphi(x)$ is the angle between the axis and the meridian normal at point x .

Note 1 to entry: See [Figure 2](#).

3.15.2**meridional radius of curvature****tangential radius of curvature**

r_m