
Medical electrical equipment —

Part 2-70:

**Particular requirements for the basic
safety and essential performance
of sleep apnoea breathing therapy
equipment**

Appareils électromédicaux —

*Partie 2-70: Exigences particulières pour la sécurité de base et les
performances essentielles de l'équipement de thérapie respiratoire
pour l'apnée du sommeil*

ISO 80601-2-70:2020

<https://standards.iteh.ai/catalog/standards/iso/79be123f-c3ec-475f-87e8-19ef223de22e/iso-80601-2-70-2020>



Reference number
ISO 80601-2-70:2020(E)

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 80601-2-70:2020

<https://standards.iteh.ai/catalog/standards/iso/79be123f-c3ec-475f-87e8-19ef223de22e/iso-80601-2-70-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	vi
Introduction.....	vii
201.1 * Scope, object and related standards.....	1
201.1.1 Scope.....	1
201.1.2 Object.....	2
201.1.3 Collateral standards.....	2
201.1.4 Particular standards.....	2
201.2 Normative references.....	3
201.3 Terms and definitions.....	4
201.4 General requirements.....	7
201.4.3 <i>Essential performance</i>	7
201.4.3.101 * Additional requirements for <i>essential performance</i>	7
201.4.6 * <i>ME equipment</i> or <i>ME system</i> parts that contact the <i>patient</i>	7
201.5 General requirements for testing of <i>ME equipment</i>.....	8
201.5.101 Additional requirements for general requirements for testing of <i>ME equipment</i>	8
201.5.101.1 Gas flowrate and pressure specifications.....	8
201.5.101.2 * <i>Sleep apnoea breathing therapy equipment</i> testing errors.....	8
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>.....	8
201.7 <i>ME equipment</i> identification, marking and documents.....	8
201.7.1.2 * Legibility of markings.....	8
201.7.2.4.101 Additional requirements for <i>accessories</i>	9
201.7.2.13.101 Additional requirements for physiological effects.....	9
201.7.2.17.101 * Additional requirements for protective packaging.....	9
201.7.2.101 Additional requirements for marking on the outside of <i>ME equipment</i> or <i>ME equipment</i> parts.....	10
201.7.4.3 Units of measurement.....	10
201.7.9.1 * Additional general requirements.....	10
201.7.9.2 Instructions for use.....	11
201.7.9.2.1.101 Additional general requirements.....	11
201.7.9.2.2.101 Additional requirements for warnings and safety notices.....	11
201.7.9.2.5.101 Additional requirements for <i>ME equipment</i> description.....	12
201.7.9.2.9.101 Additional requirements for operating instructions.....	12
201.7.9.2.12 <i>Cleaning, disinfection, and sterilization</i>	12
201.7.9.2.14.101 Additional requirements for <i>accessories</i> , supplementary equipment, used material.....	13
201.7.9.3.1.101 * Additional general requirements.....	13
201.8 Protection against electrical <i>hazards</i> from <i>ME equipment</i>.....	14
201.9 Protection against <i>mechanical hazards</i> of <i>ME equipment</i> and <i>ME systems</i>.....	14
201.9.6.2.1.101 * Additional requirements for audible acoustic energy.....	14
201.10 Protection against unwanted and excessive radiation <i>hazards</i>.....	16
201.11 Protection against excessive temperatures and other <i>hazards</i>.....	16
201.11.1.2.2 <i>Applied parts</i> not intended to supply heat to a <i>patient</i>	16

201.11.6.6 * <i>Cleaning and disinfection of ME equipment or ME system</i>	17
201.11.7 <i>Biocompatibility of ME equipment and ME systems</i>	17
201.11.8 Additional requirements for interruption of the power supply/supply mains to ME equipment	18
201.12 Accuracy of controls and instruments and protection against hazardous outputs	18
201.12.1 * Accuracy of controls and instruments	18
201.12.1.101 Stability of static <i>airway pressure accuracy</i> (long-term accuracy)	19
201.12.1.102 Stability of dynamic <i>airway pressure accuracy</i> (short-term accuracy)	20
201.12.1.102.1 <i>CPAP mode</i>	20
201.12.1.102.2 <i>Bi-level positive airway pressure mode, pressure stability</i>	22
201.12.1.103 * Maximum flowrate	24
201.12.4 Protection against hazardous output.....	25
201.12.4.101 Measurement of <i>airway pressure</i>	25
201.12.4.102 * <i>Maximum limited pressure protection device</i>	25
201.12.4.103 * <i>CO₂ rebreathing</i>	26
201.13 Hazardous situations and fault conditions	26
201.14 Programmable electrical medical systems (PEMS)	26
201.15 Construction of ME equipment	26
201.15.101 Mode of operation.....	26
201.16 ME systems	27
201.17 Electromagnetic compatibility of ME equipment and ME systems	27
201.101 Breathing gas pathway connectors	27
201.101.1 General	27
201.101.2 Other named ports	27
201.101.2.1 <i>Patient-connection port</i>	27
201.101.2.2 <i>Gas output port</i>	27
201.101.2.3 <i>Flow-direction-sensitive components</i>	28
201.101.2.4 Ancillary port.....	28
201.101.2.5 Monitoring probe port.....	28
201.101.2.6 Oxygen inlet port	28
201.102 Requirements for the breathing gas pathway and accessories	28
201.102.1 * General	28
201.102.2 Labelling	29
201.102.3 Humidification.....	29
201.102.4 <i>Breathing system filter (BSF)</i>	29
201.103 Functional connection	29
201.103.1 General	29
201.103.2 * <i>Functional connection</i> to support remote supervision.....	30
201.104 Training	30
202 Electromagnetic disturbances — Requirements and tests	30
202.4.3.1 Configurations.....	30
202.5.2.2.1 Requirements applicable to all <i>ME equipment and ME systems</i>	30
202.8.1.101 Additional general requirements.....	30
206 Usability	31
211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	31

211.4.2.3.1 Continuous operating conditions	31
Annex C (informative) Guide to marking and labelling requirements for <i>ME equipment</i> and <i>ME systems</i>.....	33
Annex D (informative) Symbols on marking	38
Annex AA (informative) Particular guidance and rationale	39
Annex BB (informative) Data interface requirements	48
Annex CC (informative) Reference to the IMDRF <i>essential principles</i> and labelling guidances ..	52
Annex DD (informative) Reference to the <i>essential principles</i>.....	56
Annex EE (informative) Reference to the general safety and performance requirements	59
Annex FF (informative) Terminology — alphabetized index of defined terms	63
Bibliography	66

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 80601-2-70:2020](https://standards.iteh.ai/catalog/standards/iso/79be123f-c3ec-475f-87e8-19ef223de22e/iso-80601-2-70-2020)

<https://standards.iteh.ai/catalog/standards/iso/79be123f-c3ec-475f-87e8-19ef223de22e/iso-80601-2-70-2020>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electromedical equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 80601-2-70:2015), which has been technically revised.

The main changes compared to the previous edition are as follows:

- modification of the bi-level positive airway pressure mode stability test method;
- modification of the *biocompatibility* requirements;
- reformatting to provide a unique identifier for each requirement;
- harmonization with the 'A2 project' of the general standard.

A list of all parts in the ISO 80601 series and the IEC 80601 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Sleep apnoea is a chronic medical condition where the *patient* repeatedly stops breathing during sleep. These episodes typically last 10 s or more and cause the oxygen levels in the blood to drop. It can be caused by obstruction of the upper airway (obstructive sleep apnoea or OSA) or by a failure of the brain to initiate a breath (central sleep apnoea).

NOTE *Sleep apnoea breathing therapy equipment* is intended for the treatment of obstructive sleep apnoea and not central sleep apnoea.

Sleep apnoea, if untreated, can cause and worsen other medical conditions, including hypertension, heart failure and diabetes^[22].

Hypopnoea refers to a transient reduction of airflow, often while the *patient* is asleep, that lasts for at least 10 s, shallow breathing. It also results in arousal or can cause oxygen saturation to drop. Hypopnoea is less severe than apnoea. It is commonly due to partial obstruction of the upper airway^[20].

Awareness of the *risks* associated with obstructive sleep apnoea has grown significantly. As a result, the use of *sleep apnoea breathing therapy equipment* to treat obstructive sleep apnoea has become common.

This document covers *basic safety* and *essential performance* requirements needed to protect *patients* in the use of this *ME equipment*.

This document covers *sleep apnoea breathing therapy equipment* for *patient* use. ISO 17510 applies to *masks* and *accessories* used to connect *sleep apnoea breathing therapy equipment* to the *patient*. Figure AA.1 shows this diagrammatically.

In this document, the following print types are used:

- Requirements and definitions: roman type
- *Test specifications and terms defined in clause 3 of the general standard, in this document or as noted: italic type;*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;

In referring to the structure of this document, the term.

- “clause” means one of the four numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.1, 201.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 201.101, 201.102 and 201.102.1 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

ISO 80601-2-70:2020(E)

- “shall” means that conformance with a requirement or a test is mandatory for conformance with this document;
- “should” means that conformance with a requirement or a test is recommended but is not mandatory for conformance with this document;
- “may” is used to describe a permission (e.g. a permissible way to achieve conformance with a requirement or test);
- “can” is used to describe a possibility or capability; and
- “must” is used to express an external constraint.

Annex C contains a guide to the marking and labelling requirements in this document.

Annex D contains a summary of the symbols referenced in this document.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[ISO 80601-2-70:2020](https://standards.iteh.ai/catalog/standards/iso/79be123f-c3ec-475f-87e8-19ef223de22e/iso-80601-2-70-2020)

<https://standards.iteh.ai/catalog/standards/iso/79be123f-c3ec-475f-87e8-19ef223de22e/iso-80601-2-70-2020>

Medical electrical equipment —

Part 2-70:

Particular requirements for the basic safety and essential performance of sleep apnoea breathing therapy equipment

201.1 * Scope, object and related standards

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 1 applies, except as follows:

NOTE The general standard is IEC 60601-1:2005+AMD1:2012+AMD2:2020.

201.1.1 Scope

IEC 60601-1:2005+Amendment 1:2012, 1.1 is replaced by:

This document is applicable to the *basic safety* and *essential performance* of *sleep apnoea breathing therapy equipment*, hereafter referred to as *ME equipment*, intended to alleviate the symptoms of *patients* who suffer from obstructive sleep apnoea by delivering a therapeutic breathing pressure to the respiratory tract of the *patient*. *Sleep apnoea breathing therapy equipment* is intended for use in the *home healthcare environment* by *lay operators* as well as in professional healthcare institutions.

* *Sleep apnoea breathing therapy equipment* is not considered to utilize a *physiologic closed-loop-control system* unless it uses a physiological *patient* variable to adjust the therapy settings.

This document excludes *sleep apnoea breathing therapy equipment* intended for use with neonates.

This document is applicable to *ME equipment* or an *ME system* intended for those *patients* who are not dependent on mechanical ventilation.

This document is not applicable to *ME equipment* or an *ME system* intended for those *patients* who are dependent on mechanical ventilation such as *patients* with central sleep apnoea.

This document is also applicable to those *accessories* intended by their *manufacturer* to be connected to *sleep apnoea breathing therapy equipment*, where the characteristics of those *accessories* can affect the *basic safety* or *essential performance* of the *sleep apnoea breathing therapy equipment*.

Masks and application *accessories* intended for use during sleep apnoea breathing therapy are additionally addressed by ISO 17510. Refer to Figure AA.1 for items covered further under this document.

If a clause or subclause is specifically intended to be applicable to *ME equipment* only, or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.

Hazards inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This document is not applicable to high-frequency jet ventilators (HFJVs) or high-frequency oscillatory ventilators (HFOVs), which are given in ISO 80601-2-87^[13].

This document does not specify the requirements for ventilators or *accessories* intended for critical care ventilators for ventilator-dependent *patients*, which are given in ISO 80601-2-12.

This document does not specify the requirements for ventilators or *accessories* intended for anaesthetic applications, which are given in ISO 80601-2-13^[8].

This document does not specify the requirements for ventilators or *accessories* intended for home care ventilators for ventilator-dependent *patients*, which are given in ISO 80601-2-72^[9].

This document does not specify the requirements for ventilators or *accessories* intended for emergency and transport, which are given in ISO 80601-2-84^[12].

This document does not specify the requirements for ventilators or *accessories* intended for home-care ventilatory support, which are given in ISO 80601-2-79^[10] and ISO 80601-2-80^[11].

201.1.2 Object

IEC 60601-1:2005, 1.2 is replaced by:

The object of this document is to establish particular *basic safety* and *essential performance* requirements for *sleep apnoea breathing therapy equipment* (as defined in 201.3.215).

NOTE 1 This document has been prepared to address the relevant *essential principles*^[17] and *labelling*^[18] guidances of the International Medical Devices Regulators Forum (IMDRF) as indicated in Annex CC.

NOTE 2 This document has been prepared to address the relevant *essential principles of safety and performance* of ISO 16142-1:2016 as indicated in Annex DD.

NOTE 3 This document has been prepared to address the relevant general safety and performance requirements of European regulation (EU) 2017/745^[16] as indicated in Annex EE.

201.1.3 Collateral standards

IEC 60601-1:2005+AMD1:2012+AMD2:2020, 1.3 applies with the following addition:

IEC 60601-1-2:2014+AMD1:2020 and IEC 60601-1-6:2010+AMD1:2013+AMD2:2020 apply as modified in Clauses 202 and 206 respectively. IEC 60601-1-3:2008+AMD1:2013 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards define *basic safety* and *essential performance* requirements, and may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular *ME equipment* under consideration.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1+AMD1:2012+AMD2:2020 is referred to in this document as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this document corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x", where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this document.

"Addition" means that the text of this document is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 211 for IEC 60601-1-11, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this document taken together.

Where there is no corresponding clause or subclause in this document, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005+AMD1:2012+AMD2:2020, Clause 2 applies, except as follows:

Replacement:

ISO 7010:2019, *Graphical symbols — Safety colours and safety signs — Registered safety signs*

Addition:

ISO 3744:2010, *Acoustics — Determination of sound power levels and sound energy levels of noise sources using sound pressure — Engineering methods for an essentially free field over a reflecting plane*

ISO 5356-1:2015, *Anaesthetic and respiratory equipment — Conical connectors — Cones and sockets*

ISO 14937:2009, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ISO 16142-1:2016, *Medical devices — Recognized essential principles of safety and performance of medical devices — Part 1: General essential principles and additional specific essential principles for all non-IVD medical devices and guidance on the selection of standards*

ISO 17510:2015, *Medical devices — Sleep apnoea breathing therapy — Masks and application accessories*

ISO 17664:2017, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices*

ISO 18562-1:2017, *Biocompatibility evaluation of breathing gas pathways in healthcare applications — Part 1: Evaluation and testing within a risk management process*

ISO 19223:2019, *Lung ventilators and related equipment — Vocabulary and semantics*

ISO 23328-1:2003, *Breathing system filters for anaesthetic and respiratory use — Part 1: Salt test method to assess filtration performance*

ISO 23328-2:2002, *Breathing system filters for anaesthetic and respiratory use — Part 2: Non-filtration aspects*

ISO 80369-1:2018, *Small-bore connectors for liquids and gases in healthcare applications — Part 1: General requirements*

ISO 80601-2-12:2020, *Medical electrical equipment — Part 2-12: Particular requirements for basic safety and essential performance of critical care ventilators*

ISO 80601-2-74:2017, *Medical electrical equipment — Part 2-74: Particular requirements for basic safety and essential performance of respiratory humidifying equipment*

IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

IEC 62366-1:2015+AMD1:2020, *Medical devices — Application of usability engineering to medical devices*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 3744:2010, ISO 16142-1:2016, ISO 17510:2015, ISO 17664:2017, ISO 18562-1:2017, ISO 19223:2019, ISO 23328-2:2002, ISO 80601-2-12:2020, ISO 80601-2-74:2017,

IEC 60601-1:2005+AMD1:2012+AMD2:2020, IEC 60601-1-2:2014+AMD1:2020,
 IEC 60601-1-6:2010+AMD1:2013+AMD2:2020, IEC 60601-1-8:2006+AMD1:2012+AMD2:2020,
 IEC 60601-1-10:2007+AMD2:2020, IEC 60601-1-11:2015, IEC 62366-1:2015+AMD1:2020 and the
 following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

NOTE An index of defined terms is found in Annex FF.

Addition:

201.3.201

airway pressure accuracy

degree of correspondence between the pressure set on the *sleep apnoea breathing therapy equipment* and the actual (true) *airway pressure*

201.3.202

*** auto CPAP**

operating mode in which the *sleep apnoea breathing therapy equipment* automatically adjusts the applied gas pressure level within a clinically predetermined range to prevent disturbances of the *patient's* breathing at the lowest necessary pressure

Note 1 to entry: These disturbances are typically obstructive apnoea (cessation of air flow) or hypopnoea (shallow breathing) events.

Note 2 to entry: An *auto CPAP* mode may be bi-level.

201.3.203

*** automatic start/stop function**

feature by which the *sleep apnoea breathing therapy equipment* automatically starts or stops the therapy based on whether the *patient* is connected or disconnected from the equipment

Note 1 to entry: *Sleep apnoea breathing therapy equipment* may provide an *automatic start* function, an *automatic stop* function, or both.

201.3.204

breathing gas pathway

gas pathways through which gas flows at respiratory pressures between the *intake* and the *patient-connection port*

201.3.205

breathing tube

non-rigid tube used to convey gases between components of a breathing system

201.3.206

diagnostic apnoea

A_d

decrease in *airway* flow greater than ≥ 90 % of pre-event baseline for at least 10 s

Note 1 to entry: Airflow is typically measured by oronasal thermal airflow sensor or nasal pressure transducer normally determined using full polysomnography in an attended sleep study or home sleep test.

Note 2 to entry: Reference [19] was used to determine this definition.

201.3.207

diagnostic apnoea-hypopnoea index

AHI_d

average number of *diagnostic apnoea* events and *diagnostic hypopnoea* events occurring per hour of sleep

$$AHI_d = \frac{(A_d + H_d) \cdot 60}{t_s}$$

where

A_d is the total number of *diagnostic apnoea* events

H_d is the total number of *diagnostic hypopnoea* events

t_s is the total sleep time in minutes

Note 1 to entry: A *diagnostic apnoea-hypopnoea index* is normally determined using full polysomnography in an attended sleep study or home sleep test.

201.3.208

diagnostic hypopnoea

H_d

decrease in *airway* flow of at least 30 % of the pre-event baseline for at least 10 s; and at least 3 % oxygen desaturation from pre-event baseline or an arousal event, excluding *diagnostic apnoea*

Note 1 to entry: For monitoring oxygen saturation, a pulse oximeter should have a maximum signal averaging time of ≤ 3 s.

Note 2 to entry: An oxygen desaturation of 4 % from pre-event baseline is also used.

Note 3 to entry: Airflow is typically measured by oronasal thermal airflow sensor or nasal pressure transducer normally determined using full polysomnography in an attended sleep study or home sleep test.

Note 4 to entry: Reference [19] was used to determine this definition.

201.3.209

equipment apnoea

A_{flow}

diagnostic apnoea as estimated by *sleep apnoea breathing therapy equipment*

201.3.210

equipment apnoea-hypopnoea index

AHI_{flow}

average number of residual *equipment apnoea* and *equipment hypopnoea* events occurring per hour of the therapy session

Note 1 to entry: References [19] and [23] were used to determine this definition.