



SLOVENSKI STANDARD
oSIST prEN IEC 62372:2022

01-november-2022

Jedrska merilna oprema - Scintilatorji v ohišju - Metode preskušanja svetlobnega toka in svojstvene ločljivosti

Nuclear instrumentation - Housed scintillators - Test methods of light output and intrinsic resolution

Strahlungsmessgeräte - Ummantelte Szintillatoren - Prüfverfahren für Lichtabgabe und intrinsische Auflösung

Instrumentation nucléaire - Scintillateurs montés - Méthodes de mesure de la lumière sortante et de résolution intrinsèque

Ta slovenski standard je istoveten z: prEN IEC 62372

ICS:

27.120.01 Jedrska energija na splošno Nuclear energy in general

oSIST prEN IEC 62372:2022

en

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

DRAFT
prEN IEC 62372

August 2022

ICS 27.120.01

English Version

**Nuclear instrumentation - Housed scintillators - Test methods of
light output and intrinsic resolution
(IEC 62372:2021)**

Instrumentation nucléaire - Scintillateurs montés -
Méthodes de mesure de la lumière sortante et de résolution
intrinsèque
(IEC 62372:2021)

Strahlungsmessgeräte - Ummantelte Szintillatoren -
Prüfverfahren für Lichtabgabe und intrinsische Auflösung
(IEC 62372:2021)

This draft European Standard is submitted to CENELEC members for enquiry.
Deadline for CENELEC: 2022-11-18.

The text of this draft consists of the text of IEC 62372:2021.

If this draft becomes a European Standard, CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CENELEC in three official versions (English, French, German).
A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

prEN IEC 62372:2022 (E)

European foreword

This document (prEN IEC 62372:2022) consists of the text of IEC 62372:2021 prepared by IEC/SC 45B "Radiation protection instrumentation" of IEC/TC 45 "Nuclear instrumentation".

This document is currently submitted to the Enquiry.

The following dates are proposed:

- latest date by which the existence of this document (doa) dor + 6 months has to be announced at national level
- latest date by which this document has to be (dop) dor + 12 months implemented at national level by publication of an identical national standard or by endorsement
- latest date by which the national standards (dow) dor + 36 months conflicting with this document have to be withdrawn (to be confirmed or modified when voting)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN IEC 62372:2022
https://standards.iteh.ai/catalog/standards/sist/c5177b7b-ff26-4383-bd37-
b7ecfe7e0741/osist-pren-iec-62372-2022](https://standards.iteh.ai/catalog/standards/sist/c5177b7b-ff26-4383-bd37-b7ecfe7e0741/osist-pren-iec-62372-2022)



IEC 62372

Edition 2.0 2021-02

INTERNATIONAL STANDARD

**Nuclear instrumentation – Housed scintillators – Test methods of light output
and intrinsic resolution**

(standards.iteh.ai)

[oSIST prEN IEC 62372:2022](https://standards.iteh.ai/catalog/standards/sist/c5177b7b-ff26-4383-bd37-b7ecfe7e0741/osist-pren-iec-62372-2022)

[https://standards.iteh.ai/catalog/standards/sist/c5177b7b-ff26-4383-bd37-
b7ecfe7e0741/osist-pren-iec-62372-2022](https://standards.iteh.ai/catalog/standards/sist/c5177b7b-ff26-4383-bd37-b7ecfe7e0741/osist-pren-iec-62372-2022)

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 27.120.01

ISBN 978-2-8322-9322-5

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	3
1 Scope.....	5
2 Normative references	5
3 Terms, definitions, symbols and abbreviated terms.....	5
3.1 Terms and definitions.....	5
3.2 Symbols and abbreviated terms	7
4 Test methods of basic parameters of housed scintillators	9
4.1 General.....	9
4.1.1 Test conditions	9
4.1.2 The sources of ionizing radiation	9
4.1.3 The assembly test conditions.....	10
4.2 Test methods of nonlinearity and instability of the assembly	10
4.2.1 Nonlinearity measurement	10
4.2.2 Instability measurement.....	12
4.3 Test methods of the intrinsic resolution, light output of the housed scintillator and PMT spectrometric constant.....	13
4.3.1 Equipment	13
4.3.2 Measurements.....	13
4.3.3 Processing of results	14
4.4 Test methods of the light output.....	15
4.4.1 General	15
4.4.2 Measurements.....	15
4.4.3 Processing of results	16
4.5 Test methods of the intrinsic resolution	16
4.5.1 Determination of PMT spectrometric constant.....	16
4.5.2 Test method of the intrinsic resolution for housed scintillator	17
Bibliography.....	19
Table 1 – Source of ionizing radiation	9

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**NUCLEAR INSTRUMENTATION – HOUSED SCINTILLATORS –
TEST METHODS OF LIGHT OUTPUT AND INTRINSIC RESOLUTION**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62372 has been prepared by IEC technical committee 45: Nuclear instrumentation.

This second edition cancels and replaces the first edition published in 2006. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- Title has been modified.
- To review the existing requirements and to update the terminology, definitions and normative references.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
45/913/FDIS	45/915/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN IEC 62372:2022](https://standards.iteh.ai/catalog/standards/sist/c5177b7b-ff26-4383-bd37-b7ecfe7e0741/osist-pren-iec-62372-2022)

<https://standards.iteh.ai/catalog/standards/sist/c5177b7b-ff26-4383-bd37-b7ecfe7e0741/osist-pren-iec-62372-2022>

NUCLEAR INSTRUMENTATION – HOUSED SCINTILLATORS – TEST METHODS OF LIGHT OUTPUT AND INTRINSIC RESOLUTION

1 Scope

This document is applicable to housed scintillators for registration and spectrometry of alpha-, beta-, gamma-, X-ray and neutron radiation.

The main parameters, such as a light output and intrinsic resolution are established. This document specifies the requirements for the testing equipment and test methods of the basic parameters of housed scintillators, such as:

- the direct method is applicable to measure the light output of housed scintillators based on scintillation material. The housed scintillators certified by this method can be used as working standard of housed scintillators (hereinafter: working standard) when performing measurements by a relative method of comparison.
- the relative method of comparison with the working standard is applicable to housed scintillators based on the same scintillation material as the working standard.

This document does not apply to gas or liquid scintillators and scintillators for counting and current modes.

The numerical values of the parameters are set to the specific type of scintillators in the specifications.

2 Normative references

There are no normative references in this document.

3 Terms, definitions, symbols and abbreviated terms

3.1 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1.1

scintillator

luminescent material, usually liquid or solid, showing radioluminescence with a short afterglow

[SOURCE: IEC 60050-845:1987, 845-04-37]

3.1.2

housed scintillator

scintillator, housed in a container with a reflector and optical window

3.1.3 scintillation detector

radiation detector consisting of a scintillator that is usually optically coupled to a photosensitive device, directly or through light guides

Note 1 to entry: The scintillator consists of a scintillating material in which the ionizing particle produces a burst of luminescence radiation along its path. A common scintillator is NaI(Tl).

[SOURCE: IEC 60050-395:2014, 395-03-01]

3.1.4 assembly

light protective chamber containing a housed scintillator, photomultiplier (PMT), PMT voltage divider

Note 1 to entry: Assembly is used for testing of the housed scintillator.

3.1.5 light yield

η

ratio of scintillation photons summed energy (E_p) to energy (E) lost by ionizing particles in the scintillator

$$\eta = \frac{E_p}{E}$$

Note 1 to entry: Value of η depends on type and energy of ionizing particle.

3.1.6 light output

C

ratio of total energy (L_{ph}) of scintillation photons, which pass through the output window of the housed scintillator of ionizing radiation, to energy (E), lost by ionizing particles in the scintillator

$$C = \frac{L_{ph}}{E}$$

3.1.7 intrinsic resolution of housed scintillator of ionizing radiation

R_d

component, given by housed scintillator of ionizing radiation to energy resolution of the scintillation detector

Note 1 to entry: The intrinsic resolution R_d is defined from the relation:

$$R_d = \sqrt{R_a^2 - R_{pm}^2},$$

where

R_a is the energy resolution of the scintillation detector;

R_{pm} is PMT intrinsic resolution.

3.1.8 total absorption peak

portion of the spectral response curve corresponding to the total absorption of photon energy in a radiation detector

Note 1 to entry: This peak represents the total absorption of photon energy from all interactive processes, namely, a) photoelectric absorption, b) Compton effect, and c) pair production.

[SOURCE: IEC 60050-395:2014, 395-03-94]

3.1.9 photomultiplier tube spectrometric constant

A
parameter, characterizing properties of the photomultiplier tube

Note 1 to entry: Defined by the following formula:

$$A = (R_a^2 - R_d^2) \times C_{ph}$$

where C_{ph} is light output, photons/MeV.

3.1.10 working standard

working standard of housed scintillator that is used to check the measuring system and to measure the light output by a method of comparison

3.1.11 full width at half maximum FWHM

in a distribution curve comprising a single peak, the distance between the abscissa of two points on the curve whose ordinates are half of the maximum ordinate of the peak

Note 1 to entry: If the curve considered comprises several peaks, a full width at half maximum exists for each peak.

3.1.12 expanded uncertainty

expanded uncertainty quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

Note 1 to entry: The fraction may be viewed as the coverage probability or level of confidence of the interval.

Note 2 to entry: To associate a specific level of confidence with the interval defined by the expanded uncertainty requires explicit or implicit assumptions regarding the probability distribution characterized by the measurement result and its combined standard uncertainty. The level of confidence that may be attributed to this interval can be known only to the extent to which such assumptions may be justified.

[SOURCE: JCGM 100:2008, 2.35]

3.1.13 relative expanded uncertainty

ratio of the expanded uncertainty of a measurement to average value of quantity. It expresses the relative size of the uncertainty of a measurement (its precision)

3.2 Symbols and abbreviated terms

A	the photomultiplier tube spectrometric constant;
a	the assembly conversion factor with the housed scintillator;
a_i	the value of conversion response, measured at energy value of E_i ;
a_{\max}	the maximal value of a ;
a_{\min}	the minimal value of a ;
Δa	the value of nonlinearity;
C	the light output, in relative units;