TECHNICAL REPORT

ISO/TR 22930-2

First edition 2020-05

Evaluating the performance of continuous air monitors —

Part 2:

Air monitors based on flow-through sampling techniques without accumulation

Évaluation de la performance des dispositifs de surveillance de l'air en continu —

Partie 2: Moniteurs d'air basés sur des techniques d'échantillonnage par circulation sans accumulation

ISO/TR 22930-2:2020

https://standards.1teh.ai/catalog/standards/1so/e/722e2c2-14f2-4be2-b21c-cce/b0960009/1so-tr-22930-2-2020



iTeh Standards (https://standards.iteh.ai) Document Preview

ISO/TR 22930-2:2020

https://standards.iteh.ai/catalog/standards/iso/e722e2c2-f4f2-4be2-b21c-cce7b0960009/iso-tr-22930-2-2020



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Fax: +41 22 749 09 47 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Co	ntents	Page
For	eword	iv
Intr	oduction	v
1	Scope	1
2	Normative references	
_		
3	Terms and definitions	
4	Symbols	4
5	Measuring principle	6
6	Study of dynamic behaviour	7
7	Evaluation of the characteristic limits	13
	7.1 General	
	7.2 Single detector	
	7.2.1 General	
	7.2.2 Definition of the model	
	7.2.3 Standard uncertainty	
	7.2.4 Decision threshold	
	7.2.6 Limits of the coverage interval	
	7.3 Double detector	
	7.3.1 General Statistical Technology	
	7.3.2 Definition of the model	18
	7.3.3 Standard uncertainty	
	7.3.4 Decision threshold	
	7.3.5 Detection limit	20
	7.3.6 Limits of the coverage interval	
8	Alarms setup, minimum detectable concentration and potential missed exposure	
	ex A (informative) Application example: Single detector with a proportional counter	
	ex B (informative) Application example: Double detector in current mode	
Bib	liography	32

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

A list of all the parts in the ISO/TR 22930 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Sampling and monitoring of airborne activity concentration in workplaces are critically important for maintaining worker safety at facilities where dispersible radioactive substances are used.

The first indication of a radioactive substance dispersion event comes, in general, from a continuous air monitor (CAM) and its associated alarm levels. In general, the response of a CAM is delayed in time compared to the actual situation of release.

The knowledge of a few factors is needed to interpret the response of a CAM and to select the appropriate CAM type and its operating parameters.

The role of the radiation protection officer is to select the appropriate CAM, to determine when effective release of radioactive substances occurs, to interpret measurement results and to take corrective action appropriate to the severity of the release.

The objective of ISO/TR 22930 series is to assist radiation protection officer in evaluating the performance of a CAM.

ISO/TR 22930 series describes the factors and operating parameters and how they influence the response of a CAM.

This document deals with monitoring systems based on flow-through sampling techniques without accumulation.

iTeh Standards (https://standards.iteh.ai) Document Preview

ISO/TR 22930-2:2020

https://standards.iteh.ai/catalog/standards/iso/e722e2c2-f4f2-4be2-b21c-cce7b0960009/iso-tr-22930-2-2020

iTeh Standards (https://standards.iteh.ai) Document Preview

ISO/TR 22930-2:2020

https://standards.iteh.ai/catalog/standards/iso/e722e2c2-f4f2-4be2-b21c-cce7b0960009/iso-tr-22930-2-2020

Evaluating the performance of continuous air monitors —

Part 2:

Air monitors based on flow-through sampling techniques without accumulation

1 Scope

The use of a continuous air monitor (CAM) is mainly motivated by the need to be alerted quickly and in the most accurate way possible with an acceptable false alarm rate when significant activity concentration value is exceeded, in order to take appropriate measures to reduce exposure of those involved.

The performance of this CAM does not only depend on the metrological aspect characterized by the decision threshold, the limit of detection and the measurement uncertainties but also on its dynamic capacity characterized by its response time as well as on the minimum detectable activity concentration corresponding to an acceptable false alarm rate.

The ideal performance is to have a minimum detectable activity concentration as low as possible associated with a very short response time, but unfortunately these two criteria are in opposition. It is therefore important that the CAM and the choice of the adjustment parameters and the alarm levels be in line with the radiation protection objectives.

This document describes

- **Document Preview**
- the dynamic behaviour and the determination of the response time,
- the determination of the characteristic limits (decision threshold, detection limit, limits of the coverage interval), and mandalds/iso/e722e2c2-[412-4be2-b21c-cce7b0960009/iso-u-22930-2-2020]
- a possible way to determine the minimum detectable activity concentration and the alarms setup.

Finally the annexes of this document show actual examples of CAM data which illustrate how to quantify the CAM performance by determining the response time, the characteristics limits, the minimum detectable activity concentration and the alarms setup.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11929-1, Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation — Fundamentals and application — Part 1: Elementary applications

ISO 16639, Surveillance of the activity concentrations of airborne radioactive substances in the workplace of nuclear facilities

IEC 60761-1, Equipment for continuous monitoring of radioactivity in gaseous effluents — Part 1: General requirements

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11929-1, ISO 16639, IEC 60761-1 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at http://www.electropedia.org/

3.1

annual limit on intake

AI.I

derived limit for the amount of radioactive substance (in Bq) taken into the body of an adult worker by inhalation or ingestion in a year

[SOURCE: ISO 16639:2017, 3.7]

3.2

continuous air monitor

CAM

instrument that continuously monitors the airborne activity concentration on a near real-time basis

[SOURCE: ISO 16639:2017, 3.10]

3.3

decision threshold

value of the estimator of the measurand, which when exceeded by the result of an actual measurement using a given measurement procedure of a measurand quantifying a physical effect, it is decided that the physical effect is present

Note 2 to entry: If the result, y, is below the decision threshold, y^* , it is decided to conclude that the result cannot be attributed to the physical effect; nevertheless, it cannot be concluded that it is absent.

[SOURCE: ISO 11929-1:2019, 3.12]

3.4

derived air concentration

DAC

concentration of a radionuclide in air that, if breathed over the period of a work year, would result in the intake of one ALI for that radionuclide

Note 1 to entry: The DAC is calculated by dividing the ALI by the volume of air breathed by reference man under light-activity work during a working year (in Bq m^{-3}).

Note 2 to entry: The parameter values recommended by the International Commission on Radiological Protection for calculating the DAC are a breathing rate of 1,2 $\text{m}^3 \cdot \text{h}^{-1}$ and a working year of 2 000 h (i.e. 2 400 m^3).

Note 3 to entry: The air concentration can be expressed in terms of a number of DAC. For example, if the DAC for a given radionuclide in a particular form is 0.2 Bq m^{-3} and the observed concentration is 1.0 Bq m^{-3} , then the observed concentration can also be expressed as 5 DAC (i.e. 1.0 divided by 0.2).

Note 4 to entry: The derived air concentration-hour (DAC-hour) is an integrated exposure and is the product of the concentration of a radioactive substance in air (expressed as a fraction or multiple of DAC for each radionuclide) and the time of exposure to that radionuclide, in hours.

[SOURCE: ISO 16639:2017, 3.12]

3.5

detection alarm level

SO

value of time-integrated activity concentration activity concentration corresponding to an acceptable false alarm rate

Note 1 to entry: When S0 increases false alarm rate decreases.

Note 2 to entry: Others values of alarm level higher than S0 can also be set up for operational reasons.

3.6

detection limit

smallest true value of the measurand which ensures a specified probability of being detectable by the measurement procedure

Note 1 to entry: With the decision threshold according to 3.3, the detection limit is the smallest true value of the measurand for which the probability of wrongly deciding that the true value of the measurand is zero is equal to a specified value, β , when, in fact, the true value of the measurand is not zero. The probability of being detectable is consequently $(1-\beta)$.

Note 2 to entry: The terms detection limit and decision threshold are used in an ambiguous way in different standards (e.g. standards related to chemical analysis or quality assurance). If these terms are referred to one has to state according to which standard they are used.

[SOURCE: ISO 11929-1:2019, 3.13]

3.7

limits of the coverage interval

values which define a coverage interval

Note 1 to entry: The limits are calculated in the ISO 11929 series to contain the true value of the measurand with a specified probability $(1-\gamma)$.

Note 2 to entry: The definition of a coverage interval is ambiguous without further stipulations. In this standard two alternatives, namely the probabilistically symmetric and the shortest coverage interval are used.

Note 3 to entry: the coverage interval is defined in ISO 11929-1:2019, 3.4, as the interval containing the set of true quantity values of a measurand with a stated probability, based on the information available.

[SOURCE: ISO 11929-1:2019, 3.16 modified – Note 3 to entry has been added]

3.8

measurand

quantity intended to be measured

[SOURCE: ISO 11929-1:2019, 3.3]

3.9

minimum detectable concentration

time-integrated activity concentration or activity concentration measurements and their associated coverage intervals for a given probability $(1-\gamma)$ corresponding to the first alarm level S0

3.10

model of evaluation

set of mathematical relationships between all measured and other quantities involved in the evaluation of measurements

[SOURCE: ISO 11929-1:2019, 3.11]

3.11

potential missed exposure

PME

time-integrated activity concentration or maximum activity concentration, as applicable, that can acceptably be missed

Note 1 to entry: the value of PME is defined according to ALARA/ALARP principles, and below legal limits.

Note 2 to entry: In order to be alerted when a measurement is likely to exceed the value of PME, an alarm level S1 is set up. The PME is then the upper limit of the coverage interval for a given probability $(1-\gamma)$ of time-integrated activity concentration or activity concentration measurements corresponding to S1.

[SOURCE: ISO 16639:2017, 3.18]

3.12

response time

time required after a step variation in the measured quantity for the output signal variation to reach a given percentage for the first time, usually 90 %, of its final value

Note 1 to entry: The intrinsic response time is related to the measurement principle and its associated model of evaluation of an ideal detector (without taking account of the counting time of the detector).

[SOURCE: IEC 60761-1:2002, 3.15]

4 Symbols

a(t)	Activity going through the detection volume at a time t , in Bq			
c	Activity concentration, in Bq·m ⁻³			
c_{ac}	Actual activity concentration, in Bq·m ⁻³ TeVIEW			
c^*	Decision threshold of the activity concentration, in Bq⋅m ⁻³			
$c^{\text{\#lps://standards.itel}}$ Detection limit of the activity concentration, in Bq·m ⁻³ e7b0960009/iso-tr-22930-2-20				
c^{\triangleleft}	Lower limit of the coverage interval of the activity concentration for a given probability (1– γ), in Bq·m ⁻³			
$c^{ hd}$	Upper limit of the coverage interval of the activity concentration for a given probability (1– γ), in Bq·m ⁻³			
c(t)	Activity concentration measured at a time t , in Bq·m $^{-3}$			
$c_{\rm ac}(t)$	Actual activity concentration measured at a time t , in $\mathrm{Bq}\cdot\mathrm{m}^{-3}$			
$c_{ m g}$	Gross primary measurement of the activity concentration, in $\text{Bq}{\cdot}\text{m}^{-3}$			
c_{\min}	Minimum detectable activity concentration, in $\text{Bq}{\cdot}\text{m}^{-3}$			
c_{\min}^{\triangleleft}	Lower limit of the coverage interval of the minimum detectable activity concentration for a given probability (1– γ), in Bq·m ⁻³			
$c_{\min}^{\triangleright}$	Upper limit of the coverage interval the minimum detectable activity concentration for a given probability (1– γ), in Bq·m ⁻³			
$c_{0,i}$	Activity concentration of the i^{th} measurement of a series of gross measurements (with $i=1,,n$) which represent a background situation, in Bq·m ⁻³			
$\overline{c_0}$	Mean value of $c_{0,i}$, in Bq·m ⁻³			

I_{\min}	Minimum amount of current registered by the measuring detector (with $I_{\min} = \frac{Q_{\min}}{t_{n}}$ in A
$I_{\min, cd}$	Minimum amount of current registered by the compensating detector (with $I_{\min, \mathrm{cd}} = \frac{Q_{\min, \mathrm{cd}}}{t_{\mathrm{C.cd}}}$), in A
$I_{g}(t)$	Instantaneous gross current of the measuring detector at a time t , in A
$I_{\rm g}(t,t_{\rm C}),I_{\rm g}$	Gross current during the counting time $t_{\mathbb{C}}$ of the measuring detector at a time t , in A
$I_{\mathrm{g,cd}}(t)$	Instantaneous gross current of the compensating detector at a time t , in A
$I_{g,cd}(t,t_{C,cd}),I_{g,cd}$	Gross current during the counting time $t_{\mathrm{C,cd}}$ of the compensating detector at a time t , in A
I_0	Background current of the measuring detector, in A
$I_{0,\mathrm{cd}}$	Background current of the compensating detector, in A
K	Detection alarm setup parameter corresponding to the chosen acceptable false alarm rate level, dimensionless
k	Quantile of a standard normal distribution, if $k_{1-\alpha} = k_{1-\beta}$, dimensionless
$k_{1-\alpha}$	Quantile of a standard normal distribution for a probability (1– α), dimensionless
$k_{1-\beta}$	Quantile of a standard normal distribution for a probability (1– β), dimensionless
$k_{1-\frac{\gamma}{2}}$	Quantile of a standard normal distribution for a probability $\left(1-\frac{\gamma}{2}\right)$, dimensionless
N	Number of atoms on the media filter, dimensionless
$n_{\rm g}(t,t_{\rm C})$ s. iteh. ai/ca	Gross count during the counting time $t_{\rm C}$ of the measuring detector at a time t , dimensionless
Q_{\min}	Minimum amount of electric charge that induces a pulse registered by the measuring detector, in \ensuremath{C}
$Q_{\min,cd}$	Minimum amount of electric charge that induces a pulse registered by the compensating detector, in \ensuremath{C}
q	Flow rate, in m ³ ⋅s ⁻¹
$r_{\mathrm{g}}\left(t ight)$	Instantaneous gross count rate of the measuring detector at a time t , in $\rm s^{-1}$
$r_{\rm g}\left(t,t_{ m C}\right),r_{ m g}$	Gross count rate during the counting time $t_{\mathbb{C}}$ of the measuring detector at a time t , in s ⁻¹
$r_{\rm g,cd}(t)$	Instantaneous gross count rate of the compensating detector at a time t , in ${\rm s}^{-1}$
$r_{\rm g,cd}(t,t_{\rm C,cd}),r_{\rm g,cd}$	Gross count rate during the counting time $t_{\rm C,cd}$ of the compensating detector at a time t , in $\rm s^{-1}$
r_0	Background count rate of the measuring detector, in s ⁻¹
$r_{0,\mathrm{cd}}$	Background count rate of the compensating detector, in s ⁻¹
s_0	Standard deviation of the activity concentration at a series of \emph{i} measurements which represent a background situation

ISO/TR 22930-2:2020(E)

 $t_{\rm C}$ Counting time of the measuring detector, in s

 $t_{\mathrm{C.cd}}$ Counting time of the compensating detector, in s

 $t_{\rm F}$ Duration of airborne release, in s

 $t_{\rm R}$ Response time, in s

 $t_{\rm RI}$ Intrinsic response time, in s

 t_0 Counting time of the measuring detector for background measurement, in s

 $t_{
m 0.\,cd}$ Counting time of the compensating detector for background measurement, in s

 $t_{1/2}$ Half-life, in s

V Detection volume, in m³

w Calibration factor, in $Bq \cdot m^{-3} \cdot s$ or $Bq \cdot m^{-3} \cdot A^{-1}$

 δ Correction factor related to sampling (sampling point representativity, radioactive

decay, ...), dimensionless

 $arepsilon_{
m D}$ Detector efficiency, in Bq-1·s-1 or A·Bq-1

λ Decay constant, in s⁻¹ eh Standards

5 Measuring principle https://standards.iteh.ai)

A representative sample of ambient air to be monitored containing an activity concentration $c_{\rm ac}(t)$ at a time t is continuously captured through a transport line then goes through a detection volume without being retained. In parallel, a detector continuously measures the activity going through the detection volume. Then a processing algorithm calculates the activity concentration c(t) and the suited alarms on the basis of the evolution of the activity going through the detection volume of air sampled and the installation or not of an ambient compensating detector. The processing algorithm can also, if necessary, take into account influence quantities which may perturb the measurement result (see Figure 1).