

---

---

**Petroleum and related products —  
Precision of measurement methods  
and results —**

**Part 1:  
Determination of precision data in  
relation to methods of test**

**AMENDMENT 1**

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

*Produits pétroliers et connexes — Fidélité des méthodes de mesure et  
de leurs résultats*

<https://standards.iteh.ai/catalog/standards/sist/7692bd2c-6467-41a9-a077-3ad2b7070181/iso-4259-1:2017/amd.1:2019>

*Partie 1: Détermination des valeurs de fidélité relatives aux  
méthodes d'essai*

*AMENDEMENT 1*



**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

<https://standards.iteh.ai/catalog/standards/sist/7692bd2c-6467-41a9-a077-f3ad2b6667f4/iso-4259-1-2017-amd-1-2019>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html) (standards.iteh.ai)

This document was prepared by Technical Committee ISO/TC 28, *Petroleum and related products, fuels and lubricants from natural or synthetic sources*.

A list of all parts in the ISO 4259 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 4259-1:2017/Amd 1:2019](https://standards.iteh.ai/catalog/standards/sist/7692bd2c-6467-41a9-a077-f3ad2b6667f4/iso-4259-1-2017-amd-1-2019)

<https://standards.iteh.ai/catalog/standards/sist/7692bd2c-6467-41a9-a077-f3ad2b6667f4/iso-4259-1-2017-amd-1-2019>

# Petroleum and related products — Precision of measurement methods and results —

## Part 1: Determination of precision data in relation to methods of test

### AMENDMENT 1

#### 5.3.1, third paragraph in total

Replace

"If the points so plotted can be considered as lying about a pair of lines parallel to the  $m$ -axis, then no transformation is necessary. If, however, the plotted points describe non-horizontal straight lines or curves of the form  $D = f_1(m)$  and  $d = f_2(m)$ , then a transformation is necessary"

with:

"Perform linear regression of  $D$  versus  $m$  and of  $d$  versus  $m$  to obtain the following linear relationship:

$$D = b_0 + b_1 \times m; d = b_0 + b_1 \times m \quad \text{ISO 4259-1:2017/Amd 1:2019} \quad (3)$$

where

$b_0$  represents the constant term and

$b_1$  represents the slope.

In both cases, test whether the value of  $b_1$  is statistically different from zero (0) at 5 % significance level. If  $b_1$  from each regression is not statistically different from zero, no transformation is required. Proceed to section 5.3.2 directly and continue.

If, however, at least one of the values for  $b_1$  is significant, or if the plotted points are curves of the form  $D = f_1(m)$  and  $d = f_2(m)$ , then a transformation is necessary. Proceed as follows:"

*And renumber all formulae and references thereto beyond this subclause.*

#### Formula (3)

Replace


"where  $K$  is a constant"

with:

"where  $K$  is a constant".

Figure 1

Replace the Figure with the following:



## 6.3.2, explanation under the three formulae

Replace

$$"I_2 = 2 \frac{(L_N - S')}{L' - 1} \quad \beta = 2 \frac{(K - S')}{(L' - 1)}"$$

with:

$$\beta = 2 \frac{(L_N - S')}{L' - 1} \quad \text{or otherwise} \quad \beta = 2 \frac{(K - S')}{(L' - 1)}"$$

6.3.3.2, explanation of  $r_1$ ,  $r_2$  and  $r_3$  under Formula (19):

Replace

"are the three successive terms in Formula (18), i.e.:"

with

"are the three successive terms in Formula (19), i.e.:"

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

## 6.4.3, second sentence

Replace:

[ISO 4259-1:2017/Amd.1:2019](https://standards.iteh.ai/catalog/standards/sist/7692bd2c-6467-4819-a077-f2a12b6667fd/iso-4259-1-2017-amd-1-2019)

"At minimum, the summary description shall include the number of laboratories, number and type of materials studied, and range of the measured average property levels"

with:

"At minimum, the summary description shall include the number of laboratories, number and type of materials studied, range of the measured average property levels, and the achieved degrees of freedom for  $r$  and  $R$ ".

6.5, first sentence of the 2<sup>nd</sup> paragraph:

Replace

"The lower limit of the scope of the test method shall be the larger of lowest sample mean tested in the ILS or lowest achievable result +  $2R$ , where  $R$  is evaluated at the lowest sample mean"

with:

"The lower limit of the scope of the test method shall be the larger of lowest retained sample mean tested after completion of the statistical analysis, or lowest achievable result +  $2R$ , where  $R$  is evaluated at the lowest retained sample mean."

6.5, end of the 2<sup>nd</sup> paragraph the whole sentence:

Delete

"Due to testing variation, the lowest acceptable single result that is deemed as a valid result of the test method shall be lower method scope limit –  $1,2 \cdot R$ , where  $R$  is evaluated at the low method scope limit value".

6.5, first sentence of the 3<sup>rd</sup> paragraph:

Replace

"Similarly, the upper limit of the scope of a test method shall be the lesser of highest sample mean tested in the ILS or highest achievable result –  $2R$ , where  $R$  is evaluated at the highest sample mean"

with:

"Similarly, the upper limit of the scope of a test method shall be the lesser of highest retained sample mean or highest achievable result –  $2R$ , where  $R$  is evaluated at the highest retained sample mean.

6.5, end of the 3<sup>rd</sup> paragraph the whole sentence:

Delete "Due to testing variation, the highest acceptable single result that is deemed as a valid result of the test method shall be higher method scope limit +  $1,2 \cdot R$ , where  $R$  is evaluated at the high method scope limit value".

(standards.iteh.ai)

New subclause 6.6

ISO 4259-1:2017/Amd 1:2019  
<https://standards.iteh.ai/catalog/standards/sist/7692bd2c-6467-41a9-a077-f3ad2b6667f4/iso-4259-1-2017-amd-1-2019>

Introduce subclause 6.6. Reporting limits instruction for the test method:

"Due to testing variation, the lowest acceptable single result that is deemed as a valid result of the test method shall be lower method scope limit –  $1,2 \cdot R$ , where  $R$  is evaluated at the lowest retained sample mean value (see 6.5).

Similarly, the highest acceptable single result that is deemed as a valid result of the test method shall be higher method scope limit +  $1,2R$ , where  $R$  is evaluated at the highest retained sample mean value (see 6.5).

A statement shall be included in the test method indicating the range of acceptable test result values that are deemed to be valid based on the above."

Table F.2, second column all indicators for variables

Replace

"B"

with:

"b".



**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 4259-1:2017/Amd 1:2019](https://standards.iteh.ai/catalog/standards/sist/7692bd2c-6467-41a9-a077-f3ad2b6667f4/iso-4259-1-2017-amd-1-2019)

<https://standards.iteh.ai/catalog/standards/sist/7692bd2c-6467-41a9-a077-f3ad2b6667f4/iso-4259-1-2017-amd-1-2019>