



SLOVENSKI STANDARD
oSIST prEN ISO 14146:2023

01-oktober-2023

**Radiološka zaščita - Merila in meje učinkovitosti za periodično ovrednotenje
dozimetričnih storitev (ISO/DIS 14146:2023)**

Radiological protection - Criteria and performance limits for the periodic evaluation of
dosimetry services (ISO/DIS 14146:2023)

Strahlenschutz - Kriterien und Mindestanforderungen bei der wiederkehrenden
Überprüfung von Dosismessstellen (ISO/DIS 14146:2023)

Radioprotection - Critères et limites de performance pour l'évaluation périodique des
services de dosimétrie (ISO/DIS 14146:2023)

Ta slovenski standard je istoveten z: prEN ISO 14146

ICS:

13.280 Varstvo pred sevanjem Radiation protection

oSIST prEN ISO 14146:2023 **en,fr,de**

DRAFT INTERNATIONAL STANDARD

ISO/DIS 14146

ISO/TC 85/SC 2

Secretariat: AFNOR

Voting begins on:
2023-08-11Voting terminates on:
2023-11-03

Radiological protection — Criteria and performance limits for the periodic evaluation of dosimetry services

Radioprotection — Critères et limites de performance pour l'évaluation périodique des services de dosimétrie

ICS: 13.280

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[oSIST prEN ISO 14146:2023](https://standards.iteh.ai/catalog/standards/sist/b6b21e99-0f94-4cce-890b-c024db24f7d2/osist-pren-iso-14146-2023)

<https://standards.iteh.ai/catalog/standards/sist/b6b21e99-0f94-4cce-890b-c024db24f7d2/osist-pren-iso-14146-2023>

This document is circulated as received from the committee secretariat.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

ISO/CEN PARALLEL PROCESSING



Reference number
ISO/DIS 14146:2023(E)

© ISO 2023

iTeh STANDARD PREVIEW (standards.iteh.ai)

[oSIST prEN ISO 14146:2023](https://standards.iteh.ai/catalog/standards/sist/b6b21e99-0f94-4cce-890b-c024db24f7d2/osist-pren-iso-14146-2023)

<https://standards.iteh.ai/catalog/standards/sist/b6b21e99-0f94-4cce-890b-c024db24f7d2/osist-pren-iso-14146-2023>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 Quantities measured.....	7
5 Frequency of evaluation.....	8
6 Test conditions.....	8
6.1 Standard test conditions and special handling conditions.....	8
6.2 Radiation qualities.....	9
6.3 Dose range.....	10
6.4 Irradiation of dosimeters.....	10
7 Performance limits.....	11
7.1 Limits to the response.....	11
7.1.1 General requirements.....	11
7.1.2 Requirements at reference conditions.....	11
7.2 Approval criterion.....	12
8 Operational procedures.....	12
8.1 Evaluation sample size.....	12
8.2 Evaluation procedure.....	12
8.3 Evaluation sequence.....	12
9 Test report.....	13
Annex A (normative) Reference conditions and standard test conditions.....	15
Annex B (informative) Graphical illustrations of the performance limits.....	17
Bibliography.....	19

ISO/DIS 14146:2023(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see <https://www.iso.org/patents>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of the standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This third edition of ISO 14146 cancels and replaces the second edition (ISO 14146:2018) of which it constitutes a technical revision. The main changes with respect to the previous edition are the addition and clarification of several definitions, the modification of the requirements to environmental dosimeters and the addition of a requirement at reference conditions.

Radiological protection — Criteria and performance limits for the periodic evaluation of dosimetry services

1 Scope

The quality of a supplier of a dosimetry service depends on both the characteristics of the approved (type-tested) dosimetry system¹⁾ and the training and experience of the staff, together with the calibration procedures and quality assurance programmes.

This document specifies the dosimetric and organizational criteria and the test procedures to be used for the periodic verification of the performance of dosimetry services supplying personal and/or workplace and/or environmental dosimeters.

The performance evaluation can be carried out as a part of the approval procedure for a dosimetry system or as an independent check to verify that a dosimetry service fulfils specified national or international type test performance requirements under representative exposure conditions that are expected or mimic workplace fields from the radiological activities being monitored.

This document applies to personal and area dosimeters for the assessment of external photon radiation with a fluence-weighted mean energy between 8 keV and 10 MeV, beta radiation with a fluence-weighted mean energy between 60 keV and 1,2 MeV, and neutron radiation with a fluence-weighted mean energy between 25,3 meV, i.e., thermal neutrons with a Maxwellian energy distribution with $kT = 25,3$ meV, and 200 MeV.

It covers all types of personal and area dosimeters needing laboratory processing (e.g., thermoluminescent, optically stimulated luminescence, radiophotoluminescent, track detectors or photographic-film dosimeters) and involving continuous measurements or measurements repeated regularly at fixed time intervals (e.g., several weeks, one month).

Active dosimeters (for dose measurement) may also be treated according to this document. Then, they should be treated as if they were passive, i.e., the dosimetry service reads their indicated values and reports them to the evaluation organization.

In this document, the corrected indicated value is the one given by the dosimetry systems as the final result of the evaluation algorithm (for example, display of the software, printout) in units of dose equivalent (Sv).

Environmental dosimeters for the quantity air kerma, K_a , or absorbed dose, D , may also be treated according to this document. All dose values stated in Sv shall then be interpreted as equivalent values in Gy.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4037-1, *X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 1: Radiation characteristics and production methods*

1) If this document is applied to a dosimetry system for which no approval (pattern or type test) has been provided, then in the following text approval or type test should be read as the technical data sheet provided by the manufacturer or as the data sheet required by the regulatory authority.

ISO/DIS 14146:2023(E)

ISO 4037-2, *X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy — Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV*

ISO 4037-3, *X and gamma reference radiation for calibrating dosimeters and dose rate meters and for determining their response as a function of photon energy — Part 3: Calibration of area and personal dosimeters and the measurement of their response as a function of energy and angle of incidence*

ISO 6980-1, *Nuclear energy — Reference beta-particle radiation — Part 1: Methods of production*

ISO 6980-2, *Nuclear energy — Reference beta-particle radiation — Part 2: Calibration fundamentals related to basic quantities characterizing the radiation field*

ISO 6980-3, *Nuclear energy — Reference beta-particle radiation — Part 3: Calibration of area and personal dosimeters and the determination of their response as a function of beta radiation energy and angle of incidence*

ISO 8529-1, *Reference radiation fields — Part 1: Characteristics and methods of production*

ISO 8529-2, *Reference neutron radiations — Part 2: Calibration fundamentals of radiation protection devices related to the basic quantities characterizing the radiation field*

ISO 8529-3, *Reference neutron radiations — Part 3: Calibration of area and personal dosimeters and determination of response as a function of energy and angle of incidence*

ISO 12789-1, *Reference radiation fields — Simulated workplace neutron fields — Part 1: Characteristics and methods of production*

ISO 12789-2, *Reference radiation fields — Simulated workplace neutron fields — Part 2: Calibration fundamentals related to the basic quantities*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

ISO/IEC 17043, *Conformity assessment — General requirements for proficiency testing providers*

ISO/TS 18090-1, *Radiological protection — Characteristics of reference pulsed radiation — Part 1: Photon radiation*

ISO 29661, *Reference radiation fields for radiation protection — Definitions and fundamental concepts*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

IEC 61267, *Medical diagnostic X-ray equipment — Radiation conditions for use in the determination of characteristics*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 29661 and the following apply.

3.1

approved dosimetry system

dosimetry system that is used by a dosimetry service that has been approved or authorized for use by the qualification body

Note 1 to entry: Several dosimeter designs can be operated using the same associated processing system (dosimeter reader, etc.). Then, they are regarded as separate dosimeters/dosimetry systems.

3.2

area dosimeter

dosimeter used for area monitoring

3.3**area monitoring**

monitoring in which a workplace or an area in the environment is monitored by taking dose (rate) measurements

Note 1 to entry: Area monitoring is usually performed in terms of $H'(0,07)$, $H'(3)$ or $H^*(10)$.

Note 2 to entry: Definition orientated at ICRP 103 and ICRP 116.

[SOURCE: IEC 62387:2020, 3.46]

3.4**background (radiation dose)**

dose (or an observed measure related to the dose) attributable to all sources other than the one(s) specified

Note 1 to entry: Strictly, this applies to measurements of dose or counts from a sample, where the background dose or counts must be considered (usually subtracted) from all measurements. However, background is used more generally to refer to the effects of other sources in any situation in which a particular source (or group of sources) is under consideration. It is also applied to quantities other than doses, such as activity concentrations in environmental media.

Note 2 to entry: Although it should not, the background dose can contain dose fractions from transportation and/or other events such as X-ray screening for security checks.

Note 3 to entry: To determine the background dose, usually, a group of control (background) dosimeters is used.

[SOURCE: IAEA Safety Glossary 2018, modified: "(radiation dose)" added to the term; "dose rate" removed; second sentence in note 1 rearranged; notes 2 and 3 added]

3.5**control (background) dosimeter**

personal, area or environmental dosimeter that provides an estimate of any radiation dose received by the evaluation sample apart from that given by the irradiating laboratory or by a controlled exposure to environmental radiation

Note 1 to entry: The control dosimeter provides a means of estimating and eliminating the contribution to the dose from natural background radiation and that received during the time between zeroing and read out, i.e., the dose during handling, transportation, etc.

Note 2 to entry: The control dosimeters are used to determine the background radiation dose.

3.6**corrected indication****corrected indicated value**

G_{corr}

indication of a dosimeter corrected for any differences of the values of the influence quantities from reference conditions

Note 1 to entry: The corrected indication G_{corr} can be calculated with the correction factor k_n for non-constant response, the q correction factors, k_f for the influence quantities of type F and the p correction summands, G_w , for the influence quantities of type S. It is given by

$$G_{corr} = k_n \cdot \left(G - \sum_{w=1}^p G_w \right) \cdot \prod_{f=1}^q k_f$$

The equation above is a model function of the measurement necessary for any determination of the uncertainty according to the ISO/IEC Guide 98-3 (GUM).

ISO/DIS 14146:2023(E)

Note 2 to entry: Some dosimetry systems, especially such for neutron radiation, can have different correction factors (or functions) k_w for different workplace categories w , each with its own reference radiation quality (e.g., ^{252}Cf (bare) for one workplace category and $^{252}\text{Cf}(\text{D}_2\text{O}$ moderated) for another workplace category). Then, to obtain the correspondingly corrected indicated dose value G_{corr} , the uncorrected indicated values G , needs to be multiplied by $k_w \neq 1$: $G_{\text{corr}} = G \cdot k_w$. Further information on the use of different correction factors (or functions) for different workplace categories can be found in ISO 21909-2.^[2] In ISO 21909-2, the symbol for the correction factor (or function) is $k_{n,E,\Omega}$ instead of k_w .

[SOURCE: ISO 29661:2012 + AMD.1:2015, 3.1.11, modified: term “corrected indicated value” added and original note 2 deleted and new note 2 added]

3.7 dosimeter

device having a reproducible, measurable response to radiation that can be used to measure absorbed dose or personal, ambient or directional dose equivalent

Note 1 to entry: In a wider sense, this term is used for meters designed to measure other quantities related to radiation such as exposure, fluence, etc. Such use is deprecated.

Note 2 to entry: This apparatus may require a separate reader to read out and software to evaluate and display the indicated value of the absorbed dose or dose equivalent.

[SOURCE: ISO 12749-2:2022, 3.4.12, modified: “ambient or directional” added and notes 1 and 2 added]

3.8 dosimetry service

organization that operates a personal, area and/or environmental dosimetry system which includes the evaluation of the reading of dosimeters after their use and includes:

- providing the user with dosimeters;
- recording the results;
- reporting the results to the user.

Note 1 to entry: The dosimetry service fulfils basic quality management and independency requirements if it fulfils the requirements stated in ISO/IEC 17025.

Note 2 to entry: The user includes not only external clients but also internal personnel who wear dosimeters provided by their own organization and are engaged in radiation protection activities inside or outside the organization. The same quality of dosimetry service which is provided to external users is also provided to organizations' employees (internal users), in accordance with their own quality management system.

3.9 dosimetry system

dosimeter, reader and all associated equipment and procedures including software used for assessing and displaying the indicated value

[SOURCE: IEC 62387:2020, 3.12, modified: “including software” and “and displaying” added]

3.10 environmental dosimeter

dosimeter used for environmental monitoring

3.11 environmental monitoring

area monitoring by the measurement of external dose (rate) in the environment

Note 1 to entry: Environmental monitoring is usually performed in terms of $H'(0,07)$, $H'(3)$ or $H^*(10)$.

[SOURCE: IEC 62387:2020, 3.48]

3.12**evaluation sample**

randomly selected representative group of personal, area or environmental dosimeters used to evaluate the performance of a dosimetry service

Note 1 to entry: The evaluation sample includes dosimeters that are irradiated, remain unirradiated or serve as control dosimeters for the evaluation procedure.

3.13**independent evaluation organization
evaluation organization**

independent organization that administers the performance evaluation of dosimetry services and assesses the results

Note 1 to entry: The evaluation organization may include the irradiating laboratory.

Note 2 to entry: The evaluation organization fulfils basic quality management and independency requirements if it fulfils the requirements stated in ISO/IEC 17043.

3.14**independent irradiating laboratory
irradiating laboratory**

independent laboratory possessing radiation sources, calibration equipment and associated facilities traceable to national, i.e., to primary or secondary, standards able to irradiate dosimeters from the evaluation sample to a high degree of accuracy

Note 1 to entry: The irradiating laboratory fulfils basic quality management and independency requirements if it fulfils the requirements stated in ISO/IEC 17025. Accreditation according to ISO/IEC 17025 independently confirms the competence of the irradiating laboratory.

3.15**independent qualification body
qualification body**

independent organization empowered by a governmental, regulatory or advisory agency to approve a dosimetry service or authorize the use of a dosimetry system

Note 1 to entry: The qualification body may include the evaluation organization (see definition [3.8](#)).

Note 2 to entry: The qualification body fulfils basic quality management and independency requirements if it fulfils the requirements stated in ISO/IEC 17025.

3.16**indication
indicated value**

G

quantity value provided by a measuring instrument or a measuring system

Note 1 to entry: The units of the indication of the dosimeter are not necessarily the same as that of the measurand. For example, for measurements with ionization chambers the instrument indication is, in general, the value of the current I or of the charge Q . It is necessary to document whether the indication is normalized to the reference conditions to account for influence quantities and is corrected for intrinsic background and other influences. The corrected indication is named G_{corr}

Note 2 to entry: It may be necessary that a measured dose (e.g., by control dosimeters) or a calculated transport and/or background dose be considered (usually subtracted) by the dosimetry service or by the evaluating organization, see notes to the definition of [3.17](#), irradiated dose.

[SOURCE: ISO 29661:2012 + AMD.1:2015, 3.1.15, modified: term “indicated value” added and original notes 1 and 3 deleted and new note 2 added]

3.17**individual dosimeter**

dosimeter used for individual monitoring