

INTERNATIONAL
STANDARD

ISO
5165

Fifth edition
2020-07

**Petroleum products — Determination
of the ignition quality of diesel fuels —
Cetane engine method**

*Produits pétroliers — Détermination de la qualité d'inflammabilité
des carburants pour moteurs diesel — Méthode cétane*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 5165:2020](#)

[https://standards.iteh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-
d55163fdb1ce/iso-5165-2020](https://standards.iteh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-d55163fdb1ce/iso-5165-2020)



Reference number
ISO 5165:2020(E)

© ISO 2020

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 5165:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-d55163fdb1ce/iso-5165-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle	3
5 Reagents and reference materials	4
6 Apparatus	5
6.1 Test engine assembly	5
6.2 Instrumentation	5
6.3 Reference fuels dispensing equipment	7
6.4 Injector nozzle tester	7
6.5 Special maintenance tools	8
7 Sampling and sample preparation	8
8 Basic engine and instrument settings and standard operating conditions	8
8.1 Installation of engine equipment and instrumentation	8
8.2 Engine speed	8
8.3 Valve timing	8
8.4 Valve lift	9
8.5 Fuel pump timing	9
8.6 Fuel pump inlet pressure	9
8.7 Direction of engine rotation	9
8.8 Injection timing	9
8.9 Injector nozzle opening pressure	9
8.10 Injection flow rate	9
8.11 Injector coolant passage temperature	9
8.12 Valve clearances	9
8.13 Oil pressure	9
8.14 Oil temperature	9
8.15 Cylinder jacket coolant temperature	10
8.16 Intake air temperature	10
8.17 Basic ignition delay	10
8.18 Cylinder jacket coolant level	10
8.19 Engine-crankcase lubricating oil level	10
8.20 Crankcase internal pressure	10
8.21 Exhaust back-pressure	10
8.22 Exhaust and crankcase breather system resonance	10
8.23 Piston over-travel	10
8.24 Belt tension	10
8.25 Injector opening or release pressure	11
8.26 Injector spray pattern	11
8.27 Indexing handwheel reading	11
8.27.1 General	11
8.27.2 Basic setting of variable compression plug	11
8.27.3 Setting handwheel micrometer drum and scale	11
8.27.4 Setting handwheel reading	11
8.28 Basic compression pressure	12
8.29 Fuel pump lubricating oil level	12
8.30 Fuel pump timing gear-box oil level	12
8.31 Setting instrumentation reference pickups	13
8.32 Setting injector pickup gap	13
9 Engine qualification	13

9.1	Engine conformity.....	13
9.2	Checking performance on check fuels.....	13
9.3	Check in the case of nonconformity.....	13
10	Procedure.....	14
10.1	General.....	14
10.2	Sample introduction.....	14
10.3	Fuel flow rate.....	14
10.4	Fuel injection timing.....	14
10.5	Ignition delay.....	14
10.6	Equilibration.....	14
10.7	Handwheel reading.....	14
10.8	Reference fuel no. 1.....	14
10.9	Reference fuel no. 2.....	15
10.10	Number of blends of reference fuels.....	16
10.11	Repeat readings.....	16
11	Calculation.....	16
12	Expression of results.....	18
13	Precision.....	18
13.1	General.....	18
13.2	Repeatability, r	18
13.3	Reproducibility, R	18
13.4	Precision basis.....	18
14	Test report.....	19
Bibliography	(standards.iteh.ai)	20

[ISO 5165:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-d55163fdb1ce/iso-5165-2020>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

The STANDARD REVIEW (standards.iteh.ai)

This document was prepared by Technical Committee ISO/TC 28, *Petroleum and related products, fuels and lubricants from natural or synthetic sources*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 19, *Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This fifth edition cancels and replaces the fourth edition (ISO 5165:2017), which has been technically revised. It is aligned with ASTM D613. The main changes to the previous edition are as follows:

- requirements for primary reference fuels (PRFs), secondary reference fuels (SRFs) and check fuels have been added;
- new low cetane primary reference fuel, pentamethylheptane (PMH), as an alternative to heptamethylnonane (HMN), have been added;
- new reporting requirements.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 5165:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-d55163fdb1ce/iso-5165-2020>

Petroleum products — Determination of the ignition quality of diesel fuels — Cetane engine method

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all of the safety problems associated with its use. It is the responsibility of users of this document to take appropriate measures to ensure the safety and health of personnel prior to the application of the document.

1 Scope

This document establishes the rating of diesel fuel oil in terms of an arbitrary scale of cetane numbers (CNs) using a standard single cylinder, four-stroke cycle, variable compression ratio, indirect injected diesel engine. The CN provides a measure of the ignition characteristics of diesel fuel oil in compression ignition engines. The CN is determined at constant speed in a pre-combustion chamber-type compression ignition test engine. However, the relationship of test engine performance to full scale, variable speed and variable load engines is not completely understood.

This document is applicable for the entire scale range from 0 CN to 100 CN but typical testing is in the range of 30 CN to 65 CN. An interlaboratory study executed by CEN in 2013 (10 samples in the range 52,4 CN to 73,8 CN)^[3] confirmed that paraffinic diesel from synthesis or hydrotreatment, containing up to a volume fraction of 17% fatty acid methyl ester (FAME), can be tested by this test method and that the precision is comparable to conventional fuels.

This test can be used for unconventional fuels such as synthetics or vegetable oils. However, the precision for those fuels has not been established and the relationship to the performance of such materials in full-scale engines is not completely understood.

Samples with fluid properties that interfere with the gravity flow of fuel to the fuel pump or delivery through the injector nozzle are not suitable for rating by this method.

NOTE This document specifies operating conditions in SI units but engine measurements are specified in inch-pound units or Fahrenheit because these are the historical units used in the manufacture of the equipment, and thus some references in this document include these and other non-SI units in parenthesis.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3170, *Petroleum liquids — Manual sampling*

ISO 3171, *Petroleum liquids — Automatic pipeline sampling*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 4787, *Laboratory glassware — Volumetric instruments — Methods for testing of capacity and for use*

ASTM D613, *Standard Test Method for Cetane Number of Diesel Fuel Oil*

ASTM D3703, *Test Method for Hydroperoxide Number of Aviation Turbine Fuels, Gasoline and Diesel Fuels*

ASTM E832-81, *Standard Specification for Laboratory Filter Papers*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

cetane number

CN

measure of the ignition performance of a diesel fuel oil obtained by comparing it to reference fuels in a standardized engine test

Note 1 to entry: Ignition performance is understood to mean the *ignition delay* (3.3) of the fuel as determined when the standard test engine is operated under controlled conditions of fuel flow rate, *injection timing* (3.4) and *compression ratio* (3.2).

3.2

compression ratio

ratio of the volume of the combustion chamber including the pre-combustion chamber with the piston at bottom-dead-centre (b.d.c.) to the comparable volume with the piston at top-dead-centre (t.d.c.)

3.3

ignition delay

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 5165:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-d55163fdb1ce/iso-5165-2020>

period of time between the start of fuel injection and the start of combustion, expressed in degrees of crank angle rotation

3.4

injection timing

injection advance

[ISO 5165:2020](#)

time in the combustion cycle at which fuel injection into the combustion chamber is initiated, expressed in degrees of crank angle

3.5

handwheel reading

arbitrary numerical value, related to *compression ratio* (3.2), obtained from a micrometer scale that indicates the position of the variable compression plug in the pre-combustion chamber of the engine

3.6

cetane meter

ignition delay meter

electronic instrument which displays *injection timing* (3.4) and *ignition delay* (3.3) derived from input pulses of multiple transducers (pickups)

Note 1 to entry: Three generations of apparatus have been approved for use as cetane meters. These are (with the year of introduction in parenthesis) the Mark II Ignition Delay Meter (1974), the Dual Digital Cetane Meter (1990) and the XCP Cetane Panel (2014).

3.7

injector nozzle opening pressure

fuel pressure that overcomes the resistance of the spring which normally holds the injector nozzle pintle closed, and thus forces the pintle to lift and release an injection spray from the nozzle

3.8

reference pickup

transducers or optical sensors mounted over the flywheel of the engine, triggered by a flywheel pointer, used to establish a t.d.c. reference and a time base for calibration of the *cetane meter* (3.6)

3.9**injector pickup**

transducer to detect motion of the injector pintle, thereby indicating the beginning of injection

3.10**combustion pickup**

pressure transducer exposed to cylinder pressure to indicate the start of combustion

3.11**primary reference fuel****PRF**

hexadecane (*n*-cetane), heptamethylnonane (HMN), pentamethylheptane (PMH) and volumetrically proportioned binary mixture of *n*-cetane with either HMN or PMH

Note 1 to entry: These PRFs now define the CN scale by the relationships given in [Formulae \(1\)](#) and [\(2\)](#).

$$CN = P_r + 0,15 P_{HMN} \quad (1)$$

$$CN = P_r + 0,163 P_{PMH} \quad (2)$$

where

P_r is the percentage *n*-cetane in the binary mixture;

P_{HMN} is the percentage of HMN in the binary mixture;

P_{PMH} is the percentage of PMH in the binary mixture.

Note 2 to entry: Alphamethylnaphthalene (1-methylnaphthalene) in its pure form, was originally defined as 0 and *n*-cetane (hexadecane) as 100 for the CN scale. With blends of the two chemicals being used for the intervening values, alphamethylnaphthalene was subsequently replaced in 1962 by HMN as the low reference material, with an assigned value of 15, as it was more readily available and experience had shown that it had better storage stability. Pentamethylheptane, a second low cetane ingredient as an alternative to HMN with an assigned value of 16,3, was introduced in 2018 to utilize a material of higher purity and better availability.

3.12**secondary reference fuel****SRF**

volumetrically proportioned blend of two selected, numbered and paired hydrocarbon mixtures designated "T fuel" (high CN) and "U fuel" (low CN) where each numbered paired set of "T fuel" and "U fuel" is rated by the ASTM Diesel National Exchange Group (NEG) in various combinations by comparison to *primary reference fuel* ([3.11](#)) blends

3.13**check fuel**

diesel fuel oil having a *cetane number* ([3.1](#)) value determined by an interlaboratory comparison which provides a guide for an individual laboratory to check the cetane rating performance of a specific engine unit

4 Principle

The CN of a diesel fuel oil is determined by comparing its combustion characteristics in a test engine with those for blends of reference fuels of known CN under standard operating conditions. This is accomplished using the bracketing handwheel procedure, which varies the compression ratio (handwheel reading) for the sample and each of two bracketing reference fuels to obtain a specific ignition delay permitting the interpolation of CN in terms of handwheel reading.

5 Reagents and reference materials

5.1 Cylinder jacket coolant

water conforming to grade 3 of ISO 3696.

Water shall be used in the cylinder jacket for laboratory locations where the resultant boiling temperature is $100\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$ ($212\text{ }^{\circ}\text{F} \pm 3\text{ }^{\circ}\text{F}$). Water with commercial glycol-based antifreeze added in a sufficient quantity to meet the boiling temperature requirement shall be used when the laboratory altitude dictates. A commercial multi-functional water-treatment material should be used in the coolant to minimize corrosion and mineral scale that can alter heat transfer and rating results.

5.2 Engine crankcase lubricating oil

an SAE 30 viscosity grade oil^[2] meeting current American Petroleum Institute (API) service classification or compatible previous API service classification for engines shall be used. It shall contain a detergent additive and have a kinematic viscosity of $9,3\text{ mm}^2/\text{s}$ to $12,5\text{ mm}^2/\text{s}$ at $100\text{ }^{\circ}\text{C}$ ($212\text{ }^{\circ}\text{F}$) and a viscosity index of not less than 85. Oils containing viscosity index improvers shall not be used. Multi-graded lubricating oils shall not be used.

The suggested oil change interval is 50 engine-running hours.

5.3 Cetane PRF

hexadecane meeting the specifications in [Table 1](#) shall be used as the designated 100 CN component.

IMPORTANT — Store and use PRFs at temperatures of $20\text{ }^{\circ}\text{C}$ or higher to avoid solidification of hexadecane, which has a melting point of $18\text{ }^{\circ}\text{C}$.

5.4 Heptamethylnonane PRF

iTeh STANDARD PREVIEW
2,2,4,4,6,8,8-heptamethylnonane meeting the specifications in [Table 1](#)
shall be used as the designated 15 CN component.
<https://standards.iteh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-d55163fdb1ce/iso-5165-2020>

5.5 Pentamethylheptane PRF

2,2,4,6,6-pentamethylheptane meeting the specifications in [Table 1](#)
shall be used as the designated 16,3 CN component.
<https://standards.iteh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-d55163fdb1ce/iso-5165-2020>

WARNING — PRFs are combustible and the vapours are harmful.

Table 1 — Specifications for primary reference fuels

Property	Hexadecane	Heptamethylnonane	Pentamethylheptane	Test method
Purity g/kg, minimum	990	980	995	Gas Chromatography
Hydroperoxide number, mg/kg as 0, maximum	5,0	5,0	5,0	ASTM D3703

5.6 SRFs

volumetric blends of two diesel fuels meeting the specifications in [Table 2](#) that have been round-robin engine calibrated by a recognized exchange testing group.

Storage and use of “T fuel” and “U fuel” should be at temperatures above $0\text{ }^{\circ}\text{C}$ ($32\text{ }^{\circ}\text{F}$) to avoid potential solidification, particularly of “T fuel”. Before a container that has been stored at low temperature is placed in service, it should be warmed to a temperature of at least $14\text{ }^{\circ}\text{C}$ ($57\text{ }^{\circ}\text{F}$) above its cloud point. It should be held at this temperature for a period of at least 30 min and then the container should be thoroughly remixed.

NOTE 1 Cloud point is normally determined in accordance with ISO 3015.

SRF blends are rated in numbered pairs and are not interchangeable with SRF blends from other batches.

WARNING — SRFs are combustible and the vapours harmful.

NOTE 2 Blends of "T fuel" and "U fuel" that have been engine calibrated by the ASTM Diesel National Exchange Group can be, and typically are, used for routine testing. The calibration data are incorporated in blend tables that list the CNs assigned for various volume percentage blends of "T fuel" and "U fuel".

5.7 Check fuels, diesel fuel oils typical of the middle distillate type having a CN value determined by an interlaboratory comparison and meeting the specifications in [Table 2](#).

WARNING — Check fuels are combustible and the vapours harmful.

Table 2 — Specifications for SRFs and check fuels

Property	SRFs		Check fuels		Test method
	T-Fuel	U-Fuel	Low	High	
CN, minimum	73	19	38	50	ISO 5165
CN, maximum	76	22	42	55	ISO 5165
Hydroperoxide number, mg/kg as 0 maximum	5,0	5,0	5,0	5,0	ASTM D3703

6 Apparatus

6.1 Test engine assembly

As shown in [Figure 1](#). It comprises a single cylinder engine consisting of a standard crankcase with fuel pump assembly, a cylinder with separate head assembly of the pre-combustion type (see [Figure 2](#)), thermal-siphon recirculating jacket coolant system, multiple fuel tank system with selector valving, injector assembly with specific injector nozzle, electrical controls and a suitable exhaust pipe. The engine shall be belt connected to a special electric power-absorption motor, which acts as a motor driver to start the engine and as a means to absorb power at constant speed when combustion is occurring (engine firing).

https://standards.itelh.ai/catalog/standards/sist/86e1a8a4-9fc3-462a-8b22-d5163fdb1ce/iso-5165-2020

6.2 Instrumentation

An electronic instrument to measure injection and ignition delay timing as well as conventional thermometry, gauges and general-purpose meters.

NOTE Engine equipment and instrumentation are available from the single source manufacturer, CFR Engines Inc.¹⁾, N8 W22577 Johnson Drive, Pewaukee WI 53186, USA. CFR Engines Inc. also has authorized sales and service organizations in selected geographic areas.

1) This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used if they can be shown to lead to the same results.