INTERNATIONAL STANDARD

Third edition

Meat and meat products — Determination of L-(+)-glutamic acid content — Reference method

Viande et produits à base de viande — Détermination de la teneur en acide L-(+)-glutamique — Méthode de référence

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/PRF 4134 https://standards.iteh.ai/catalog/standards/sist/fa1946fc-a311-4798-b899-68cfd69da41f/iso-prf-4134

PROOF/ÉPREUVE



Reference number ISO 4134:2021(E)

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/PRF 4134 https://standards.iteh.ai/catalog/standards/sist/fa1946fc-a311-4798-b899-68cfd69da41f/iso-prf-4134



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword			iv
1	Scope		
2	Normative references		
3	Terms and definitions		
4	Princi	nle	1
5	Samn	μησ	2
5	Dropo	ration of tost sample	2 ງ
5	пера		
7	Test n	letnod of spectrophotometer	Z
	7.1 7.2	Apparatus	Z
	7.2 7.2	Appaiatus.	
	7.5	721 Conoral	
		7.5.1 General	
		7.5.2 Test portion of extract	
		7.5.5 Preparation	
	7 4	7.3.4 Determination	
	7.4	Calculation and results	
		7.4.1 Absorbance difference of L-(+)-glutamic acid standard solution	
		7.4.2 Absorbance difference of L-(+)-glutamic acid test solution	
		7.4.3 Formula of linear regression of L-(+)-grutamic acid standard curve	
		7.4.4 L-(+)-glutamic acid concentration of the test solution	
		7.4.5 L-(+)-glutamic acid contendoi test sample	8
	7.5	Precision	
		7.5.1 Interlaboratory test <u>SO/PRF 4134</u>	8
		7.5.2 http://www.catalog/standards/sist/fa1946fc-a311-4798-b899-	8
		7.5.3 Reproducibility _{8cfd69da41ffiso-prf-4134}	8
	7.6	Detection limit	
8	Test method of light absorption microplate reader		9
	8.1	Reagents	
	8.2	Apparatus	9
	8.3	Procedure	9
		8.3.1 General	9
		8.3.2 Extraction of L-(+)-glutamic acid in the test portion	9
		8.3.3 Determination	9
	8.4	Calculation and results	
		8.4.1 Absorbance difference for L-(+)-glutamic acid standard solution	
		8.4.2 Absorbance difference for L-(+)-glutamic acid test solution	10
		8.4.3 Formula of linear regression for L-(+)-glutamic acid standard curve	11
		8.4.4 L-(+)-glutamic acid concentration of the test solution	11
		8.4.5 L-(+)-glutamic acid content of test sample	11
	8.5	Precision	
		8.5.1 Interlaboratory test	
		8.5.2 Repeatability	
		8.5.3 Reproducibility	
	8.6	Detection limit	
9	Test r	enort	12
Annos	· A (inf	or of the second s	1/
Annex	Thinks It (Informative) safety practices		
Bibliography			

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 6, *Meat, poultry, fish, eggs and their products*. ISO/PRF 4134 https://standards.iteh.ai/catalog/standards/sist/fa1946fc-a311-4798-b899-

This third edition cancels and replaces the second edition (15014134:1999), which has been technically revised. The main changes compared with the previous edition are as follows:

- a new test method, the light absorption microplate reader method, has been added;
- the order of the clauses has been rearranged;
- the Scope (<u>Clause 1</u>) has been revised to specify free L-(+)-glutamic acid in meat and meat products;
- the Normative references (<u>Clause 2</u>) have been updated;
- the Terms and definitions (<u>Clause 3</u>) have been modified by adding the term "free L-(+)-glutamic acid";
- in <u>Clause 4</u>, the description of "extraction of L-(+)-glutamic acid of test portion" has been modified and the detection wavelength has been changed from "492 nm" to "490 nm";
- in <u>7.1</u>, the identification of enzyme activity units for diaphorase and glutamate dehydrogenase has been supplemented; the concentration of KOH, NAD has been modified; the NAD and diaphorase have been mixed into a solution; and the buffer, NAD and enzymes have been labelled with R1, R2, and R3;
- the apparatus list (7.2) has been updated;
- in <u>7.3</u>, the procedure of the test method of spectrophotometer has been modified by halving the sample mass and solution volume;
- in <u>7.3.4</u>, the method of judging the absorbance of the reaction end point has been modified and, as a result, the previous Annex B "Example of plotting and extrapolation of absorbance values" has been deleted;

- in <u>8.4</u>, the formula and symbol description of spectrophotometer has been modified;
- the previous Annex C "Derivation of equation for calculation of L-(+)-glutamic acid content" has been deleted;
- the Bibliography has been updated.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at <u>www.iso.org/members.html</u>.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/PRF 4134 https://standards.iteh.ai/catalog/standards/sist/fa1946fc-a311-4798-b899-68cfd69da41f/iso-prf-4134

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO/PRF 4134 https://standards.iteh.ai/catalog/standards/sist/fa1946fc-a311-4798-b899-68cfd69da41f/iso-prf-4134

Meat and meat products — Determination of L-(+)glutamic acid content — Reference method

1 Scope

This document specifies the spectrophotometer method and the light absorption microplate reader method for the determination of the free L-(+)-glutamic acid content of meat and meat products.

This document is applicable to meat and meat products, including livestock and poultry products.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 648, Laboratory glassware — Single-volume pipettes

ISO 1042, Laboratory glassware — One-mark volumetric flasks

ISO 1442, Meat and meat products — Determination of moisture content (Reference method)

ISO 3696, Water for analytical laboratory use - Specification and test methods

ISO 8655-2, Piston-operated volumetric apparatus ____Part 2: piston pipettes

https://standards.iteh.ai/catalog/standards/sist/fa1946fc-a311-4798-b899-

3 Terms and definitions 68cfd69da41f/iso-prf-4134

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at <u>http://www.electropedia.org/</u>

3.1

free L-(+)-glutamic acid

L-(+)-glutamic acid and glutamate existing in meat and meat products in the form of free state

4 Principle

The free L-(+)-glutamic acid present in a test portion is extracted with perchloric acid solution. The extract is centrifuged, decanted and filtered, and diluted to appropriate concentration with water, and the pH is adjusted to 10. Nicotinamide adenine dinucleotide (NAD) is reduced by the L-(+)-glutamic acid in the presence of glutamate dehydrogenase, see Formula (1). The resultant reduced nicotinamide adenine dinucleotide (NADH) reacts with iodonitrotetrazolium chloride in the presence of diaphorase, see Formula (2). The resulting formazane is measured at a wavelength of 490 nm and the free L-(+)-glutamic acid content of the test sample is calculated.

L-(+)-glutamic acid/glutamate + NAD⁺ +
$$H_2O \leftarrow \frac{glutamate dehydrogenase}{}$$

 α -ketoglutamate + NADH + NH₃ + H⁺

PROOF/ÉPREUVE

(1)

 $\xrightarrow{\text{diaphorase}} \text{NAD}^+ + \text{formazane}$ NADH + iodonitrotetrazolium chloride +H⁺ \leftarrow (2)

Sampling 5

Sampling is not part of the method specified in this document. A recommended sampling method is given in CAC/GL 50-2004.

It is important that the sample received by the laboratory is truly representative and has not been damaged or changed during transport or storage.

Start from a representative sample of at least 200 g. Store the sample at a temperature of 4 °C or frozen at –18 °C if not immediately analysed, so that deterioration and change in composition are prevented.

Preparation of test sample 6

Homogenize the laboratory sample with the appropriate equipment (see 7.2.1). Take care that the temperature of the sample does not exceed 25 °C. If a mincer is used, pass the sample at least twice through the equipment.

Fill a suitable airtight container with the prepared sample. Close the container and store at a temperature of 4 °C or frozen at -18 °C if not immediately analysed, so that deterioration and change in composition are prevented. Analyse the sample as soon as practicable, but always within 24 h after homogenization. iTeh STANDARD PREVIEW

(standards.iteh.ai)

Test method of spectrophotometer 7

ISO/PRF 4134

7.1 Reagents https://standards.iteh.ai/catalog/standards/sist/fa1946fc-a311-4798-b899-

68cfd69da41f/iso-prf-4134 Only reagents of recognized analytical grade and only water of at least grade 2 purity as defined in ISO 3696 shall be used. Except for the solutions of inorganic compounds (7.1.1 and 7.1.2), store all solutions in stoppered brown glass bottles which have been scrupulously cleaned and steamed or sterilized.

7.1.1 Dilute perchloric acid, c = 1.0 mol/l.

WARNING — Contact with oxidizable or combustible materials or with dehydrating or reducing agents can result in fire or explosion. Persons using this acid should be thoroughly familiar with its hazards. See the safety practices listed in Annex A.

Add 8,6 ml of the perchloric acid (70 g/ 100 g, ρ_{20} = 1,67 g/ml) to the bulk of water, diluting to 100 ml.

7.1.2 Potassium hydroxide solution, $c = 4 \mod 1$, $2 \mod 1$, $0,5 \mod 1$ and $0,02 \mod 1$.

Dissolve 22,4 g of potassium hydroxide in water. Dilute the solution to 100 ml, c = 4 mol/l, and mix evenly after cooling.

Dissolve 11,2 g of potassium hydroxide in water. Dilute the solution to 100 ml, c = 2 mol/l, and mix evenly after cooling.

Transfer 2,5 ml of 2 mol/l potassium hydroxide solution to 10 ml volumetric flask, dilute to the mark with water and mix, c = 0.5 mol/l.

Transfer 0,1 ml of 2 mol/l potassium hydroxide solution to 10 ml volumetric flask, dilute to the mark with water and mix, c = 0.02 mol/l.

7.1.3 Solution R1, triethanolamine phosphate buffer solution, pH = 8,6.

Dissolve 1,86 g of triethanolamine hydrochloride in approximately 25 ml of water, adjust the pH to 8,6 with 2 mol/l potassium hydroxide solution (7.1.2), detected by a pH-meter. Add 0,68 g of octylphenol decaethyleneglycol ether (e.g. Triton X-100). Dilute to 100 ml with water and mix (solution A).

Dissolve 0,86 g of dipotassium hydrogen phosphate (K_2HPO_4) and 7 mg of potassium dihydrogen phosphate (KH_2PO_4) in water. Dilute to 100 ml with water and mix evenly (solution B).

Mix 20 ml of solution A with 5 ml of solution B.

The solution is stable for two months when stored at a temperature of between 0 °C and 6 °C.

7.1.4 Solution R2, the mixed solution of NAD and diaphorase (lipoamide dehydrogenase EC¹) **1.8.1.4**), $\rho_{\text{NAD}} = 11 \text{ mg/ml}$, diaphorase, approximately 4 IU/ml.

Weigh 110 mg of NAD, and approximately 8 mg (approximately 40 IU) of diaphorase in a stoppered flask. Add 10,0 ml water and mix evenly.

The solution is stable for one week when stored in the dark at a temperature of between 0 °C and 6 °C.

7.1.5 Solution R3, iodonitrotetrazolium chloride (INT) solution, 2-(4-iodophenyl)-3-(4-introphenyl)-5-phenyltetrazolium chloride, $\rho = 0.6$ mg/ml.

Weigh 6 mg of INT in a small, stoppered brown flask. Add 10 ml of water and mix evenly.

The solution is stable for four weeks when stored in the dark at a temperature of between 0 °C and 6 °C. (standards.iteh.ai)

7.1.6 Solution R123, the mixed solution of solution R1, solution R2 and solution R3.

Pipette 6 ml of the solution R1, 2 ml of the solution R2 and 2 ml of the solution R3 in a stoppered brown glass bottle, and mix evenly before the test

The mixed solution is stable for 1 h in stoppered brown glass bottles at room temperature.

7.1.7 Glutamate dehydrogenase (GLDH) solution (EC¹ 1.4.1.3), approximately 900 IU/ml.

Weigh 10 mg (approximately 900 IU) of lyophilized glutamate dehydrogenase (GLDH) in a small stoppered flask. Add 1 ml water and mix.

Free from ammonium sulfate, ethylene-dinitrilotetraacetic acid (EDTA) and glutaminase, this solution is stable for 12 months when stored at a temperature of between 0 °C and 6 °C.

7.1.8 L-(+)-glutamic acid standard stock solution, $\rho = 1000 \text{ mg/l}$.

Weigh, to the nearest 0,000 1 g, approximately 50,0 mg of L-(+)-glutamic acid ($C_5H_9O_4N$). Dissolve it in approximately 25 ml of water.

Adjust the pH to 5 to 6 with a few drops of 2 mol/l potassium hydroxide solution (7.1.2). Then adjust the pH to 7,0 slowly with 0,02 mol/l potassium hydroxide solution (7.1.2). Dilute to 50 ml with water and mix evenly.

The solution is stable for six months when stored at a temperature of between 0 °C and 6 °C.

¹⁾ The EC number refers to the Enzyme Classification number as given in: The International Union of Biochemistry, Enzymenomenclature, Elsevier, Amsterdam, 1965.

7.1.9 L-(+)-glutamic acid standard solution, $\rho = 100 \text{ mg/l}$.

Pipette accurately 5,0 ml of L-(+)-glutamic acid standard stock solution (7.1.8) into a 50 ml volumetric flask (7.2.8), dilute to the mark with water and mix evenly.

The solution is with the current use.

7.1.10 L-(+)-glutamic acid series standard solution, $\rho = 5 \text{ mg/l}$, 10 mg/l, 15 mg/l, 20 mg/l, 30 mg/l and 40 mg/l.

Pipette accurately 0,50 ml, 1,00 ml, 1,50 ml, 2,00 ml, 3,00 ml and 4,00 ml of L-(+)-glutamic acid standard solution (7.1.9) into each of six 10 ml volumetric flasks (7.2.8) separately, dilute to the mark with water and mix evenly.

The solution is with the current use.

7.2 Apparatus

The usual laboratory equipment and, in particular, the following shall be used.

7.2.1 Mechanical or electrical equipment, capable of homogenizing the laboratory sample.

This includes a high-speed rotational cutter, or a mincer fitted with a plate with apertures not exceeding 4,0 mm in diameter.

7.2.2 Laboratory mixer, stirrer or oscillator.

(standards.iteh.ai)

7.2.3 Laboratory centrifuge, with 50 ml or 100 ml centrifuge tubes, operating at a radial acceleration of about 2 000 gn or equivalent speed (e.g. 3 500 n/min/to/4 0/00 r/min).

https://standards.iteh.ai/catalog/standards/sist/fa1946fc-a311-4798-b899-

7.2.4 Analytical balance, capable of weighing to the nearest 0,001 g, 0,000 1 g.

- 7.2.5 Constant temperature drying box.
- 7.2.6 pH-meter.

7.2.7 Filter papers, diameter of about 15 cm, high or moderate speed.

7.2.8 One-mark volumetric flasks, capacities of 10 ml, 50 ml and 100 ml, conforming to ISO 1042, class B standard.

7.2.9 Single-volume pipettes, capacities of 50 ml, 25 ml and 1 ml conforming to ISO 648, class B standard.

7.2.10 Single channel or multi-channel transferring pipettes and tips, of 5 ml, 1 000 µl, 200 µl and 100 µl, conforming to ISO 8655-2.

7.2.11 Small plastic spatula or lid, for mixing the content evenly by stirring with the spatula in the cuvette or by shaking the cuvette covered with the lid.

7.2.12 Photoelectric colorimeter, provided with a filter having a transmittance maximum at a wavelength of 490 nm, or **spectrometer**.

7.2.13 Cuvettes, of 10 mm optical path length.

7.3 Procedure

7.3.1 General

If it is required to check whether the repeatability requirement is met, two individual determinations should be performed.

7.3.2 Test portion

Weigh, to the nearest 0,001 g, approximately 25 g, or other appropriate mass (m_1) of the test sample (see <u>Clause 6</u>).

7.3.3 Preparation of extract

7.3.3.1 Add 50 ml of dilute perchloric acid solution (7.1.1) to the test sample, homogenize the mixture with the laboratory mixer (7.2.2).

7.3.3.2 Transfer the homogenized sample to centrifuge tube (7.2.3). Centrifuge for 10 min at 10 °C, 2 000 gn or equivalent speed (e.g. 3 500 r/min to 4 000 r/min). Carefully move aside the fat layer and decant all the supernatant liquid through a filter paper (7.2.7) into a 100 ml conical flask. Discard the first 10 ml of the filtrate.

7.3.3.3 Transfer 25 ml of the solution (which should be only slightly turbid) with pipette (7.2.9) into centrifuge tube (7.2.3). With detected by the pH meter (7.2.6), adjust the pH to 7 to 8 with 4 mol/l potassium hydroxide solution (7.1.2), then adjust the pH to 10,0 slowly with 2 mol/l and 0,5 mol/l potassium hydroxide solution (7.1.2). Centrifuge for 3 min at 2 000 gn or equivalent speed (e.g. 3 500 r/min to 4 000 r/min).

ISO/PRF 4134

NOTE If the pH is slightly above 10,0 ait can be adjusted with dilute perchloric acid back to the required pH value (7.1.1). 68cfd69da41f/iso-prf-4134

7.3.3.4 Transfer all the supernatant into a 50 ml volumetric flask (7.2.8). Dilute to the mark with water and mix.

7.3.3.5 Cool the solution in ice for 10 min, and filter through a filter paper (<u>7.2.7</u>). Discard the first 10 ml of the filtrate.

7.3.3.6 Pipet 5 ml, or some other appropriate volume (V_1) of the filtrate into a 50 ml volumetric flask (7.2.8). Dilute to the mark with water and mix. The solution obtained will be used to determine the content of free L-(+)-glutamic acid in the test portion.

The volume V_1 should be chosen so that the L-(+)-glutamic acid content of the solution is between 8 mg/l and 40 mg/l.

7.3.4 Determination

7.3.4.1 Preparation of detection instrument

Set up the spectrophotometer (7.2.12) and preheat the instrument according to the instrument specification until equilibrium conditions are achieved. Set the detection wavelength to 490 nm. Adjust the baseline of the equipment to zero with pure water.