



**SLOVENSKI STANDARD**  
**oSIST prEN ISO 15192:2024**  
**01-februar-2024**

---

**Tla in odpadki - Določanje kroma Cr (VI) v trdnem mediju z alkalnim razklopom in ionsko kromatografijo s spektrofotometrijsko detekcijo (ISO/DIS 15192:2023)**

Soil and waste - Determination of Chromium(VI) in solid material by alkaline digestion and ion chromatography with spectrometric detection (ISO/DIS 15192:2023)

Boden und Abfall - Bestimmung von sechswertigem Chrom in Feststoffen durch alkalischen Aufschluss und Ionenchromatographie mit photometrischer Detektion (ISO/DIS 15192:2023)

Déchets et sols - Dosage du chrome(VI) dans les matériaux solides par digestion alcaline et chromatographie ionique avec détection spectrométrique (ISO/DIS 15192:2023)

<https://standards.iteh.ai>

<https://standards.iteh.ai> Ta slovenski standard je istoveten z: [6aff](https://standards.iteh.ai) prEN ISO 15192 [57bcf2b039/osist-pren-iso-15192-2024](https://standards.iteh.ai)

---

**ICS:**

13.080.10	Kemijske značilnosti tal	Chemical characteristics of soils
71.040.50	Fizikalnokemijske analitske metode	Physicochemical methods of analysis

**oSIST prEN ISO 15192:2024**

**en,fr,de**



# DRAFT INTERNATIONAL STANDARD

## ISO/DIS 15192

ISO/TC 190/SC 3

Secretariat: DIN

Voting begins on:  
2023-12-25Voting terminates on:  
2024-03-18

---

---

## Soil and waste — Determination of Chromium(VI) in solid material by alkaline digestion and ion chromatography with spectrometric detection

*Déchets et sols — Dosage du chrome(VI) dans les matériaux solides par digestion alcaline et chromatographie ionique avec détection spectrométrique*

ICS: 13.080.10

### iTeh Standards (<https://standards.iteh.ai>) Document Preview

[oSIST prEN ISO 15192:2024](https://standards.iteh.ai/catalog/standards/sist/11bc6aff-420d-40d8-ae00-0f57bcf2b039/osist-pren-iso-15192-2024)<https://standards.iteh.ai/catalog/standards/sist/11bc6aff-420d-40d8-ae00-0f57bcf2b039/osist-pren-iso-15192-2024>

This document is circulated as received from the committee secretariat.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

### ISO/CEN PARALLEL PROCESSING



Reference number  
ISO/DIS 15192:2023(E)

© ISO 2023

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[oSIST prEN ISO 15192:2024](https://standards.iteh.ai/catalog/standards/sist/11bc6aff-420d-40d8-ae0-0f57bcf2b039/osist-pren-iso-15192-2024)

<https://standards.iteh.ai/catalog/standards/sist/11bc6aff-420d-40d8-ae0-0f57bcf2b039/osist-pren-iso-15192-2024>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

Foreword.....	iv
Introduction.....	v
<b>1 Scope.....</b>	<b>1</b>
<b>2 Normative references.....</b>	<b>1</b>
<b>3 Terms and definitions.....</b>	<b>1</b>
<b>4 Safety remarks.....</b>	<b>1</b>
<b>5 Principle.....</b>	<b>2</b>
5.1 Digestion.....	2
5.2 Determination.....	2
5.3 Interferences and sources of error.....	2
<b>6 Apparatus.....</b>	<b>3</b>
<b>7 Reagents.....</b>	<b>3</b>
<b>8 Sample pretreatment.....</b>	<b>6</b>
<b>9 Alkaline digestion procedure.....</b>	<b>6</b>
9.1 General.....	6
9.2 Preparation of test solutions using a hotplate or heating block.....	6
<b>10 Analytical procedure.....</b>	<b>7</b>
10.1 General information.....	7
10.2 Instrumental set-up.....	7
10.3 Calibration.....	7
10.4 Test solution measurement.....	7
10.5 Quality control.....	8
10.5.1 General.....	8
10.5.2 Blank test solution.....	8
10.5.3 Verification of method.....	8
10.5.4 Duplicate samples.....	8
10.5.5 Soluble Cr(VI) spiked samples.....	9
10.5.6 Cr(III) spiked samples.....	9
10.5.7 Interpretation of quality control data.....	9
<b>11 Calculation.....</b>	<b>9</b>
<b>12 Expression of results.....</b>	<b>10</b>
<b>13 Test report.....</b>	<b>10</b>
<b>Annex A (informative) Ion chromatographic system.....</b>	<b>11</b>
<b>Annex B (informative) Requirements for test portion preparation.....</b>	<b>13</b>
<b>Annex C (informative) Validation.....</b>	<b>14</b>
<b>Annex D (informative) Background on methods for the determination of Cr(VI) in solid samples.....</b>	<b>19</b>
<b>Bibliography.....</b>	<b>23</b>

## ISO/DIS 15192:2023(E)

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical and physical characterization*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 444, *Environmental Characterization*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 15192:2010), which has been technically revised.

The main changes compared to the previous edition are as follows:

- integration of the determination of the total chromium in the alkaline digestion solution;
- Addition of barium chromate as an alternative to lead chromate for the verification of the method in [clause 10.5.3](#) ;
- the text has been editorially revised, including updating of references.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Under environmental conditions chromium in compounds exists in the trivalent, Cr(III), or the hexavalent, Cr(VI) state. Chromium is an essential trace element for mammals, including man, whereas it is presumed that Cr(VI) compounds are genotoxic and potentially carcinogenic in humans. Interconversion of trivalent and hexavalent chromium species can occur during sample preparation and analysis, but these processes are minimised, to the extent possible, by the sample preparation methods prescribed by this document.

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[oSIST prEN ISO 15192:2024](https://standards.iteh.ai/catalog/standards/sist/11bc6aff-420d-40d8-ae00-0f57bcf2b039/osist-pren-iso-15192-2024)

<https://standards.iteh.ai/catalog/standards/sist/11bc6aff-420d-40d8-ae00-0f57bcf2b039/osist-pren-iso-15192-2024>





# Soil and waste — Determination of Chromium(VI) in solid material by alkaline digestion and ion chromatography with spectrometric detection

## 1 Scope

This document specifies the determination of Cr(VI) in solid waste material and soil by alkaline digestion and ion chromatography with spectrometric detection. This method can be used to determine Cr(VI)-mass fractions in solids higher than 0,1 mg/kg.

NOTE In case of reducing or oxidising waste matrix no valid Cr(VI) content can be reported.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods — Part 1: Linear calibration function*

ISO 11464, *Soil quality — Pretreatment of samples for physico-chemical analysis*

ISO 11465, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

EN 15002, *Characterization of waste — Preparation of test portions from the laboratory sample*

EN 15934, *Sludge, treated biowaste, soil and waste — Calculation of dry matter fraction after determination of dry residue or water content*

## 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

## 4 Safety remarks

Anyone dealing with waste and soil analysis shall be aware of the typical risks of the material irrespective of the parameters determined. Waste and soil samples may contain hazardous (e.g. toxic, reactive, flammable, infectious) substances, which can be liable to biological and/or chemical reaction. Consequently, these samples should be handled with special care. The gases which may be produced by microbiological or chemical activity are potentially flammable and can pressurise sealed bottles. Bursting bottles are likely to result in hazardous shrapnel, dust and/or aerosol. It is presupposed that national regulations are followed with respect to all hazards associated with this method.

## ISO/DIS 15192:2023(E)

Avoid any contact with the skin, ingestion or inhalation of Cr(VI) compounds. Cr(VI) compounds are genotoxic and potentially carcinogenic to humans.

## 5 Principle

### 5.1 Digestion

This document describes an alkaline digestion procedure for extracting Cr(VI) from soluble, adsorbed and precipitated forms of chromium compounds in solid waste materials and soil. To quantify the content of Cr(VI) in a solid matrix, three criteria shall be satisfied:

- a) digestion solution shall solubilize all species of Cr(VI);
- b) conditions of the digestion shall not induce reduction of native Cr(VI) to Cr(III);
- c) method shall not cause oxidation of native Cr(III) contained in the sample to Cr(VI).

The alkaline digestion described in this document meets these criteria for a wide spectrum of soils and wastes. Under the alkaline conditions of the digestion, negligible reduction of Cr(VI) or oxidation of native Cr(III) is expected. The addition of  $Mg^{2+}$  in a phosphate buffer to the alkaline solution minimises air oxidation of trivalent chromium<sup>[1][5][8]</sup>.

NOTE Background on methods for the determination of Cr(VI) in solid samples is given in [Annex D](#).

### 5.2 Determination

Quantification of Cr(VI) in the alkaline digestion solution should be performed using a suitable technique with appropriate accuracy. For this purpose, e.g., ion chromatography is used to separate Cr(VI) from interferences. Following this ion chromatographic separation, Cr(VI) is measured spectrophotometrically either at 365 nm (direct UV detection) or after post-column derivatisation with 1,5-diphenylcarbazide in acid solution at 540 nm. Post-column derivatisation involves reaction of 1,5-diphenylcarbazide with Cr(VI) to produce trivalent chromium and diphenylcarbazone. These then combine to form a trivalent chromium-diphenylcarbazone complex containing the characteristic magenta chromagen ( $\lambda_{max} = 540 \text{ nm}$ ).

NOTE The choice of detection method is based upon the required sensitivity. Direct UV detection is less sensitive than detection after post-column derivatisation with 1,5-diphenylcarbazide (see [Annex C](#)).

Hyphenated methods with suitable chromatographic separation and detection techniques, such as inductively coupled plasma mass spectrometry (ICP-MS) or inductively coupled plasma atomic emission spectroscopy (ICP-OES), may be used once validation of the chosen analytical method has been performed.

A preliminary determination of the total chromium in the alkaline digestion solution by such techniques as ICP-MS, ICP-OES, or graphite furnace atomic absorption spectrometry could reduce the number of samples where the determination of Cr(VI) is required. If the content of alkaline-extractable total chromium is lower than the legislative limit for Cr(VI) then the determination of Cr(VI) may be omitted.

NOTE Due to the high element concentrations, e.g. sodium, in the alkaline digestion solutions, the calibration strategy must be adapted appropriately. In many cases, matrix matching of the calibration solutions and dilution of the sample together with addition of internal standards or using the standard addition method is necessary. The analytical method needs to be validated on alkaline digestion solutions prior to routine use.

### 5.3 Interferences and sources of error

- Use of ion chromatography is necessary for the separation of Cr(VI) from possible interferences in the alkaline digestion solution from solid material<sup>[6]</sup>.

- For waste materials or soils, where the Cr(III)/Cr(VI) ratio is expected to be high, Cr(VI) results may be biased due to method induced oxidation. This can be particularly expected in soils high in Mn content and amended with soluble Cr(III) salts or freshly precipitated Cr(OH)<sub>3</sub>[3].
- Cr(VI) can be reduced to Cr(III) during digestion from the sample due to reaction with reducing agents such as e.g. divalent iron. This problem is minimised in the described procedure using alkaline digestion solution[5].
- Cr(III) can be oxidised to Cr(VI) in hot alkaline solutions. This problem is minimised in the described procedure by adding magnesium to the alkaline digestion solution[2][3][5][8].
- Overloading the analytical column capacity with high concentrations of anionic species (e.g. chloride) may cause underestimation of Cr(VI)[9].

## 6 Apparatus

### 6.1 Digestion equipment.

**6.1.1 Hotplate with a magnetic stirrer**, thermostatically controlled with a digestion vessel of 250 ml covered with a watch glass, or

**6.1.2 Heating block with a magnetic stirrer**, thermostatically controlled with a digestion vessel of 250 ml covered with a watch glass.

NOTE Other thermostatically controlled digestion equipment with a magnetic stirrer can be used once validation has been performed.

**6.2 Filtration equipment**, suitable for using 0,45-µm membrane filters.

**6.3 Membrane filters**, 0,45-µm pore size, chemically inert.

**6.4 Ion chromatographic system.**

All components which come into contact with the sample or eluent stream shall be comprised of inert materials, e.g. polyetherether ketone (PEEK), as shall all connecting tubing (see [Annex B](#)).

**6.5 Ion chromatographic column**, suitable for chromate separation with a sufficient ion exchange capacity.

**6.6 Detection system.**

**6.6.1 UV-VIS spectrophotometer**, at 365 nm, or

**6.6.2 VIS spectrophotometer**, at 540 nm after post column derivatisation.

## 7 Reagents

### 7.1 General.

During the analysis, only use reagents of recognised analytical grade, and water as specified in [7.2](#).

### 7.2 Water.

Water with an electrical conductivity less than 0,1 mS m<sup>-1</sup> (equivalent to resistivity greater than 0,01 MΩ m at 25 °C). It is recommended that the water used is obtained from a purification system