INTERNATIONAL STANDARD

ISO 23783-2

First edition 2022-08

Automated liquid handling systems —

Part 2:

Measurement procedures for the determination of volumetric performance

Systèmes automatisés de manipulation de liquides —

Partie 2: Procédures de mesure pour la détermination des performances volumétriques

ISO 23783-2:2022

https://standards.iteh.ai/catalog/standards/sist/f79dd62c-9363-4813-87c0/5a4d4864a81d/iso-23783-2-2022



iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 23783-2:2022 https://standards.iteh.ai/catalog/standards/sist/f79dd62c-9363-4813-87c0-5a4d4864a81d/iso-23783-2-2022



COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

Co	ntent	is a second of the second of t	Page
Fore	word		iv
Intr	oductio	on	v
1	Scor	De	1
2	-	mative references	
3		ns and definitions	
4		reviated terms	
5		surement methods	
J	5.1 5.2	Overview of methods suitable for measuring ALHS performance Photometric methods 5.2.1 Dual-dye ratiometric photometric method 5.2.2 Single-dye photometric method	
	5.3 5.4 5.5	5.2.3 Fluorescence method Gravimetric methods 5.3.1 Single channel method 5.3.2 Regression analysis Hybrid photometric/gravimetric method Dimensional methods 5.5.1 Optical image analysis of droplets 5.5.2 Optical image analysis of capillaries	9 9 10 10
6	Equi 6.1 6.2 6.3	ipment and preparation Test equipment Manually operated single- and multi-channel pipettes Preparation for testing	11 11 12
7	The	rmal expansion ISO 23783-2:2022	13
8		ttps://standards.iteh.ai/catalog/standards/sist/f79dd62c-9363-4813-87c0-ceability and measuring system uncertainty Traceability Estimation of measuring system uncertainty 8.2.1 Whole system approach 8.2.2 Measurement model approach	
9	Rep	orting	14
Ann	ex A (n	ormative) Calculation of liquid volumes from balance readings	15
Ann	ex B (n	ormative) Dual-dye ratiometric photometric procedure	18
Ann	ex C (no	ormative) Single dye photometric procedure	24
	-	ormative) Gravimetric procedure, single channel measurement	
		ormative) Gravimetric regression procedure	
		ormative) Photometric/gravimetric hybrid procedure	
		ormative) Optical image analysis of droplets	
	-	ormative) Fluorescence procedure	
		ormative) Optical image analysis of capillaries	
		hv	
	IVEL GILL	41 V	/ U

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*.

This first edition of ISO 23783-2, together with ISO 23783-1 and ISO 23783-3, cancels and replaces IWA 15:2015. https://standards.iteh.ai/catalog/standards/sist/179dd62c-9363-4813-87c0-

A list of all parts in the ISO 23783 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Globalization of laboratory operations requires standardized practices for operating automated liquid handling systems (ALHS), communicating test protocols, as well as analysing and reporting of performance parameters. IWA 15:2015 was developed to provide standardized terminology, test protocols, and analytical methods for reporting test results. The concepts developed for, and described in, IWA 15 form the foundation of the ISO 23783 series.

Specifically, this document addresses the needs of:

- users of ALHS, as a basis for calibration, verification, validation, optimization, and routine testing of trueness and precision;
- manufacturers of ALHS, as a basis for quality control, communication of acceptance test specifications and conditions, and issuance of manufacturer's declarations (where appropriate);
- test houses and other bodies, as a basis for certification, calibration, and testing.

The tests established in this document should be carried out by trained personnel.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 23783-2:2022 https://standards.iteh.ai/catalog/standards/sist/f79dd62c-9363-4813-87c0-5a4d4864a81d/iso-23783-2-2022

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 23783-2:2022

https://standards.iteh.ai/catalog/standards/sist/f79dd62c-9363-4813-87c0-5a4d4864a81d/iso-23783-2-2022

Automated liquid handling systems —

Part 2:

Measurement procedures for the determination of volumetric performance

1 Scope

This document specifies procedures for the determination of volumetric performance of automated liquid handling systems (ALHS), including traceability and estimations of measurement uncertainty of measurement results.

This document is applicable to all ALHS with complete, installed liquid handling devices, including tips and other essential parts needed for delivering a specified volume, which perform liquid handling tasks without human intervention into labware.

NOTE For terminology and general requirements of automated liquid handling systems, see ISO 23783-1. Determination, specification, and reporting of volumetric performance of automated liquid handling systems is described in ISO 23783-3.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, Water for analytical laboratory use — Specification and test methods

ISO 8655-6, Piston-operated volumetric apparatus – Part 6: Gravimetric reference measurement procedure for the determination of volume

ISO 23783-1, Automated liquid handling systems — Part 1: Terminology and general requirements

ISO 23783-3, Automated liquid handling systems — Part 3: Determination, specification, and reporting of volumetric performance

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 23783-1 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at https://www.electropedia.org/

4 Abbreviated terms

For the purposes of this document, the abbreviated terms given in ISO 23783-1 apply.

5 Measurement methods

5.1 Overview of methods suitable for measuring ALHS performance

When choosing a test method for an ALHS, its suitability for the specific test situation shall be evaluated. This evaluation shall consider the systematic and random error requirements of the ALHS to which the test method is being applied. The selected test method shall be adequate to evaluate whether the ALHS performance is fit for its intended purpose.

NOTE 1 Fitness for purpose is a foundational concept and closely related to the process of metrological confirmation as described in ISO 9000 and ISO 9001.

The test method shall have a sufficiently small measuring system uncertainty (MSU) for the specific test situation. The MSU should be determined in accordance with a suitable approach (see 8.2 for more detail).

NOTE 2 The measurement model approach for estimating MSU is described in ISO/IEC Guide 98-3 and the measurement system approach is described in EURACHEM/CITAG Guide CG 4 [4].

Table 1 is intended to provide an overview of methods suitable for determining the volumetric performance of ALHS. It provides cross-references between the method abstracts from 5.2 to 5.5, and the corresponding procedures in Annexes B to I. It further describes the volume ranges, plate and liquid types which can be used for testing ALHS performance with a given method. It also lists typical systematic and random errors achievable if a test procedure is exactly followed as described in its respective annex. The suitability of a method for a given test situation may also be determined by the required equipment or environmental conditions under which it needs to be carried out.

Only key test equipment is listed in <u>Table 1</u>, while test equipment to monitor liquid and air temperatures, relative humidity, and barometric pressure is required for each procedure, as specified in the corresponding annexes.

ISO 23783-2:2022 https://standards.iteh.ai/catalog/standards/sist/f79dd62c-9363-4813-87c0-55444864581d/iso-23783-2-2022

Table 1 — Test methods for ALHS

Method Ref.	Method	Liquid type	Plate type	Volume range	Typical systematic error ^a	Typical random error a	Environmental conditions b	Test equipment
			wells	μl	1%	%		
Photome	Photometric methods				:://s	ĺ		
5.2.1		Aqueous, 96	96	0,1 to 350,0	2,0 to 3,0	0,15 to 0,25	0,15 to 0,25 Temperature:	— Microplate absorbance reader capable
Annex B	etric photometric method	DMS0 c	384	0,01 to 55,0	2,5 to 5,5	0,35 to 0,55	Aqueous:	of measuring absorbance at 520 nm and 730 nm;
					ds.i		15 °C to 30 °C	
					teh. 5		DMSO c:	— dimensionally characterized 96- or 384-well microplates with optically
					ai/c a4d	A ta	19 °C to 30 °C	clear bottom;
					<u>I:</u> atal 1486	n	RH d: 20 % to 90 %	 calibration plate for plate reader;
					<u>SO 2</u> og/si 4a81	D)		— microplate shaker;
					2 <u>378</u> tand l d/is	A		— balance;
					3-2:20 ards/si o-2378	kD ls.it		— spectrophotometer capable of measuring absorbance at 520 nm and
					<u>22</u> st/f 83-2	te		730 nm;
					79d 2-20	'K h.		— pH meter;
					d62)22	ai		 volumetric flasks.
a Typica	Typically, larger test volumes lead to smaller errors.	s lead to sm	aller errors.		c-9	i)		
b The m humidity o	b The minimum temperature of the test environment sh humidity of the test environment shall be non-condensing.	of the test er shall be no	nvironment sh n-condensing.	all be above the m	elting point of t	the test liquid,	ensuring that it will no	The minimum temperature of the test environment shall be above the melting point of the test liquid, ensuring that it will not solidify at any point during the test. The relative idity of the test environment shall be non-condensing.
c Dimet	Dimethylsulfoxide.							
d Relativ	Relative humidity.							

Table 1 (continued)

Method Ref.	Method	Liquid type	Plate type	Volume range	Typical systematic error ^a	Typical ic random error a	Environmental conditions ^b	Test equipment
			wells	μl	%	%		
5.2.2	Single-dye photo-	Aqueous	96	1,0 to 100,0	3	1,5	Temperature:	— Microplate absorbance reader capable
Annex C	metric method		384	0,25 to 20,0	http ന	1,5	15 °C to 30 °C	of measuring absorbance at 492 nm
)S:/		RH d. 40 % to 70 %	and Ozo min,
					//st		0/ 0/ 0/ 0/ 0/ 0/ 0/ 0/ 0/ 0/ 0/ 0/ 0/ 0	— 96- or 384-well microplates with
					and	- C		optically clear bottom;
					lard	11		— balance;
					s.ite	(Q.	— magnetic stirrer;
					h.a 5a	st		micronlato chalzor.
					i/c 4d	A	<u></u>	— iiiici upiate siiakei,
					<u>l</u> ata 148 <i>6</i>	ın		— pH meter;
					. <u>5U</u> log/ 54a8	da		— manual pipettes;
					<u>23</u> stai 31d			0.000
					/ 8 nd /is			— Volumetric Hasks.
a Typic	Typically, larger test volumes lead to smaller errors.	s lead to sn	naller errors.					
b The n humidity	b The minimum temperature of the test environment sh humidity of the test environment shall be non-condensing.	of the test e t shall be no	environment short-condensing.	all be above the m	elting point	of the test liquic	l, ensuring that it will not	The minimum temperature of the test environment shall be above the melting point of the test liquid, ensuring that it will not solidify at any point during the test. The relative idity of the test environment shall be non-condensing.
c Dime	Dimethylsulfoxide.)					
d Relati	Relative humidity.							
					62 22	ai		

Table 1 (continued)

Method Ref.	Method	Liquid type	Plate type	Volume range	Typical systematic error ^a	Typical random error a	Environmental conditions ^b	Test equipment
			wells	μl	%	%		
5.2.3	Fluorescence meth- Aqueous,	Aqueous,	384	0,001 to 0,015		8>	Temperature:	- Microplate fluorescence reader with
Annex H	po	DMS0 с	1 536	0,001 to 0,015	stan &	8>	17 °C to 27 °C	excitation wavelength at 494 nm and emission analysis at 521 nm:
					dar		RH d: non-condensing	
					ds.it	5		— 384- or 1536-well fluorescence microplates;
					eh.ai 5a	st		— balance;
					<u>IS</u> /catal 4d486	and		 bulk liquid dispenser or multi-channel pipette;
					SO 2. og/st 4a81	da	D	microplate shaker;
					3783 anda d/isc	A.F.		— pH meter;
					3-2:2 rds/s 3-23	S.i		— manual pipettes;
					022 sist/ 783	ite		volumetric flasks.
a Typic	Typically, larger test volumes lead to smaller errors.	s lead to sm	naller errors.		f79 2-2	h		
b The n humidity	b The minimum temperature of the test environment sh humidity of the test environment shall be non-condensing.	of the test e t shall be no	nvironment shor-condensing.	all be above the me	elting point of	the test liquid	ensuring that it will not	The minimum temperature of the test environment shall be above the melting point of the test liquid, ensuring that it will not solidify at any point during the test. The relative idity of the test environment shall be non-condensing.
c Dimet	Dimethylsulfoxide.							
d Relati	Relative humidity.				363			
					-4813-87	L W		

Table 1 (continued)

Method Ref.	Method	Liquid type	Plate type	Volume range	Typical systematic error a	Typical random error a	Environmental conditions b	Test equipment
			wells	μ	%	%		
Gravime	Gravimetric methods							
5.3.1	Single channel anal- Any	Any	n/a	0,5 to <20	>1,4 pttp	9,0≥	Temperature:	— Balance;
Annex D	ysis			20 to <200	s:/o>	≤0,3	17 °C to 30 °C	density meter.
				200 to 1 000	stand 6′0≤	€0,3 —	RH ^d : 45 % to 80 %	anti-electrostatic equipment;
					lards			— anti-vibration table;
					.iteh.a 5a			 temperature- and humidity control for environment;
					<u>IS</u> i/catalo 4d486		A 78.T	 draft shield or draft-free environment for balance.
5.3.2	Regression analysis Any	Any	n/a	<0,015	20 to 50	<10	Temperature:	— Balance;
Annex E				0,015 to <0,050	2 to 5	2,5 to 5	17 °C to 27 °C	density meter with 6 decimal places:
				0,050 to 1	0,5 to 2	<0,5	RH d: 45 % to 80 %	anti-oloctroctotic ominment:
					2:202 ds/sis 2378		Barometric pressure: 600 hPa to 1 100 hPa	anti-vibration table;
					<u>:2</u> t/f79dd 3-2-202			temperature- and humidity control for environment;
					162c-9. 22			 draft shield or draft-free environment for balance.
a Typica	Typically, larger test volumes lead to smaller errors	s lead to sm	naller errors.		363			
b The m	b The minimum temperature of the test environment shumidity of the test environment shall he non-condensing	of the test e	nvironment sh	all be above the mo	elting point of t	the test liquid	ensuring that it will not s	The minimum temperature of the test environment shall be above the melting point of the test liquid, ensuring that it will not solidify at any point during the test. The relative idity of the test environment shall be non-condensing
tominary of	the cost cirent comments							
	Dimethylsunoxide.							
d Relati	Relative humidity.				c0			

Table 1 (continued)

					Table T (concuraca)	nemacaj		
Method Ref.	Method	Liquid type	Plate type	Volume range	Typical systematic error ^a	Typical random error a	Environmental conditions ^b	Test equipment
			wells	μl	%	%		
Photome	Photometric/gravimetric hybrid method	brid met	hod		5://s	1	• 17	
5.4	Tartrazine as	Aqueous	96	1,0 to 300,0	0,2 to 0,8	0,5 to 1,0	Temperature:	— Balance;
<u>Annex F</u>					dards	n		— microplate absorbance reader
					s.iteh. 5	(S)		following wavelengths, depending on the chromophore used:
			384	1,0 to 20,0	0,4 to 1,0	0,9 to 1,5	17 °C to 30 °C	— 4-nitrophenol: 405 nm and 620 nm,
	4-nitrophenol as	Aqueous	96	10 to 1 000	<1 to 5	1 to 2	Temperature stability:	— Tartrazine: 450 nm and 620 nm,
	chromophore			5 to 250	[<u>SO</u> log/ 54a	d	2° €.0± >	
			96, 384	1 to 60	<1 to 5	1 to 2	RH d: 45 % to 80 %	— Orange G: 492 nm and 620 nm
				0,5 to 25	783 nda /iso	d'd	RH ^d stability:	
					-2:: rds. -23	S.	< ±10 %	
			384, 1 536	0,1 to 8	<2 to 10	2 to 5		— microplate shaker;
					<u>2</u> :/f79 3-2-2	eh		— 96- or 384-well microplates with
					dd6 2022	.a		optically clear bottom;
					52c- 2	i)		— manual pipettes;
					936.	1		— centrifuge tubes 1,5 ml;
					3-48	L.		— anti-vibration table;
	Orange Gas chromo- Aqueous phore	Aqueous	96	1 to 100	<1 to 5	1,5		 temperature- and humidity control for environment;
			384	1 to 50	<1 to 5	1,5		 draft shield or draft-free environment for balance.
a Typica	Typically, larger test volumes lead to smaller errors.	s lead to sm	naller errors.					

b The minimum temperature of the test environment shall be above the melting point of the test liquid, ensuring that it will not solidify at any point during the test. The relative humidity of the test environment shall be non-condensing.

Dimethylsulfoxide.

Relative humidity.

Table 1 (continued)

Method Ref.	Method	Liquid type	Plate type	Volume range	Typical systematic error ^a	Typical random error a	Environmental conditions ^b	Test equipment
			wells	μl	%	%		
Dimensi	Dimensional methods							
5.5.1 Annex G	Optical image anal- Any ysis of droplets	Any	n/a	Free flying droplets of $V < 0.5 \mu l$	https://	<2	Temperature: (20±3) °C or (27±3) °C	 Stroboscopic camera or high-speed camera;
					'star		RH d: 50 % to 80 %	 automatic image detection software.
5.5.2	Optical image analy- Any	Any	n/a	0,1 to 1,0	10a	<7	Temperature:	— Flatbed scanner;
Annex I	sis of capillaries			>1,0 to 1 000	rds.i	<4	15 °C to 35 °C	— image analysis software;
					iteh.a 5a	(S1	RH d: 15 % to 90 %	 specialized plates with capillaries.
a Typica	Typically, larger test volumes lead to smaller errors.	s lead to sm	naller errors.		ıi/c ı4d	a		
b The m	b The minimum temperature of the test environment sh	of the test e	nvironment sh	all be above the me	elting point of t	the test liquid,	ensuring that it will not	The minimum temperature of the test environment shall be above the melting point of the test liquid, ensuring that it will not solidify at any point during the test. The relative idity of the test environment shall be non-condensing
c Dimet	Dimethylsulfoxide.							
d Relati	Relative humidity.							
					3-2:2022 ards/sist/f79dd62c-9363-48 o-23783-2-2022	RD PREVIE ls.iteh.ai)		

5.2 Photometric methods

5.2.1 Dual-dye ratiometric photometric method

This method allows the determination of volumes of aqueous test liquids from 0,1 μ l to 350 μ l in 96-well plates, and from 0,01 μ l to 55 μ l in 384-well plates. Volumes of dimethylsulfoxide (DMSO)-based test liquids can be determined from 0,11 μ l to 10 μ l in 96-well plates, and from 0,01 μ l to 2,5 μ l in 384-well plates.

This method is suitable to determine the performance of ALHS with up to 384 channels. The operating environment for this method is 15 °C to 30 °C (19 °C to 30 °C for DMSO liquids), and it is not dependent on the ambient relative humidity and barometric pressure at the test location. Further information on the effect of relative humidity and barometric pressure on this method can be found in Reference [5].

Traceability of the measurement results to the International System of Units (SI) is achieved through the use of a calibrated microplate absorbance reader, dimensionally characterized microplates, calibrated balance, and calibrated volumetric glassware.

The procedure for the dual-dye ratiometric photometric method specified in <u>Annex B</u> shall be followed.

5.2.2 Single-dye photometric method

This method is suitable for evaluating the volumetric performance of ALHS with up to 384 channels using aqueous test liquids. Volumes from 1 μ l to 100 μ l can be measured in 96-well plates, and from 0,25 μ l to 20 μ l in 384-well plates.

Traceability of the measurement results to the SI is achieved through the use of a calibrated balance, calibrated pipettes, a calibrated microplate absorbance reader, and calibrated volumetric glassware.

The procedure for the single-dye photometric method specified in Annex C shall be followed.

5.2.3 Fluorescence method .ai/catalog/standards/sist/f79dd62c-9363-4813-87c0-

This method is suitable to evaluate the volumetric performance of ALHS delivering volumes smaller than 15 nl. The fluorescence of the test liquid of fluorescein in DMSO is measured in 384-well or 1536-well microplates, which are specifically suited for fluorescence measurements.

This method is intended to be used for non-contact liquid delivery devices (e.g. acoustic, dispensing valves, or inkjet-type technology) that deliver the liquid volume as free flying droplets or jets into the wells of the microplate.

Traceability of the measurement results to the SI is achieved through the use of a calibrated fluorescence microplate reader, calibrated balance, calibrated pipettes, and calibrated volumetric glassware.

The procedure for the fluorescence method specified in Annex H shall be followed.

5.3 Gravimetric methods

5.3.1 Single channel method

This method describes the apparatus, procedure and reference material for recording measurements with the gravimetric method. A single pan balance is used to take a measurement from a single channel at a time. The following accommodations shall be made:

- placement of the balance and the weighing vessel which reduce draft and vibrations to a suitable level;
- control of the environmental conditions affecting the mass to volume conversion of the measurement (temperature and relative humidity);