

---

---

**Fireworks — Test methods for  
determination of specific chemical  
substances —**

**Part 9:  
Mercury content by hydride  
generation atomic fluorescence  
spectrometry**

*Artifices de divertissement — Méthodes d'essai pour la détermination  
de substances chimiques spécifiques —*

*Partie 9: Teneur en mercure par spectrométrie de fluorescence  
atomique par génération d'hydrures*

ISO 22863-9:2021

<https://standards.itech.ai/catalog/standards/iso/a2b097a8-29d7-44c5-97fe-8d071ad87dde/iso-22863-9-2021>



iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 22863-9:2021](https://standards.iteh.ai/catalog/standards/iso/a2b097a8-29d7-44c5-97fe-8d071ad87dde/iso-22863-9-2021)

<https://standards.iteh.ai/catalog/standards/iso/a2b097a8-29d7-44c5-97fe-8d071ad87dde/iso-22863-9-2021>



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Principle of the method</b> .....	<b>1</b>
<b>5 Reagents</b> .....	<b>1</b>
<b>6 Apparatus</b> .....	<b>2</b>
<b>7 Test procedure</b> .....	<b>3</b>
7.1 Sample pre-treatment, digestion and preparation of the solution to be tested .....	3
7.2 Test conditions.....	3
<b>8 Calculations</b> .....	<b>3</b>
<b>9 Accuracy</b> .....	<b>4</b>
<b>10 Other</b> .....	<b>4</b>
<b>11 Test report</b> .....	<b>4</b>
<b>Annex A (informative) Standard addition method</b> .....	<b>5</b>

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 22863-9:2021](https://standards.iteh.ai/catalog/standards/iso/a2b097a8-29d7-44c5-97fe-8d071ad87dde/iso-22863-9-2021)

<https://standards.iteh.ai/catalog/standards/iso/a2b097a8-29d7-44c5-97fe-8d071ad87dde/iso-22863-9-2021>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 264, *Fireworks*.

A list of all the parts in the ISO 22863 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

# Fireworks — Test methods for determination of specific chemical substances —

## Part 9:

## Mercury content by hydride generation atomic fluorescence spectrometry

### 1 Scope

This document specifies the test method for the determination of the mercury content in pyrotechnic compositions by hydride generation atomic fluorescence spectrometry.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 22863-1, *Fireworks — Test methods for determination of specific chemical substances — Part 1: General*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 22863-1 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <http://www.electropedia.org/>

### 4 Principle of the method

The sample is heated and digested in a boiling water bath using a nitric acid - hydrochloric acid mixed reagent. In the acidic medium, the mercury in the sample is reduced to atomic mercury by potassium borohydride, and then loaded into the atomic fluorescence photometer by a carrier gas (argon). Under the irradiation of a mercury hollow cathode lamp, the mercury atoms will emit fluorescence with a characteristic wavelength when they transit from high energy state to ground state. The fluorescence intensity is proportional to the mercury concentration in the liquid to be measured and is quantitatively compared with mercury standard solutions.

### 5 Reagents

#### 5.1 Hydrochloric acid (GR)

#### 5.2 Nitric acid (GR)

#### 5.3 Potassium dichromate (GR)

#### 5.4 Sodium hydroxide (GR)

## 5.5 Potassium borohydride (AR)

## 5.6 Nitric acid solution (volume fraction 5 %):

Take 50 ml of nitric acid (5.2) with a pipette and dilute it to 1 000 ml with water.

## 5.7 Nitric diluted solution of potassium dichromate (0,5 g/l):

Weigh and dissolve 0,5 g of potassium dichromate (5.3) in 1 000 ml of nitric acid solution (5.6).

## 5.8 Hydrochloric acid-nitric acid mixed reagent [(1 + 1) aqua regia]:

Mix 150 ml of hydrochloric acid (5.1) with 50 ml of nitric acid (5.2) and then dilute it with water to double.

## 5.9 Sodium hydroxide solution (mass fraction 0,2 %):

Weigh and dissolve 1,0 g of sodium hydroxide (5.4) in 500 ml of water.

## 5.10 Potassium borohydride solution (mass fraction 2 %):

Weigh and dissolve 10,0 g of potassium borohydride (5.5) in 500 ml of sodium hydroxide solution (5.9).

## 5.11 Mercury standard solution (1 000 mg/l)

## 5.12 Mercury standard intermediate solution (1 µg/ml):

Take 100 µl of mercury standard solution (5.11) with a pipette, dilute it to 100 ml by adding nitric diluted potassium dichromate solution (5.7), shake well.

## 5.13 Mercury standard use solution (20 µg/l):

Take 1 ml of mercury standard intermediate solution (5.12) with a pipette and dilute it to 50 ml by adding nitric diluted potassium dichromate solution (5.7), shake well.

## 5.14 Preparation of mercury standard working curve solutions:

Separately place 0,0 ml, 0,5 ml, 1,0 ml, 2,0 ml, 3,0 ml, 5,0 ml mercury standard use solution (5.13) in a 50 ml volumetric flask (6.5), dilute with the nitric diluted potassium dichromate solution (5.7) to 50 ml so that the concentrations of the mercury standard working curve solutions are 0,0 µg/l, 0,2 µg/l, 0,4 µg/l, 0,8 µg/l, 1,2 µg/l and 2,0 µg/l. Shake well.

# 6 Apparatus

## 6.1 Agate mortar

## 6.2 80 mesh standard sample sieve

## 6.3 Water bath

## 6.4 Atomic fluorescence photometer: equipped with a mercury hollow cathode lamp

## 6.5 Volumetric flasks (50 ml)

## 6.6 Capped test tubes: volume 100 ml