



## Standard Test Method for Application and Analysis of Radiometric Monitors for Reactor Vessel Surveillance, E 706(IIIA)<sup>1</sup>

This standard is issued under the fixed designation E 1005; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### 1. Scope

1.1 This method describes general procedures for measuring the specific activities of radioactive nuclides produced in radiometric monitors (RMs) by nuclear reactions induced during surveillance exposures for reactor vessels and support structures. More detailed procedures for individual RMs are provided in separate standards identified in 2.1 and in Refs **11, 24-27**. The measurement results can be used to define corresponding neutron induced reaction rates which can in turn be used to characterize the irradiation environment of the reactor vessel and support structure. The principal measurement technique is high resolution gamma-ray spectrometry, although X-ray photon spectrometry and Beta particle counting are used to a lesser degree for specific RMs **(1-29)**.<sup>2</sup>

1.1.1 The measurement procedures include corrections for detector background radiation, random and true coincidence summing losses, differences in geometry between calibration source standards and the RMs, self absorption of radiation by the RM, other absorption effects, and radioactive decay corrections **(1-10, 12-22)**.

1.1.2 Specific activities are calculated by taking into account the time duration of the count, the elapsed time between start of count and the end of the irradiation, the half life, the mass of the target nuclide in the RM, and the branching intensities of the radiation of interest. Using the appropriate half life and known conditions of the irradiation, the specific activities may be converted into corresponding reaction rates **(24-30)**.

1.1.3 Procedures for calculation of reaction rates from the radioactivity measurements and the irradiation power time history are included. A reaction rate can be converted to neutron fluence rate (flux density) and fluence using the appropriate integral cross section and effective irradiation time values, and, with other reaction rates can be used to define the neutron spectrum through the use of suitable computer programs **(24-30)**.

1.1.4 The use of benchmark neutron fields for calibration of RMs can reduce significantly or eliminate systematic errors

since many parameters, and their respective uncertainties, required for calculation of absolute reaction rates are common to both the benchmark and test measurements and therefore are self cancelling. The benchmark equivalent fluence rates, for the environment tested, can be calculated from a direct ratio of the measured saturated activities in the two environments and the certified benchmark fluence rate **(24-30)**.

1.2 This method is intended to be used in conjunction with ASTM Guide E 844, also referred to as E 706 (IIC)<sup>3</sup>. The following existing or proposed ASTM practices, guides, and methods are also directly involved in the physics-dosimetry evaluation of reactor vessel and support structure surveillance measurements:

E 706 (O) Master Matrix for Light-Water Reactor Pressure Vessel Surveillance Standards<sup>3</sup>

E 706 (IA), E 853 Analysis and Interpretation of Light-Water Reactor Surveillance Results<sup>3</sup>

E 706 (IC), E 560 Practice for Extrapolating Reactor Vessel Surveillance Dosimetry Results<sup>3</sup>

E 706 (ID), E 693 Practice for Characterizing Neutron Exposures in Ferritic Steels in Terms of Displacements Per Atom (DPA)<sup>3</sup>

E 706 (IE) Damage Correlation for Reactor Vessel Surveillance<sup>3</sup>

E 706 (IF), E 185 Practice for Conducting Surveillance Tests for Light-Water Nuclear Power Reactor Vessels<sup>3</sup>

E 706 (IG) Surveillance Tests for Nuclear Reactor Support Structures<sup>3</sup>

E 706 (IH), E 636 Practice for Conducting Supplemental Surveillance Tests for Nuclear Power Reactor Vessels<sup>3</sup>

E 706 (IIA), E 944 Guide for Application of Neutron Spectrum Adjustment Methods in Reactor Surveillance<sup>3</sup>

E 706 (IIB), E 1018 Guide for Application of ASTM Evaluated Cross Section and Data File<sup>3</sup>

E 706 (IID), E 482 Guide for Application of Neutron Transport Methods for Reactor Vessel Surveillance<sup>3</sup>

E 706 (IIE) Benchmark Testing of Reactor Vessel Dosimetry<sup>3</sup>

E 706 (IIIB), E 854 Test Method for Application and Analysis of Solid State Track Recorder (SSTR) Monitors for Reactor Vessel Surveillance<sup>3</sup>

<sup>1</sup> This method is under the jurisdiction of ASTM Committee E-10 on Nuclear Technology and Applications and is the direct responsibility of Subcommittee E10.05 on Nuclear Radiation Metrology.

Current edition approved June 10, 1997. Published May 1998.

<sup>2</sup> The boldface numbers in parentheses refer to the list of references appended to this method.

<sup>3</sup> The reference in parentheses refers to Section 5 as well as Figs. 1 and 2 of Matrix E 706.

 **E 1005**

E 706 (IIIC), E 910 Test Method for Application and Analysis of Helium Accumulation Fluence Monitors for Reactor Vessel Surveillance<sup>3</sup>

E 706 (IIID) Application and Analysis of Damage Monitors for Reactor Vessel Surveillance<sup>3</sup>

E 706 (IIIE) Application and Analysis of Temperature Monitors for Reactor Vessel Surveillance<sup>3</sup>

1.3 The general procedures in this method are applicable to the measurement of radioactivity in RMs which satisfy the specific constraints and conditions imposed for their analysis. More detailed procedures for individual RM monitors are identified in 2.1 and in Refs **11, 24-27** (see Table 1).

1.4 This method, along with the individual RM monitor standard methods, are intended for use by knowledgeable persons who are intimately familiar with the procedures, equipment, and techniques necessary to achieve high precision and accuracy in radioactivity measurements.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

**TABLE 1 Radiometric Monitors Proposed for Reactor Vessel Surveillance**

Dosimetry Reactions	Residual Nucleus			Target Atom Natural Abundance <sup>a</sup> [31]	Detector Response <sup>B</sup>	ASTM Standard or Ref.
	Half-life <sup>C,A,D</sup>	E <sub>γ</sub> <sup>D</sup> (keV)	Yield <sup>D</sup> (%) γ/Reaction			
<sup>23</sup> Na(n,γ) <sup>24</sup> Na	0.62356 (17) d	1368.633 2754.030	99.9936 99.855	1.00	NTR	<b>(24-30)</b>
<sup>27</sup> Al(n,α) <sup>24</sup> Na	0.62356 (17) d	1368.633 2754.030	99.9936 99.855	1.00	TR	E 266
<sup>32</sup> S(n,p) <sup>32</sup> P	14.262 (14) d	<E <sub>β</sub> >=694.9	100.	0.9502 (9)	TR	E 265
<sup>45</sup> Sc(n,γ) <sup>46</sup> Sc	83.79 (4) d	889.277 1120.545	99.9844 99.9874	1.00	NTR	<b>(24-30)</b>
<sup>46</sup> Ti(n,p) <sup>46</sup> Sc	83.79 (4) d	889.277 1120.545	99.9844 99.9874	0.0825 (3)	TR	E 526
<sup>47</sup> Ti(n,p) <sup>47</sup> Sc	3.3492 (16) d	159.381	68.3	0.0744 (2)	TR	E 526
<sup>48</sup> Ti(n,p) <sup>48</sup> Sc	43.67 (9) h	983.526 1037.522 1312.120	100.0 97.5 100.0	0.7372 (3)	TR	E 526
<sup>55</sup> Mn(n,2n) <sup>54</sup> Mn	312.3 (4) d	834.843	99.9758	1.00	TR	E 261, E 263 <b>(24-30)</b>
<sup>54</sup> Fe(n,p) <sup>54</sup> Mn	312.3 (4) d	834.843	99.9758	0.05845 (35)	TR	E 263
<sup>54</sup> Fe(n,γ) <sup>55</sup> Fe	2.73 (3) y	5.888 5.899 6.490	8.2 16.2 2.86	0.05845 (35)	NTR	<b>(24-30)</b>
<sup>56</sup> Fe(n,p) <sup>56</sup> Mn	2.5785 (2) hr	846.754 1810.72 2113.05	98.87 27.18925 14.33615	0.91754 (36)	TR	<b>(24-30)</b>
<sup>58</sup> Fe(n,γ) <sup>59</sup> Fe	44.503 (6) d	1099.251 1291.596 1481.7	56.5 43.2 0.059	0.00282 (4)	NTR	<b>(24-30)</b>
<sup>59</sup> Co(n,γ) <sup>60</sup> Co	1925.5 (5) d 10.467 (6) m (meta)	1173.238 1332.502 58.603 826.28 1332.501 2158.77	99.857 99.983 2.01 0.00768 0.24 0.00072	1.00	NTR	E 262, E 481
<sup>58</sup> Ni(n,p) <sup>58</sup> Co	70.82 (3) d 9.15 (10) h (meta)	810.775 863.959 1674.730 24.889	99.45 0.69 0.519 0.0369	0.68077 (9)	TR	E 264
<sup>60</sup> Ni(n,p) <sup>60</sup> Co	1925.5 (5) d 10.467 (6) m (meta)	1173.238 1332.502 58.603 826.28	99.857 99.983 2.01 0.00768	0.26223 (8)	TR	<b>(24-30)</b>

 **E 1005**

**TABLE 1** *Continued*

Dosimetry Reactions	Residual Nucleus			Target Atom Natural Abundance <sup>A</sup> [31]	Detector Response <sup>B</sup>	ASTM Standard or Ref.
	Half-life <sup>C,A,D</sup>	E <sub>γ</sub> <sup>D</sup> (keV)	Yield <sup>D</sup> (%) γ/Reaction			
		1332.501 2158.77	0.24 0.00072			
<sup>63</sup> Cu(n,γ) <sup>64</sup> Cu	12.701 (2) h	1345.77	0.47336	0.6917 (3)	NTR	(24-30)
<sup>63</sup> Cu(n,α) <sup>60</sup> Co	1925.5 (5) d  10.467 (6) m (meta)	1173.238	99.857	0.6917 (3)	TR	E 523
		1332.502	99.983			
		58.603	2.01			
		826.33	0.0058			
		1332.501 2158.86	0.25 0.00088			
<sup>93</sup> Nb(n,n') <sup>93m</sup> Nb	5.89 (5) × 10 <sup>3</sup> d	30.77 16.52 (K <sub>α1,2</sub> )	0.000549 9.25	1.00	TR	(11, 24-30)
<sup>103</sup> Rh(n,n') <sup>103m</sup> Rh	56.12 (1) m	39.755	0.0684	1.00	TR	(24-30)
<sup>109</sup> Ag(n,γ) <sup>110m</sup> Ag	249.76 (20) d	116.48	0.00799	0.48161 (7)	NTR	E 481
		884.685	72.192			
		937.493	34.1314			
		1384.300	24.1204			
		1505.040	12.9532			
		1475.788	3.96868			
<sup>115</sup> In(n,γ) <sup>116m</sup> In	54.41 (17) m	1293.54	84.4	0.9571 (2)	NTR	E 261, E 262
		1097.3	56.2104			
		818.7	11.4784			
		2112.1	15.5296			
<sup>115</sup> In(n,n') <sup>115m</sup> In	4.486 (4) h	336.241	45.9	0.9571 (2)	TR	(24-30)
		497.370	0.047			
<sup>181</sup> Ta(n,γ) <sup>182</sup> Ta	114.43 (3) d	1121.3008	34.9	0.99988 (2)	NTR	E 262
		1189.0503	16.225			
		1221.4066	26.9777			
<sup>197</sup> Au(n,γ) <sup>198</sup> Au	2.6943 (8) d	1087.6904	0.159045	1.00	NTR	E 261, E 262 (24-30)
		675.8874	0.8038278			
		411.804	95.57			
<sup>232</sup> Th(n,γ) <sup>233</sup> Th	22.3 (1) m	890.1	0.14	1.00	NTR	(24-30)
		490.80	0.17			
		499.02	0.21			
		699.901	0.68			
		764.4	0.120			
... → <sup>233</sup> Pa	26.967 (2) d	312.17	38.6			
FM(n,f) <sup>144</sup> Ce	284.893 (8) d	133.515	11.09	— <sup>E</sup>	NTR, TR	E 704, E 705 (24-30)
		80.120	1.36407 (see Table 2)			
FM(n,f) <sup>140</sup> Ba	12.752 (3) d	537.261	24.4 (see Table 2)	— <sup>E</sup>	NTR, TR	E 393, E 704, E 705
<sup>140</sup> Ba → <sup>140</sup> La	1.6781 (3) d	1596.21	95.4			(24-30)
		815.772	23.2776			
		487.021	45.5058			
			(see Table 2)			
FM(n,f) <sup>137</sup> Cs	30.07 (16) y	661.660	85.1 (see Table 2)	— <sup>E</sup>	NTR, TR	E 320, E 704, E 705
<sup>137</sup> Cs → <sup>137m</sup> Ba	2.552 (1) m	661.660	90.11 (see Table 2)			(24-30)
FM(n,f) <sup>106</sup> Ru	373.59 (15) d	—	— (see Table 2)	— <sup>E</sup>	NTR, TR	E 704, E 705 (24-30)
<sup>106</sup> Ru → <sup>106</sup> Rh	29.80 (8) s	511.8605	20.4			

**TABLE 1** *Continued*

Dosimetry Reactions	Residual Nucleus			Target Atom Natural Abundance <sup>A</sup> [31]	Detector Response <sup>B</sup>	ASTM Standard or Ref.
	Half-life <sup>C,A,D</sup>	E <sub>γ</sub> <sup>D</sup> (keV)	Yield <sup>D</sup> (%) γ/Reaction			
			(see Table 2)			
FM(n,f) <sup>103</sup> Ru	39.26 (2) d	497.084	91.0 (see Table 2)	— <sup>E</sup>	NTR, TR	E 704, E 705 (24-30)
FM(n,f) <sup>95</sup> Zr	64.02 (5) d	756.729 724.199	54.46 44.1725 (see Table 2)	— <sup>E</sup>	NTR, TR	E 704, E 705 (24-30)
<sup>95</sup> Zr→ <sup>95</sup> Nb	34.975 (7) d	765.807	99.81 (see Table 2)			

<sup>A</sup>The numbers in parentheses following some given values is the uncertainty in the last digit(s) of the value: 0.729 (8) means 0.729 ± 0.008, 70.8 (1) means 70.8 ± 0.1.

<sup>B</sup>NTR = Non-Threshold Response, TR = Threshold Response.

<sup>C</sup>The time units listed for half-life are years (y), days (d), hours (h), minutes (m), and seconds (s).

<sup>D</sup>The nuclear data has been drawn from several primary sources including References (31), (33) and (34). Reference (32) summarizes the source of the selected nuclear constants.

<sup>E</sup>FM = Fission Monitor: <sup>235</sup>U and <sup>239</sup>Pu (NTR) and <sup>238</sup>U, <sup>237</sup>Np, and <sup>232</sup>Th (TR) target isotope or weight fraction varies with material batch.

## 2. Referenced Documents

2.1 ASTM Standards not identified in 1.2, including those for individual RM monitors:

### 2.2 ASTM Standards:

E 170 Terminology Relating to Radiation Measurements and Dosimetry<sup>4</sup>

E 177 Practice for Use of the Terms Precision and Bias in ASTM Test Methods<sup>5</sup>

E 181 General Methods for Detector Calibration and Analysis of Radionuclides<sup>4</sup>

E 219 Test Method for Atom Percent Fission in Uranium Fuel (Radiochemical Method)<sup>4</sup>

E 261 Practice for Determining Neutron Fluence Rate, Fluence, and Spectra by Radioactivation Techniques<sup>4</sup>

E 262 Test Method for Determining Thermal Neutron Reaction and Fluence Rates by Radioactivation Techniques<sup>4</sup>

E 263 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Iron<sup>4</sup>

E 264 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Nickel<sup>4</sup>

E 265 Test Method for Measuring Reaction Rates and Fast-Neutron Fluences by Radioactivation of Sulfur-32<sup>4</sup>

E 266 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Aluminum<sup>4</sup>

E 267 Test Method for Uranium and Plutonium Concentrations and Isotopic Abundances<sup>4</sup>

E 320 Test Methods for Cesium-137 in Nuclear Fuel Solutions by Radiochemical Analysis<sup>4</sup>

E 321 Test Method for Atom Percent Fission in Uranium and Plutonium Fuel (Neodymium-148 Method)<sup>4</sup>

E 343 Test Method for Measuring Reaction Rates by Analysis of Molybdenum-99 Radioactivity from Fission Dosimeters<sup>4</sup>

E 393 Test Method for Measuring Reaction Rates by Analysis of Barium-140 from Fission Dosimeters<sup>4</sup>

E 481 Test Method for Measuring Neutron Fluence Rate by Radioactivation of Cobalt and Silver<sup>4</sup>

E 523 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Copper<sup>4</sup>

E 526 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Titanium<sup>4</sup>

E 704 Test Method for Measuring Reaction Rates by Radioactivation of Uranium-238<sup>4</sup>

E 705 Test Method for Measuring Reaction Rates by Radioactivation of Neptunium-237<sup>4</sup>

E 844 Guide for Sensor Set Design and Irradiation for Reactor Surveillance, E 706 (IIC)<sup>4</sup>

### 2.3 ANSI Standard:

N42.14 Calibration and Usage of Germanium Detectors for Measurement of Gamma-Ray Emission Rates of Radionuclides<sup>6</sup>

## 3. Terminology

### 3.1 Definitions:

3.1.1 *radiometric monitor (RM), dosimeter, foil*—a small quantity of material consisting of or containing an accurately known mass of a specific target nuclide. Usually fabricated in a specified and consistent geometry and used to determine neutron fluence rate (flux density), fluence and spectra by measuring a specific radioactive neutron-induced reaction product. A single RM may contain more than one target nuclide or have more than one specific reaction product.

3.1.2 *calibration standard*—a calibrated radioactive source standardized using an absolute calibration method or by rigorous comparison to a national or certified radioactivity standard source.

3.1.3 *national radioactivity standard source*—a calibrated radioactive source prepared and distributed as a standard reference material by the National Institute of Standards and Technology (NIST).

<sup>4</sup> Annual Book of ASTM Standards, Vol 12.02.

<sup>5</sup> Annual Book of ASTM Standards, Vol 14.02.

<sup>6</sup> Available from American National Standards Institute, 11 W. 42nd St., 13th Floor, New York, NY 10036.



**TABLE 2 Recommended Fission Yield Data<sup>A</sup>**

Fissile Isotope	Reaction Product	Cumulative Fission Yield (Energy Dependent)		Independent Fission Yield (Energy Dependent)		
		0.5 MeV	Thermal	0.5 MeV	Thermal	
<sup>232</sup> Th(n,f)	<sup>95</sup> Zr	5.67313 ± 2.8 %	—	3.84804 × 10 <sup>-3</sup> ± 64 %	—	
	<sup>95</sup> Nb	5.67313 ± 2.8 %	—	7.49008 × 10 <sup>-7</sup> ± 64 %	—	
	<sup>103</sup> Ru	0.156332 ± 4.0 %	—	6.12007 × 10 <sup>-8</sup> ± 64 %	—	
	<sup>106</sup> Ru	0.0537306 ± 11 %	—	1.05001 × 10 <sup>-4</sup> ± 64 %	—	
	<sup>106</sup> Rh	0.0537306 ± 11 %	—	1.33001 × 10 <sup>-8</sup> ± 64 %	—	
	<sup>137</sup> Cs	5.84355 ± 4 %	—	8.32609 × 10 <sup>-3</sup> ± 64 %	—	
	<sup>137m</sup> Ba	5.528 ± 4 %	—	7.63008 × 10 <sup>-6</sup> ± 64 %	—	
	<sup>140</sup> Ba	7.87647 ± 2.8 %	—	4.82795 × 10 <sup>-2</sup> ± 64 %	—	
	<sup>140</sup> La	7.87649 ± 2 %	—	2.71003 × 10 <sup>-5</sup> ± 64 %	—	
	<sup>144</sup> Ce	7.94699 ± 4 %	—	4.80505 × 10 <sup>-3</sup> ± 64 %	—	
	<sup>235</sup> U(n,f)	<sup>95</sup> Zr	6.42554 ± 1 %	6.49458 ± 1 %	2.93502 × 10 <sup>-2</sup> ± 64 %	1.2208 × 10 <sup>-1</sup> ± 16 %
		<sup>95</sup> Nb	6.42557 ± 0.7 %	6.49471 ± 1 %	2.21002 × 10 <sup>-5</sup> ± 64 %	1.02003 × 10 <sup>-4</sup> ± 64 %
		<sup>103</sup> Ru	3.26185 ± 1 %	3.03144 ± 1.4 %	2.35002 × 10 <sup>-5</sup> ± 64 %	2.31006 × 10 <sup>-5</sup> ± 64 %
		<sup>106</sup> Ru	0.535398 ± 1 %	0.44108 ± 1.4 %	7.55005 × 10 <sup>-3</sup> ± 64 %	8.90023 × 10 <sup>-7</sup> ± 64 %
<sup>106</sup> Rh		0.535404 ± 1 %	0.400983 ± 1.4 %	6.06004 × 10 <sup>-6</sup> ± 64 %	0	
<sup>137</sup> Cs		6.22335 ± 0.5 %	6.18682 ± 0.5 %	1.84879 × 10 <sup>-1</sup> ± 32 %	5.93635 × 10 <sup>-2</sup> ± 11 %	
<sup>137m</sup> Ba		5.88768 ± 0.5 %	5.85286 ± 0.5 %	3.93003 × 10 <sup>-4</sup> ± 64 %	1.30003 × 10 <sup>-4</sup> ± 64 %	
<sup>140</sup> Ba		5.98741 ± 1 %	6.19595 ± 1 %	4.72071 × 10 <sup>-1</sup> ± 64 %	4.64072 × 10 <sup>-1</sup> ± 32 %	
<sup>140</sup> La		5.98872 ± 1 %	6.2012 ± 1 %	1.31401 × 10 <sup>-5</sup> ± 64 %	5.25214 × 10 <sup>-3</sup> ± 64 %	
<sup>144</sup> Ce		5.26689 ± 1.4 %	5.49709 ± 0.7 %	4.35013 × 10 <sup>-2</sup> ± 64 %	6.27386 × 10 <sup>-2</sup> ± 64 %	
<sup>237</sup> Np(n,f)		<sup>95</sup> Zr	5.68896 ± 2.8 %	—	6.1647 × 10 <sup>-2</sup> ± 64 %	—
		<sup>95</sup> Nb	5.68907 ± 2 %	—	8.5 × 10 <sup>-5</sup> ± 64 %	—
		<sup>103</sup> Ru	5.56212 ± 2.8 %	—	1.44 × 10 <sup>-4</sup> ± 64 %	—
		<sup>106</sup> Ru	2.18067 ± 11 %	—	6.0445 × 10 <sup>-2</sup> ± 64 %	—
	<sup>106</sup> Rh	2.18077 ± 11 %	—	9.29 × 10 <sup>-5</sup> ± 64 %	—	
	<sup>137</sup> Cs	6.16977 ± 2.8 %	—	2.74311 × 10 <sup>-1</sup> ± 64 %	—	
	<sup>137m</sup> Ba	5.83804 ± 2.8 %	—	1.438 × 10 <sup>-3</sup> ± 64 %	—	
	<sup>140</sup> Ba	5.47246 ± 1.4 %	—	5.83709 × 10 <sup>-1</sup> ± 64 %	—	
	<sup>140</sup> La	5.47688 ± 2.8 %	—	4.421 × 10 <sup>-3</sup> ± 64 %	—	
	<sup>144</sup> Ce	4.12987 ± 2 %	—	9.1496 × 10 <sup>-2</sup> ± 64 %	—	
	<sup>238</sup> U(n,f)	<sup>95</sup> Zr	5.15126 ± 1 %	—	7.88021 × 10 <sup>-4</sup> ± 64 %	—
		<sup>95</sup> Nb	5.15126 ± 1 %	—	8.34022 × 10 <sup>-8</sup> ± 64 %	—
		<sup>103</sup> Ru	6.26113 ± 1 %	—	3.06008 × 10 <sup>-7</sup> ± 64 %	—
		<sup>106</sup> Ru	2.48759 ± 1.4 %	—	3.8101 × 10 <sup>-7</sup> ± 64 %	—
<sup>106</sup> Rh		2.48759 ± 1.4 %	—	0	—	
<sup>137</sup> Cs		6.02075 ± 1 %	—	9.28724 × 10 <sup>-3</sup> ± 64 %	—	
<sup>137m</sup> Ba		5.69564 ± 1 %	—	4.10011 × 10 <sup>-6</sup> ± 64 %	—	
<sup>140</sup> Ba		5.84596 ± 1 %	—	2.57317 × 10 <sup>-2</sup> ± 64 %	—	
<sup>140</sup> La		5.84597 ± 1 %	—	1.38004 × 10 <sup>-5</sup> ± 64 %	—	
<sup>144</sup> Ce		4.55034 ± 1.4 %	—	1.30603 × 10 <sup>-3</sup> ± 64 %	—	
<sup>239</sup> Pu(n,f)		<sup>95</sup> Zr	4.65412 ± 1.4 %	4.80834 ± 1.4 %	7.94561 × 10 <sup>-2</sup> ± 64 %	1.25228 × 10 <sup>-1</sup> ± 64 %
		<sup>95</sup> Nb	4.65431 ± 1.4 %	4.80904 ± 1.4 %	1.45002 × 10 <sup>-4</sup> ± 64 %	5.64006 × 10 <sup>-4</sup> ± 64 %
		<sup>103</sup> Ru	6.85701 ± 1.4 %	7.0047 ± 2 %	4.41006 × 10 <sup>-4</sup> ± 64 %	1.01101 × 10 <sup>-3</sup> ± 64 %
		<sup>106</sup> Ru	4.33807 ± 1.4 %	4.32901 ± 2 %	2.00083 × 10 <sup>-1</sup> ± 64 %	3.2276 × 10 <sup>-1</sup> ± 32 %
	<sup>106</sup> Rh	4.33845 ± 1.4 %	4.32923 ± 2 %	3.76005 × 10 <sup>-4</sup> ± 64 %	2.23002 × 10 <sup>-4</sup> ± 64 %	
	<sup>137</sup> Cs	6.57917 ± 0.7 %	6.61591 ± 0.5 %	9.9008 × 10 <sup>-1</sup> ± 16 %	5.94641 × 10 <sup>-1</sup> ± 16 %	
	<sup>137m</sup> Ba	6.22844 ± 0.7 %	6.26233 ± 0.5 %	4.54506 × 10 <sup>-3</sup> ± 64 %	3.68004 × 10 <sup>-3</sup> ± 64 %	
	<sup>140</sup> Ba	5.31538 ± 1 %	5.37475 ± 2 %	9.4334 × 10 <sup>-1</sup> ± 64 %	1.51518 × 10 <sup>-0</sup> ± 23 %	
	<sup>140</sup> La	5.32713 ± 1 %	5.38286 ± 1.4 %	1.17572 × 10 <sup>-2</sup> ± 64 %	8.11109 × 10 <sup>-3</sup> ± 64 %	
	<sup>144</sup> Ce	3.67369 ± 0.7 %	3.74236 ± 1 %	1.56328 × 10 <sup>-1</sup> ± 64 %	1.06259 × 10 <sup>-1</sup> ± 64 %	

<sup>A</sup>All yield data is given as a percentage.

3.1.4 *certified radioactivity standard source*—a calibrated radioactive source, with stated accuracy, whose calibration is traceable to a national radioactivity measurements system.

3.1.5 *check source, control standard*—a radioactivity source, not necessarily calibrated, which is used as a working reference to verify the continuing satisfactory operation of an instrument.

3.1.6 *FWHM (full width at half maximum)*—a measure of detector/system gamma-ray energy resolution expressed as the width of the gamma-ray peak distribution, in units of energy, measured at one-half the maximum peak height above the background.

3.1.7 *FWTM (full width at tenth maximum)*—identical to FWHM except the width is measured at one tenth the maximum peak height above the background.

3.1.8 *resolution, gamma-ray*—usually expressed as the FWHM and often including a specification for the FWTM.

3.1.9 *peak-to-compton-ratio*—the ratio of the net height of a Gaussian fit of the gamma-ray peak to average net counts in channels in the relatively flat portion of the Compton continuum.

#### 4. Summary of Method

4.1 Appropriate radiation detection-measurement instruments shall be used in conjunction with suitable calibration standards, nuclear parameters, and test data to quantitatively determine the decay rate of selected radioactive nuclides produced in RMs during test and surveillance irradiations in neutron fields. These results together with established cross sections, spectral response data, and known test parameters allow the determination of the neutron flux, fluence, and spectrum. Conversely, by using well-characterized controlled neutron fields to irradiate the selected target foils, cross sections and spectral response data can be determined from the radioactivity measurements.

4.2 The appropriate standard method of analysis identified in Section 2 for the individual RMs shall be followed as the individual problems that may be encountered and the precision and bias of the analysis for that particular RM are more fully discussed in these standards.

4.3 The neutron fluence rate (flux density), fluence, and spectral data shall be correlated to radiation induced change and damage in reactor materials through the use of appropriate analytical/calculational codes (see Matrix E 706 (IIA), (IIB), (IIC), (IID), and (IIE)).

#### 5. Significance and Use

5.1 Radiometric monitors shall provide a proven passive dosimetry technique for the determination of neutron fluence rate (flux density), fluence, and spectrum in a diverse variety of neutron fields. These data are required to evaluate and estimate probable long-term radiation-induced damage to nuclear reactor structural materials such as the steel used in reactor pressure vessels and their support structures.

5.2 A number of radiometric monitors, their corresponding neutron activation reactions, and radioactive reaction products and some of the pertinent nuclear parameters of these RMs and products are listed in Table 1. Table 2 provides data (35) on the cumulative and independent fission yields of the important fission monitors. Additional fission product reactions that may provide in situ photo fission information will be added to Table 1 as information is developed and verified (23-29, 36-39).

#### 6. Apparatus

6.1 A high resolution gamma-ray spectrometry system consisting of, but not limited to the following items:

6.1.1 *Gamma-Ray Detector*—A high purity germanium or lithium drifted germanium diode with its cryostat, preamplifier and high-voltage (bias) power supply, and liquid nitrogen dewar. The detector (incorporated into the complete spectrometry system) shall have a resolution of  $\leq 2.5$  keV (FWHM) measured at the 1332 keV  $^{60}\text{Co}$  peak with the FWTM no larger than 2 times the FWHM. The peak-to-Compton ratio shall be 25 to 1 or greater. The relative full-energy peak counting efficiency shall be about 0.1 (1).

6.1.1.1 If more than one detector is available, the specifications can be advantageously tailored to optimize performance over the range of radioactivity levels and gamma-ray energies to be measured.

6.1.2 *Linear Amplifier*, for nuclear spectroscopy—multichannel pulse-height analyzer with at least 4000 channels,

live time correction, and a data read out device such as a printer, teletype, or other hard copy terminal. A CRT display is extremely useful and in many cases essential for efficient operations, as are computer compatible storage devices as punch paper tape, magnetic tape, diskettes, and hard disks. A built in computerized data handling and reduction system is also useful for processing large numbers of samples and to reduce possibility of human error.

6.2 *Thallium Activated Sodium Iodide Scintillation Crystal*—[NaI(Tl)], optically coupled to a photomultiplier tube with preamplifier, high voltage power supply, linear amplifier, multichannel analyzer with at least 400 channel capacity and a suitable data readout device. It is often feasible and advantageous to use a portion of the multichannel analyzer used for the high resolution germanium detector system for the NaI(Tl) detector through use of multiplexing techniques. A 3 by 3-in. integrally mounted NaI(Tl) detector is a good choice for general use.

6.3 *Beta Particle Counting System*, consisting of a suitable detector ranging from a thin end-window Geiger-Mueller type detector, proportional counter, scintillation counter to partially depleted silicon diodes; electronic components such as preamplifiers, amplifiers, discriminator-drivers, scalars, timers and high voltage power supplies to complete the system. Refer to General Methods E 181 for preparation of apparatus and counting procedures.

6.4 *X-ray Spectrometry System*, utilizing high resolution lithium drifted silicon, Si(Li), or germanium X-ray detector with liquid nitrogen cryostat, preamplifier, amplifier and multichannel analyzer system with at least 1000 channel capacity and suitable data readout and display devices. Multiplexing could permit use of the same multichannel analyzer used for the high resolution germanium gamma spectrometer if adequate capacity exists or the analyzer could be dedicated to one use or the other to suit analysis schedules and requirements.

6.5 *High-Density Shielding* (usually lead) around the detectors to reduce interferences from background radiations.

6.6 *Sample Positioning Hardware*, to provide a number of reproducible fixed positions which can be calibrated for each detector as appropriate to accommodate different sample activities and sizes.

6.7 *National and Certified Radioactivity Standard Sources*.

6.8 *Calibration and Control Standards*.

6.9 Apparatus and reagents as listed in applicable ASTM standards for RM analysis.

#### 7. Precautions

7.1 Refer to General Methods E 181 and Guide E 844. For high fluence irradiations, burn-in or burn-out of target nuclides in the RM must be considered. For decay chains, such as  $^{140}\text{Ba}$ - $^{140}\text{La}$ , decay corrections must take into account formation of a radioactive daughter by a radioactive parent. When appropriate, round-robin intercalibration tests such as those conducted by NBS, the LWR Pressure Vessel Surveillance Dosimetry Improvement Program, or under the Interlaboratory Reaction Rate (ILRR) Program shall be undertaken to detect and eliminate unforeseen sources of error (24-30).