
**Milk, milk products and infant
formula — Determination of
melamine and cyanuric acid by liquid
chromatography and tandem mass
spectrometry (LC-MS/MS)**

*Lait, produits laitiers et formule infantile — Détermination de la
teneur en mélamine et en acide cyanurique par chromatographie en
phase liquide couplée à la spectrométrie de masse en tandem (CL-SM/
SM)*

(<https://standards.iteh.ai>)
Document Preview

ISO 23970:2021

<https://standards.iteh.ai/catalog/standards/iso/9cbdec64-e8bc-4a0c-ac2b-e97fabbad568/iso-23970-2021>



Reference numbers
ISO 23970:2021(E)
IDF 252:2021(E)

© ISO and IDF 2021

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 23970:2021

<https://standards.iteh.ai/catalog/standards/iso/9cbdec64-e8bc-4a0c-ac2b-e97fabbad568/iso-23970-2021>



COPYRIGHT PROTECTED DOCUMENT

© ISO and IDF 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11

Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

International Dairy Federation
Silver Building • Bd Auguste Reyers 70/B
B-1030 Brussels
Phone: +32 2 325 67 40
Fax: +32 2 325 67 41
Email: info@fil-idf.org
Website: www.fil-idf.org

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	1
5.1 List of reagents	1
5.2 Preparation of stock solutions	2
5.3 Preparation of standard solutions	3
5.4 Preparation of calibration solutions	4
6 Apparatus	5
7 Procedure	6
7.1 Sample preparation	6
7.2 Extraction	6
7.2.1 General	6
7.2.2 Extraction procedure	6
7.3 Determination	6
8 System settings	7
8.1 HPLC parameters	7
8.2 HPLC-MS/MS parameters	7
8.3 UHPLC parameters	9
8.4 UHPLC-MS/MS parameters	9
8.5 Preparation of the LC-MS/MS system and samples	10
8.5.1 Tuning of the LC-MS/MS system	10
8.5.2 Checking of the instrument settings	10
8.5.3 Checking of the sensitivity of the system	10
8.5.4 Sample list	11
9 Calculation and expression of results	11
9.1 General	11
9.2 Calibration criteria	11
9.3 Identification criteria	11
9.4 Recovery	12
9.5 Calculation of results	13
10 Precision data	14
10.1 General	14
10.2 Repeatability	14
10.3 Reproducibility	14
11 Test report	15
Annex A (informative) Precision data	16
Bibliography	18

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF). It is being published jointly by ISO and IDF.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

IDF (the International Dairy Federation) is a non-profit private sector organization representing the interests of various stakeholders in dairying at the global level. IDF members are organized in National Committees, which are national associations composed of representatives of dairy-related national interest groups including dairy farmers, dairy processing industry, dairy suppliers, academics and governments/food control authorities.

ISO and IDF collaborate closely on all matters of standardization relating to methods of analysis and sampling for milk and milk products. Since 2001, ISO and IDF jointly publish their International Standards using the logos and reference numbers of both organizations.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. IDF shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

This document was prepared by the IDF *Standing Committee on Analytical Methods for Additives and Contaminants* and ISO Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*. It is being published jointly by ISO and IDF.

The work was carried out by the ISO/IDF Action Team A05 of the *Standing Committee on Analytical Methods for Additives and Contaminants* under the aegis of its project leader, Dr S. Holroyd.

ITEH Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 23970:2021

<https://standards.iteh.ai/catalog/standards/iso/9cbdec64-e8bc-4a0c-ac2b-e97fabbad568/iso-23970-2021>

Milk, milk products and infant formula — Determination of melamine and cyanuric acid by liquid chromatography and tandem mass spectrometry (LC-MS/MS)

1 Scope

This document specifies a method for the determination of melamine and cyanuric acid with liquid chromatography in combination with tandem mass spectrometry (LC-MS/MS). The method has been validated in an interlaboratory study via the analysis of spiked samples of milk-based infant formula, soy-based infant formula, milk powder, whole milk, soy drink and milk chocolate ranging from 0,71 mg/kg to 1,43 mg/kg for melamine and 0,57 mg/kg to 1,45 mg/kg for cyanuric acid. The limits of quantification (LOQ) for melamine and cyanuric acid in food are 0,05 mg/kg and 0,25 mg/kg, respectively. The upper limit of the working range is up to 10 mg/kg for melamine and up to 25 mg/kg for cyanuric acid.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

4 Principle

A test portion of the homogenous food sample is fortified with ¹³C labelled internal standards (melamine and cyanuric acid). After incubation for at least 1 h, water is added to the sample and, after shaking, the slurry is dissolved in a mixture of acetonitrile and water. The sample is shaken and centrifuged. After the separation of supernatant from sediments, benzoguanamine is added as a recovery standard. An aliquot of the aqueous supernatant is injected into a LC-MS/MS system. The triple quadrupole mass spectrometer is coupled either to high performance liquid chromatography (HPLC) or to ultra high performance liquid chromatography (UHPLC). Chromatography is based on hydrophilic interaction liquid chromatography (HILIC). Ionization is achieved by electrospray ionization (ESI) in multiple reaction monitoring (MRM).

5 Reagents

5.1 List of reagents

Use only reagents of recognized analytical grade and water conforming to grade 1 of ISO 3696, unless otherwise specified. Use only reagents with purity suitable for melamine and cyanuric acid analysis.

Check the purity of the reagents and reference materials (e.g. standard solutions) by performing a blank test under the same conditions as used in the method. The chromatogram shall not show any interfering impurity at the retention time of compounds of interest.

5.1.1 Formic acid, mass fraction of 98 % to 100 % (CAS RN®¹⁾ 64-18-6).

5.1.2 Acetonitrile, HPLC gradient grade (CAS RN® 75-05-8).

5.1.3 Ammonium acetate, mass fraction of approximately 98 % (CAS RN® 631-61-8).

5.1.4 Methanol, Ultra LC-MS grade (CAS RN® 67-56-1).

5.1.5 Melamine, mass fraction ≥ 99 %, solid (CAS RN® 108-78-1).

5.1.6 Cyanuric acid, solid (CAS RN® 108-80-5).

5.1.7 Benzoguanamine, solid (CAS RN® 91-76-9).

5.1.8 ¹³C melamine, ¹³C₃ (99 %), Amino-¹⁵N₃ (98 %), solution with mass concentration $\rho = 1\,000\ \mu\text{g/ml}$. Other isotope marked melamine/cyanuric acid solutions can be used if they lead to similar results; this would need to be appropriately validated.

EXAMPLE 1,3,5-Triazine-2,4,6-triamine-¹³C₃ (melamine), 1,3,5-Triazine-2,4,6-triol-¹³C₃, 2,4,6-Trihydroxy-1,3,5-triazine-¹³C (cyanuric acid).

5.1.9 ¹³C cyanuric acid, ¹³C₃ (99 %), ¹⁵N₃ (98 %), solution with $\rho = 1\,000\ \mu\text{g/ml}$.

NOTE Alternatives can be used as per [5.1.8](#).

5.2 Preparation of stock solutions

ISO 23970:2021

<https://standards.iteh.ai/catalog/standards/iso/9cbdec64-e8bc-4a0c-ac2b-e97fabbad568/iso-23970-2021>

5.2.1 Melamine stock solution, $\rho = 1\,000\ \mu\text{g/ml}$.

Weigh, to the nearest 0,1 mg, approximately 100 mg of melamine ([5.1.5](#)) into a 100 ml glass flask ([6.2](#)) and add by weighing an amount of water to achieve a concentration of $1\,000\ \mu\text{g/ml}$. Store the solution in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 year.

5.2.2 Cyanuric acid stock solution, $\rho = 1\,000\ \mu\text{g/ml}$.

Weigh, to the nearest 0,1 mg, approximately 100 mg of cyanuric acid ([5.1.6](#)) into a 100 ml glass flask ([6.2](#)) and add by weighing an amount of water to achieve a concentration of $1\,000\ \mu\text{g/ml}$. Store the solution in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 year.

5.2.3 Benzoguanamine stock solution, $\rho = 1\,000\ \mu\text{g/ml}$.

Weigh, to the nearest 0,1 mg, approximately 100 mg of benzoguanamine ([5.1.7](#)) into a 100 ml glass flask ([6.2](#)) and add by weighing an amount of methanol ([5.1.4](#)) to achieve a concentration of $1\,000\ \mu\text{g/ml}$. Store the solution in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 year.

1) CAS Registry Number® (CAS RN®) is the trademark of a product supplied by CAS corporation. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO or IDF of the product named. Equivalent products may be used if they can be shown to lead to the same results.

5.2.4 ¹³C melamine stock solution, $\rho = 20 \mu\text{g/ml}$.

Pipette 1,0 ml of ¹³C melamine (5.1.8) into a 50 ml volumetric flask (6.3). Make up to the mark with water and mix. The final concentration is 20 $\mu\text{g/ml}$. Transfer this solution into a 100 ml glass flask (6.2). The solution is stable under these conditions during at least 1 year.

5.2.5 ¹³C cyanuric acid stock solution, $\rho = 20 \mu\text{g/ml}$.

Pipette 1,0 ml of ¹³C cyanuric acid (5.1.9) into a 50 ml volumetric flask (6.3). Make up to the mark with water and mix. The final concentration is 20 $\mu\text{g/ml}$. Transfer this solution into a 100 ml glass flask (6.2). The solution is stable under these conditions during at least 1 year.

5.3 Preparation of standard solutions

5.3.1 Melamine standard solution I, $\rho = 20 \mu\text{g/ml}$.

Pipette 1,0 ml of the melamine stock solution (5.2.1) into a 50 ml volumetric flask (6.3). Make up to the mark with the dilution solution (5.4.6) and mix. The concentration is 20 $\mu\text{g/ml}$. Transfer this solution into a 100 ml glass flask (6.2) and store it in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 month.

5.3.2 Cyanuric acid standard solution I, $\rho = 20 \mu\text{g/ml}$.

Pipette 1,0 ml of the cyanuric acid stock solution (5.2.2) into a 50 ml volumetric flask (6.3). Make up to the mark with the dilution solution (5.4.6) and mix. The concentration is 20 $\mu\text{g/ml}$. Transfer this solution into a 100 ml glass flask (6.2) and store it in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 month.

5.3.3 Melamine standard solution II, $\rho = 0,2 \mu\text{g/ml}$.

Pipette 1,0 ml of the melamine standard solution I (5.3.1) into a 100 ml volumetric flask (6.3). Make up to the mark with the dilution solution (5.4.6) and mix. The concentration is 0,2 $\mu\text{g/ml}$. Transfer this solution into a 100 ml glass flask (6.2) and store it in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 month.

5.3.4 Cyanuric acid standard solution II, $\rho = 0,2 \mu\text{g/ml}$.

Pipette 1,0 ml of the cyanuric acid standard solution I (5.3.2) into a 100 ml volumetric flask (6.3). Make up to the mark with the dilution solution (5.4.6) and mix. The concentration is 0,2 $\mu\text{g/ml}$. Transfer this solution into a 100 ml glass flask (6.2) and store it in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 month.

5.3.5 Benzoguanamine standard solution I, $\rho = 20 \mu\text{g/ml}$.

Pipette 1,0 ml of the benzoguanamine stock solution (5.2.3) into a 50 ml volumetric flask (6.3). Make up to the mark with the dilution solution (5.4.6) and mix. The concentration is 20 $\mu\text{g/ml}$. Transfer this solution into a 100 ml glass flask (6.2) and store it in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 month.

5.3.6 Benzoguanamine standard solution II, $\rho = 2 \mu\text{g/ml}$.

Pipette 1,0 ml of the benzoguanamine standard solution I (5.3.5) into a 10 ml volumetric flask (6.3). Make up to the mark with the dilution solution (5.4.6) and mix. The concentration is 2 $\mu\text{g/ml}$. Transfer this solution into a 20 ml glass flask (6.2) and store it in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 month.

5.3.7 ¹³C melamine and ¹³C cyanuric acid standard solution, $\rho = 2 \mu\text{g/ml}$.

Pipette 0,5 ml of the ¹³C melamine stock solution (5.2.4) and 0,5 ml of the ¹³C cyanuric acid stock solution (5.2.5) into a 5 ml volumetric flask (6.3). Make up to the mark with water and mix. The concentration for both compounds is 2 $\mu\text{g/ml}$. Transfer this solution into a 20 ml glass flask (6.2) and store it in a refrigerator at 4 °C (± 3 °C). The solution is stable under these conditions during at least 1 month if the mass is carefully monitored.

5.4 Preparation of calibration solutions

5.4.1 Pipette the volumes given in Table 1 into a 10 ml volumetric flask (6.3) and make up to the mark with the dilution solution (5.4.6).

Table 1 — Preparation of the calibration solutions

Calibration solution	Calibration solution number						
	1	2	3	4	5	6	7
Melamine standard solution I (5.3.1), in ml	0	0	0	0	0	0,05	0,25
Cyanuric acid standard solution I (5.3.2), in ml	0	0	0	0	0	0,05	0,25
Melamine standard solution II (5.3.3), in ml	0	0,05	0,25	1,25	2,5	0	0
Cyanuric acid standard solution II (5.3.4), in ml	0	0,05	0,25	1,25	2,5	0	0
Benzoguanamine standard solution II (5.3.6), in ml	0,5	0,5	0,5	0,5	0,5	0,5	0,5
¹³ C melamine and ¹³ C cyanuric acid standard solution (5.3.7), in ml	0,5	0,5	0,5	0,5	0,5	0,5	0,5
Concentration ¹³ C melamine – ¹³ C cyanuric acid, in $\mu\text{g/ml}$	0,1	0,1	0,1	0,1	0,1	0,1	0,1
Concentration native melamine – cyanuric acid, in $\mu\text{g/ml}$	0	0,001	0,005	0,025	0,05	0,1	0,5

5.4.2 **Mobile phase A for HPLC**, substance concentration $c(\text{ammonium acetate}) = 10 \text{ mmol/l}$.

Dissolve 0,77 g of ammonium acetate (5.1.3) in 1 000 ml of water.

5.4.3 **Mobile phase B for HPLC.**

Acetonitrile (5.1.2).

5.4.4 **Mobile phase A for UHPLC.**

Mix 3 parts per volume of formic acid (5.1.1) with 97 parts per volume of water.

5.4.5 **Mobile phase B for UHPLC**, $c(\text{ammonium acetate}) = 20 \text{ mmol/l}$ in a mixture of 3 parts per volume of water and 97 parts per volume of acetonitrile.

Dissolve 1,54 g of ammonium acetate (5.1.3) in 30 ml of water. Add 970 ml of acetonitrile (5.1.2) to the mixture and shake firmly. The turbid mixture will clear overnight.

The optimal choice for the mobile phases A and B can depend on the instrument configuration (see 6.10 to 6.12), particularly the type of column used. Equivalent products may be used if they can be shown to lead to the same results.

5.4.6 **Dilution solution.**

Transfer, using measuring cylinders, 70 ml of acetonitrile (5.1.2) and 30 ml of water into a 100 ml volumetric flask (6.3) and mix. Store at room temperature for no longer than 1 month.