
**Urine-absorbing aids for
incontinence — Polyacrylate
superabsorbent powders —**

**Part 3:
Test method for determination of
the particle size distribution by sieve
fractionation**

*Aides pour absorption d'urine — Méthodes d'essai pour caractériser
les matériaux absorbants à base de polymères —*

*Partie 3: Détermination de la distribution granulométrique des
particules au moyen du fractionnement par tamisage*

ISO 17190-3:2020

<https://standards.iteh.ai/catalog/standards/iso/20082b72-277c-4fb5-9af3-83890125ae63/iso-17190-3-2020>



iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 17190-3:2020

<https://standards.iteh.ai/catalog/standards/iso/20082b72-277c-4fb5-9af3-83890125ae63/iso-17190-3-2020>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

| | |
|---------------------------------------|----------|
| Foreword | iv |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 1 |
| 4 Principle | 2 |
| 5 Reagents and materials | 2 |
| 6 Apparatus | 2 |
| 7 Conditioning | 2 |
| 8 Sampling | 2 |
| 9 Procedure | 3 |
| 10 Calculation | 4 |
| 11 Report | 4 |
| 12 Precision | 5 |
| Bibliography | 6 |

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 17190-3:2020

<https://standards.iteh.ai/catalog/standards/iso/20082b72-277c-4fb5-9af3-83890125ae63/iso-17190-3-2020>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 173, *Assistive products*, Subcommittee SC 3, *Aids for ostomy and incontinence*.

This second edition cancels and replaces the first edition (ISO 17190-3:2001), which has been technically revised. The main changes compared to the previous edition are as follows:

- full text review and new laboratory analysis with statistical evaluation.

A list of all parts in the ISO 17190 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Urine-absorbing aids for incontinence — Polyacrylate superabsorbent powders —

Part 3:

Test method for determination of the particle size distribution by sieve fractionation

WARNING — This document does not claim to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. It is expected that the person performing this test has been fully trained in all aspects of this procedure.

1 Scope

This document specifies a test method for measuring particle size distributions up to 850 μ of cross-linked polyacrylate superabsorbent powders. It applies only to measurements made where sieve shaking is used for the separation. Tapping equipment is not expected to deliver the same results.

This method is applicable to powdered polymeric superabsorbent materials that are free-flowing under the specified test conditions.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 187, *Paper, board and pulps — Standard atmosphere for conditioning and testing and procedure for monitoring the atmosphere and conditioning of samples*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

sample

product or portion of a product taken from a production lot for testing purposes and identifiable and traceable back to its origin

3.2

specimen

specific portion of the identified *sample* (3.1) upon which a test is performed

4 Principle

A defined amount of superabsorbent powder is split into specific particle size fractions upon passing through a sequence of standard sieves. Each fraction is weighed, and the value reported as a percentage of the total amount of material.

5 Reagents and materials

None.

6 Apparatus

6.1 Analytical balance, capable of weighing the entire sieve assembly, or each individual sieve, to an accuracy of $\pm 0,01$ g.

6.2 Beaker, glass or plastic, having a capacity of 250 ml.

6.3 Sieve shaker, capable of holding standard sieves of 200 mm diameter, with bottom receiving pan, grounded to avoid static electricity.

EXAMPLE Retsch™¹⁾ AS 200 type or equivalent.

6.4 Stainless steel sieves of 200 mm diameter. Example hole sizes: 45, 106, 150, 300, 600 and 850 μ , with bottom receiving pan and top lid.

6.5 Brush, made of camel hair or vacuum cleaner with HEPA filter.

The brush/vacuum cleaner may only be used to clean powder from the standard sieves with hole sizes greater than 150 μ .

7 Conditioning

Samples shall be delivered in a closed container, to prevent absorption of atmospheric moisture. Allow the closed container to equilibrate to the laboratory conditions. The preferred test conditions are (23 ± 2) °C and (45 ± 15) % relative humidity. If these conditions are not available, test at ambient conditions and report the temperature and relative humidity. Measure these laboratory conditions in accordance with ISO 187.

8 Sampling

WARNING — Powder Handling – The German Commission for the Investigation of Health Hazards of Chemical Compounds in the Work Area (MAK Commission) has provided a guideline value for long-term exposure to the respirable portion of superabsorbent polyacrylate dust of $0,05 \text{ mg}\cdot\text{m}^{-3}$. The respirable portion is defined as those particles of less than 10 μ diameter. Commercial superabsorbent polymers typically contain less than 0,1 % of such particles. Precautions should be taken to avoid routine exposure to atmospheric respirable particles above this guideline value.

1) Retsch™ is an example of a suitable product available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this product.