

SLOVENSKI STANDARD SIST EN ISO 10519:2015/oprA1:2023

01-oktober-2023

Seme oljne repice - Določevanje vsebnosti klorofila - Spektrometrijska metoda -Dopolnilo A1: Priprava kalibracijske krivulje za določitev k-faktorja (ISO 10519:2015/DAM 1:2023)

Rapeseed - Determination of chlorophyll content - Spectrometric method - Amendment 1: Preparation of the calibration curve to determine the k factor (ISO 10519:2015/DAM 1:2023)

Rapssamen - Bestimmung des Chlorophyllgehaltes - Spektrometrisches Verfahren -Änderung 1: Erstellung der Kalibrierkurve zur Bestimmung des k-Faktors (ISO 10519:2015/DAM 1:2023)

https://standards.iteh.ai/catalog/standards/sist/e79f5463-bab6-48b7-af52-

Graines de colza - Détermination de la teneur en chlorophylle - Méthode spectrométrique - Amendement 1: Préparation de la courbe d'étalonnage pour déterminer le facteur k (ISO 10519:2015/DAM 1:2023)

Ta slovenski standard je istoveten z: EN ISO 10519:2015/prA1

ICS:

67.200.20 Oljnice

Oilseeds

SIST EN ISO 10519:2015/oprA1:2023 en,fr,de

SIST EN ISO 10519:2015/oprA1:2023

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN ISO 10519:2015/oprA1:2023</u> https://standards.iteh.ai/catalog/standards/sist/e79f5463-bab6-48b7-af52-6044773aa62b/sist-en-iso-10519-2015-opra1-2023

DRAFT AMENDMENT ISO 10519:2015/DAM 1

ISO/TC 34/SC 2

Secretariat: AFNOR

Voting begins on: **2023-08-18**

Voting terminates on: 2023-11-10

Rapeseed — Determination of chlorophyll content — Spectrometric method

AMENDMENT 1: Preparation of the calibration curve to determine the k factor

ICS: 67.200.20

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN ISO 10519:2015/oprA1:2023</u> https://standards.iteh.ai/catalog/standards/sist/e79f5463-bab6-48b7-af52-6044773aa62b/sist-en-iso-10519-2015-opra1-2023

This document is circulated as received from the committee secretariat.

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENT AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

ISO/CEN PARALLEL PROCESSING



Reference number ISO 10519:2015/DAM 1:2023(E) ISO 10519:2015/DAM 1:2023(E)

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 10519:2015/oprA1:2023 https://standards.iteh.ai/catalog/standards/sist/e79f5463-bab6-48b7-af52-6044773aa62b/sist-en-iso-10519-2015-opra1-2023



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office CP 401 • Ch. de Blandonnet 8 CH-1214 Vernier, Geneva Phone: +41 22 749 01 11 Email: copyright@iso.org Website: www.iso.org

Published in Switzerland

ISO 10519:2015/DAM 1:2023(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 34, *Food products,* Subcommittee SC 2, *Oleaginous seeds and fruits and oilseed meals.*

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at <u>www.iso.org/members.html</u>.

SIST EN ISO 10519:2015/oprA1:2023

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 10519:2015/oprA1:2023 https://standards.iteh.ai/catalog/standards/sist/e79f5463-bab6-48b7-af52-6044773aa62b/sist-en-iso-10519-2015-opra1-2023

Rapeseed — Determination of chlorophyll content — Spectrometric method

AMENDMENT 1: Preparation of the calibration curve to determine the k factor

10

Add the following text after the last paragraph:

k is a constant depending of the pathlenght of the spectrophotometer. As the spectrophotometer band width is variable, therefore k of 13 used in this standard, might need to be determined. The preparation of the calibration curve to determine the k factor is presented in Annex B.

After Annex A

Add the following as Annex B: A NDARD PREVIEW

(standards.iteh.ai)

SIST EN ISO 10519:2015/oprA1:2023 https://standards.iteh.ai/catalog/standards/sist/e79f5463-bab6-48b7-af52-6044773aa62b/sist-en-iso-10519-2015-opra1-2023