



SLOVENSKI STANDARD
oSIST prEN ISO 15708-2:2024
01-julij-2024

**Neporušitvene preiskave - Sevalne metode za računalniško tomografijo - 2. del:
Načela, oprema in vzorci (ISO/DIS 15708-2:2024)**

Non-destructive testing - Radiation methods for computed tomography - Part 2:
Principles, equipment and samples (ISO/DIS 15708-2:2024)

Zerstörungsfreie Prüfung - Durchstrahlungsverfahren für Computertomographie - Teil 2:
Grundlagen, Geräte und Proben (ISO/DIS 15708-2:2024)

Essais non destructifs - Méthodes par rayonnements pour la tomographie informatisée -
Partie 2: Principes, équipements et échantillons (ISO/DIS 15708-2:2024)

Ta slovenski standard je istoveten z: prEN ISO 15708-2

[oSIST prEN ISO 15708-2:2024](https://standards.sist.si/catalog/standards/sist/9c490000-1d07-450a-bc30-c0b71c190000/osist-pr-en-iso-15708-2-2024)

ICS:

19.100 Neporušitveno preskušanje Non-destructive testing

oSIST prEN ISO 15708-2:2024

en,fr,de



DRAFT International Standard

ISO/DIS 15708-2

Non-destructive testing — Radiation methods for computed tomography —

Part 2: Principles, equipment and samples

*Essais non destructifs — Méthodes par rayonnements pour la
tomographie informatisée —*

Partie 2: Principes, équipements et échantillons

ICS: 19.100

ISO/TC 135/SC 5

Secretariat: DIN

Voting begins on:
2024-04-08

Voting terminates on:
2024-07-01

<https://standards.iteh.ai/catalog/standards/sist/9c4960b0-fdc7-458a-be3c-ecb71e19b068/osist-pren-iso-15708-2-2024>

This document is circulated as received from the committee secretariat.

ISO/CEN PARALLEL PROCESSING

Reference number
ISO/DIS 15708-2:2024(en)

THIS DOCUMENT IS A DRAFT CIRCULATED FOR COMMENTS AND APPROVAL. IT IS THEREFORE SUBJECT TO CHANGE AND MAY NOT BE REFERRED TO AS AN INTERNATIONAL STANDARD UNTIL PUBLISHED AS SUCH.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

© ISO 2024

ISO/DIS 15708-2:2024(en)

iTeh Standards (<https://standards.iteh.ai>) Document Preview

[oSIST prEN ISO 15708-2:2024](https://standards.iteh.ai/catalog/standards/sist/9c4960b0-fdc7-458a-be3c-ecb71e19b068/osist-pren-iso-15708-2-2024)

<https://standards.iteh.ai/catalog/standards/sist/9c4960b0-fdc7-458a-be3c-ecb71e19b068/osist-pren-iso-15708-2-2024>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

© ISO 2024 – All rights reserved

ISO/DIS 15708-2:2024(en)

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General principles	1
4.1 Basic principles.....	1
4.2 Advantages of CT.....	2
4.3 Limitations of CT.....	2
4.4 Main CT process steps.....	3
4.4.1 Acquisition.....	3
4.4.2 Reconstruction.....	3
4.4.3 Visualization and analysis.....	4
4.5 Artefacts in CT images.....	4
5 Equipment and apparatus	4
5.1 General.....	4
5.2 Radiation sources.....	5
5.3 Detectors.....	6
5.4 Manipulation.....	6
5.5 Acquisition, reconstruction, visualization and storage system.....	6
6 CT system stability	7
6.1 General.....	7
6.2 X-Ray Stability.....	7
6.3 Manipulator stability.....	7
7 Geometric alignment	8
8 Sample considerations	8
8.1 Size and shape of sample.....	8
8.2 Materials (including table voltage/thickness of penetration).....	8
Annex A (informative) CT system components	10
Bibliography	16

<https://standards.iteh.ai/catalog/standards/sist/9c4960b0-fdc7-458a-be3c-ecb71e19b068/osist-pren-iso-15708-2-2024>

ISO/DIS 15708-2:2024(en)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by European Technical Committee CEN/TC 138 *Non-destructive testing* and was adopted under a special "fast-track procedure", by Technical Committee ISO/TC 135, *Non-destructive testing*, Subcommittee SC 5, *Radiographic testing*.

This third edition cancels and replaces the second edition (ISO 15708-2:2017), which has been technically revised.

The main changes are as follows:

- Correction of the vacuum level for activating the turbo pump in [A.1.1](#);
- Addition of photon counting as an example under semiconductors in [A.2.3](#);
- editorial changes.

A list of all parts in the ISO 15708 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Non-destructive testing — Radiation methods for computed tomography —

Part 2: Principles, equipment and samples

1 Scope

This document specifies the general principles of X-ray computed tomography (CT), the equipment used and basic considerations of sample, materials and geometry.

It is applicable to industrial imaging (i.e. non-medical applications) and gives a consistent set of CT performance parameter definitions, including how those performance parameters relate to CT system specifications.

This document deals with computed axial tomography and excludes other types of tomography such as translational tomography and tomosynthesis.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15708-1:2024, *Non-destructive testing — Radiation methods for computed tomography — Part 1: Terminology*

ISO 15708-3:2024, *Non-destructive testing — Radiation methods for computed tomography — Part 3: Operation and interpretation*

ISO 15708-4:2024, *Non-destructive testing — Radiation methods for computed tomography — Part 4: Qualification*

ISO 9712, *Non-destructive testing — Qualification and certification of NDT personnel*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 15708-1 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

4 General principles

4.1 Basic principles

Computed tomography (CT) is a radiographic inspection method which delivers three-dimensional information on an object from a number of radiographic projections either over cross-sectional planes (CT

ISO/DIS 15708-2:2024(en)

slices) or over the complete volume. Radiographic imaging is possible because different materials have different X-ray attenuation coefficients. In CT images, the linear X-ray attenuation coefficients are displayed as different CT grey values (or in false colour). For conventional radiography the three-dimensional object is X-rayed from one direction and an X-ray projection is produced with the corresponding information aggregated over the ray path. In contrast, multiple X-ray-projections of an object are acquired at different projection angles during a CT scan. From these projection images the actual slices or volumes are reconstructed. The fundamental advantage compared to radiography is the preservation of full volumetric information. The resulting CT image (2D-CT slice or 3D-CT volume), is a quantitative representation of the X-ray linear attenuation coefficient averaged over the finite volume of the corresponding volume element (voxel) at each position in the sample.

The linear attenuation coefficient characterizes the local instantaneous rate at which X-rays are attenuated as they propagate through the object during the scan. The attenuation of the X-rays as they interact with matter is the result of several different interaction mechanisms: Compton scattering and photoelectric absorption being the predominant ones for X-ray CT. The linear attenuation coefficient depends on the atomic numbers of the corresponding materials and is proportional to the material density. It also depends on the energy of the X-ray beam.

4.2 Advantages of CT

Among the radiographic techniques, CT can be an excellent examination technique whenever the primary goal is to locate and quantify volumetric details in three dimensions. In addition, since CT is X-ray based, it can be used on metallic and non-metallic samples, solid and fibrous materials and smooth and irregularly surfaced objects.

In contrast to conventional radiography, in which the internal features of a sample are projected onto a single image plane and thus are superposed on each other, in CT images the individual features of the sample appear separately from each other, preserving the full spatial information.

With proper calibration, dimensional inspections and material density determinations can also be carried out.

Complete three-dimensional representations of examined objects can be obtained either by reconstructing and assembling successive CT slices (2D-CT) or by direct 3D CT image (3D-CT) reconstruction. Computed tomography is therefore valuable in the industrial application areas of non-destructive testing, 2D and 3D metrology and reverse engineering.

CT has several advantages over conventional metrology methods:

- acquisition without contact;
- access to internal and external dimensional information;
- a direct input to 3D modelling especially of internal structures.

In some cases, dual energy (DE) CT acquisitions can help to obtain information on the material density and the average atomic number of certain materials. For known materials, the additional information can be used for improved discrimination or improved characterization.

4.3 Limitations of CT

CT is an indirect test procedure and measurements (e.g. of the size of material inhomogeneities or of wall thicknesses) must be compared with another absolute measurement procedure, see ISO 15708-3. Another potential drawback of CT imaging is the possible occurrence of artefacts (see 4.5) in the data. Artefacts limit the ability to quantitatively extract information from an image. Therefore, as with any examination technique, the user shall be able to recognize and discount common artefacts subjectively.

Like any imaging system, a CT system can never reproduce an exact image of the scanned object. The accuracy of the CT image is dictated largely by the competing influences of the imaging system, namely spatial resolution, statistical noise and artefacts. Each of these aspects is discussed in 4.4.1. See ISO 15708-3 for a more detailed description.

ISO/DIS 15708-2:2024(en)

CT grey values cannot be used to identify unknown materials unambiguously unless a priori information is available, since a given experimental value measured at a given position may correspond to a broad range of materials.

Furthermore, there shall be sufficient X-ray transmission through the sample at all projection angles (see [8.2](#)) without saturating any part of the detector.

4.4 Main CT process steps

4.4.1 Acquisition

Multiple projections are systematically recorded during a CT scan: the images are acquired from a number of different viewing angles. Feature recognition depends, among other factors, on the number of angles from which the individual projections are acquired. The CT image quality can be improved by increasing the number of projections in a scan.

As all image capture systems contain inherent artefacts, CT scans usually begin with the capture of offset and gain reference images to allow flat field correction; using black (X-rays off) and white (X-rays on with the sample out of the field of view) images to correct for detector anomalies. The capture of reference images for distortion correction (pin cushion distortion in the case of camera-based detector systems with optical distortion), and centre of rotation correction can also take place at this stage. These corrections are applied to each subsequently acquired image of the CT data set. Some systems can be configured to enhance either the X-ray settings or the image to ensure that the background intensity level of the captured images remains constant throughout the duration of the CT scan.

The quality of a CT image depends on a number of system-level performance factors, with one of the most important being spatial resolution.

Spatial resolution is generally quantified in terms of the smallest separation at which two features can be distinguished as separate entities. The limits of spatial resolution are determined by the design and construction of the system and by the resolution and number of CT projections. The resolution of the CT projection is limited by the maximum magnification that can be used while still imaging all parts of the sample at all rotation angles.

It is important to note that the smallest feature that can be detected in a CT image is not the same as the smallest that can be resolved spatially. A feature considerably smaller than a single voxel can affect the voxel to which it corresponds to such an extent that it appears with a visible contrast so that it can be easily detected with respect to adjacent voxels. This phenomenon is due to the "partial-volume effect".

Although region-of-interest CT (local tomography) can improve spatial resolution in certain regions of larger objects, it introduces artefacts (due to incomplete data) which can sometimes be reduced by special processing.

Radiographic imaging, as used for CT examination, is always affected by noise. In radiography this noise arises from two sources: (1) intrinsic variation corresponding to photon statistics in the emission and detection of photons and (2) variations specific to instruments and processing used. Noise in CT projections is often amplified by the reconstruction algorithm. In the CT images statistical noise appears as random variation superimposed on the CT grey value of each voxel, limiting the density resolution.

Although statistical noise is unavoidable, the signal-to-noise ratio can be improved by increasing the number of projections and/or time of exposure for each of them, the intensity of the X-ray source or the voxel size. However, some of these measures will decrease spatial resolution. This trade-off between spatial resolution and statistical noise is inherent in computed tomography.

4.4.2 Reconstruction

A CT scan initially produces a number of projections of an object. The subsequent reconstruction of the CT image from these individual projections is the main step in computed tomography, which distinguishes this examination technique from other radiographic techniques.

ISO/DIS 15708-2:2024(en)

The reconstruction software may apply additional corrections to the CT projections during reconstruction, e.g. reduction of noise, correction of beam hardening and/or scattered radiation.

Depending on the CT system, either individual CT slices or 3D CT images are reconstructed.

4.4.3 Visualization and analysis

This step includes all operations and data manipulations, for extracting the desired information from the reconstructed CT image.

Visualisation can either be performed in 2D (slice views) or in 3D (volume). 2D visualisation allows the user to examine the data slice-wise along a defined axis (generally it can be an arbitrary path).

For 3D imaging, the CT volume or selected surfaces derived from it, are used for generating the desired image according to the optical model underlying the algorithm. The main advantage of this type of visualisation is that the visual perception of the image corresponds well with the natural appearance of the object for the human eye, although features may appear superimposed in the 2D-representation on a screen.

During visualisation, additional artefacts of different origin can occur, especially in the 3D imaging of the CT volume. Such artefacts due to sampling, filtering, classification and blending within the visualisation software depend on the hardware and software used, as well as the visualisation task at hand. Therefore such artefacts are not included in the definition of artefacts as found in 4.5. Nevertheless, the user should be aware that data may be misinterpreted in this process step.

To highlight features of interest during visualisation, different digital filter operations can be performed. It is characteristic of all these operations that although they enhance one or more properties of the data, they simultaneously deteriorate other properties (for example: highlighting the edges deteriorates recognition of inner structures of an object). Therefore digital filters should always be used cautiously for specific tasks, being aware which benefits and which detriments they are associated with.

A computer used for 3D visualisation should be able to process the complete volume of interest in its main memory. The corresponding monitor should have a resolution, a dynamic range and settings sufficient for the given visualisation task. Adequate vision of the personnel shall be ensured in accordance with ISO 9712.

4.5 Artefacts in CT images

An artefact is an artificial feature which appears on the CT image but does not correspond to a physical feature of the sample. Artefacts result from different origins; they can be classified into artefacts arising from the measurement itself and the equipment (artefacts due to a finite beam width, scattered radiation, instabilities and detector peculiarities), and artefacts inherent to the technique (e.g. beam hardening). Artefacts can also be divided into acquisition artefacts (e.g. scattered radiation, ring artefacts) and reconstruction artefacts (e.g. cone beam artefacts). Some artefacts can be eliminated by using an appropriate measurement technique with suitable parameters, while others can only be reduced in their extent. Artefacts may be detrimental for specific measurement or analysis tasks, but may have no impact on certain other analyses. With this fact in mind, the type and extent of artefacts in a data set has to be evaluated in the context of the corresponding analysis task.

Noise and the partial volume effect are not considered as artefacts in this standard.

More details are given in ISO 15708-3:2017, 5.5.

5 Equipment and apparatus

5.1 General

In relation to performance, a CT system can be considered as comprising four main components: the X-ray source, detector, sample manipulation stages (including any mechanical structures that influence image stability) and reconstruction/visualisation system.