

**ISO/FDIS 24384:2023(E)**

Date: 2023-11-16

ISO/TC 147/SC 2/WG 82 ~~N11~~

Secretariat: DIN

**Water quality — Determination of chromium(VI) and chromium(III) in water — Method using liquid chromatography with inductively coupled plasma mass spectrometry (LC-ICP-MS) after chelating pretreatment**

**iTeh Standards**  
**(<https://standards.itih.ai>)**  
**Document Preview**

**First edition**

Date: 2023-07-28

ISO/FDIS 24384

<https://standards.itih.ai/catalog/standards/sist/d38850b5-8b00-4e4f-8b76-7d09444b37af/iso-fdis-24384>

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO Copyright Office  
CP 401 • CH-1214 Vernier, Geneva  
Phone: + 41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)  
Published in Switzerland.

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO/FDIS 24384](#)

<https://standards.iteh.ai/catalog/standards/sist/d38850b5-8b00-4e4f-8b76-7d09444b37af/iso-fdis-24384>

<b>Contents</b>	<b>Page</b>
Foreword.....	iv
Introduction.....	v
1 Scope .....	1
2 Normative references .....	1
3 Terms and definitions.....	1
4 Principle.....	1
5 Interferences.....	2
5.1 General .....	2
5.2 Samples .....	2
5.3 Sample storage and sample preparation .....	2
5.4 Chelating pretreatment.....	2
5.5 LC-ICP-MS measurements.....	2
6 Reagents and standards .....	3
7 Apparatus .....	5
8 Sampling, preservation and storage of samples.....	7
9 Procedure .....	7
9.1 Sample preparation .....	7
9.1.1 pH-adjustment of water sample .....	7
9.1.2 Chelating pretreatment.....	7
9.2 Optimization of operating condition for LC-ICP-MS.....	8
9.3 Identification of Cr(VI) and Cr(III) on LC-ICP-MS.....	8
9.4 Blank value measurements.....	8
10 Calibration.....	9
10.1 General requirements.....	9
10.2 Calibration covering the total procedure .....	9
10.3 Recovery test of target substances .....	10
11 Calculation.....	11
11.1 Use of the calibration curve to determine the result.....	11
11.2 Calculation of results after calibration .....	11
11.3 Treatment of results lying outside the calibration range .....	11
12 Expression of results .....	11
13 Test report .....	11
Annex A (informative) Example of an operating condition of LC-ICP-MS and chromatogram of Cr(VI) and Cr(III) in the case of PDCA-chelating pretreatment.....	13
Annex B (informative) Example of operating condition of LC-ICP-MS and chromatograms of Cr(VI) and Cr(III) in the case of EDTA-chelating pretreatment .....	15
Annex C (informative) Performance data .....	17
Bibliography .....	19

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement. (<https://standards.iteh.ai>)

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Chromium (Cr) exists in natural resources and is also widely used in industries as plating agents, paints, dyes, catalysts, and dietary supplements. The Cr(VI) compounds are highly harmful and recognized to be a human carcinogen. The Cr(III) compounds are recently used as a substitute for Cr(VI) compounds in industries, e.g. plating. In wastewater, surface water, or drinking water, chromium mainly exists in two oxidation states: +3 [Cr(III)] and +6 [Cr(VI)]. However, the proportion between Cr(VI) and Cr(III) is quite variable. Therefore, the determination of the individual oxidation states of chromium is crucial to evaluate and control the risk of chromium to human and environmental health. This document will be beneficial to perform a robust, simple, and rapid determination of chromium of the individual oxidation states.

# iTeh Standards (<https://standards.iteh.ai>) Document Preview

[ISO/FDIS 24384](https://standards.iteh.ai/catalog/standards/sist/d38850b5-8b00-4e4f-8b76-7d09444b37af/iso-fdis-24384)

<https://standards.iteh.ai/catalog/standards/sist/d38850b5-8b00-4e4f-8b76-7d09444b37af/iso-fdis-24384>



# Water quality — Determination of chromium(VI) and chromium(III) in water — Method using liquid chromatography with inductively coupled plasma mass spectrometry (LC-ICP-MS) after chelating pretreatment

WARNING — Persons using this document should be familiar with normal laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices.

IMPORTANT — It is absolutely essential that tests conducted according to this document be carried out by suitably qualified staff.

## 1 Scope

This document specifies a method for the determination of hexavalent chromium [Cr(VI)] and trivalent chromium [Cr(III)] in water by liquid chromatography with inductively coupled plasma mass spectrometry (LC-ICP-MS) after chelating pretreatment.

This method is applicable to the determination of Cr(VI) and Cr(III) dissolved in wastewater, surface water, groundwater, or drinking water from 0,20 µg/l to 500 µg/l of each compound as chromium (Cr) mass. Samples containing Cr at concentrations higher than the working range can be analysed following appropriate dilution of the sample.

## 2 Normative references

ISO/FDIS 24384

<https://standards.iteh.ai/catalog/standards/sist/d38850b5-8b00-4e4f-8b76-7d09444b37af/iso-fdis-24384>

The following **referenced document is indispensable for the application** documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods — Part 1: Linear calibration function*

## 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

## 4 Principle

The chemical forms of various Cr(III) species in water samples are unified to a stable Cr(III) complex by a chelating pretreatment with 2,6-pyridinedicarboxylic acid (PDCA) or ethylenediaminetetraacetic acid (EDTA) after adjusting the sample solution pH to  $6,9 \pm 0,1$ .<sup>[1][2]</sup> Liquid chromatography combined with inductively coupled plasma mass spectrometry (LC-ICP-MS) determines the chromatographically

separated Cr(VI) and the Cr(III)-PDCA complex or Cr(III)-EDTA complex in the pretreated sample solutions.

## 5 Interferences

### 5.1 General

If any of the interferences described in 5.2 to 5.5 are recognized or can be expected due to additional information about the sample, the sum of Cr(VI) and Cr(III) concentrations determined by the proposed method should be compared with the total Cr concentration which can be determined by ICP-MS.

If the Cr(III) and Cr(VI) species are properly determined by the proposed method, the sum of Cr(III) and Cr(VI) should agree with the total Cr with a difference of  $\leq 30\%$ . An exceedance of this limit (30 %) indicates the occurrence of interferences.

**NOTE** The water sample can contain Cr<sup>0</sup> or Cr(III) species that cannot be complexed via EDTA or PDCA (e.g. nanoparticles, colloids). These stable species can be so small that they are not retained even by filtration through a 0,2  $\mu\text{m}$  filter membrane. However, as they are uncomplexed, they will not be detected by the LC-ICP-MS analysis and can thus result in a greater difference between the sum of Cr species determined by LC-ICP-MS and the total Cr determination.

### 5.2 Samples

Reductants or oxidants in the sample may lead to false results for the Cr(VI) and/or Cr(III) concentration through the reduction of Cr(VI) to Cr(III) or oxidation of Cr(III) to Cr(VI), respectively. For example, divalent iron ions or ascorbic acid at 1 mg/l or 10 mg C/l, respectively, cause the reduction within 10 min.<sup>[1]</sup> Organic matter at high concentrations may slowly reduce Cr(VI) to Cr(III). For example, tartaric acid at 100 mg C/l reduces Cr(VI) to Cr(III) in one night at room temperature, however, the same compound at 10 mg C/l does not cause the reduction. The sample pH also changes the redox equilibrium between Cr(VI) and Cr(III) and the complexation equilibrium between Cr species with coexisting inorganic and organic substances.<sup>[3]</sup>

### 5.3 Sample storage and sample preparation

Potential sources of Cr contamination during sampling, sample storage, and sample preparation include: labware, containers, sampling equipment, reagents, water and human contact. Potential unexpected redox reaction between Cr(III) and Cr(VI) may occur during these operations. All apparatus and labware shall be cleaned using the cleaning procedure (see Clause 7).

### 5.4 Chelating pretreatment

Transition metal cations at high concentrations in the sample solution may lead to a negative biased value for the Cr(III) concentration because these metals decrease the concentration of free chelating agents due to complex formation with chelating agents. For example, neither cobalt(II) nor nickel(II) up to 10 mg/l interfere, however, these metal ions<sup>[1]</sup> and also calcium ion interfere above 100 mg/l.

Large amounts of organic matter which strongly binds to Cr(III) species in the sample solution may lead to a negatively biased value for the Cr(III) concentration. For example, Cr(III)-PDCA complex formation in the PDCA chelating pretreatment is depressed by the presence of EDTA as an interfering substance at 100 mg/l level, although not depressed at 10 mg/l.<sup>[1]</sup> The opposite is also true and Cr(III)-EDTA complex formation in the pretreatment of Cr(III) with EDTA is inhibited by the presence of large amounts of PDCA as an interfering substance. The depression at the chelating procedure can be evaluated by recovery experiments (10.3).

### 5.5 LC-ICP-MS measurements

Polyatomic ions are formed in the argon plasma of the ICP-MS by the reaction among argon, water, reagents and sample matrix, etc. The formation of  $^{40}\text{Ar}^{12}\text{C}^+$  and  $^{37}\text{Cl}^{16}\text{O}^+$  may interfere with the ICP-MS



detection of Cr at mass-to-charge ratio ( $m/z$ ) 52 and at  $m/z$  53, respectively. Some of the interferences such as  $^{40}\text{Ar}^{12}\text{C}^+$  can be reduced by using a collision-reaction cell of the ICP-MS. On the other hand, some of the interferences such as  $^{37}\text{Cl}^{16}\text{O}^+$  can be reduced by chromatographic separation of chloride (Cl) ions from the Cr species.<sup>[4]</sup> High-resolution or tandem mass spectrometers, e.g. sector-field or MS/MS-type mass spectrometers, can also reduce the polyatomic interferences very effectively.

LC retention time of Cr species may shift for water samples containing salts or organic matter at high concentrations. If this causes the peak of Cr(III)-PDCA complex [or Cr(III)-EDTA complex] to overlap with Cr(VI), it may cause positive and negative interferences to Cr(VI) and Cr(III), respectively. In such cases, adjust elution conditions such as composition of eluent to ensure adequate separation.

The metallic parts of LC-ICP-MS instruments, e.g. LC columns, tubes, and connectors, potentially contaminate the eluent if in contact. Therefore, metallic parts or pathways should be avoided or reduced to a minimum.

## 6 Reagents and standards

Unless otherwise indicated, reagents of purity grade “for analysis” or “for trace analysis” are used as reagents. If available, use only reagents of pro analysis grade (or purer) free of compounds containing Cr. Weigh the reagents with an accuracy of  $\pm 1\%$  of the nominal mass, unless stated otherwise.

Prepare alternative concentrations and volumes of solutions as described hereafter, if necessary. Alternatively, use commercially available stock solutions of the required concentration.

**6.1 Water**, with an electrical resistivity of  $\geq 18,2 \text{ M}\Omega \text{ cm}$  (25 °C).

The water shall not contain any measurable quantity of Cr(III) and Cr(VI) or interfering compounds at or above one-third the method quantification limit.

**6.2 Nitric acid**,  $w(\text{HNO}_3) = 650 \text{ g/kg}$ ,  $\rho(\text{HNO}_3) = 1,4 \text{ g/ml}$ .

NOTE Nitric acid is available as  $\rho(\text{HNO}_3) = 1,38 \text{ g/ml}$  [ $w(\text{HNO}_3) = 610 \text{ g/kg}$ ] and  $\rho(\text{HNO}_3) = 1,42 \text{ g/ml}$  [ $w(\text{HNO}_3) = 690 \text{ g/kg}$ ] as well as  $\rho(\text{HNO}_3) = 1,40 \text{ g/ml}$  [ $w(\text{HNO}_3) = 650 \text{ g/kg}$ ].

**6.3 Nitric acid stock solution**,  $c(\text{HNO}_3) = 6 \text{ mol/l}$ .

Transfer 25 ml of water (6.1) to 100 ml volumetric flask (7.14), and add 27 ml of nitric acid (6.2) and then fill up to mark with water (6.1).

**6.4 Nitric acid solution for pH adjustment**,  $c(\text{HNO}_3) = 2 \text{ mol/l}$ .

Transfer 33 ml of nitric acid stock solution (6.3) to a 100 ml volumetric flask (7.14) and fill up to mark with water (6.1).

**6.5 Sodium hydroxide**, NaOH.

**6.6 Sodium hydroxide solution for pH adjustment**,  $c(\text{NaOH}) = 2 \text{ mol/l}$ .

Weigh 8 g of sodium hydroxide pellets (6.5) and transfer them to a 100 ml beaker. Then add approximately 50 ml water (6.1) and stir with a stirrer (7.12) until the pellets have dissolved. Transfer to a 100 ml volumetric flask (7.14) and fill up to the mark with water (6.1). Commercially available solution of sodium hydroxide can be used to dilute them to the required concentration.

**6.7 2,6-Pyridinedicarboxylic acid (PDCA)**, CAS [Registry Number<sup>1</sup>](#) 499-83-2, C<sub>7</sub>H<sub>5</sub>NO<sub>4</sub>.

**6.8 Disodium hydrogenphosphate**, Na<sub>2</sub>HPO<sub>4</sub>.

**6.9 Ammonium acetate**, CH<sub>3</sub>COONH<sub>4</sub>.

**6.10 Ethylenediamine-*N,N,N',N'*-tetraacetic acid (EDTA) disodium salt dihydrate** (CAS 6381-92-6, C<sub>10</sub>H<sub>14</sub>Na<sub>2</sub>N<sub>2</sub>O<sub>8</sub>·2H<sub>2</sub>O) or **EDTA dipotassium salt dihydrate** (CAS 25102-12-9, C<sub>10</sub>H<sub>14</sub>K<sub>2</sub>N<sub>2</sub>O<sub>8</sub>·2H<sub>2</sub>O).

**6.11 Ammonia solution, mass fraction**,  $w(\text{NH}_4\text{OH}) = 280 \text{ g/kg}$ .

**6.12 Ammonia solution for pH adjustment**,  $c(\text{NH}_3) = 1 \text{ mol/l}$ .

**6.13 Potassium dichromate**, K<sub>2</sub>Cr<sub>4</sub>O<sub>7</sub>.

**6.14 Chromium(III) nitrate nonahydrate**, Cr(NO<sub>3</sub>)<sub>3</sub>·9H<sub>2</sub>O.

**6.15 Cr(VI) stock solution**,  $\rho[\text{Cr(VI)}] = 1\,000 \text{ mg/l}$ .

WARNING — Potassium chromate can be carcinogenic.

Heat 5 g of potassium dichromate (6.13) at 150 °C for 1 h and then cool at room temperature in a dried desiccator. Dissolve 2,829 g of the dried potassium dichromate (6.13) with water (6.1) in a 1 000 ml volumetric flask (7.14) and fill up to mark with water (6.1). Commercially available Cr(VI) stock solution of the required concentration can be used.

**6.16 Cr(VI) standard solution**,  $\rho[\text{Cr(VI)}] = 10 \text{ mg/l}$ .

Transfer 1,00 ml of the Cr(VI) stock solution (6.15) to a 100 ml volumetric flask (7.14) and fill up to the mark with water (6.1). Prepare this solution on the day of use.

**6.17 Cr(III) stock solution**,  $\rho[\text{Cr(III)}] = 1\,000 \text{ mg/l}$ .

Dissolve 7,696 g of chromium(III) nitrate nonahydrate (6.14) in 250 ml water and transfer to a 1 000 ml volumetric flask (7.14). Add 50 ml of nitric acid stock solution (6.3) and fill up to the mark with water (6.1). Commercially available Cr(III) stock solution of the required concentration can be used.

**6.18 Cr(III) standard solution**,  $\rho[\text{Cr(III)}] = 10 \text{ mg/l}$ .

Transfer 1,00 ml of the Cr(III) stock solution (6.17) to a 100 ml volumetric flask (7.14) and fill up to the mark with water (6.1). Prepare this solution on the day of use.

**6.19 PDCA solution**,  $c(\text{PDCA}) 0,02 \text{ mol/l}$ .

Dissolve 3,35 g of 2,6-pyridinedicarboxylic acid (6.7), 2,85 g of disodium hydrogenphosphate (6.8), and 38,5 g of ammonium acetate (6.9) to 900 ml of water (6.1) in a bottle (7.16) of 1 000 ml. Adjust the pH of the solution to pH 6,9 ± 0,1 using a pH meter (7.11) by adding the sodium hydroxide solution (6.6). Transfer the adjusted PDCA solution to a 1 000 ml volumetric flask and fill up to the mark with water (6.1).

---

<sup>1</sup> [Chemical Abstracts Service \(CAS\) Registry Number<sup>®</sup>](#) is a trademark of the [American Chemical Society \(ACS\)](#). This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used if they can be shown to lead to the same results.