



Designation: E682 – 92 (Reapproved 2011)

## Standard Practice for Liquid Chromatography Terms and Relationships<sup>1</sup>

This standard is issued under the fixed designation E682; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

*This standard has been approved for use by agencies of the U.S. Department of Defense.*

### 1. Scope

1.1 This practice deals primarily with the terms and relationships used in liquid column chromatography. However, most of the terms should also apply to other kinds of liquid chromatography, notably planar chromatography such as paper or thin-layer chromatography.

NOTE 1—Although electrophoresis can also be considered a liquid chromatographic technique, it and its associated terms have not been included in this practice.

1.2 Since most of the basic terms and definitions also apply to gas chromatography, this practice uses, whenever possible, symbols identical to Practice E355.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

### 2. Referenced Documents

2.1 *ASTM Standards*:<sup>2</sup>

D3016 Practice for Use of Liquid Exclusion Chromatography Terms and Relationships

E355 Practice for Gas Chromatography Terms and Relationships

E1151 Practice for Ion Chromatography Terms and Relationships

### 3. Names of Techniques

NOTE 2—In the chromatographic literature one may often find the term high-performance (or high-pressure) liquid chromatography, abbreviated as HPLC. This term was introduced to distinguish the present-day column chromatographic techniques employing high inlet pressures and columns containing small diameter packing from the classical methods. The utilization of this term or any derivative term (for example, HPLSC for high-performance liquid-solid chromatography) is not recommended.

Similarly, the use of the term high-performance thin-layer

<sup>1</sup> This practice is under the jurisdiction of ASTM Committee E13 on Molecular Spectroscopy and Separation Science and is the direct responsibility of Subcommittee E13.19 on Separation Science.

Current edition approved Nov. 1, 2011. Published December 2011. Originally approved in 1979. Last previous edition approved in 2006 as E682 – 92 (2006). DOI: 10.1520/E0682-92R11.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

chromatography, abbreviated as HPTLC, describing newer variations of thin-layer chromatography, is also not recommended.

3.1 *Liquid Chromatography*, abbreviated as LC, comprises all chromatographic methods in which the mobile phase is liquid under the conditions of analysis. The stationary phase may be a solid or a liquid supported by or chemically bonded to a solid.

3.2 The stationary phase may be present on or as a plane (*Planar Chromatography*), or contained in a cylindrical tube (*Column Chromatography*).

3.3 Separation is achieved by differences in the distribution of the components of a sample between the mobile and stationary phases, causing them to move along the plane surface or through the column at different rates (differential migration).

3.3.1 In *Planar Chromatography*, the differential migration process will cause the sample components to separate as a series of spots behind the mobile phase front.

3.3.2 In *Column Chromatography*, the differential migration process will cause the sample components to elute from the column at different times.

3.3.3 In *Dry-Column Chromatography*, mobile phase flow is stopped as soon as the mobile phase has reached the end of the column of *dry* medium. This column can be glass or a rigid or flexible solvent compatible plastic. Solute visualization and recovery are from the extruded or sliced column packing.

3.3.4 In *Flash Chromatography*, mobile phase flow is continued after the mobile phase has reached the end of the column of *dry* medium until elution of the desired components is achieved. Often low pressures, compatible with the materials of construction of the column, are applied to the top of the column to speed up the elution.

3.4 The basic process of selective distribution during the chromatographic process can vary depending on the type of stationary phase and the nature of the mobile phase.

3.4.1 In *Liquid-Liquid Chromatography*, abbreviated LLC, the stationary phase is a liquid and the separation is based on selective partitioning between the mobile and stationary liquid phases.

3.4.2 In *Liquid-Solid Chromatography*, abbreviated as LSC, the stationary phase is an interactive solid. Depending on the

type of the solid, separation may be based on selective adsorption on an inorganic substrate such as silica gel, or an organic gel. In this definition, *Ion-Exchange Chromatography* is considered to be a special case of LSC in which the interactive solid has ionic sites and separation is due to ionic interaction.

3.4.2.1 In this definition, *Ion Exchange Chromatography* is considered to be a special case of LSC in which the interactive solid has permanently bonded ionic sites and separation is due to electrostatic interaction.

3.4.2.2 In this definition, *Ion Pair Chromatography* is considered to be a special case of LSC in which ionic counterions are added to the mobile phase to effect the separation of ionic solutes. In this technique both electrostatic and adsorptive forces are involved in the separation.

NOTE 3—Other terminology for this technique include, but are not limited to, extraction chromatography, paired ion chromatography, soap chromatography, ion pair extraction chromatography, ion pair partition chromatography, and ion interaction chromatography, but utilization of these terms is not recommended.

3.4.2.3 In this definition, *Affinity Chromatography* is considered to be a special case of LSC in which special ligands are bonded to a stationary phase so that bio-specific interactions (for example, antibody/antigen, enzyme/substrate) may be invoked to effect the separation.

3.4.2.4 In this definition, *Ion Chromatography* is considered to be a special application of LSC in which the ion exchange mechanism is still effecting the separation. Special columns or devices, after the separating column, may be needed to remove higher concentrations of inorganic ions which might otherwise interfere with the detectability using conductivity. See Practice E1151 for further details of nomenclature for this technique.

3.4.2.5 In this definition, *Hydrophobic Interaction Chromatography*, is considered to be a special application of LSC in which the separation is based upon interaction of the hydrophobic moieties of the solutes and the hydrophobic moieties of the sites on a reversed phase packing. High to low salt gradients are used to effect this type of separation.

3.4.3 In some cases, such as with bonded stationary phases, the exact nature of the separation process is not fully established and it may be based on a combination of liquid-liquid and liquid-solid interactions.

3.4.4 In *Steric Exclusion Chromatography*, the stationary phase is a noninteractive porous solid, usually silica or an organic gel. In this case, separation is affected by the size of the sample molecules, where those which are small enough penetrate the porous matrix to varying extents and degrees while those that are largest are confined to the interstitial region of the particles. Thus, the larger molecules elute before the smaller molecules. See Practice D3016 for further details of nomenclature for this technique.

3.5 In liquid chromatography, the composition of the mobile phase may be constant or changing during a chromatographic separation.

3.5.1 The term *Isocratic* may be used when the composition of the mobile phase at the column inlet is kept constant during a chromatographic separation.

3.5.2 The term *Gradient* is used to specify the technique when a deliberate change in the mobile phase operating condition is made during the chromatographic procedure. The change is usually in mobile phase composition, flow rate, pH, or temperature. The first-named change is called *Gradient Elution*. *Flow Programming* is a technique where the mobile phase linear velocity is changed during the chromatographic procedure. The changes are made to enhance separation or to speed elution of sample components, or both. Such changes in operating conditions may be continuous or step-wise.

3.6 In the standard modes of liquid chromatography, the stationary phase is more polar than the mobile phase. This is referred to as *Normal Phase Chromatography*. The opposite case is also possible, in which the mobile phase is more polar than the stationary phase. This version of the technique is called *Reversed-Phase Chromatography*.

3.7 *Planar Chromatography* comprises two versions: paper chromatography and thin-layer chromatography.

3.7.1 In *Paper Chromatography*, the process is carried out on a sheet or strip of paper. Separation is usually based on LLC in which water held on the cellulose fibers acts as the stationary phase. Separation based on LSC may also be utilized when the paper is impregnated or loaded with an interactive solid.

3.7.2 In *Thin-Layer Chromatography*, the solid stationary phase is utilized in the form of a relatively thin layer on an inactive plate or sheet.

3.7.3 In any version of planar chromatography, the mobile phase may be applied in a number of ways. In normal usage, *Ascending*, *Descending*, and *Horizontal Development*, the mobile phase movement depends upon capillary action. In *Horizontal Development*, the mobile phase may move predominantly linearly or radially. In *Radial Development*, the mobile phase is applied as a point source. Devices have been employed which accelerate the mobile phase movement on planar layers by pressure or centrifugal force.

3.7.4 The *Mobile Phase Front* is the leading edge of mobile phase as it traverses the planar media. In all forms of development, including radial, the local tangent to the *Mobile Phase Front* is everywhere normal to the local direction of development.

3.7.5 *Consecutive Developments* of planar media may be carried out after removal of the mobile phase from a previous development. If the consecutive development is accomplished in the same direction as previously, this is *Multiple Development*. If a second development is accomplished at a right angle to the first development, this is *Two-Dimensional Development*. Continuous development of planar media is possible by allowing evaporation of the mobile phase near the *Mobile Phase Front*.

3.7.6 *Impregnation* is the technique of applying a reagent to the planar media to effect an enhanced separation or detection.

This impregnation is accomplished by dipping or spraying a reagent solution after the preparation of the medium, or by incorporating during the manufacturing process.

#### 4. Apparatus

4.1 *Pumps*—The function of the pumps is to deliver the mobile phase at a controlled flow rate to the chromatographic column.

4.1.1 *Syringe Pumps* have a piston that advances at a controlled rate within a smooth cylinder to displace the mobile phase.

4.1.2 *Reciprocating Pumps* have a single or dual chamber from which mobile phase is displaced by reciprocating piston(s) or diaphragm(s). The chamber volume is relatively small compared to the volume of the column.

4.1.3 *Pneumatic Pumps* employ a gas to displace the mobile phase either directly or through a piston or collapsible container. The volume within these pumps may be large or small as compared to the volume of the column.

4.2 *Sample Inlet Systems* represent the means for introducing samples into the column.

4.2.1 *Septum Injectors*—Sample contained in a syringe is introduced directly into the pressurized flowing mobile phase by piercing an elastomeric barrier. The syringe is exposed to pressure and defines the sample volume.

4.2.2 *Septumless Injectors*—Sample contained in a syringe is introduced into an ambient-pressure chamber, and the chamber is subsequently mechanically displaced into the pressurized flowing mobile phase. The syringe is not exposed to pressure and defines the sample volume.

4.2.3 *Valve Injectors*—Sample contained in a syringe (or contained in a sample vial) is injected into (or drawn into) an ambient-pressure chamber which is subsequently displaced into the pressurized flowing mobile phase. The displacement is by means of rotary or sliding motion. The chamber is a section (loop) of tubing or an internal chamber. The chamber can be completely filled, in which case the chamber volume defines the sample volume, or it can be partially filled, in which case the syringe calibration marks define the sample volume.

4.3 *Columns* consist of tubes that contain the stationary phase and through which the mobile phase flows.

4.3.1 *Separating Column* is the column on which the separation of the solutes is accomplished.

4.3.2 *Pre-column* is a column that has been used classically to precondition the mobile phase, placed between the pump and the injector. In the instance of its use with liquid-liquid separations involving coated stationary phases, such a column contained an excess of the coating phase to presaturate the mobile phase so it would not strip the same phase from the coated stationary phase during the separation. Its predominate use today is as a protector column for silica based column packing materials. It is filled with large particle silica which is slowly dissolved by polar, ionic mobile phases. By so doing, the silicate saturated mobile phase cannot dissolve the silica backbone of the analytical or preparative column.

4.3.3 *Guard Column* is a protector column placed between the injector and the separating column. The purpose of this column is to be the *final filter* for the sample, adsorbing

unwanted sample components that otherwise might bind irreversibly to the separating column. It has a volume of no more than  $\frac{1}{20}$  the volume of the separating column. It may be filled with any material which will effectively remove the unwanted components without interfering with subsequent chromatographic processes.

4.3.4 *Concentrator Column* is a small column placed in-line at the loop injector for introducing a dilute sample which is collected into it before elution onto the separating column.

NOTE 4—Other terminology for this technique include, but are not limited to, trace enrichment column, collector column, and sample concentration column, but utilization of these terms is not recommended.

4.3.5 Column sizes with various internal diameters (ID) and lengths can be made. Larger columns present no problems concerning nomenclature, but columns with small internal diameters are now being used. As pointed out by Basey and Oliver<sup>3</sup> as many as nine terms (capillary, microcapillary, narrow bore capillary, micro, microbore, ultramicro, narrow bore, small bore, and small diameter) have been seen in the literature and with no clear distinction between them when the actual column ID is examined. It is recommended that all descriptive terms regarding column ID be discontinued, that is, *packed column, 1000  $\mu\text{m}$  ID  $\times$  100 mm* or *open column, 250  $\mu\text{m}$  ID  $\times$  1 m*.

4.3.6 *Column Inlet* is the end of a column where the mobile phase is introduced.

4.3.7 *Column Outlet* is the end of a column where the mobile phase exits.

4.3.8 *Frit* is the porous element placed at the ends of a chromatography column, or in a special device for in-line filtration to effect the removal of particulate material in the mobile phase or the sample solution.

4.4 *Detectors* are devices that respond to the presence of eluted solutes in the mobile phase emerging from the column. Ideally, the response should be proportional to the mass or concentration of solute in the mobile phase. Detectors may be divided either according to the type of measurement or the principle of detection.

4.4.1 *Bulk Property Detectors* measure the change in a physical property of the mobile phase passing from the column. Thus a change in the refractive index, conductivity, or dielectric constant of a mobile phase can indicate the presence of eluting components.

4.4.2 *Solute Property Detectors* measure the physical or chemical characteristics of the component eluting from the column. Thus, light absorption (ultraviolet, visible, infrared), fluorescence, and polarography are examples of detectors capable of responding in such a manner.

4.4.3 *Differential Detectors* measure the instantaneous proportion of eluted sample components in the mobile phase passing through the detector or their instantaneous rate of arrival at the detector.

4.4.4 *Integral Detectors* measure the accumulated quantity of sample component(s) reaching the detector.

4.4.5 The detectors used in liquid chromatography may also be based on a variety of other physical or chemical phenomena.

<sup>3</sup> Basey, and Oliver, *Journal of Chromatography*, No. 251, 1982, p. 265.

4.5 *Fraction Collectors* are devices for recovering time-separated fractional volumes of the column effluent. The fraction collectors may be operated manually or automatically. Automatic fraction collectors consist of a series of test tubes or flasks. Column effluent is carried to one of the vessels and after a measured volume is collected or a set period of time has passed, the system automatically places the next vessel into position to receive a corresponding aliquot.

4.6 *The Developing Chamber* is a closed or open container, for either conventional or continuous development, respectively. Customarily it is of relatively large internal volume, used to enclose the media used in paper or thin-layer chromatography and also the mobile phase. It may be lined with a porous paper (*Saturated Development*) or it may be unlined (*Unsaturated Development*). Paper or plate equilibration is also possible by standing the paper or thin layer plate in the developing chamber containing the mobile phase for a given period of time before development without allowing the mobile phase to touch the paper or plate. If used for *Continuous Development*, the lid of the chamber is adjusted so the top portion of the thin layer plate can protrude past the lid allowing evaporation of the mobile phase near the solvent front. Automated instrumentation can effect this type of development by use of heated elements or air streams to force the evaporation of the mobile phase near the solvent front. A *Sandwich Chamber* has walls that are one half to one centimetre apart giving a relatively small internal volume. This type of developing chamber prohibits mobile phase vapors from getting onto the layer before the solvent front carries it throughout the layer effecting a different type of separation.

4.7 *Spotting Device* is a syringe or micropipet used to deliver a known volume of sample as a spot or streak to the paper or thin-layer media at the origin or near the beginning end of the planar media.

4.8 *Visualization Chamber* is a device in which the planar media may be viewed under ultraviolet light or sprayed with visualization reagents.

4.9 *Densitometer* is a device that allows portions of the developed paper or thin-layer media to be scanned with a beam of light of variable wavelength. The instrument in this manner is able to respond to differences in spot size and density in order to quantitate the separated compounds. The device may work in a transmission or reflectance mode.

## 5. Reagents

NOTE 5—In liquid chromatographic techniques the term “solvent” has been widely used to describe the mobile phase (that is, developing solvent, eluting solvent, solvent front). Due to the ambiguity of this term, its use is not recommended.

In various liquid chromatographic techniques the term “carrier” has been used to describe the solid on which the stationary phase is distributed or certain active groups involved in the separation process are bonded. Due to the similarity to the term “carrier gas” used as a synonym for the mobile phase in gas chromatography, the use of this expression is not recommended.

5.1 The *Mobile Phase* is the liquid used to sweep or elute the sample components along the planar surface or through the column. It may consist of a single component or a mixture of components. The term eluent is often used for the preferred *Mobile Phase*.

5.1.1 *Degassing* is the process of removing dissolved gases from the *Mobile Phase* before or during use. This can be accomplished by sparging (with helium), sonicating, heating, or applying a vacuum to the *Mobile Phase*.

5.2 The *Stationary Phase* is the active immobile material on the planar surface or within the column that retards the passage of sample components by one of a number of processes or their combination. There are three types of stationary phase: *Liquid Phases*, *Interactive Solids*, and *Bonded Phases*. Inert materials that merely provide physical support for the stationary phase are not part of the stationary phase.

5.2.1 The *Liquid Phase* is a stationary phase which has been sorbed (but not covalently bonded) to a solid support, paper sheet, or thin layer. Differences in the solubilities of the sample components in the liquid and mobile phase constitute the basis for their separation. Examples of materials that can be used as liquid phases are  $\beta,\beta'$ -oxydipropionitrile, silicone oil, and water.

5.2.2 The *Interactive Solid* is a stationary phase that comprises a relatively homogeneous surface on which the sample components sorb and desorb effecting a separation. Examples are silica, alumina, graphite, and ion exchangers.

5.2.3 The *Bonded Phase* is a stationary phase that comprises a chemical (or chemicals) that has been covalently attached to a solid support. The sample components sorb onto and off the bonded phase differentially to effect separation. Octadecylsilyl groups bonded to silica represent a typical example for a bonded phase.

5.2.3.1 A *Monomeric* phase is a bonded phase that has been attached to a support using a monofunctional silane reagent.

5.2.3.2 A *Polymeric* phase is a bonded phase that has been attached to a support using a di- or tri-functional silane reagent. The multifunctional reagent allows other cross-linking mechanisms to occur near the bonding region.

5.2.3.3 *Endcapping* is the process of bonding residual silanols not bonded by previous silanizing reactions through use of a smaller silanizing reagent such as trimethylchlorosilane.

5.2.3.4 *Coverage* is a relative measure of the amount of bonded phase on an inorganic support. It is usually described as  $\mu\text{mol}/\text{m}^2$  or in terms of percent carbon.

5.3 The *Solid Support* is the inert material to which the stationary phase is sorbed (liquid phases) or covalently attached (bonded phases). It holds the stationary phase in contact with the mobile phase.

5.4 The *Column Packing* consists of all the material used to fill packed columns. There are two types: totally porous and pellicular.

5.4.1 A *Totally Porous Packing* is one in which the stationary phase is found throughout each porous particle.